

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1352 SAS No.: Q1352 SDG NO.: Q1352

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL03 Date Analyzed: 02/17/2025

Lab Sample No.: PSTDCCC050 Data File : PL094217.D Time Analyzed: 09:43

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Decachlorobiphenyl	9.066	8.953	9.153	46.360	50.000	-7.3
Endrin	6.582	6.472	6.672	46.000	50.000	-8.0
gamma-BHC (Lindane)	4.335	4.227	4.427	46.760	50.000	-6.5
Heptachlor	4.923	4.814	5.014	48.410	50.000	-3.2
Heptachlor epoxide	5.691	5.582	5.782	47.010	50.000	-6.0
Methoxychlor	7.507	7.398	7.598	51.490	50.000	3.0
Tetrachloro-m-xylene	3.545	3.439	3.639	46.130	50.000	-7.7

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1352 SAS No.: Q1352 SDG NO.: Q1352

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL03 Date Analyzed: 02/17/2025

Lab Sample No.: PSTDCCC050 Data File : PL094217.D Time Analyzed: 09:43

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.912	7.810	8.010	47.570	50.000	-4.9
Endrin	5.638	5.536	5.736	48.030	50.000	-3.9
gamma-BHC (Lindane)	3.607	3.507	3.707	47.090	50.000	-5.8
Heptachlor	3.945	3.845	4.045	46.680	50.000	-6.6
Heptachlor epoxide	4.728	4.627	4.827	47.160	50.000	-5.7
Methoxychlor	6.612	6.509	6.709	48.870	50.000	-2.3
Tetrachloro-m-xylene	2.774	2.674	2.874	46.620	50.000	-6.8