

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1352 SAS No.: Q1352 SDG NO.: Q1352

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL04 Date Analyzed: 02/17/2025

Lab Sample No.: PSTDCCC050 Data File : PL094227.D Time Analyzed: 12:40

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	9.055	8.953	9.153	46.710	50.000	-6.6
Endrin	6.573	6.472	6.672	46.230	50.000	-7.5
gamma-BHC (Lindane)	4.326	4.227	4.427	46.330	50.000	-7.3
Heptachlor	4.914	4.814	5.014	47.470	50.000	-5.1
Heptachlor epoxide	5.683	5.582	5.782	46.840	50.000	-6.3
Methoxychlor	7.500	7.398	7.598	50.800	50.000	1.6
Tetrachloro-m-xylene	3.537	3.439	3.639	45.070	50.000	-9.9

CALIBRATION VERIFICATION SUMMARY

Contract: WEST04

Lab Code: CHEM Case No.: Q1352 SAS No.: Q1352 SDG NO.: Q1352

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/21/2025 01/21/2025

Client Sample No.: CCAL04 Date Analyzed: 02/17/2025

Lab Sample No.: PSTDCCC050 Data File : PL094227.D Time Analyzed: 12:40

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Decachlorobiphenyl	7.908	7.810	8.010	47.730	50.000	-4.5
Endrin	5.634	5.536	5.736	46.920	50.000	-6.2
gamma-BHC (Lindane)	3.604	3.507	3.707	47.410	50.000	-5.2
Heptachlor	3.942	3.845	4.045	46.900	50.000	-6.2
Heptachlor epoxide	4.724	4.627	4.827	47.290	50.000	-5.4
Methoxychlor	6.608	6.509	6.709	49.350	50.000	-1.3
Tetrachloro-m-xylene	2.771	2.674	2.874	46.830	50.000	-6.3