

Surrogate Summary

SW-846

SDG No.: Q1352

Client: Weston Solutions

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB166700TB	PB166700TB	2-Fluorophenol	150	140	93		19	119
		Phenol-d6	150	137	91		10	130
		Nitrobenzene-d5	100	91.9	92		44	120
		2-Fluorobiphenyl	100	93.1	93		44	119
		2,4,6-Tribromophenol	150	139	93		43	140
		Terphenyl-d14	100	92.4	92		50	134
PB166711BL	PB166711BL	2-Fluorophenol	150	136	91		19	119
		Phenol-d6	150	133	89		10	130
		Nitrobenzene-d5	100	93.0	93		44	120
		2-Fluorobiphenyl	100	95.8	96		44	119
		2,4,6-Tribromophenol	150	141	94		43	140
		Terphenyl-d14	100	92.6	93		50	134
PB166711BS	PB166711BS	2-Fluorophenol	150	135	90		19	119
		Phenol-d6	150	134	89		10	130
		Nitrobenzene-d5	100	91.6	92		44	120
		2-Fluorobiphenyl	100	91.0	91		44	119
		2,4,6-Tribromophenol	150	137	92		43	140
		Terphenyl-d14	100	95.2	95		50	134
Q1352-02	TAP-IDW-SOIL-021025	2-Fluorophenol	150	133	89		19	119
		Phenol-d6	150	121	81		10	130
		Nitrobenzene-d5	100	92.6	93		44	120
		2-Fluorobiphenyl	100	92.0	92		44	119
		2,4,6-Tribromophenol	150	141	94		43	140
		Terphenyl-d14	100	103	103		50	134
Q1356-04MS	CARBON-WATERMS	2-Fluorophenol	150	132	88		19	119
		Phenol-d6	150	119	80		10	130
		Nitrobenzene-d5	100	95.9	96		44	120
		2-Fluorobiphenyl	100	94.3	94		44	119
		2,4,6-Tribromophenol	150	143	95		43	140
		Terphenyl-d14	100	109	109		50	134
Q1356-04MSD	CARBON-WATERMSD	2-Fluorophenol	150	133	89		19	119
		Phenol-d6	150	120	80		10	130
		Nitrobenzene-d5	100	98.5	99		44	120
		2-Fluorobiphenyl	100	95.0	95		44	119
		2,4,6-Tribromophenol	150	142	95		43	140
		Terphenyl-d14	100	117	117		50	134