

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1352

Client: Weston Solutions

Analytical Method: SW8270E

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1356-04MSD	Client Sample ID:	CARBON-WATERMSD					DataFile:	BF141636.D		
Pyridine	500	0	250	ug/L	50		8		10	153	20
1,4-Dichlorobenzene	500	0	450	ug/L	90		2		29	112	20
2-Methylphenol	500	0	450	ug/L	90		2		30	117	20
3+4-Methylphenols	500	0	450	ug/L	90		0		29	110	20
Hexachloroethane	500	0	450	ug/L	90		2		21	115	20
Nitrobenzene	500	0	450	ug/L	90		2		45	121	20
Hexachlorobutadiene	500	0	450	ug/L	90		2		22	124	20
2,4,6-Trichlorophenol	500	0	460	ug/L	92		4		50	125	20
2,4,5-Trichlorophenol	500	0	440	ug/L	88		0		53	123	20
2,4-Dinitrotoluene	500	0	470	ug/L	94		0		57	128	20
Hexachlorobenzene	500	0	470	ug/L	94		0		53	125	20
Pentachlorophenol	1000	0	800	ug/L	80		0		35	138	20