

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1352 SAS No.: Q1352 SDG NO.: Q1352

EPA Sample No.: SSTDCCC040 Date Analyzed: 02/14/2025

Lab File ID: BF141628.D Time Analyzed: 09:45

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	103235	6.787	408887	8.07	222735	9.82
UPPER LIMIT	206470	7.287	817774	8.569	445470	10.322
LOWER LIMIT	51617.5	6.287	204444	7.569	111368	9.322
EPA SAMPLE NO.						
01 PB166711BL	91091	6.79	362070	8.07	202635	9.82
02 PB166711BS	90964	6.79	371902	8.07	212243	9.82
03 PB166700TB	90404	6.79	375424	8.07	209209	9.82
04 TAP-IDW-SOIL-021025	86790	6.79	356864	8.07	194980	9.82
05 CARBON-WATERMS	82541	6.79	333730	8.07	183625	9.82
06 CARBON-WATERMSD	84972	6.79	344114	8.07	188709	9.82

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.