

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: WEST04

Lab Code: CHEM Case No.: Q1352 SAS No.: Q1352 SDG No.: Q1352

Instrument ID: MSVOA_X Calibration Date/Time: 02/14/2025 10:04

Lab File ID: VX044948.D Init. Calib. Date(s): 02/10/2025 02/10/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:25 12:28

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.827	0.803		-2.9	20
1,1-Dichloroethene	0.644	0.637		-1.09	20
2-Butanone	0.478	0.507		6.07	20
Carbon Tetrachloride	0.460	0.493		7.17	20
Chloroform	1.196	1.212		1.34	20
Benzene	1.465	1.515		3.41	20
1,2-Dichloroethane	0.467	0.510		9.21	20
Trichloroethene	0.335	0.353		5.37	20
Tetrachloroethene	0.315	0.332		5.4	20
Chlorobenzene	1.074	1.118	0.3	4.1	20
1,2-Dichloroethane-d4	0.732	0.726		-0.82	20
Dibromofluoromethane	0.325	0.335		3.08	20
Toluene-d8	1.229	1.231		0.16	20
4-Bromofluorobenzene	0.414	0.440		6.28	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.