

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: WEST04

Lab Code: CHEM Case No.: Q1352 SAS No.: Q1352 SDG No.: Q1352

Instrument ID: MSVOA_X Calibration Date/Time: 02/14/2025 20:06

Lab File ID: VX044974.D Init. Calib. Date(s): 02/10/2025 02/10/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:25 12:28

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Vinyl Chloride	0.827	0.791		-4.35	50
1,1-Dichloroethene	0.644	0.650		0.93	50
2-Butanone	0.478	0.578		20.92	50
Carbon Tetrachloride	0.460	0.487		5.87	50
Chloroform	1.196	1.296		8.36	50
Benzene	1.465	1.543		5.32	50
1,2-Dichloroethane	0.467	0.534		14.35	50
Trichloroethene	0.335	0.359		7.16	50
Tetrachloroethene	0.315	0.320		1.59	50
Chlorobenzene	1.074	1.141	0.3	6.24	50
1,2-Dichloroethane-d4	0.732	0.698		-4.64	50
Dibromofluoromethane	0.325	0.301		-7.39	50
Toluene-d8	1.229	1.100		-10.5	50
4-Bromofluorobenzene	0.414	0.407		-1.69	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.