

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1352

Client: Weston Solutions

Analytical Method: SW8260-Low

Datafile : VX044951.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0214WBS01	Vinyl chloride	20	18.3	ug/L	92			58	137	
	1,1-Dichloroethene	20	18.9	ug/L	95			71	131	
	2-Butanone	100	110	ug/L	110			56	143	
	Carbon Tetrachloride	20	20.2	ug/L	101			72	136	
	Chloroform	20	20.6	ug/L	103			79	124	
	Benzene	20	20.0	ug/L	100			79	120	
	1,2-Dichloroethane	20	21.8	ug/L	109			73	128	
	Trichloroethene	20	19.8	ug/L	99			79	123	
	Tetrachloroethene	20	21.0	ug/L	105			74	129	
	Chlorobenzene	20	20.2	ug/L	101			82	118	