

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: WEST04
 Lab Code: CHEM Case No.: Q1352 SAS No.: Q1352 SDG NO.: Q1352
 Lab File ID: VX044948.D Date Analyzed: 02/14/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:04
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	79140	12.018				
UPPER LIMIT	158280	12.518				
LOWER LIMIT	39570	11.518				
EPA SAMPLE NO.						
TAP-IDW-SOIL-021025	66540	12.02				
VX0214WBL01	59196	12.02				
VX0214WBS01	79806	12.02				
VX0214WBSD01	77784	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.