

Initial and Continuing Calibration Blank Summary

Client: G Environmental

SDG No.: Q1355

Project: Amsterdam

RunNo.: LB134705

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
Sample ID: ICB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.034	2	01/20/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.011	0.6	01/20/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.057	0.4	01/20/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.011	0.6	01/20/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.0034	0.5	01/20/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.032	3	01/20/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.079	1	01/20/2025
Sample ID: CCB1							
Bromide	mg/L	< 1.0000	1.0000	U	0.034	2	02/11/2025
Chloride	mg/L	< 0.3000	0.3000	U	0.011	0.6	02/11/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.057	0.4	02/11/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.011	0.6	02/11/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.0034	0.5	02/11/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.032	3	02/11/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.079	1	02/11/2025
Sample ID: CCB2							
Bromide	mg/L	< 1.0000	1.0000	U	0.034	2	02/11/2025
Chloride	mg/L	0.065	0.3000	J	0.011	0.6	02/11/2025
Fluoride	mg/L	< 0.2000	0.2000	U	0.057	0.4	02/11/2025
Nitrite	mg/L	< 0.3000	0.3000	U	0.011	0.6	02/11/2025
Nitrate	mg/L	< 0.2500	0.2500	U	0.0034	0.5	02/11/2025
Sulfate	mg/L	< 1.5000	1.5000	U	0.032	3	02/11/2025
Orthophosphate as P	mg/L	< 0.5000	0.5000	U	0.079	1	02/11/2025

Initial and Continuing Calibration Blank Summary

Client: G Environmental

SDG No.: Q1355

Project: Amsterdam

RunNo.: LB134705

Analyte	Units	Result	Acceptance Limits	Conc Qual	MDL	RDL	Analysis Date
---------	-------	--------	----------------------	--------------	-----	-----	------------------