

Matrix Spike Summary

Client:	G Environmental	SDG No.:	Q1355
Project:	Amsterdam	Sample ID:	Q1355-01
Client ID:	RW1MS	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Bromide	mg/L	80-120	10.0		0.56	J	10	1	94		02/11/2025
Chloride	mg/L	80-120	323	OR	336	OR	3	1	-433	*	02/11/2025
Fluoride	mg/L	80-120	1.10		0.29	J	2	1	41	*	02/11/2025
Nitrite	mg/L	80-120	2.90		0.011	U	3	1	97		02/11/2025
Nitrate	mg/L	80-120	2.40		0.0034	U	2.5	1	96		02/11/2025
Sulfate	mg/L	80-120	115	OR	103	OR	15	1	80		02/11/2025
Orthophosphate as P	mg/L	80-120	0.57	J	0.079	U	5	1	11	*	02/11/2025

Matrix Spike Summary

Client:	G Environmental	SDG No.:	Q1355
Project:	Amsterdam	Sample ID:	Q1355-01
Client ID:	RW1MSD	Percent Solids for Spike Sample:	0

Analyte	Units	Acceptance Limit %R	Spiked Result	Conc. Qualifier	Sample Result	Conc. Qualifier	Spike Added	Dilution Factor	% Rec	Qual	Analysis Date
Bromide	mg/L	80-120	9.80		0.56	J	10	1	92		02/11/2025
Chloride	mg/L	80-120	323	OR	336	OR	3	1	-433	*	02/11/2025
Fluoride	mg/L	80-120	2.20		0.29	J	2	1	96		02/11/2025
Nitrite	mg/L	80-120	2.80		0.011	U	3	1	93		02/11/2025
Nitrate	mg/L	80-120	2.40		0.0034	U	2.5	1	96		02/11/2025
Sulfate	mg/L	80-120	115	OR	103	OR	15	1	80		02/11/2025
Orthophosphate as P	mg/L	80-120	2.70		0.079	U	5	1	54	*	02/11/2025