

Laboratory Control Sample Summary

Client: G Environmental

SDG No.: Q1355

Project: Amsterdam

Run No.: LB134705

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134705BSW							
Bromide	mg/L	10	10.2		102	1	90-110	02/11/2025
Chloride	mg/L	3	3.10		103	1	90-110	02/11/2025
Fluoride	mg/L	2	2.00		100	1	90-110	02/11/2025
Nitrite	mg/L	3	3.10		103	1	90-110	02/11/2025
Nitrate	mg/L	2.5	2.60		104	1	90-110	02/11/2025
Sulfate	mg/L	15	15.1		101	1	90-110	02/11/2025
Orthophosphate as P	mg/L	5	5.00		100	1	90-110	02/11/2025