



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: G Environmental SDG No.: Q1355
Contract: GENV01 Lab Code: CHEM Case No.: Q1355 SAS No.: Q1355
Initial Calibration Source: _____
Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRA	Mercury	0.22	0.2	108	40 - 160	CV	02/12/2025	14:16	LB134685
CRI01	Aluminum	95.4	100	95	40 - 160	P	02/13/2025	12:54	LB134709
	Antimony	50.3	50.0	101	40 - 160	P	02/13/2025	12:54	LB134709
	Arsenic	21.2	20.0	106	40 - 160	P	02/13/2025	12:54	LB134709
	Barium	84.9	100	85	40 - 160	P	02/13/2025	12:54	LB134709
	Beryllium	5.67	6.0	94	40 - 160	P	02/13/2025	12:54	LB134709
	Cadmium	5.67	6.0	94	40 - 160	P	02/13/2025	12:54	LB134709
	Calcium	1890	2000	95	40 - 160	P	02/13/2025	12:54	LB134709
	Chromium	10.3	10.0	103	40 - 160	P	02/13/2025	12:54	LB134709
	Cobalt	28.7	30.0	96	40 - 160	P	02/13/2025	12:54	LB134709
	Copper	21.8	20.0	109	40 - 160	P	02/13/2025	12:54	LB134709
	Iron	96.7	100	97	40 - 160	P	02/13/2025	12:54	LB134709
	Lead	10.4	12.0	86	40 - 160	P	02/13/2025	12:54	LB134709
	Magnesium	2030	2000	102	40 - 160	P	02/13/2025	12:54	LB134709
	Manganese	18.3	20.0	92	40 - 160	P	02/13/2025	12:54	LB134709
	Nickel	38.5	40.0	96	40 - 160	P	02/13/2025	12:54	LB134709
	Potassium	2000	2000	100	40 - 160	P	02/13/2025	12:54	LB134709
	Selenium	21.0	20.0	105	40 - 160	P	02/13/2025	12:54	LB134709
	Silver	10.2	10.0	102	40 - 160	P	02/13/2025	12:54	LB134709
	Sodium	1740	2000	87	40 - 160	P	02/13/2025	12:54	LB134709
	Thallium	38.1	40.0	95	40 - 160	P	02/13/2025	12:54	LB134709
	Vanadium	37.1	40.0	93	40 - 160	P	02/13/2025	12:54	LB134709
	Zinc	42.0	40.0	105	40 - 160	P	02/13/2025	12:54	LB134709