



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	G Environmental				SDG No.:	Q1355				
Contract:	GENV01	Lab Code:	CHEM		Case No.:	Q1355	SAS No.:	Q1355		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number	
ICB02	Mercury	0.20	+/-0.20	U	0.20	CV	02/12/2025	14:06	LB134685	

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental

SDG No.: Q1355

Contract: GENV01

Lab Code: CHEM

Case No.: Q1355

SAS No.: Q1355

Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB63	Mercury	0.20	+/-0.20	U	0.20	CV	02/12/2025	14:14	LB134685
CCB64	Mercury	0.20	+/-0.20	U	0.20	CV	02/12/2025	14:46	LB134685
CCB65	Mercury	0.20	+/-0.20	U	0.20	CV	02/12/2025	15:22	LB134685
CCB66	Mercury	0.20	+/-0.20	U	0.20	CV	02/12/2025	15:52	LB134685

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	G Environmental			SDG No.:	Q1355				
Contract:	GENV01	Lab Code:	CHEM	Case No.:	Q1355	SAS No.:	Q1355		
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
ICB01	Aluminum	100	+/-100	U	100	P	02/13/2025	12:50	LB134709
	Antimony	50.0	+/-50.0	U	50.0	P	02/13/2025	12:50	LB134709
	Arsenic	20.0	+/-20.0	U	20.0	P	02/13/2025	12:50	LB134709
	Barium	100	+/-100	U	100	P	02/13/2025	12:50	LB134709
	Beryllium	6.00	+/-6.00	U	6.00	P	02/13/2025	12:50	LB134709
	Cadmium	6.00	+/-6.00	U	6.00	P	02/13/2025	12:50	LB134709
	Calcium	2000	+/-2000	U	2000	P	02/13/2025	12:50	LB134709
	Chromium	10.0	+/-10.0	U	10.0	P	02/13/2025	12:50	LB134709
	Cobalt	30.0	+/-30.0	U	30.0	P	02/13/2025	12:50	LB134709
	Copper	20.0	+/-20.0	U	20.0	P	02/13/2025	12:50	LB134709
	Iron	100	+/-100	U	100	P	02/13/2025	12:50	LB134709
	Lead	12.0	+/-12.0	U	12.0	P	02/13/2025	12:50	LB134709
	Magnesium	2000	+/-2000	U	2000	P	02/13/2025	12:50	LB134709
	Manganese	20.0	+/-20.0	U	20.0	P	02/13/2025	12:50	LB134709
	Nickel	40.0	+/-40.0	U	40.0	P	02/13/2025	12:50	LB134709
	Potassium	2000	+/-2000	U	2000	P	02/13/2025	12:50	LB134709
	Selenium	20.0	+/-20.0	U	20.0	P	02/13/2025	12:50	LB134709
	Silver	10.0	+/-10.0	U	10.0	P	02/13/2025	12:50	LB134709
	Sodium	2000	+/-2000	U	2000	P	02/13/2025	12:50	LB134709
	Thallium	40.0	+/-40.0	U	40.0	P	02/13/2025	12:50	LB134709
	Vanadium	40.0	+/-40.0	U	40.0	P	02/13/2025	12:50	LB134709
	Zinc	40.0	+/-40.0	U	40.0	P	02/13/2025	12:50	LB134709

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: G Environmental		SDG No.: Q1355							
Contract: GENV01	Lab Code: CHEM	Case No.: Q1355	SAS No.: Q1355						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB01	Aluminum	100	+/-100	U	100	P	02/13/2025	13:27	LB134709
	Antimony	50.0	+/-50.0	U	50.0	P	02/13/2025	13:27	LB134709
	Arsenic	20.0	+/-20.0	U	20.0	P	02/13/2025	13:27	LB134709
	Barium	100	+/-100	U	100	P	02/13/2025	13:27	LB134709
	Beryllium	6.00	+/-6.00	U	6.00	P	02/13/2025	13:27	LB134709
	Cadmium	6.00	+/-6.00	U	6.00	P	02/13/2025	13:27	LB134709
	Calcium	2000	+/-2000	U	2000	P	02/13/2025	13:27	LB134709
	Chromium	10.0	+/-10.0	U	10.0	P	02/13/2025	13:27	LB134709
	Cobalt	30.0	+/-30.0	U	30.0	P	02/13/2025	13:27	LB134709
	Copper	20.0	+/-20.0	U	20.0	P	02/13/2025	13:27	LB134709
	Iron	100	+/-100	U	100	P	02/13/2025	13:27	LB134709
	Lead	12.0	+/-12.0	U	12.0	P	02/13/2025	13:27	LB134709
	Magnesium	2000	+/-2000	U	2000	P	02/13/2025	13:27	LB134709
	Manganese	20.0	+/-20.0	U	20.0	P	02/13/2025	13:27	LB134709
	Nickel	40.0	+/-40.0	U	40.0	P	02/13/2025	13:27	LB134709
	Potassium	2000	+/-2000	U	2000	P	02/13/2025	13:27	LB134709
	Selenium	20.0	+/-20.0	U	20.0	P	02/13/2025	13:27	LB134709
	Silver	10.0	+/-10.0	U	10.0	P	02/13/2025	13:27	LB134709
	Sodium	2000	+/-2000	U	2000	P	02/13/2025	13:27	LB134709
	Thallium	40.0	+/-40.0	U	40.0	P	02/13/2025	13:27	LB134709
	Vanadium	40.0	+/-40.0	U	40.0	P	02/13/2025	13:27	LB134709
	Zinc	40.0	+/-40.0	U	40.0	P	02/13/2025	13:27	LB134709
CCB02	Aluminum	100	+/-100	U	100	P	02/13/2025	14:29	LB134709
	Antimony	50.0	+/-50.0	U	50.0	P	02/13/2025	14:29	LB134709
	Arsenic	20.0	+/-20.0	U	20.0	P	02/13/2025	14:29	LB134709
	Barium	100	+/-100	U	100	P	02/13/2025	14:29	LB134709
	Beryllium	6.00	+/-6.00	U	6.00	P	02/13/2025	14:29	LB134709
	Cadmium	6.00	+/-6.00	U	6.00	P	02/13/2025	14:29	LB134709
	Calcium	2000	+/-2000	U	2000	P	02/13/2025	14:29	LB134709
	Chromium	10.0	+/-10.0	U	10.0	P	02/13/2025	14:29	LB134709
	Cobalt	30.0	+/-30.0	U	30.0	P	02/13/2025	14:29	LB134709
	Copper	20.0	+/-20.0	U	20.0	P	02/13/2025	14:29	LB134709
	Iron	100	+/-100	U	100	P	02/13/2025	14:29	LB134709
	Lead	12.0	+/-12.0	U	12.0	P	02/13/2025	14:29	LB134709
	Magnesium	2000	+/-2000	U	2000	P	02/13/2025	14:29	LB134709
	Manganese	20.0	+/-20.0	U	20.0	P	02/13/2025	14:29	LB134709
	Nickel	40.0	+/-40.0	U	40.0	P	02/13/2025	14:29	LB134709
	Potassium	2000	+/-2000	U	2000	P	02/13/2025	14:29	LB134709
	Selenium	20.0	+/-20.0	U	20.0	P	02/13/2025	14:29	LB134709

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client: <u>G Environmental</u>		SDG No.: <u>Q1355</u>							
Contract: <u>GENV01</u>	Lab Code: <u>CHEM</u>	Case No.: <u>Q1355</u>	SAS No.: <u>Q1355</u>						
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number
CCB02	Silver	10.0	+/-10.0	U	10.0	P	02/13/2025	14:29	LB134709
	Sodium	2000	+/-2000	U	2000	P	02/13/2025	14:29	LB134709
	Thallium	40.0	+/-40.0	U	40.0	P	02/13/2025	14:29	LB134709
	Vanadium	40.0	+/-40.0	U	40.0	P	02/13/2025	14:29	LB134709
	Zinc	40.0	+/-40.0	U	40.0	P	02/13/2025	14:29	LB134709
CCB03	Aluminum	100	+/-100	U	100	P	02/13/2025	14:51	LB134709
	Antimony	50.0	+/-50.0	U	50.0	P	02/13/2025	14:51	LB134709
	Arsenic	20.0	+/-20.0	U	20.0	P	02/13/2025	14:51	LB134709
	Barium	100	+/-100	U	100	P	02/13/2025	14:51	LB134709
	Beryllium	6.00	+/-6.00	U	6.00	P	02/13/2025	14:51	LB134709
	Cadmium	6.00	+/-6.00	U	6.00	P	02/13/2025	14:51	LB134709
	Calcium	2000	+/-2000	U	2000	P	02/13/2025	14:51	LB134709
	Chromium	10.0	+/-10.0	U	10.0	P	02/13/2025	14:51	LB134709
	Cobalt	30.0	+/-30.0	U	30.0	P	02/13/2025	14:51	LB134709
	Copper	20.0	+/-20.0	U	20.0	P	02/13/2025	14:51	LB134709
	Iron	100	+/-100	U	100	P	02/13/2025	14:51	LB134709
	Lead	12.0	+/-12.0	U	12.0	P	02/13/2025	14:51	LB134709
	Magnesium	2000	+/-2000	U	2000	P	02/13/2025	14:51	LB134709
	Manganese	20.0	+/-20.0	U	20.0	P	02/13/2025	14:51	LB134709
	Nickel	40.0	+/-40.0	U	40.0	P	02/13/2025	14:51	LB134709
	Potassium	2000	+/-2000	U	2000	P	02/13/2025	14:51	LB134709
	Selenium	20.0	+/-20.0	U	20.0	P	02/13/2025	14:51	LB134709
	Silver	10.0	+/-10.0	U	10.0	P	02/13/2025	14:51	LB134709
	Sodium	2000	+/-2000	U	2000	P	02/13/2025	14:51	LB134709
	Thallium	40.0	+/-40.0	U	40.0	P	02/13/2025	14:51	LB134709
	Vanadium	40.0	+/-40.0	U	40.0	P	02/13/2025	14:51	LB134709
	Zinc	40.0	+/-40.0	U	40.0	P	02/13/2025	14:51	LB134709

Metals

- 3a -

INITIAL AND CONTINUING CALIBRATION BLANK SUMMARY

Client:	<u>G Environmental</u>				SDG No.:	<u>Q1355</u>			
Contract:	<u>GENV01</u>	Lab Code:	<u>CHEM</u>		Case No.:	<u>Q1355</u>	SAS No.:	<u>Q1355</u>	
Sample ID	Analyte	Result ug/L	Acceptance Limit	Conc Qual	CRQL	M	Analysis Date	Analysis Time	Run Number