

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental **SDG No.:** Q1355
Contract: GENV01 **Lab Code:** CHEM **Case No.:** Q1355 **SAS No.:** Q1355

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166694BS							
Aluminum	ug/L	1000	888		89	80 - 120	P
Antimony	ug/L	400	369		92	80 - 120	P
Arsenic	ug/L	400	353		88	80 - 120	P
Barium	ug/L	100	81.3		81	80 - 120	P
Beryllium	ug/L	100	83.2		83	80 - 120	P
Cadmium	ug/L	100	85.7		86	80 - 120	P
Calcium	ug/L	500	462	J	92	80 - 120	P
Chromium	ug/L	200	195		98	80 - 120	P
Cobalt	ug/L	100	90.0		90	80 - 120	P
Copper	ug/L	150	143		95	80 - 120	P
Iron	ug/L	1500	1380		92	80 - 120	P
Lead	ug/L	500	425		85	80 - 120	P
Magnesium	ug/L	1000	870	J	87	80 - 120	P
Manganese	ug/L	100	88.1		88	80 - 120	P
Nickel	ug/L	250	224		90	80 - 120	P
Potassium	ug/L	5000	4530		91	80 - 120	P
Selenium	ug/L	1000	854		85	80 - 120	P
Silver	ug/L	37.5	34.9		93	80 - 120	P
Sodium	ug/L	1500	1230		82	80 - 120	P
Thallium	ug/L	1000	869		87	80 - 120	P
Vanadium	ug/L	150	133		89	80 - 120	P
Zinc	ug/L	100	93.5		94	80 - 120	P

Metals

- 7 -

LABORATORY CONTROL SAMPLE SUMMARY

Client: G Environmental

SDG No.: Q1355

Contract: GENV01

Lab Code: CHEM

Case No.: Q1355

SAS No.: Q1355

Analyte	Units	True Value	Result	C	% Recovery	Acceptance Limits	M
PB166706BS Mercury	ug/L	4.0	3.87		97	80 - 120	CV