

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q1355 SAS No.: Q1355 SDG NO.: Q1355
 Lab File ID: VX044920.D Date Analyzed: 02/12/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:07
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	120105	5.54	211946	6.75	185163	10.05
UPPER LIMIT	240210	6.044	423892	7.251	370326	10.549
LOWER LIMIT	60052.5	5.044	105973	6.251	92581.5	9.549
EPA SAMPLE NO.						
RW1	83265	5.54	169774	6.76	156232	10.05
MW2	85678	5.54	173415	6.76	155759	10.05
VX0212WBL01	96997	5.54	194457	6.76	176635	10.05
VX0212WBS01	111261	5.54	204350	6.76	179478	10.05
VX0212WBSD01	109124	5.54	195530	6.75	172961	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.