

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GENV01
 Lab Code: CHEM Case No.: Q1355 SAS No.: Q1355 SDG NO.: Q1355
 Lab File ID: VX044920.D Date Analyzed: 02/12/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:07
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	84033	12.018				
UPPER LIMIT	168066	12.518				
LOWER LIMIT	42016.5	11.518				
EPA SAMPLE NO.						
RW1	67649	12.02				
MW2	62222	12.02				
VX0212WBL01	76229	12.02				
VX0212WBS01	82301	12.02				
VX0212WBSD01	79611	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.