

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: GENV01
Lab Code: CHEM Case No.: Q1355 SAS No.: Q1355 SDG NO.: Q1355
Lab File ID: VX044919.D BFB Injection Date: 02/12/2025
Instrument ID: MSVOA_X BFB Injection Time: 09:39
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	20.4
75	30.0 - 60.0% of mass 95	52.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.9
173	Less than 2.0% of mass 174	0.7 (0.9) 1
174	50.0 - 100.0% of mass 95	77.1
175	5.0 - 9.0% of mass 174	5.6 (7.2) 1
176	95.0 - 101.0% of mass 174	74.9 (97.1) 1
177	5.0 - 9.0% of mass 176	4.9 (6.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VX044920.D	02/12/2025	10:07
VX0212WBL01	VX0212WBL01	VX044922.D	02/12/2025	10:57
VX0212WBS01	VX0212WBS01	VX044923.D	02/12/2025	11:20
VX0212WBSD01	VX0212WBSD01	VX044924.D	02/12/2025	11:46
RW1	Q1355-01	VX044937.D	02/12/2025	16:44
MW2	Q1355-02	VX044938.D	02/12/2025	17:07