

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q1356

Project: 323562

Run No.: LB134697

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134697BS							
TPH	mg/L	20.0	16.7		84	1	78-114	02/13/2025

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q1356

Project: 323562

Run No.: LB134697

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134697BSD							
TPH	mg/L	20.0	17.0		85	1	78-114	02/13/2025

Laboratory Control Sample Summary

Client: Parkway Generation Operating LLC

SDG No.: Q1356

Project: 323562

Run No.: LB134717

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134717BS							
TPH	mg/Kg	100	94.9		95	1	80-120	02/14/2025