

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q1359

Project: Pre Treatment Plant

RunNo.: LB134693

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Orthophosphate as P	mg/L	0.496	0.50	99	90-110	02/12/2025
Sample ID: CCV1 Orthophosphate as P	mg/L	0.509	0.5	102	90-110	02/12/2025
Sample ID: CCV2 Orthophosphate as P	mg/L	0.510	0.5	102	90-110	02/12/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q1359

Project: Pre Treatment Plant

RunNo.: LB134695

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV Phosphorus, Total	mg/L	0.504	0.50	101	90-110	02/12/2025
Sample ID: CCV1 Phosphorus, Total	mg/L	0.507	0.50	101	90-110	02/12/2025
Sample ID: CCV2 Phosphorus, Total	mg/L	0.505	0.50	101	90-110	02/12/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q1359

Project: Pre Treatment Plant

RunNo.: LB134721

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
Sample ID: ICV1 Ammonia as N	mg/L	1	1	100	90-110	02/14/2025
Sample ID: CCV1 Ammonia as N	mg/L	0.93	1	93	90-110	02/14/2025
Sample ID: CCV2 Ammonia as N	mg/L	1	1	100	90-110	02/14/2025
Sample ID: CCV3 Ammonia as N	mg/L	1	1	100	90-110	02/14/2025
Sample ID: CCV4 Ammonia as N	mg/L	0.99	1	99	90-110	02/14/2025
Sample ID: CCV5 Ammonia as N	mg/L	0.99	1	99	90-110	02/14/2025

Initial and Continuing Calibration Verification

Client: Holland Manufacturing Co.

SDG No.: Q1359

Project: Pre Treatment Plant

RunNo.: LB134721

Analyte	Units	Result	True Value	% Recovery	Acceptance Window (%R)	Analysis Date
---------	-------	--------	------------	---------------	---------------------------	------------------