

Extraction and Analytical Summary Report

Analysis Method: 1664A
Test: TPH
Run Number: LB134697
Analysis Date: 02/13/2025
BalanceID: WC SC-6
OvenID: EXT OVEN-3

ANALYST: jignesh
REVIEWED BY: Iwona
Extraction Date: 02/13/2025
Extraction IN Time: 08:00
Extraction OUT Time: 08:40
Thermometer ID: EXT OVEN#3

Dish #	Lab ID	Client ID	Matrix	pH	Sample Vol (ml)	Final Volume (ml)	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Silica Gel Weight (g)	Weight After Drying (g)	Final Weight After Drying (g)	Change Weight (g)	Result in ppm
1	LB134697BL	LB134697BL	WATER	1.3	1000	100	2.7463	2.7463	3.01	2.7464	2.7464	0.0001	0.1
2	LB134697BS	LB134697BS	WATER	1.3	1000	100	3.0158	3.0158	3.00	3.0325	3.0325	0.0167	16.7
3	LB134697BSD	LB134697BSD	WATER	1.3	1000	100	2.9866	2.9866	3.03	3.0036	3.0036	0.0170	17
4	Q1356-04	CARBON-WATER	WATER	1.3	1000	100	3.0112	3.0112	3.02	3.0116	3.0116	0.0004	0.4
5	Q1356-05	CARBON-FB	WATER	1.3	1000	100	2.9952	2.9952	3.03	2.9954	2.9954	0.0002	0.2
6	Q1356-06	WATER-A	WATER	1.6	1000	100	3.0018	3.0018	3.04	3.0021	3.0021	0.0003	0.3
7	Q1356-07	WATER-B	WATER	1.6	1000	100	3.1056	3.1056	3.03	3.1063	3.1063	0.0007	0.7
8	Q1356-08	WATER-FB	WATER	1.3	1000	100	3.0892	3.0892	3.02	3.0895	3.0895	0.0003	0.3
9	Q1356-09	SOIL-FB	WATER	1.3	1000	100	3.0374	3.0374	3.01	3.0377	3.0377	0.0003	0.3
10	Q1361-01	GRAB	WATER	1.6	1000	100	3.0406	3.0406	18.00	3.1782	3.1782	0.1376	137.6

QC Batch# LB134697

Test: TPH

Analysis Date: 02/13/2025

Chemicals Used:

Chemical Name	Chemical Lot #
HEXANE	W3177
pH Paper 0-14	M6069
Sodium Sulfate	EP2585
1:1 HCL	WP110826
Silica Gel	W3079
Sand	NA

Standards Used:

Standard Name	Amount Used	Standard Lot #
LCSW	5.00 ML	WP100827
LCSWD	5.00 ML	WP100828
MS/MSD	NA	NA

BALANCE CALIBRATION / OVEN Dessicator Data

Analytical Balance ID # : WC SC-6

Before Analysis

0.0020 gram Balance: 0.0018 (0.0018-0.0022) In OVEN TEMP1 : 70 °C Dessicator Time In1 : 10:41

1.0000 gram Balance: 1.0004 (0.9950-1.0050) In Time1: 09:45

Bal Check Time: 08:10 Out OVEN TEMP1: 70 °C Dessicator Time Out1: 11:30

Out Time1: 10:40

After Analysis

0.0020 gram Balance: 0.0021 (0.0018-0.0022) In OVEN TEMP2 : 71 °C Dessicator Time In2 : 12:46

1.0000 gram Balance: 1.0005 (0.9950-1.0050) In Time2: 12:10

Bal Check Time: 13:30 Out OVEN TEMP2: 71 °C Dessicator Time Out2: 13:25

Out Time2: 12:45

WORKLIST(Hardcopy Internal Chain)

UB 134697

WorkList Name : TPH W Q1364

WorkList ID : 187686

Department : Wet-Chemistry

Date : 02-13-2025 07:47:25

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1356-04 M	CARBON-WATER	Water	TPH	Conc H2SO4 to pH < 2	PSEG04	N51	02/11/2025	1664A
Q1356-05 M	CARBON-FB	Water	TPH	Conc H2SO4 to pH < 2	PSEG04	N51	02/11/2025	1664A
Q1356-06 M	WATER-A	Water	TPH	Conc H2SO4 to pH < 2	PSEG04	N51	02/11/2025	1664A
Q1356-07 M	WATER-B	Water	TPH	Conc H2SO4 to pH < 2	PSEG04	N51	02/11/2025	1664A
Q1356-08 M	WATER-FB	Water	TPH	Conc H2SO4 to pH < 2	PSEG04	N51	02/11/2025	1664A
Q1356-09 M	SOIL-FB	Water	TPH	Conc H2SO4 to pH < 2	PSEG04	N51	02/11/2025	1664A
Q1361-01	GRAB	Water	TPH	Conc H2SO4 to pH < 2	ARAM01	N51	02/12/2025	1664A

Date/Time 02/13/25 07:55

Raw Sample Received by: JG woc

Raw Sample Relinquished by: AP Sn

Date/Time 02/13/25 13:30

Raw Sample Received by: AP Sn

Raw Sample Relinquished by: JG woc