

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q1361

Project: Monthly 2025

Run No.: LB134697

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134697BS							
TPH	mg/L	20.0	16.7		84	1	78-114	02/13/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q1361

Project: Monthly 2025

Run No.: LB134697

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134697BSD							
TPH	mg/L	20.0	17.0		85	1	78-114	02/13/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q1361

Project: Monthly 2025

Run No.: LB134698

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134698BS							
TSS	mg/L	550	532		97	1	90-110	02/13/2025

Laboratory Control Sample Summary

Client: Aramark Uniforms

SDG No.: Q1361

Project: Monthly 2025

Run No.: LB134708

Analyte	Units	True Value	Result	Conc. Qualifier	% Recovery	Dilution Factor	Acceptance Limit %R	Analysis Date
Sample ID	LB134708BS							
BOD5	mg/L	198	199		100	1	84.6-115.4	02/13/2025