

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TECH05  
Lab Code: CHEM Case No.: Q1363 SAS No.: Q1363 SDG NO.: Q1363  
Lab File ID: VX044976.D Date Analyzed: 02/18/2025  
Instrument ID: MSVOA\_X Time Analyzed: 10:45  
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                | IS1<br>AREA # | RT #  | IS2<br>AREA # | RT #  | IS3<br>AREA # | RT #   |
|----------------|---------------|-------|---------------|-------|---------------|--------|
| 12 HOUR STD    | 118005        | 5.54  | 207185        | 6.75  | 183597        | 10.05  |
| UPPER LIMIT    | 236010        | 6.038 | 414370        | 7.251 | 367194        | 10.549 |
| LOWER LIMIT    | 59002.5       | 5.038 | 103593        | 6.251 | 91798.5       | 9.549  |
| EPA SAMPLE NO. |               |       |               |       |               |        |
| EB-01-021125   | 84342         | 5.54  | 169783        | 6.76  | 153945        | 10.05  |
| VX0218WBL01    | 92895         | 5.55  | 191012        | 6.76  | 185140        | 10.05  |
| VX0218WBS01    | 109899        | 5.54  | 202747        | 6.76  | 177300        | 10.05  |

IS1 = Pentafluorobenzene  
IS2 = 1,4-Difluorobenzene  
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area  
AREA LOWER LIMIT = -50% of internal standard area  
RT UPPER LIMIT = +0.50 minutes of internal standard RT  
RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
\* Values outside of QC limits.