

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TECH05
Lab Code: CHEM Case No.: Q1363 SAS No.: Q1363 SDG NO.: Q1363
Lab File ID: VX044948.D Date Analyzed: 02/14/2025
Instrument ID: MSVOA_X Time Analyzed: 10:04
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	79140	12.018				
UPPER LIMIT	158280	12.518				
LOWER LIMIT	39570	11.518				
EPA SAMPLE NO.						
MW-30D-021125	58167	12.02				
TB-01-021125	74015	12.02				
MW-33D-021225	66405	12.02				
MW-33D-021225MS	80350	12.02				
MW-33D-021225MSD	71520	12.02				
VFCC-2-021225	68646	12.02				
VFCC-3-021225	65055	12.02				
MW-33S-021225	72129	12.02				
MW-33S-FD-021225	79529	12.02				
VX0214WBL01	59196	12.02				
VX0214WBS01	79806	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.