

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TECH05  
 Lab Code: CHEM Case No.: Q1363 SAS No.: Q1363 SDG NO.: Q1363  
 Lab File ID: VX044976.D Date Analyzed: 02/18/2025  
 Instrument ID: MSVOA\_X Time Analyzed: 10:45  
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

|                | IS4<br>AREA # | RT #   |  |  |  |  |
|----------------|---------------|--------|--|--|--|--|
| 12 HOUR STD    | 88801         | 12.018 |  |  |  |  |
| UPPER LIMIT    | 177602        | 12.518 |  |  |  |  |
| LOWER LIMIT    | 44400.5       | 11.518 |  |  |  |  |
| EPA SAMPLE NO. |               |        |  |  |  |  |
| EB-01-021125   | 67101         | 12.02  |  |  |  |  |
| VX0218WBL01    | 91305         | 12.02  |  |  |  |  |
| VX0218WBS01    | 83521         | 12.02  |  |  |  |  |

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area  
 AREA LOWER LIMIT = -50% of internal standard area  
 RT UPPER LIMIT = +0.50 minutes of internal standard RT  
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

# Column used to flag values outside QC limits with an asterisk.  
 \* Values outside of QC limits.