

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1370 SAS No.: Q1370 SDG NO.: Q1370

Continuing Calib Date: 02/24/2025 Initial Calibration Date(s): 02/24/2025 02/24/2025

Continuing Calib Time: 21:27 Initial Calibration Time(s): 17:17 19:59

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.84	4.84	4.77	4.91	0.00
Aroclor-1016-2	(2)	4.87	4.87	4.80	4.94	0.01
Aroclor-1016-3	(3)	4.93	4.93	4.86	5.00	0.00
Aroclor-1016-4	(4)	5.03	5.03	4.96	5.10	0.00
Aroclor-1016-5	(5)	5.35	5.35	5.28	5.42	0.01
Aroclor-1260-1	(1)	6.56	6.56	6.49	6.63	0.00
Aroclor-1260-2	(2)	6.84	6.84	6.77	6.91	0.00
Aroclor-1260-3	(3)	7.23	7.23	7.16	7.30	0.00
Aroclor-1260-4	(4)	7.48	7.48	7.41	7.55	0.00
Aroclor-1260-5	(5)	7.82	7.82	7.75	7.89	0.00
TCX		3.62	3.62	3.57	3.67	0.00
DCB		9.51	9.51	9.41	9.61	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1370 SAS No.: Q1370 SDG NO.: Q1370

Continuing Calib Date: 02/24/2025 Initial Calibration Date(s): 02/24/2025 02/24/2025

Continuing Calib Time: 21:27 Initial Calibration Time(s): 17:17 19:59

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.07	4.07	4.00	4.14	0.00
Aroclor-1016-2	(2)	4.09	4.09	4.02	4.16	0.00
Aroclor-1016-3	(3)	4.27	4.27	4.20	4.34	0.00
Aroclor-1016-4	(4)	4.32	4.32	4.25	4.39	0.00
Aroclor-1016-5	(5)	4.54	4.54	4.47	4.61	0.00
Aroclor-1260-1	(1)	5.63	5.63	5.56	5.70	0.00
Aroclor-1260-2	(2)	5.84	5.84	5.77	5.91	0.00
Aroclor-1260-3	(3)	6.00	6.00	5.93	6.07	0.00
Aroclor-1260-4	(4)	6.50	6.51	6.44	6.58	0.01
Aroclor-1260-5	(5)	6.77	6.77	6.70	6.84	0.00
TCX		2.95	2.96	2.91	3.01	0.01
DCB		8.22	8.23	8.13	8.33	0.01