

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1370 SAS No.: Q1370 SDG NO.: Q1370

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL01 Date Analyzed: 02/24/2025

Lab Sample No.: AR1660CCC400 Data File : PR070349.D Time Analyzed: 20:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.843	4.743	4.943	29693480.000	28497100.000	-4.0
Aroclor-1016-2	4.865	4.765	4.965	48477110.000	45379700.000	-6.4
Aroclor-1016-3	4.929	4.829	5.029	31336600.000	29684900.000	-5.3
Aroclor-1016-4	5.031	4.930	5.130	24345700.000	22892600.000	-6.0
Aroclor-1016-5	5.346	5.245	5.445	26148880.000	24478200.000	-6.4
Aroclor-1260-1	6.558	6.457	6.657	46086410.000	43307900.000	-6.0
Aroclor-1260-2	6.841	6.741	6.941	46325500.000	42759200.000	-7.7
Aroclor-1260-3	7.232	7.132	7.332	39782920.000	37351000.000	-6.1
Aroclor-1260-4	7.474	7.376	7.576	42468210.000	39500900.000	-7.0
Aroclor-1260-5	7.815	7.717	7.917	65645630.000	60033900.000	-8.5
DCB	9.511	9.411	9.611	597291800.000	553051000.000	-7.4
TCX	3.623	3.523	3.723	1259296000.000	1182630000.000	-6.1

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1370 SAS No.: Q1370 SDG NO.: Q1370

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL01 Date Analyzed: 02/24/2025

Lab Sample No.: AR1660CCC400 Data File : PR070349.D Time Analyzed: 20:28

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.073	3.973	4.173	160311700.000	153728000.000	-4.1
Aroclor-1016-2	4.090	3.991	4.191	248437200.000	234008000.000	-5.8
Aroclor-1016-3	4.270	4.171	4.371	134062100.000	126748000.000	-5.5
Aroclor-1016-4	4.320	4.221	4.421	120912200.000	114536000.000	-5.3
Aroclor-1016-5	4.539	4.439	4.639	146431000.000	139062000.000	-5.0
Aroclor-1260-1	5.633	5.534	5.734	236629100.000	221787000.000	-6.3
Aroclor-1260-2	5.840	5.741	5.941	251132000.000	234789000.000	-6.5
Aroclor-1260-3	5.998	5.899	6.099	256654200.000	234309000.000	-8.7
Aroclor-1260-4	6.505	6.406	6.606	207477000.000	192391000.000	-7.3
Aroclor-1260-5	6.772	6.674	6.874	418556100.000	390070000.000	-6.8
DCB	8.224	8.126	8.326	3231215000.000	2979030000.000	-7.8
TCX	2.955	2.855	3.055	6283122000.000	5936920000.000	-5.5