

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1370 SAS No.: Q1370 SDG NO.: Q1370

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL02 Date Analyzed: 02/24/2025

Lab Sample No.: AR1660CCC400 Data File : PR070353.D Time Analyzed: 21:27

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.844	4.743	4.943	29693480.000	29249800.000	-1.5
Aroclor-1016-2	4.865	4.765	4.965	48477110.000	46104800.000	-4.9
Aroclor-1016-3	4.929	4.829	5.029	31336600.000	29666000.000	-5.3
Aroclor-1016-4	5.030	4.930	5.130	24345700.000	23000300.000	-5.5
Aroclor-1016-5	5.345	5.245	5.445	26148880.000	24620900.000	-5.8
Aroclor-1260-1	6.557	6.457	6.657	46086410.000	43945600.000	-4.6
Aroclor-1260-2	6.841	6.741	6.941	46325500.000	44246900.000	-4.5
Aroclor-1260-3	7.233	7.132	7.332	39782920.000	38495700.000	-3.2
Aroclor-1260-4	7.476	7.376	7.576	42468210.000	40986800.000	-3.5
Aroclor-1260-5	7.817	7.717	7.917	65645630.000	63878800.000	-2.7
DCB	9.510	9.411	9.611	597291800.000	577457000.000	-3.3
TCX	3.623	3.523	3.723	1259296000.000	1182600000.000	-6.1

CALIBRATION VERIFICATION SUMMARY

Contract: CHEM02

Lab Code: CHEM Case No.: Q1370 SAS No.: Q1370 SDG NO.: Q1370

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 02/24/2025 02/24/2025

Client Sample No.: CCAL02 Date Analyzed: 02/24/2025

Lab Sample No.: AR1660CCC400 Data File : PR070353.D Time Analyzed: 21:27

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.071	3.973	4.173	160311700.000	155857000.000	-2.8
Aroclor-1016-2	4.089	3.991	4.191	248437200.000	234388000.000	-5.7
Aroclor-1016-3	4.269	4.171	4.371	134062100.000	127516000.000	-4.9
Aroclor-1016-4	4.320	4.221	4.421	120912200.000	113290000.000	-6.3
Aroclor-1016-5	4.538	4.439	4.639	146431000.000	136379000.000	-6.9
Aroclor-1260-1	5.632	5.534	5.734	236629100.000	221332000.000	-6.5
Aroclor-1260-2	5.839	5.741	5.941	251132000.000	237757000.000	-5.3
Aroclor-1260-3	5.997	5.899	6.099	256654200.000	239447000.000	-6.7
Aroclor-1260-4	6.504	6.406	6.606	207477000.000	194775000.000	-6.1
Aroclor-1260-5	6.771	6.674	6.874	418556100.000	401121000.000	-4.2
DCB	8.223	8.126	8.326	3231215000.000	3068010000.000	-5.1
TCX	2.954	2.855	3.055	6283122000.000	5869790000.000	-6.6