

Surrogate Summary

SDG No.: Q1370

Client: Chemtech Consulting Group

Analytical Method: SFAM_PCB

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PR070335.D	PIBLK-PR070335.D	TCX	1	20	0.21	104		30	150
		DCB	1	40	0.44	111		30	150
		TCX	2	20	0.22	108		30	150
		DCB	2	40	0.43	109		30	150
I.BLK-PR070348.D	PIBLK-PR070348.D	TCX	1	20	0.21	104		30	150
		DCB	1	40	0.43	109		30	150
		TCX	2	20	0.22	110		30	150
		DCB	2	40	0.44	109		30	150
Q1370-13	PCB-GPC2-BLANK	TCX	1	20	0	0	*	30	150
		DCB	1	40	0	0	*	30	150
		TCX	2	20	0	0	*	30	150
		DCB	2	40	0	0	*	30	150
Q1370-14	PCB-GPC2-BLANK-SPIKE	TCX	1	20	0	0	*	30	150
		DCB	1	40	0	0	*	30	150
		TCX	2	20	0	0	*	30	150
		DCB	2	40	0	0	*	30	150
I.BLK-PR070352.D	PIBLK-PR070352.D	TCX	1	20	0.21	104		30	150
		DCB	1	40	0.44	109		30	150
		TCX	2	20	0.22	108		30	150
		DCB	2	40	0.44	111		30	150