

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1370 SAS No.: Q1370 SDG NO.: Q1370

EPA Sample No.: SSTD020662 Date Analyzed: 02/26/2025

Lab File ID: BP024110.D Time Analyzed: 11:21

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	453164	7.763	1863920	10.53	1092430	14.39
UPPER LIMIT	906328	8.263	3727840	11.034	2184860	14.892
LOWER LIMIT	226582	7.263	931960	10.034	546215	13.892
EPA SAMPLE NO.						
01 SVOC-GPC2-BLANK	375433	7.76	1533100	10.53	929199	14.39

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.