

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: CHEM02
 Lab Code: CHEM Case No.: Q1370 SAS No.: Q1370 SDG No.: Q1370
 Instrument ID: MSVOA_X Calibration Date/Time: 02/18/2025 10:45
 Lab File ID: VX044976.D Init. Calib. Date(s): 02/10/2025 02/10/2025
 Heated Purge: (Y/N) N Init. Calib. Time(s): 10:25 12:28
 GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Chloroethane	50.000	69.84	-39.67	20