

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: CHEM02
 Lab Code: CHEM Case No.: Q1370 SAS No.: Q1370 SDG NO.: Q1370
 Lab File ID: VX044976.D Date Analyzed: 02/18/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:45
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	88801	12.018				
UPPER LIMIT	177602	12.518				
LOWER LIMIT	44400.5	11.518				
EPA SAMPLE NO.						
Storage-Blank-SOIL-REF-6	67038	12.02				
Storage-Blank-WATER-REF-5	65829	12.02				
Storage-Blank-WATER-REF-4	63533	12.02				
Storage-Blank-SAMPLE-MANAGEMENT	69682	12.02				
VX0218WBL01	91305	12.02				
VX0218WBS01	83521	12.02				
VX0218WBSD01	77597	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.