

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: CHEM02
Lab Code: CHEM Case No.: Q1370 SAS No.: Q1370 SDG NO.: Q1370
Lab File ID: VX044867.D BFB Injection Date: 02/10/2025
Instrument ID: MSVOA_X BFB Injection Time: 09:35
GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	19.7
75	30.0 - 60.0% of mass 95	52.3
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.9
173	Less than 2.0% of mass 174	0.5 (0.6) 1
174	50.0 - 100.0% of mass 95	75.9
175	5.0 - 9.0% of mass 174	5.7 (7.5) 1
176	95.0 - 101.0% of mass 174	72.6 (95.7) 1
177	5.0 - 9.0% of mass 176	4.5 (6.2) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC001	VSTDIC001	VX044868.D	02/10/2025	10:25
VSTDIC005	VSTDIC005	VX044869.D	02/10/2025	10:48
VSTDIC020	VSTDIC020	VX044870.D	02/10/2025	11:11
VSTDIC050	VSTDIC050	VX044871.D	02/10/2025	11:34
VSTDIC100	VSTDIC100	VX044872.D	02/10/2025	12:05
VSTDIC150	VSTDIC150	VX044873.D	02/10/2025	12:28