

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: GFEL01
 Lab Code: CHEM Case No.: Q1376 SAS No.: Q1376 SDG No.: Q1376
 Instrument ID: MSVOA_X Calibration Date(s): 02/10/2025 02/10/2025
 Heated Purge: (Y/N) N Calibration Time(s): 10:25 12:28
 GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:	RRF001 = VX044868.D			RRF005 = VX044869.D		RRF020 = VX044870.D		
	RRF050 = VX044871.D			RRF100 = VX044872.D		RRF150 = VX044873.D		
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
Dichlorodifluoromethane	0.723	0.669	0.716	0.700	0.706	0.693	0.701	2.7
Chloromethane	0.958	0.860	0.862	0.843	0.805	0.793	0.854	6.9
Vinyl Chloride	0.839	0.846	0.847	0.808	0.807	0.814	0.827	2.3
Bromomethane		0.248	0.252	0.249	0.242	0.246	0.247	1.4
Chloroethane	0.500	0.288	0.280	0.341	0.249	0.181	0.307	35.4
Trichlorofluoromethane	1.062	1.066	1.096	1.029	1.013	1.008	1.046	3.3
1,1,2-Trichlorotrifluoroethane	0.583	0.647	0.668	0.626	0.631	0.637	0.632	4.5
1,1-Dichloroethene	0.647	0.639	0.661	0.630	0.632	0.657	0.644	2
Acetone	0.305	0.292	0.298	0.293	0.285	0.292	0.294	2.3
Carbon Disulfide	1.689	1.732	1.786	1.762	1.789	1.846	1.767	3
Methyl tert-butyl Ether	1.941	2.065	2.130	2.046	2.011	2.110	2.050	3.4
Methyl Acetate	0.882	0.901	0.926	0.946	0.922	0.995	0.928	4.2
Methylene Chloride	0.747	0.717	0.741	0.704	0.695	0.720	0.721	2.8
trans-1,2-Dichloroethene	0.608	0.622	0.657	0.640	0.633	0.644	0.634	2.7
1,1-Dichloroethane	1.155	1.257	1.292	1.227	1.209	1.257	1.233	3.9
Cyclohexane		1.154	1.174	1.121	1.107	1.127	1.137	2.4
2-Butanone	0.422	0.472	0.504	0.506	0.477	0.487	0.478	6.4
Carbon Tetrachloride	0.457	0.466	0.478	0.453	0.445	0.459	0.460	2.4
cis-1,2-Dichloroethene	0.680	0.783	0.812	0.758	0.758	0.779	0.762	5.9
Bromochloromethane	0.634	0.579	0.607	0.584	0.572	0.579	0.593	4
Chloroform	1.167	1.209	1.268	1.169	1.153	1.208	1.196	3.5
1,1,1-Trichloroethane	1.014	1.003	1.051	1.005	0.984	1.028	1.014	2.3
Methylcyclohexane	0.509	0.571	0.667	0.622	0.634	0.635	0.606	9.4
Benzene	1.370	1.488	1.577	1.470	1.429	1.453	1.465	4.7
1,2-Dichloroethane	0.417	0.465	0.502	0.472	0.462	0.482	0.467	6.1
Trichloroethene	0.293	0.340	0.367	0.335	0.332	0.343	0.335	7.2
1,2-Dichloropropane	0.343	0.354	0.389	0.367	0.360	0.372	0.364	4.3
Bromodichloromethane	0.428	0.481	0.514	0.500	0.500	0.513	0.489	6.6
4-Methyl-2-Pentanone	0.439	0.514	0.562	0.554	0.506	0.498	0.512	8.6
Toluene	0.776	0.872	0.957	0.898	0.866	0.864	0.872	6.7

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: GFEL01
 Lab Code: CHEM Case No.: Q1376 SAS No.: Q1376 SDG No.: Q1376
 Instrument ID: MSVOA_X Calibration Date(s): 02/10/2025 02/10/2025
 Heated Purge: (Y/N) N Calibration Time(s): 10:25 12:28
 GC Column: DB-624UI ID: 0.18 (mm)

LAB FILE ID:		RRF001 = VX044868.D		RRF005 = VX044869.D		RRF020 = VX044870.D			
		RRF050 = VX044871.D		RRF100 = VX044872.D		RRF150 = VX044873.D			
COMPOUND	RRF001	RRF005	RRF020	RRF050	RRF100	RRF150	RRF	% RSD	
t-1,3-Dichloropropene	0.417	0.451	0.518	0.514	0.528	0.543	0.495	10	
cis-1,3-Dichloropropene	0.452	0.511	0.587	0.577	0.587	0.599	0.552	10.5	
1,1,2-Trichloroethane	0.307	0.342	0.362	0.341	0.331	0.329	0.335	5.4	
2-Hexanone	0.313	0.360	0.406	0.404	0.369	0.362	0.369	9.3	
Dibromochloromethane	0.317	0.342	0.381	0.373	0.368	0.370	0.359	6.7	
1,2-Dibromoethane	0.302	0.330	0.367	0.350	0.345	0.345	0.340	6.5	
Tetrachloroethene	0.306	0.311	0.343	0.310	0.307	0.314	0.315	4.4	
Chlorobenzene	0.969	1.093	1.140	1.096	1.071	1.076	1.074	5.3	
Ethyl Benzene	1.690	1.873	2.021	1.935	1.923	1.929	1.895	5.9	
m/p-Xylenes	0.616	0.700	0.754	0.724	0.706	0.694	0.699	6.6	
o-Xylene	0.661	0.721	0.747	0.707	0.691	0.681	0.701	4.4	
Styrene	0.909	1.124	1.249	1.199	1.161	1.139	1.130	10.4	
Bromoform	0.186	0.247	0.272	0.280	0.276	0.287	0.258	14.7	
Isopropylbenzene	3.735	4.012	4.347	4.045	3.940	4.076	4.026	4.9	
1,1,2,2-Tetrachloroethane	1.429	1.403	1.438	1.366	1.305	1.360	1.383	3.6	
1,3-Dichlorobenzene	1.616	1.669	1.741	1.679	1.663	1.703	1.678	2.5	
1,4-Dichlorobenzene	1.662	1.712	1.762	1.686	1.660	1.701	1.697	2.2	
1,2-Dichlorobenzene	1.512	1.713	1.763	1.666	1.604	1.639	1.650	5.3	
1,2-Dibromo-3-Chloropropane	0.202	0.236	0.268	0.258	0.261	0.289	0.252	11.9	
1,2,4-Trichlorobenzene	0.860	0.934	1.013	1.013	1.057	1.112	0.998	9	
1,2,3-Trichlorobenzene	0.858	0.952	1.042	1.031	1.043	1.109	1.006	8.7	
1,2-Dichloroethane-d4		0.764	0.718	0.723	0.707	0.747	0.732	3.2	
Dibromofluoromethane		0.335	0.322	0.320	0.320	0.328	0.325	2	
Toluene-d8		1.239	1.249	1.239	1.208	1.212	1.229	1.5	
4-Bromofluorobenzene		0.404	0.410	0.431	0.415	0.412	0.414	2.5	

* Compounds with required minimum RRF and maximum %RSD values.
 All other compounds must meet a minimum RRF of 0.010.
 RRF of 1,4-Dioxane = Value should be divide by 1000.