

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: GFEL01

Lab Code: CHEM Case No.: Q1376 SAS No.: Q1376 SDG No.: Q1376

Instrument ID: MSVOA_X Calibration Date/Time: 02/19/2025 10:52

Lab File ID: VX044996.D Init. Calib. Date(s): 02/10/2025 02/10/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 10:25 12:28

GC Column: DB-624UI ID: 0.18 (mm)

COMPOUND	Amount	Conc.	%Dev	MAX%D
Chloroethane	50.000	97.75	-95.506	20