

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1376

Client: GFE LLC

Analytical Method: SW8260-Low

Datafile : VX044979.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VX0218WBS01	Dichlorodifluoromethane	20	19.1	ug/L	96			69	116	
	Chloromethane	20	18.9	ug/L	95			65	116	
	Vinyl chloride	20	18.8	ug/L	94			65	117	
	Bromomethane	20	30.4	ug/L	152		*	58	125	
	Chloroethane	20	27.2	ug/L	136		*	56	128	
	Trichlorofluoromethane	20	20.7	ug/L	104			73	115	
	1,1,2-Trichlorotrifluoroethane	20	20.4	ug/L	102			80	112	
	1,1-Dichloroethene	20	19.6	ug/L	98			74	110	
	Acetone	100	110	ug/L	110			60	125	
	Carbon disulfide	20	18.8	ug/L	94			64	112	
	Methyl tert-butyl Ether	20	20.3	ug/L	102			78	114	
	Methyl Acetate	20	21.8	ug/L	109			67	125	
	Methylene Chloride	20	20.0	ug/L	100			72	114	
	trans-1,2-Dichloroethene	20	20.3	ug/L	102			75	108	
	1,1-Dichloroethane	20	20.2	ug/L	101			78	112	
	Cyclohexane	20	19.5	ug/L	98			75	110	
	2-Butanone	100	110	ug/L	110			65	122	
	Carbon Tetrachloride	20	20.6	ug/L	103			77	113	
	cis-1,2-Dichloroethene	20	20.1	ug/L	101			77	110	
	Bromochloromethane	20	21.5	ug/L	108			70	124	
	Chloroform	20	20.6	ug/L	103			79	113	
	1,1,1-Trichloroethane	20	20.4	ug/L	102			80	108	
	Methylcyclohexane	20	20.0	ug/L	100			72	115	
	Benzene	20	19.9	ug/L	100			82	109	
	1,2-Dichloroethane	20	21.3	ug/L	106			80	115	
	Trichloroethene	20	20.2	ug/L	101			77	113	
	1,2-Dichloropropane	20	20.3	ug/L	102			83	111	
	Bromodichloromethane	20	21.5	ug/L	108			83	110	
	4-Methyl-2-Pentanone	100	110	ug/L	110			74	118	
	Toluene	20	20.5	ug/L	103			82	110	
	t-1,3-Dichloropropene	20	19.6	ug/L	98			79	110	
	cis-1,3-Dichloropropene	20	20.5	ug/L	103			82	110	
	1,1,2-Trichloroethane	20	20.6	ug/L	103			83	112	
	2-Hexanone	100	110	ug/L	110			73	117	
	Dibromochloromethane	20	21.3	ug/L	106			82	110	
	1,2-Dibromoethane	20	20.5	ug/L	103			81	110	
	Tetrachloroethene	20	21.0	ug/L	105			67	123	
	Chlorobenzene	20	20.6	ug/L	103			82	109	
	Ethyl Benzene	20	20.1	ug/L	101			83	109	
	m/p-Xylenes	40	40.9	ug/L	102			82	110	
	o-Xylene	20	20.3	ug/L	102			83	109	
	Styrene	20	21.0	ug/L	105			80	111	
	Bromoform	20	21.4	ug/L	107			79	109	
	Isopropylbenzene	20	19.4	ug/L	97			83	112	
	1,1,2,2-Tetrachloroethane	20	19.2	ug/L	96			76	118	
	1,3-Dichlorobenzene	20	20.2	ug/L	101			82	108	
	1,4-Dichlorobenzene	20	20.0	ug/L	100			82	107	
	1,2-Dichlorobenzene	20	20.8	ug/L	104			82	109	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1376

Client: GFE LLC

Analytical Method: SW8260-Low

Datafile : VX044979.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits	RPD
									High	
VX0218WBS01	1,2-Dibromo-3-Chloropropane	20	19.3	ug/L	97			68	112	
	1,2,4-Trichlorobenzene	20	20.2	ug/L	101			75	113	
	1,2,3-Trichlorobenzene	20	20.1	ug/L	101			76	114	