

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GFEL01
 Lab Code: CHEM Case No.: Q1376 SAS No.: Q1376 SDG NO.: Q1376
 Lab File ID: VX044976.D Date Analyzed: 02/18/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:45
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	118005	5.54	207185	6.75	183597	10.05
UPPER LIMIT	236010	6.038	414370	7.251	367194	10.549
LOWER LIMIT	59002.5	5.038	103593	6.251	91798.5	9.549
EPA SAMPLE NO.						
MW32	91141	5.54	179653	6.76	166286	10.05
MW33	81057	5.54	162106	6.76	147456	10.05
VX0218WBL01	92895	5.55	191012	6.76	185140	10.05
VX0218WBS01	109899	5.54	202747	6.76	177300	10.05
VX0218WBSD01	102760	5.54	182711	6.75	163817	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.