

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GFEL01
 Lab Code: CHEM Case No.: Q1376 SAS No.: Q1376 SDG NO.: Q1376
 Lab File ID: VX044996.D Date Analyzed: 02/19/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:52
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	101434	5.54	178680	6.75	161969	10.05
UPPER LIMIT	202868	6.038	357360	7.251	323938	10.549
LOWER LIMIT	50717	5.038	89340	6.251	80984.5	9.549
EPA SAMPLE NO.						
MW32DL	87497	5.54	177846	6.76	162028	10.05
VX0219WBL01	98956	5.54	201046	6.76	186227	10.05
VX0219WBS01	108230	5.55	197953	6.75	168480	10.05

IS1 = Pentafluorobenzene
 IS2 = 1,4-Difluorobenzene
 IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.