

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: GFEL01
 Lab Code: CHEM Case No.: Q1376 SAS No.: Q1376 SDG NO.: Q1376
 Lab File ID: VX044976.D Date Analyzed: 02/18/2025
 Instrument ID: MSVOA_X Time Analyzed: 10:45
 GC Column: DB-624UI ID: 0.18 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	88801	12.018				
UPPER LIMIT	177602	12.518				
LOWER LIMIT	44400.5	11.518				
EPA SAMPLE NO.						
MW32	76215	12.02				
MW33	59734	12.02				
VX0218WBL01	91305	12.02				
VX0218WBS01	83521	12.02				
VX0218WBSD01	77597	12.02				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
 AREA LOWER LIMIT = -50% of internal standard area
 RT UPPER LIMIT = +0.50 minutes of internal standard RT
 RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
 * Values outside of QC limits.