

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1378 SAS No.: Q1378 SDG NO.: Q1378

Continuing Calib Date: 02/19/2025 Initial Calibration Date(s): 01/28/2025 01/28/2025

Continuing Calib Time: 15:10 Initial Calibration Time(s): 09:37 17:30

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	5.68	5.69	5.59	5.79	0.01
Aroclor-1016-2	(2)	5.70	5.71	5.61	5.81	0.01
Aroclor-1016-3	(3)	5.76	5.78	5.68	5.88	0.02
Aroclor-1016-4	(4)	5.86	5.87	5.77	5.97	0.01
Aroclor-1016-5	(5)	6.15	6.17	6.07	6.27	0.02
Aroclor-1260-1	(1)	7.27	7.29	7.19	7.39	0.02
Aroclor-1260-2	(2)	7.53	7.54	7.44	7.64	0.01
Aroclor-1260-3	(3)	7.88	7.90	7.80	8.00	0.02
Aroclor-1260-4	(4)	8.11	8.13	8.03	8.23	0.02
Aroclor-1260-5	(5)	8.43	8.45	8.35	8.55	0.02
Tetrachloro-m-xylene		4.52	4.54	4.44	4.64	0.02
Decachlorobiphenyl		10.25	10.28	10.18	10.38	0.03

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1378 SAS No.: Q1378 SDG NO.: Q1378

Continuing Calib Date: 02/19/2025 Initial Calibration Date(s): 01/28/2025 01/28/2025

Continuing Calib Time: 15:10 Initial Calibration Time(s): 09:37 17:30

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1016-1	(1)	4.92	4.93	4.83	5.03	0.01
Aroclor-1016-2	(2)	4.94	4.95	4.85	5.05	0.01
Aroclor-1016-3	(3)	5.12	5.13	5.03	5.23	0.01
Aroclor-1016-4	(4)	5.16	5.17	5.07	5.27	0.01
Aroclor-1016-5	(5)	5.37	5.39	5.29	5.49	0.02
Aroclor-1260-1	(1)	6.41	6.43	6.33	6.53	0.02
Aroclor-1260-2	(2)	6.60	6.62	6.52	6.72	0.02
Aroclor-1260-3	(3)	6.75	6.77	6.67	6.87	0.02
Aroclor-1260-4	(4)	7.23	7.24	7.14	7.34	0.01
Aroclor-1260-5	(5)	7.47	7.48	7.38	7.58	0.01
Tetrachloro-m-xylene		3.83	3.84	3.74	3.94	0.01
Decachlorobiphenyl		8.89	8.91	8.81	9.01	0.02