

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1379 SAS No.: Q1379 SDG NO.: Q1379

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL01 Date Analyzed: 02/18/2025

Lab Sample No.: AR1660CCC500 Data File : PP069808.D Time Analyzed: 12:38

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	5.678	5.592	5.792	506.230	500.000	1.2
Aroclor-1016-2	5.701	5.614	5.814	491.720	500.000	-1.7
Aroclor-1016-3	5.763	5.676	5.876	478.610	500.000	-4.3
Aroclor-1016-4	5.861	5.774	5.974	478.750	500.000	-4.3
Aroclor-1016-5	6.152	6.067	6.267	495.050	500.000	-1.0
Aroclor-1260-1	7.274	7.187	7.387	485.390	500.000	-2.9
Aroclor-1260-2	7.527	7.441	7.641	464.790	500.000	-7.0
Aroclor-1260-3	7.885	7.800	8.000	479.020	500.000	-4.2
Aroclor-1260-4	8.110	8.025	8.225	459.410	500.000	-8.1
Aroclor-1260-5	8.430	8.347	8.547	472.280	500.000	-5.5
Decachlorobiphenyl	10.253	10.179	10.379	49.080	50.000	-1.8
Tetrachloro-m-xylene	4.525	4.438	4.638	50.950	50.000	1.9

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1379 SAS No.: Q1379 SDG NO.: Q1379

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL01 Date Analyzed: 02/18/2025

Lab Sample No.: AR1660CCC500 Data File : PP069808.D Time Analyzed: 12:38

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.920	4.834	5.034	500.750	500.000	0.2
Aroclor-1016-2	4.939	4.853	5.053	514.360	500.000	2.9
Aroclor-1016-3	5.116	5.030	5.230	509.710	500.000	1.9
Aroclor-1016-4	5.158	5.072	5.272	504.520	500.000	0.9
Aroclor-1016-5	5.373	5.288	5.488	524.290	500.000	4.9
Aroclor-1260-1	6.412	6.326	6.526	526.270	500.000	5.3
Aroclor-1260-2	6.600	6.515	6.715	528.420	500.000	5.7
Aroclor-1260-3	6.754	6.670	6.870	468.810	500.000	-6.2
Aroclor-1260-4	7.227	7.142	7.342	518.410	500.000	3.7
Aroclor-1260-5	7.467	7.384	7.584	534.690	500.000	6.9
Decachlorobiphenyl	8.889	8.811	9.011	48.300	50.000	-3.4
Tetrachloro-m-xylene	3.829	3.740	3.940	51.520	50.000	3.0