

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1379 SAS No.: Q1379 SDG NO.: Q1379

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL02 Date Analyzed: 02/18/2025

Lab Sample No.: AR1660CCC500 Data File : PP069821.D Time Analyzed: 17:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.674	5.592	5.792	417.940	500.000	-16.4
Aroclor-1016-2	5.697	5.614	5.814	477.870	500.000	-4.4
Aroclor-1016-3	5.759	5.676	5.876	442.420	500.000	-11.5
Aroclor-1016-4	5.857	5.774	5.974	463.590	500.000	-7.3
Aroclor-1016-5	6.149	6.067	6.267	429.810	500.000	-14.0
Aroclor-1260-1	7.270	7.187	7.387	452.050	500.000	-9.6
Aroclor-1260-2	7.523	7.441	7.641	447.940	500.000	-10.4
Aroclor-1260-3	7.880	7.800	8.000	447.400	500.000	-10.5
Aroclor-1260-4	8.105	8.025	8.225	427.460	500.000	-14.5
Aroclor-1260-5	8.425	8.347	8.547	439.700	500.000	-12.1
Decachlorobiphenyl	10.250	10.179	10.379	45.550	50.000	-8.9
Tetrachloro-m-xylene	4.522	4.438	4.638	50.390	50.000	0.8

CALIBRATION VERIFICATION SUMMARY

Contract: ATCE02

Lab Code: CHEM Case No.: Q1379 SAS No.: Q1379 SDG NO.: Q1379

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 01/28/2025 01/28/2025

Client Sample No.: CCAL02 Date Analyzed: 02/18/2025

Lab Sample No.: AR1660CCC500 Data File : PP069821.D Time Analyzed: 17:46

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.919	4.834	5.034	468.230	500.000	-6.4
Aroclor-1016-2	4.939	4.853	5.053	479.740	500.000	-4.1
Aroclor-1016-3	5.116	5.030	5.230	498.020	500.000	-0.4
Aroclor-1016-4	5.158	5.072	5.272	487.710	500.000	-2.5
Aroclor-1016-5	5.373	5.288	5.488	520.430	500.000	4.1
Aroclor-1260-1	6.411	6.326	6.526	515.760	500.000	3.2
Aroclor-1260-2	6.599	6.515	6.715	520.500	500.000	4.1
Aroclor-1260-3	6.753	6.670	6.870	456.660	500.000	-8.7
Aroclor-1260-4	7.225	7.142	7.342	488.370	500.000	-2.3
Aroclor-1260-5	7.466	7.384	7.584	510.670	500.000	2.1
Decachlorobiphenyl	8.887	8.811	9.011	46.270	50.000	-7.5
Tetrachloro-m-xylene	3.829	3.740	3.940	50.990	50.000	2.0