

Surrogate Summary

SDG No.: Q1379

Client: ATC Group Services LLC

Analytical Method: 8082A

Lab Sample ID	Client ID	Parameter	Column	Spike	Result	Rec	Qual	Limits	
								Low	High
I.BLK-PP069264.D	PIBLK-PP069264.D	Tetrachloro-m-xylene	1	20	25.7	128		60	140
		Decachlorobiphenyl	1	20	25.4	127		60	140
		Tetrachloro-m-xylene	2	20	26.0	130		60	140
		Decachlorobiphenyl	2	20	25.4	127		60	140
I.BLK-PP069812.D	PIBLK-PP069812.D	Tetrachloro-m-xylene	1	20	21.8	109		60	140
		Decachlorobiphenyl	1	20	21.3	106		60	140
		Tetrachloro-m-xylene	2	20	22.0	110		60	140
		Decachlorobiphenyl	2	20	21.0	105		60	140
PB166750BS	PB166750BS	Tetrachloro-m-xylene	1	20	23.0	115		32	144
		Decachlorobiphenyl	1	20	22.2	111		32	175
		Tetrachloro-m-xylene	2	20	22.5	113		32	144
		Decachlorobiphenyl	2	20	21.5	108		32	175
I.BLK-PP069825.D	PIBLK-PP069825.D	Tetrachloro-m-xylene	1	20	22.2	111		60	140
		Decachlorobiphenyl	1	20	20.6	103		60	140
		Tetrachloro-m-xylene	2	20	21.9	110		60	140
		Decachlorobiphenyl	2	20	20.6	103		60	140
I.BLK-PP069831.D	PIBLK-PP069831.D	Tetrachloro-m-xylene	1	20	23.3	116		60	140
		Decachlorobiphenyl	1	20	22.6	113		60	140
		Tetrachloro-m-xylene	2	20	23.0	115		60	140
		Decachlorobiphenyl	2	20	21.9	109		60	140
PB166750BL	PB166750BL	Tetrachloro-m-xylene	1	20	24.6	123		32	144
		Decachlorobiphenyl	1	20	23.6	118		32	175
		Tetrachloro-m-xylene	2	20	24.3	121		32	144
		Decachlorobiphenyl	2	20	22.6	113		32	175
Q1379-01	1A-1B-1C-GRAY-CAULK	Tetrachloro-m-xylene	1	20	29.5	148	*	32	144
		Decachlorobiphenyl	1	20	25.6	128		32	175
		Tetrachloro-m-xylene	2	20	27.1	135		32	144
		Decachlorobiphenyl	2	20	25.3	126		32	175
I.BLK-PP069846.D	PIBLK-PP069846.D	Tetrachloro-m-xylene	1	20	23.1	116		60	140
		Decachlorobiphenyl	1	20	22.2	111		60	140
		Tetrachloro-m-xylene	2	20	22.9	115		60	140
		Decachlorobiphenyl	2	20	22.9	114		60	140