

**DATA PACKAGE
SEMI-VOLATILE ORGANICS**

PROJECT NAME : NWIRP Bethpage CTO WE13 - VPB-192 #112G08005

**TETRA TECH NUS, INC.
661 Andersen Drive
Suite 200
Pittsburgh, PA - 15220-2745
Phone No: 412-921-7090**

**ORDER ID : Q1380
ATTENTION : Ernie Wu**



Laboratory Certification ID # 20012



1) SEMI-VOLATILE DATA	2
2) Signature Page	4
3) Case Narrative	5
4) Qualifier Page	8
5) Conformance/Non Conformance	9
6) QA Checklist	12
7) Chronicle	13
8) Hit Summary	14
9) QC Data Summary For SVOC-SIMGroup1	15
9.1) Deuterated Monitoring Compound Summary	16
9.2) MS/MSD Summary	17
9.3) LCS/LCSD Summary	19
9.4) Method Blank Summary	20
9.5) GS/MS Tune Summary	21
9.6) Internal Standard Area and RT Summary	23
10) Sample Data	25
10.0.1) BP-VPB-192-EB-20250212	26
10.0.2) BP-VPB-192-GW-780-782	37
10.1) VPB192-HYD-20250214	48
11) Calibration Data Summary	59
11.1) Initial Calibration Data	60
11.1.1) BN021025	60
11.2) Continued Calibration Data	247
11.2.1) BN036473.D	247
11.2.2) BN036482.D	273
12) QC Sample Data	299
12.1) Tune Raw Data	300
12.2) Method Blank Data	308
12.3) LCS Data	318
12.4) MS Data	342
12.5) MSD Data	366
13) Manual Integration	390
14) Analytical Runlogs	392
15) Percent Solid	396
16) Extraction Logs	400

Table Of Contents for Q1380

16.1) PB166758.pdf	400
16.2) PB166758IC.pdf	402
17) Standard Prep Logs	403
18) Shipping Document	471
18.1) Chain Of Custody	472
18.2) Lab Certificate	473
18.3) Internal COC	474
19) Not Reviewed Data	476

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

Cover Page

Order ID : Q1380

Project ID : NWIRP Bethpage CTO WE13 - VPB-192 #112G08005

Client : Tetra Tech NUS, Inc.

Lab Sample Number

Q1380-01
Q1380-02
Q1380-06
Q1380-07
Q1380-08
Q1380-09
Q1380-10
Q1380-11
Q1380-12

Client Sample Number

BP-VPB-192-EB-20250212
BP-VPB-192-TB-20250210
BP-VPB-192-GW-725-727
BP-VPB-192-GW-725-727MS
BP-VPB-192-GW-725-727MSD
BP-VPB-192-GW-780-782
BP-VPB-192-GW-780-782MS
BP-VPB-192-GW-780-782MSD
VPB192-HYD-20250214

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the laboratory manager or his designee, as verified by the following signature.

Signature :

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 12:33 pm, Mar 04, 2025

Date: 2/26/2025

NYDOH CERTIFICATION NO - 11376

NJDEP CERTIFICATION NO - 20012

CASE NARRATIVE

Tetra Tech NUS, Inc.

Project Name: NWIRP Bethpage CTO WE13 - VPB-192 #112G08005

Project Manager: Ernie Wu

Chemtech Project # Q1380

Test Name: SVOC-SIMGroup1

A. Number of Samples and Date of Receipt:

3 Solid samples were received on 02/17/2025.

6 Water samples were received on 02/17/2025.

B. Parameters

According to the Chain of Custody document, the following analyses were requested: SVOC-SIMGroup1 and VOCMS Group1. This data package contains results for SVOC-SIMGroup1.

C. Analytical Techniques:

The samples were analyzed on instrument BNA_N using GC Column ZB-SemiVolatiles Guardian which is 30 meters, 0.25 mm ID, 0.5 um df, Catalog # 7HG-G027-17-GGAThe analysis of SVOC-SIMGroup1 was based on method 8270-Modified and extraction was done based on method 3510.

D. QA/ QC Samples:

The Holding Times were met for all analysis.

The Surrogate recoveries met the acceptable criteria except for PB166758BS [2-Fluorobiphenyl - 118%], BP-VPB-192-EB-20250212 [Terphenyl-d14 - 146%], BP-VPB-192-GW-780-782 [2-Fluorobiphenyl - 118%], BP-VPB-192-GW-780-782MS [2-Fluorobiphenyl - 110%, Terphenyl-d14 - 135%], BP-VPB-192-GW-780-782MSD [2-Fluorobiphenyl - 113%, Terphenyl-d14 - 138%], failure surrogates are not associated with the client list, as per criteria affected surrogates were passing, therefore no corrective action was taken while, VPB192-HYD-20250214 [2-Methylnaphthalene-d10 - 6%] surrogate failing biased low also Due to the limited volume of this sample, it will not be re-extracted therefore no corrective action taken.

The Internal Standards Areas met the acceptable requirements.

The Retention Times were acceptable for all samples.

The MS {Q1380-10MS} with File ID: BN036477.D recoveries met the requirements for all compounds except for 1,4-Dioxane[64%] due to matrix interference.

The MSD recoveries met the acceptable requirements .

The sample # BP-VPB-192-GW-780-782MS is failing for 1,4-Dioxane and the original sample(BP-VPB-192-GW-780-782) is reported with M flag for this compound.

The RPD met criteria .

The Blank Spike met requirements for all samples .

The Blank analysis did not indicate the presence of lab contamination.

The Initial Calibration met the requirements .

The Continuous Calibration met the requirements .

The Tuning criteria met requirements.

E. Additional Comments:

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is).

Less volume was taken for samples #BP-VPB-192-GW-780-782, BP-VPB-192-GW-780-782MS,BP-VPB-192-GW-780-782MSD at the time of extraction due to Limited volume received.

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

F. Manual Integration Comments:

Please refer to the Manual integration Report included with the Run Logs for information on the manual integrations performed.

I certify that the data package is in compliance with the terms and conditions of the contract, both technically and for completeness, for other than the conditions detailed above. The laboratory manager or his designee, as verified by the following signature has authorized release of the data contained in this hard copy data package.

Signature_____

APPROVED

By Nimisha Pandya, QA/QC Supervisor at 12:35 pm, Mar 04, 2025



284 Sheffield Street, Mountainside, NJ 07092
Phone: 908 789 8900 Fax: 908 789 8922

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18
- 19

DATA REPORTING QUALIFIERS- ORGANIC

For reporting results, the following “ Results Qualifiers” are used:

Value	If the result is a value greater than or equal to the detection limit, report the value
U	Indicates the compound was analyzed for but was not detected. Report the minimum detection limit for the sample with the U, i.e. “10 U”. This is not necessarily the instrument detection limit attainable for this particular sample based on any concentration or dilution that may have been required.
ND	Indicates the analyte was analyzed for, but not detected
J	Indicates an estimated value. This flag is used: (1) When estimating a concentration for a tentatively identified compound (library search hits, where a 1:1 response is assumed.) (2) When the mass spectral data indicated the identification, however the result was less than the specified detection limit greater than zero. If the detection limit was 10ug/L and a concentration of 3 ug/L was calculated report as 3 J. This flag is used when similar situation arise on any organic parameter i.e. Pest, PCB and others.
B	Indicates the analyte was found in the blank as well as the sample report as “12 B”.
E	Indicates the analyte ‘s concentration exceeds the calibrated range of the instrument for that specific analysis.
D	This flag identifies all compounds identified in an analysis at a secondary dilution factor.
P	This flag is used for Pesticide/PCB target analyte when there is >25% difference for detected concentrations between the two GC columns. The lower of the two values is reported on Form 1 and flagged with a “P”.
N	This flag indicates presumptive evidence of a compound. This is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It applies to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the flag is not used.
A	This flag indicates that a Tentatively Identified Compound is a suspected aldol-condensation product.
Q	Indicates the LCS did not meet the control limits requirements

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: Q1380

MATRIX: /Water

METHOD: 8270-Modified/3510

		NA	NO	YES
1.	Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)			✓
2.	GC/MS Tuning Specifications. DFTPP Meet Criteria. (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)			✓
3.	GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 Series.			✓
4.	GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series.			✓
5.	GC/MS Calibration Requirements.			✓
	The Initial Calibration met the requirements .			
	The Continuous Calibration met the requirements .			
6.	Blank Contamination - If yes, list compounds and concentrations in each blank:		✓	
7.	Surrogate Recoveries Meet Criteria		✓	
	If not met, list those compounds and their recoveries which fall outside the acceptable ranges.			
	The Surrogate recoveries met the acceptable criteria except for PB166758BS [2- Fluorobiphenyl - 118%], BP-VPB-192-EB-20250212 [Terphenyl-d14 - 146%], BP- VPB-192-GW-780-782 [2-Fluorobiphenyl - 118%], BP-VPB-192-GW-780-782MS [2- Fluorobiphenyl - 110%, Terphenyl-d14 - 135%], BP-VPB-192-GW-780-782MSD [2- Fluorobiphenyl - 113%, Terphenyl-d14 - 138%], failure surrogates are not associated with the client list, as per criteria affected surrogates were passing, therefore no corrective action was taken while,			
	VPB192-HYD-20250214 [2-Methylnaphthalene-d10 - 6%]surrogate failing biased low also Due to the limited volume of this sample, it will not be re-extracted therefore no corrective action taken.			

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

(CONTINUED)

		NA	NO	YES
8.	Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria			✓
	If not met, list those compounds and their recoveries which fall outside the acceptable range.			
	The MS {Q1380-10MS} with File ID: BN036477.D recoveries met the requirements for all compounds except for 1,4-Dioxane[64%] due to matrix interference.			
	The MSD recoveries met the acceptable requirements .			
	The sample # BP-VPB-192-GW-780-782MS is failing for 1,4-Dioxane and the original sample(BP-VPB-192-GW-780-782) is reported with M flag for this compound..			
	The Blank Spike met requirements for all samples .			
9.	Internal Standard Area/Retention Time Shift Meet Criteria			✓
	Comments:			
10.	Extraction Holding Time Met			✓
	If not met, list number of days exceeded for each sample:			
11.	Analysis Holding Time Met			✓
	If not met, list number of days exceeded for each sample:			

ADDITIONAL COMMENTS:

The laboratory certifies that the all-electronic diskette deliverable exactly match the data summary forms (i.e. Form Is).

Less volume was taken for samples #BP-VPB-192-GW-780-782, BP-VPB-192-GW-780-782MS,BP-VPB-192-GW-780-782MSD at the time of extraction due to Limited volume received.

The Form 6 is not included in the data package because the Initial Calibration was performed using 7 points.

The not QT review data is reported in the Miscellaneous.

Please use %D calculated based on Avg RF and CCRF for all compounds using Average Response Factor when the %RSD value for a compound is <20% for the Initial Calibration curve and use %D calculated based on Amount added and Calculated amount for all compounds using Linear Regression when the %RSD value for a compound is > 20% for the Initial Calibration curve for SW-846 analysis.

REVIEWED

QA REVIEW

By *Sohil Jodhani*, QA/QC Director at 12:13 pm, Mar 04, 2025

ALLIANCE 284 Sheffield Street, Mountainside New Jersey 07092

NEW JERSEY LAB ID#: 20012: NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

(CONTINUED)

NA NO YES

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18
- 19

APPENDIX A

QA REVIEW GENERAL DOCUMENTATION

Project #: Q1380

Completed

For thorough review, the report must have the following:

GENERAL:

Are all original paperwork present (chain of custody, record of communication,airbill, sample management lab chronicle, login page)

✓

Check chain-of-custody for proper relinquish/return of samples

✓

Is the chain of custody signed and complete

✓

Check internal chain-of-custody for proper relinquish/return of samples /sample extracts

✓

Collect information for each project id from server. Were all requirements followed

✓

COVER PAGE:

Do numbers of samples correspond to the number of samples in the Chain of Custody on login page

✓

Do lab numbers and client Ids on cover page agree with the Chain of Custody

✓

CHAIN OF CUSTODY:

Do requested analyses on Chain of Custody agree with form I results

✓

Do requested analyses on Chain of Custody agree with the log-in page

✓

Were the correct method log-in for analysis according to the Analytical Request and Chain of Custody

✓

Were the samples received within hold time

✓

Were any problems found with the samples at arrival recorded in the Sample Management Laboratory Chronicle

✓

ANALYTICAL:

Was method requirement followed?

✓

Was client requirement followed?

✓

Does the case narrative summarize all QC failure?

✓

All runlogs and manual integration are reviewed for requirements

✓

All manual calculations and /or hand notations verified

✓

QA Review Signature: SOHIL JODHANI

Date: 02/26/2025

LAB CHRONICLE

OrderID:	Q1380	OrderDate:	2/17/2025 3:59:00 PM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005
Contact:	Ernie Wu	Location:	H31,VOA Ref. #2 Soil,VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1380-01	BP-VPB-192-EB-2025 0212	Water			02/12/25			02/17/25
			SVOC-SIMGroup1	8270-Modified		02/18/25	02/20/25	
Q1380-09	BP-VPB-192-GW-780- 782	Water			02/13/25			02/17/25
			SVOC-SIMGroup1	8270-Modified		02/18/25	02/20/25	
Q1380-12	VPB192-HYD-202502 14	Water			02/14/25			02/17/25
			SVOC-SIMGroup1	8270-Modified		02/18/25	02/20/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet SW-846

SDG No.: Q1380

Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID :				0.000					
Total Svoc :					0.00				
Total Concentration:					0.00				



QC SUMMARY

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18
- 19

Surrogate Summary

SW-846

SDG No.: Q1380

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB166758BL	PB166758BL	2-Methylnaphthalene-d10	0.4	0.35	87		30	150
		Fluoranthene-d10	0.4	0.42	104		30	150
		Nitrobenzene-d5	0.4	0.37	93		55	111
		2-Fluorobiphenyl	0.4	0.35	88		53	106
		Terphenyl-d14	0.4	0.42	104		58	132
PB166758BS	PB166758BS	2-Methylnaphthalene-d10	0.4	0.52	129		30	150
		Fluoranthene-d10	0.4	0.41	101		30	150
		Nitrobenzene-d5	0.4	0.39	97		55	111
		2-Fluorobiphenyl	0.4	0.47	118	*	53	106
		Terphenyl-d14	0.4	0.43	108		58	132
Q1380-01	BP-VPB-192-EB-20250212	2-Methylnaphthalene-d10	0.4	0.31	76		30	150
		Fluoranthene-d10	0.4	0.43	108		30	150
		Nitrobenzene-d5	0.4	0.30	76		55	111
		2-Fluorobiphenyl	0.4	0.38	96		53	106
		Terphenyl-d14	0.4	0.58	146	*	58	132
Q1380-09	BP-VPB-192-GW-780-782	2-Methylnaphthalene-d10	0.4	0.35	88		30	150
		Fluoranthene-d10	0.4	0.41	103		30	150
		Nitrobenzene-d5	0.4	0.36	91		55	111
		2-Fluorobiphenyl	0.4	0.47	118	*	53	106
		Terphenyl-d14	0.4	0.50	126		58	132
Q1380-10MS	BP-VPB-192-GW-780-782MS	2-Methylnaphthalene-d10	0.4	0.32	80		30	150
		Fluoranthene-d10	0.4	0.45	112		30	150
		Nitrobenzene-d5	0.4	0.32	80		55	111
		2-Fluorobiphenyl	0.4	0.44	110	*	53	106
		Terphenyl-d14	0.4	0.54	135	*	58	132
Q1380-11MSD	BP-VPB-192-GW-780-782MSD	2-Methylnaphthalene-d10	0.4	0.32	80		30	150
		Fluoranthene-d10	0.4	0.47	116		30	150
		Nitrobenzene-d5	0.4	0.32	80		55	111
		2-Fluorobiphenyl	0.4	0.45	113	*	53	106
		Terphenyl-d14	0.4	0.55	138	*	58	132
Q1380-12	VPB192-HYD-20250214	2-Methylnaphthalene-d10	0.4	0.023	6	*	30	150
		Fluoranthene-d10	0.4	0.34	86		30	150
		Nitrobenzene-d5	0.4	0.31	79		55	111
		2-Fluorobiphenyl	0.4	0.39	97		53	106
		Terphenyl-d14	0.4	0.45	112		58	132

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1380

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8270-Modified

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1380-10MS	Client Sample ID:	BP-VPB-192-GW-780-782MS					DataFile:	BN036477.D		
1,4-Dioxane	1.3	0	0.83	ug/L	64	*			70	130	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1380

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8270-Modified

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1380-11MSD	Client Sample ID:	BP-VPB-192-GW-780-782MSD					DataFile:	BN036478.D		
1,4-Dioxane	1.3	0	0.92	ug/L	71		10		70	130	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1380

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified DataFile: BN036480.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB166758BS	1,4-Dioxane	0.4	0.35	ug/L	88				70	130	

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB166758BL

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM Case No.: Q1380

SAS No.: Q1380 SDG NO.: Q1380

Lab File ID: BN036474.D

Lab Sample ID: PB166758BL

Instrument ID: BNA_N

Date Extracted: 02/18/2025

Matrix: (soil/water) Water

Date Analyzed: 02/20/2025

Level: (low/med) LOW

Time Analyzed: 10:47

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB166758BS	PB166758BS	BN036480.D	02/20/2025
BP-VPB-192-EB-20250212	Q1380-01	BN036475.D	02/20/2025
BP-VPB-192-GW-780-782	Q1380-09	BN036476.D	02/20/2025
BP-VPB-192-GW-780-782MS	Q1380-10MS	BN036477.D	02/20/2025
BP-VPB-192-GW-780-782MSD	Q1380-11MSD	BN036478.D	02/20/2025
VPB192-HYD-20250214	Q1380-12	BN036479.D	02/20/2025

COMMENTS: _____



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECHContract: TETR06Lab Code: CHEMSAS No.: Q1380SDG NO.: Q1380Lab File ID: BN036408.DDFTPP Injection Date: 02/10/2025Instrument ID: BNA_NDFTPP Injection Time: 11:46

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	51.4
68	Less than 2.0% of mass 69	0.3 (0.7) 1
69	Mass 69 relative abundance	47.7
70	Less than 2.0% of mass 69	0.3 (0.6) 1
127	10.0 - 80.0% of mass 198	48.3
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
275	10.0 - 60.0% of mass 198	24.7
365	Greater than 1% of mass 198	3.3
441	Present, but less than mass 443	7.1
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	10.5 (20.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN036409.D	02/10/2025	12:25
SSTDICC0.2	SSTDICC0.2	BN036410.D	02/10/2025	13:01
SSTDICCC0.4	SSTDICCC0.4	BN036411.D	02/10/2025	13:36
SSTDICC0.8	SSTDICC0.8	BN036412.D	02/10/2025	14:12
SSTDICC1.6	SSTDICC1.6	BN036413.D	02/10/2025	14:48
SSTDICC3.2	SSTDICC3.2	BN036414.D	02/10/2025	15:24
SSTDICC5.0	SSTDICC5.0	BN036415.D	02/10/2025	16:00



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECHContract: TETR06Lab Code: CHEMSAS No.: Q1380SDG NO.: Q1380Lab File ID: BN036472.DDFTPP Injection Date: 02/20/2025Instrument ID: BNA_NDFTPP Injection Time: 09:32

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	57.4
68	Less than 2.0% of mass 69	0.8 (1.5) 1
69	Mass 69 relative abundance	52.1
70	Less than 2.0% of mass 69	0.3 (0.5) 1
127	10.0 - 80.0% of mass 198	50.3
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
275	10.0 - 60.0% of mass 198	24.3
365	Greater than 1% of mass 198	3.5
441	Present, but less than mass 443	8.4
442	Greater than 50% of mass 198	100
443	15.0 - 24.0% of mass 442	9.6 (19.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN036473.D	02/20/2025	10:11
PB166758BL	PB166758BL	BN036474.D	02/20/2025	10:47
BP-VPB-192-EB-20250212	Q1380-01	BN036475.D	02/20/2025	11:24
BP-VPB-192-GW-780-782	Q1380-09	BN036476.D	02/20/2025	12:00
BP-VPB-192-GW-780-782MS	Q1380-10MS	BN036477.D	02/20/2025	12:36
BP-VPB-192-GW-780-782MSD	Q1380-11MSD	BN036478.D	02/20/2025	13:12
VPB192-HYD-20250214	Q1380-12	BN036479.D	02/20/2025	13:48
PB166758BS	PB166758BS	BN036480.D	02/20/2025	14:24
SSTDCCC0.4EC	SSTDCCC0.4	BN036482.D	02/20/2025	16:08

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG NO.: Q1380

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 02/20/2025

Lab File ID: BN036473.D Time Analyzed: 10:11

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	1766	7.746	4292	10.54	2899	14.39
UPPER LIMIT	3532	8.246	8584	11.041	5798	14.887
LOWER LIMIT	883	7.246	2146	10.041	1449.5	13.887
EPA SAMPLE NO.						
01 PB166758BL	1874	7.75	3905	10.55	2305	14.40
02 BP-VPB-192-EB-20250212	1491	7.75	3323	10.54	2119	14.39
03 PB166758BS	1827	7.75	4127	10.54	2357	14.39
04 BP-VPB-192-GW-780-782	1676	7.75	3840	10.54	2412	14.39
05 BP-VPB-192-GW-780-782MS	1709	7.75	3995	10.54	2495	14.39
06 BP-VPB-192-GW-780-782MSD	1787	7.75	4250	10.53	2636	14.39
07 VPB192-HYD-20250214	1727	7.75	4024	10.54	2593	14.39

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG NO.: Q1380

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 02/20/2025

Lab File ID: BN036473.D Time Analyzed: 10:11

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	6167	17.136	5687	21.313	5826	23.586
UPPER LIMIT	12334	17.636	11374	21.813	11652	24.086
LOWER LIMIT	3083.5	16.636	2843.5	20.813	2913	23.086
EPA SAMPLE NO.						
01 PB166758BL	4889	17.15	4485	21.33	3907	23.60
02 BP-VPB-192-EB-20250212	4759	17.14	4677	21.32	4457	23.59
03 PB166758BS	5256	17.14	4823	21.32	4362	23.59
04 BP-VPB-192-GW-780-782	5513	17.12	5354	21.32	4859	23.59
05 BP-VPB-192-GW-780-782MS	5565	17.12	5523	21.31	5300	23.59
06 BP-VPB-192-GW-780-782MSD	5810	17.12	5865	21.31	5479	23.58
07 VPB192-HYD-20250214	5756	17.14	5466	21.32	4501	23.59

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



SAMPLE DATA

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	02/12/25
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005	Date Received:	02/17/25
Client Sample ID:	BP-VPB-192-EB-20250212	SDG No.:	Q1380
Lab Sample ID:	Q1380-01	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	890 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036475.D	1	02/18/25 11:40	02/20/25 11:24	PB166758

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.22	U	0.080	0.22	0.22	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.31		30 - 150		76%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.43		30 - 150		108%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.30		55 - 111		76%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.38		53 - 106		96%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.58	*	58 - 132		146%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1490		7.746			
1146-65-2	Naphthalene-d8	3320		10.541			
15067-26-2	Acenaphthene-d10	2120		14.388			
1517-22-2	Phenanthrene-d10	4760		17.136			
1719-03-5	Chrysene-d12	4680		21.322			
1520-96-3	Perylene-d12	4460		23.592			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
Data File : BN036475.D
Acq On : 20 Feb 2025 11:24
Operator : RC/JU
Sample : Q1380-01
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-EB-20250212

Quant Time: Feb 21 00:09:41 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 01:17:14 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.746	152	1491	0.400	ng	0.00
7) Naphthalene-d8	10.541	136	3323	0.400	ng	0.00
13) Acenaphthene-d10	14.388	164	2119	0.400	ng	0.00
19) Phenanthrene-d10	17.136	188	4759	0.400	ng	0.00
29) Chrysene-d12	21.322	240	4677	0.400	ng	0.00
35) Perylene-d12	23.592	264	4457	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.341	112	488	0.138	ng	0.00
5) Phenol-d6	6.930	99	321	0.078	ng	0.00
8) Nitrobenzene-d5	8.907	82	997	0.304	ng	0.00
11) 2-Methylnaphthalene-d10	12.131	152	1557	0.305	ng	-0.01
14) 2,4,6-Tribromophenol	15.883	330	365	0.347	ng	0.00
15) 2-Fluorobiphenyl	13.009	172	3049	0.383	ng	-0.01
27) Fluoranthene-d10	19.164	212	5702	0.431	ng	0.00
31) Terphenyl-d14	19.764	244	5814	0.582	ng	0.00
Target Compounds						Qvalue
34) Bis(2-ethylhexyl)phtha...	21.241	149	1068	0.111	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
Data File : BN036475.D
Acq On : 20 Feb 2025 11:24
Operator : RC/JU
Sample : Q1380-01
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-EB-20250212

Quant Time: Feb 21 00:09:41 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 01:17:14 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.746 min Scan# 64
Delta R.T. -0.007 min
Lab File: BN036475.D
Acq: 20 Feb 2025 11:24




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-EB-20250212


Tgt Ion:152 Resp: 1491
Ion Ratio Lower Upper
152 100
150 148.5 123.7 185.5
115 64.8 52.5 78.7



#4
2-Fluorophenol
Concen: 0.138 ng
RT: 5.341 min Scan# 313
Delta R.T. -0.007 min
Lab File: BN036475.D
Acq: 20 Feb 2025 11:24

Tgt Ion:112 Resp: 488
Ion Ratio Lower Upper
112 100
64 67.4 53.4 80.0
63 40.8 30.3 45.5









#5
Phenol-d6
Concen: 0.078 ng
RT: 6.930 min Scan# 51
Delta R.T. -0.007 min
Lab File: BN036475.D
Acq: 20 Feb 2025 11:24




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-EB-20250212


Tgt Ion: 99 Resp: 321
Ion Ratio Lower Upper
99 100
42 27.7 21.7 32.5
71 53.3 32.6 49.0#



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.541 min Scan# 933
Delta R.T. 0.000 min
Lab File: BN036475.D
Acq: 20 Feb 2025 11:24

Tgt Ion: 136 Resp: 3323
Ion Ratio Lower Upper
136 100
137 13.0 10.1 15.1
54 13.1 11.8 17.6
68 9.0 7.2 10.8









#8
Nitrobenzene-d5
Concen: 0.304 ng
RT: 8.907 min Scan# 71
Delta R.T. 0.000 min
Lab File: BN036475.D
Acq: 20 Feb 2025 11:24




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-EB-20250212


Tgt Ion:	82	Resp:	997
Ion	Ratio	Lower	Upper
82	100		
128	46.0	31.9	47.9
54	66.7	53.1	79.7



#11
2-Methylnaphthalene-d10
Concen: 0.305 ng
RT: 12.131 min Scan# 1146
Delta R.T. -0.010 min
Lab File: BN036475.D
Acq: 20 Feb 2025 11:24

Tgt Ion:	152	Resp:	1557
Ion	Ratio	Lower	Upper
152	100		
151	21.6	16.6	25.0









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.388 min Scan# 1448
Delta R.T. 0.000 min
Lab File: BN036475.D
Acq: 20 Feb 2025 11:24




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-EB-20250212


Tgt Ion:	164	Resp:	2119
Ion	Ratio	Lower	Upper
164	100		
162	107.6	84.1	126.1
160	54.1	41.4	62.0



#14
2,4,6-Tribromophenol
Concen: 0.347 ng
RT: 15.883 min Scan# 1583
Delta R.T. 0.000 min
Lab File: BN036475.D
Acq: 20 Feb 2025 11:24

Tgt Ion:	330	Resp:	365
Ion	Ratio	Lower	Upper
330	100		
332	97.8	76.6	114.8
141	46.0	37.8	56.8









#15
2-Fluorobiphenyl
Concen: 0.383 ng
RT: 13.009 min Scan# 1319
Delta R.T. -0.011 min
Lab File: BN036475.D
Acq: 20 Feb 2025 11:24




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-EB-20250212


Tgt Ion:172	Resp:	3049
Ion Ratio	Lower	Upper
172	100	
171	37.6	29.6 44.4
170	26.8	19.8 29.6



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.136 min Scan# 1684
Delta R.T. 0.000 min
Lab File: BN036475.D
Acq: 20 Feb 2025 11:24

Tgt Ion:188	Resp:	4759
Ion Ratio	Lower	Upper
188	100	
94	0.0	0.0 0.0
80	11.6	9.8 14.6









#27
Fluoranthene-d10
Concen: 0.431 ng
RT: 19.164 min Scan# 1935
Delta R.T. -0.005 min
Lab File: BN036475.D
Acq: 20 Feb 2025 11:24




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-EB-20250212


Tgt Ion	Ratio	Lower	Upper
212	100		
106	15.5	11.5	17.3
104	8.8	7.1	10.7



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.322 min Scan# 2263
Delta R.T. 0.000 min
Lab File: BN036475.D
Acq: 20 Feb 2025 11:24

Tgt Ion	Ratio	Lower	Upper
240	100		
120	14.4	13.3	19.9
236	29.5	23.0	34.6









#31
Terphenyl-d14
Concen: 0.582 ng
RT: 19.764 min Scan# 2065
Delta R.T. -0.005 min
Lab File: BN036475.D
Acq: 20 Feb 2025 11:24




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-EB-20250212


Tgt Ion	Ratio	Lower	Upper
244	100		
212	10.6	8.1	12.1
122	15.4	12.8	19.2



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.111 ng
RT: 21.241 min Scan# 2254
Delta R.T. 0.000 min
Lab File: BN036475.D
Acq: 20 Feb 2025 11:24


Tgt Ion	Ratio	Lower	Upper
149	100		
167	25.8	21.2	31.8
279	3.0	2.7	4.1







#35
Perylene-d12
Concen: 0.400 ng
RT: 23.592 min Scan# 21
Delta R.T. -0.003 min
Lab File: BN036475.D
Acq: 20 Feb 2025 11:24

Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-EB-20250212



Tgt Ion:264 Resp: 4457
Ion Ratio Lower Upper
264 100
260 26.8 20.9 31.3
265 100.0 60.7 91.1#



Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	02/13/25
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005	Date Received:	02/17/25
Client Sample ID:	BP-VPB-192-GW-780-782	SDG No.:	Q1380
Lab Sample ID:	Q1380-09	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	380 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036476.D	1	02/18/25 11:40	02/20/25 12:00	PB166758

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.53	UM	0.18	0.53	0.53	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.35		30 - 150		88%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.41		30 - 150		103%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.36		55 - 111		91%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.47	*	53 - 106		118%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.50		58 - 132		126%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1680	7.746				
1146-65-2	Naphthalene-d8	3840	10.541				
15067-26-2	Acenaphthene-d10	2410	14.388				
1517-22-2	Phenanthrene-d10	5510	17.124				
1719-03-5	Chrysene-d12	5350	21.322				
1520-96-3	Perylene-d12	4860	23.59				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
Data File : BN036476.D
Acq On : 20 Feb 2025 12:00
Operator : RC/JU
Sample : Q1380-09
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782

Quant Time: Feb 21 00:10:00 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 01:17:14 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.746	152	1676	0.400	ng	0.00
7) Naphthalene-d8	10.541	136	3840	0.400	ng	# 0.00
13) Acenaphthene-d10	14.388	164	2412	0.400	ng	0.00
19) Phenanthrene-d10	17.124	188	5513	0.400	ng	#-0.01
29) Chrysene-d12	21.322	240	5354	0.400	ng	0.00
35) Perylene-d12	23.590	264	4859	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.341	112	978	0.247	ng	0.00
5) Phenol-d6	6.923	99	814	0.175	ng	-0.01
8) Nitrobenzene-d5	8.897	82	1381	0.364	ng	-0.01
11) 2-Methylnaphthalene-d10	12.131	152	2066	0.350	ng	-0.01
14) 2,4,6-Tribromophenol	15.883	330	429	0.359	ng	0.00
15) 2-Fluorobiphenyl	13.009	172	4271	0.471	ng	-0.01
27) Fluoranthene-d10	19.164	212	6346	0.414	ng	0.00
31) Terphenyl-d14	19.764	244	5755	0.504	ng	0.00
Target Compounds						Qvalue
34) Bis(2-ethylhexyl)phtha...	21.241	149	756	0.069	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
Data File : BN036476.D
Acq On : 20 Feb 2025 12:00
Operator : RC/JU
Sample : Q1380-09
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782


Quant Time: Feb 21 00:10:00 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 01:17:14 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.746 min Scan# 64
Delta R.T. -0.007 min
Lab File: BN036476.D
Acq: 20 Feb 2025 12:00

Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782







Tgt Ion:152 Resp: 1676
Ion Ratio Lower Upper
152 100
150 153.7 123.7 185.5
115 67.3 52.5 78.7



#4
2-Fluorophenol
Concen: 0.247 ng
RT: 5.341 min Scan# 313
Delta R.T. -0.007 min
Lab File: BN036476.D
Acq: 20 Feb 2025 12:00

Tgt Ion:112 Resp: 978
Ion Ratio Lower Upper
112 100
64 66.9 53.4 80.0
63 39.8 30.3 45.5









#5
Phenol-d6
Concen: 0.175 ng
RT: 6.923 min Scan# 511
Delta R.T. -0.014 min
Lab File: BN036476.D
Acq: 20 Feb 2025 12:00




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782


Tgt Ion: 99 Resp: 814
Ion Ratio Lower Upper
99 100
42 36.9 21.7 32.5#
71 49.8 32.6 49.0#



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.541 min Scan# 933
Delta R.T. 0.000 min
Lab File: BN036476.D
Acq: 20 Feb 2025 12:00

Tgt Ion: 136 Resp: 3840
Ion Ratio Lower Upper
136 100
137 13.0 10.1 15.1
54 11.7 11.8 17.6#
68 8.3 7.2 10.8









#8
Nitrobenzene-d5
Concen: 0.364 ng
RT: 8.897 min Scan# 71
Delta R.T. -0.011 min
Lab File: BN036476.D
Acq: 20 Feb 2025 12:00




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782


Tgt Ion: 82 Resp: 1381
Ion Ratio Lower Upper
82 100
128 38.6 31.9 47.9
54 69.6 53.1 79.7



#11
2-Methylnaphthalene-d10
Concen: 0.350 ng
RT: 12.131 min Scan# 1146
Delta R.T. -0.010 min
Lab File: BN036476.D
Acq: 20 Feb 2025 12:00

Tgt Ion: 152 Resp: 2066
Ion Ratio Lower Upper
152 100
151 20.5 16.6 25.0









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.388 min Scan# 1448
Delta R.T. 0.000 min
Lab File: BN036476.D
Acq: 20 Feb 2025 12:00




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782


Tgt Ion:	164	Resp:	2412
Ion	Ratio	Lower	Upper
164	100		
162	103.2	84.1	126.1
160	52.4	41.4	62.0



#14
2,4,6-Tribromophenol
Concen: 0.359 ng
RT: 15.883 min Scan# 1583
Delta R.T. 0.000 min
Lab File: BN036476.D
Acq: 20 Feb 2025 12:00

Tgt Ion:	330	Resp:	429
Ion	Ratio	Lower	Upper
330	100		
332	98.8	76.6	114.8
141	48.0	37.8	56.8









#15
2-Fluorobiphenyl
Concen: 0.471 ng
RT: 13.009 min Scan# 11
Delta R.T. -0.011 min
Lab File: BN036476.D
Acq: 20 Feb 2025 12:00




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782


Tgt Ion:	172	Resp:	4271
Ion Ratio	Lower	Upper	
172	100		
171	37.8	29.6	44.4
170	25.1	19.8	29.6



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.124 min Scan# 1683
Delta R.T. -0.012 min
Lab File: BN036476.D
Acq: 20 Feb 2025 12:00

Tgt Ion:	188	Resp:	5513
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	17.1	9.8	14.6#









#27
Fluoranthene-d10
Concen: 0.414 ng
RT: 19.164 min Scan# 1935
Delta R.T. -0.005 min
Lab File: BN036476.D
Acq: 20 Feb 2025 12:00




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782


Tgt Ion	Ratio	Lower	Upper
212	100		
106	15.0	11.5	17.3
104	9.6	7.1	10.7



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.322 min Scan# 2263
Delta R.T. 0.000 min
Lab File: BN036476.D
Acq: 20 Feb 2025 12:00


Tgt Ion	Ratio	Lower	Upper
240	100		
120	14.3	13.3	19.9
236	28.2	23.0	34.6








#31
Terphenyl-d14
Concen: 0.504 ng
RT: 19.764 min Scan# 2064
Delta R.T. -0.005 min
Lab File: BN036476.D
Acq: 20 Feb 2025 12:00

Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782







Tgt Ion	Ratio	Lower	Upper
244	100		
212	10.8	8.1	12.1
122	16.3	12.8	19.2



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.069 ng
RT: 21.241 min Scan# 2254
Delta R.T. 0.000 min
Lab File: BN036476.D
Acq: 20 Feb 2025 12:00

Tgt Ion	Ratio	Lower	Upper
149	100		
167	28.0	21.2	31.8
279	2.8	2.7	4.1





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.590 min Scan# 21

Delta R.T. -0.006 min

Lab File: BN036476.D


Acq: 20 Feb 2025 12:00

Instrument :

BNA_N

ClientSampleId :

BP-VPB-192-GW-780-782





Tgt Ion:264 Resp: 4859

Ion Ratio Lower Upper

264 100

260 26.6 20.9 31.3

265 95.5 60.7 91.1#



Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	02/14/25
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005	Date Received:	02/17/25
Client Sample ID:	VPB192-HYD-20250214	SDG No.:	Q1380
Lab Sample ID:	Q1380-12	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	940 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036479.D	1	02/18/25 11:40	02/20/25 13:48	PB166758

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.21	U	0.070	0.21	0.21	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.023	*	30 - 150		6%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.34		30 - 150		86%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.31		55 - 111		79%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.39		53 - 106		97%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.45		58 - 132		112%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1730	7.746				
1146-65-2	Naphthalene-d8	4020	10.541				
15067-26-2	Acenaphthene-d10	2590	14.388				
1517-22-2	Phenanthrene-d10	5760	17.136				
1719-03-5	Chrysene-d12	5470	21.322				
1520-96-3	Perylene-d12	4500	23.587				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
Data File : BN036479.D
Acq On : 20 Feb 2025 13:48
Operator : RC/JU
Sample : Q1380-12
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
VPB192-HYD-20250214

Quant Time: Feb 21 00:11:17 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 01:17:14 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.746	152	1727	0.400	ng	0.00
7) Naphthalene-d8	10.541	136	4024	0.400	ng	0.00
13) Acenaphthene-d10	14.388	164	2593	0.400	ng	0.00
19) Phenanthrene-d10	17.136	188	5756	0.400	ng	0.00
29) Chrysene-d12	21.322	240	5466	0.400	ng	0.00
35) Perylene-d12	23.587	264	4501	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.341	112	254	0.062	ng	0.00
5) Phenol-d6	6.937	99	13	0.003	ng	0.00
8) Nitrobenzene-d5	8.907	82	1245	0.314	ng	0.00
11) 2-Methylnaphthalene-d10	12.131	152	144	0.023	ng	-0.01
14) 2,4,6-Tribromophenol	15.882	330	390	0.303	ng	0.00
15) 2-Fluorobiphenyl	13.008	172	3787	0.388	ng	-0.01
27) Fluoranthene-d10	19.164	212	5514	0.344	ng	0.00
31) Terphenyl-d14	19.764	244	5234	0.449	ng	0.00
Target Compounds						Qvalue
34) Bis(2-ethylhexyl)phtha...	21.232	149	2546	0.227	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
Data File : BN036479.D
Acq On : 20 Feb 2025 13:48
Operator : RC/JU
Sample : Q1380-12
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
VPB192-HYD-20250214

Quant Time: Feb 21 00:11:17 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 01:17:14 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.746 min Scan# 64
Delta R.T. -0.007 min
Lab File: BN036479.D
Acq: 20 Feb 2025 13:48




Instrument :
BNA_N
ClientSampleId :
VPB192-HYD-20250214


Tgt Ion:152 Resp: 1727
Ion Ratio Lower Upper
152 100
150 152.3 123.7 185.5
115 67.7 52.5 78.7



#4
2-Fluorophenol
Concen: 0.062 ng
RT: 5.341 min Scan# 313
Delta R.T. -0.007 min
Lab File: BN036479.D
Acq: 20 Feb 2025 13:48

Tgt Ion:112 Resp: 254
Ion Ratio Lower Upper
112 100
64 66.5 53.4 80.0
63 38.2 30.3 45.5









#5
Phenol-d6
Concen: 0.003 ng
RT: 6.937 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN036479.D
Acq: 20 Feb 2025 13:48




Instrument :
BNA_N
ClientSampleId :
VPB192-HYD-20250214


Tgt Ion: 99 Resp: 13
Ion Ratio Lower Upper
99 100
42 53.8 21.7 32.5#
71 0.0 32.6 49.0#



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.541 min Scan# 933
Delta R.T. -0.000 min
Lab File: BN036479.D
Acq: 20 Feb 2025 13:48

Tgt Ion: 136 Resp: 4024
Ion Ratio Lower Upper
136 100
137 13.9 10.1 15.1
54 12.6 11.8 17.6
68 8.4 7.2 10.8









#8
Nitrobenzene-d5
Concen: 0.314 ng
RT: 8.907 min Scan# 71
Delta R.T. -0.000 min
Lab File: BN036479.D
Acq: 20 Feb 2025 13:48




Instrument :
BNA_N
ClientSampleId :
VPB192-HYD-20250214


Tgt Ion:	82	Resp:	1245
Ion	Ratio	Lower	Upper
82	100		
128	44.9	31.9	47.9
54	63.6	53.1	79.7



#11
2-Methylnaphthalene-d10
Concen: 0.023 ng
RT: 12.131 min Scan# 1146
Delta R.T. -0.010 min
Lab File: BN036479.D
Acq: 20 Feb 2025 13:48

Tgt Ion:	152	Resp:	144
Ion	Ratio	Lower	Upper
152	100		
151	26.4	16.6	25.0#









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.388 min Scan# 1448
Delta R.T. -0.000 min
Lab File: BN036479.D
Acq: 20 Feb 2025 13:48




Instrument :
BNA_N
ClientSampleId :
VPB192-HYD-20250214


Tgt Ion:164 Resp: 2593
Ion Ratio Lower Upper
164 100
162 105.5 84.1 126.1
160 51.8 41.4 62.0



#14
2,4,6-Tribromophenol
Concen: 0.303 ng
RT: 15.882 min Scan# 1583
Delta R.T. -0.000 min
Lab File: BN036479.D
Acq: 20 Feb 2025 13:48

Tgt Ion:330 Resp: 390
Ion Ratio Lower Upper
330 100
332 101.3 76.6 114.8
141 45.9 37.8 56.8









#15
2-Fluorobiphenyl
Concen: 0.388 ng
RT: 13.008 min Scan# 1319
Delta R.T. -0.011 min
Lab File: BN036479.D
Acq: 20 Feb 2025 13:48




Instrument :
BNA_N
ClientSampleId :
VPB192-HYD-20250214


Tgt Ion:	172	Resp:	3787
Ion	Ratio	Lower	Upper
172	100		
171	36.3	29.6	44.4
170	25.6	19.8	29.6



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.136 min Scan# 1684
Delta R.T. -0.000 min
Lab File: BN036479.D
Acq: 20 Feb 2025 13:48

Tgt Ion:	188	Resp:	5756
Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	11.8	9.8	14.6





#27

Fluoranthene-d10

Concen: 0.344 ng

RT: 19.164 min Scan# 1935

Delta R.T. -0.005 min

Lab File: BN036479.D

Acq: 20 Feb 2025 13:48

Instrument :

BNA_N

ClientSampleId :

VPB192-HYD-20250214





Tgt Ion:212 Resp: 5514

Ion Ratio Lower Upper

212 100

106 15.1 11.5 17.3

104 9.9 7.1 10.7



#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.322 min Scan# 2263

Delta R.T. -0.000 min

Lab File: BN036479.D

Acq: 20 Feb 2025 13:48




Tgt Ion:240 Resp: 5466


Ion Ratio Lower Upper

240 100

120 14.1 13.3 19.9

236 28.8 23.0 34.6









#31
 Terphenyl-d14
 Concen: 0.449 ng
 RT: 19.764 min Scan# 2064
 Delta R.T. -0.005 min
 Lab File: BN036479.D
 Acq: 20 Feb 2025 13:48




Instrument :
 BNA_N
 ClientSampleId :
 VPB192-HYD-20250214


Tgt	Ion:244	Resp:	5234
Ion	Ratio	Lower	Upper
244	100		
212	10.7	8.1	12.1
122	15.8	12.8	19.2



#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.227 ng
 RT: 21.232 min Scan# 2253
 Delta R.T. -0.009 min
 Lab File: BN036479.D
 Acq: 20 Feb 2025 13:48

Tgt Ion:149	Resp:	2546	
Ion	Ratio	Lower	Upper
149	100		
167	26.0	21.2	31.8
279	3.9	2.7	4.1





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.587 min Scan# 21

Delta R.T. -0.009 min

Lab File: BN036479.D


Acq: 20 Feb 2025 13:48

Instrument :

BNA_N

ClientSampleId :

VPB192-HYD-20250214





Tgt Ion:264 Resp: 4501

Ion Ratio Lower Upper

264 100

260 26.9 20.9 31.3

265 107.1 60.7 91.1#





CALIBRATION SUMMARY

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\
Method File : 8270-SIM-BN021025.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Feb 11 01:17:14 2025
Response Via : Initial Calibration

Calibration Files

0.1 =BN036409.D 0.2 =BN036410.D 0.4 =BN036411.D 0.8 =BN036412.D 1.6 =BN036413.D 3.2 =BN036414.D 5.0 =BN036415.D

	Compound	0.1	0.2	0.4	0.8	1.6	3.2	5.0	Avg	%RSD
1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2)	1,4-Dioxane	0.555	0.437	0.433	0.414	0.411	0.433	0.381	0.438	12.66
3)	n-Nitrosodimet...	0.906	0.779	0.764	0.724	0.708	0.769	0.670	0.760	9.90
4) S	2-Fluorophenol	1.009	0.954	0.936	0.920	0.914	0.999	0.885	0.945	4.80
5) S	Phenol-d6	1.134	1.007	1.032	1.062	1.099	1.267	1.164	1.109	8.00
6)	bis(2-Chloroet...	1.382	1.070	1.086	1.129	1.120	1.225	1.107	1.160	9.48
7) I	Naphthalene-d8	-----ISTD-----								
8) S	Nitrobenzene-d5	0.500	0.363	0.365	0.370	0.367	0.417	0.381	0.395	12.70
9)	Naphthalene	1.400	1.141	1.116	1.088	1.075	1.186	1.073	1.154	10.01
10)	Hexachlorobuta...	0.319	0.293	0.283	0.272	0.264	0.282	0.253	0.281	7.67
11) SURR2	Methylnaphth...	0.647	0.583	0.602	0.588	0.597	0.668	0.618	0.615	5.19
12)	2-Methylnaphth...	0.833	0.712	0.738	0.721	0.726	0.816	0.750	0.757	6.40
13) I	Acenaphthene-d10	-----ISTD-----								
14) S	2,4,6-Tribromo...	0.196	0.181	0.186	0.184	0.195	0.226	0.219	0.198	8.90
15) S	2-Fluorobiphenyl	1.409	1.390	1.377	1.491	1.564	1.738	1.558	1.504	8.57
16)	Acenaphthylene	1.807	1.667	1.692	1.683	1.734	1.964	1.820	1.767	5.98
17)	Acenaphthene	1.245	1.125	1.146	1.128	1.175	1.273	1.169	1.180	4.89
18)	Fluorene	1.696	1.630	1.661	1.627	1.669	1.829	1.646	1.680	4.17
19) I	Phenanthrene-d10	-----ISTD-----								
20)	4,6-Dinitro-2-...	0.071	0.067	0.069	0.074	0.084	0.107		0.078	19.60
21)	4-Bromophenyl-...	0.243	0.227	0.231	0.232	0.236	0.264	0.238	0.239	5.15
22)	Hexachlorobenzene	0.305	0.296	0.284	0.287	0.289	0.317	0.285	0.295	4.11
23)	Atrazine	0.196	0.190	0.187	0.186	0.194	0.229	0.213	0.199	8.00
24)	Pentachlorophenol	0.140	0.125	0.122	0.122	0.134	0.170	0.167	0.140	14.74
25)	Phenanthrene	1.233	1.090	1.095	1.112	1.138	1.273	1.153	1.156	6.12
26)	Anthracene	0.990	0.933	0.967	0.978	1.015	1.167	1.088	1.020	7.92
27) SURR	Fluoranthene-d10	1.109	1.043	1.063	1.059	1.098	1.258	1.156	1.112	6.70
28)	Fluoranthene	1.441	1.323	1.353	1.356	1.404	1.607	1.461	1.421	6.76
29) I	Chrysene-d12	-----ISTD-----								
30)	Pyrene	1.584	1.568	1.534	1.490	1.488	1.629	1.492	1.541	3.59
31) S	Terphenyl-d14	0.860	0.847	0.852	0.829	0.834	0.913	0.843	0.854	3.27
32)	Benzo(a)anthra...	1.257	1.276	1.293	1.255	1.300	1.471	1.362	1.316	5.86
33)	Chrysene	1.449	1.456	1.360	1.414	1.404	1.527	1.366	1.425	4.08
34)	Bis(2-ethylhex...	0.902	0.875	0.777	0.745	0.761	0.861	0.819	0.820	7.45
35) I	Perylene-d12	-----ISTD-----								

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\

Method File : 8270-SIM-BN021025.M

36)	Indeno(1,2,3-c...	1.182	1.289	1.378	1.390	1.446	1.630	1.471	1.398	10.13
37)	Benzo(b)fluora...	1.174	1.220	1.260	1.290	1.333	1.529	1.416	1.317	9.24
38)	Benzo(k)fluora...	1.258	1.253	1.363	1.326	1.347	1.532	1.413	1.356	7.08
39) C	Benzo(a)pyrene	1.091	1.081	1.102	1.114	1.145	1.309	1.206	1.150	7.12
40)	Dibenzo(a,h)an...	0.906	1.021	1.075	1.087	1.154	1.304	1.176	1.103	11.40
41)	Benzo(g,h,i)pe...	1.140	1.212	1.254	1.230	1.269	1.400	1.249	1.250	6.27

(#) = Out of Range

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
 Data File : BN036409.D
 Acq On : 10 Feb 2025 12:25
 Operator : RC/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: Feb 11 00:34:59 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 00:33:05 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.753	152	2370	0.400	ng	0.00
7) Naphthalene-d8	10.552	136	5687	0.400	ng	# 0.01
13) Acenaphthene-d10	14.398	164	3837	0.400	ng	0.01
19) Phenanthrene-d10	17.136	188	8539	0.400	ng	0.00
29) Chrysene-d12	21.331	240	8027	0.400	ng	# 0.00
35) Perylene-d12	23.598	264	8068	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	598	0.099	ng	0.00
5) Phenol-d6	6.937	99	672	0.095	ng	0.00
8) Nitrobenzene-d5	8.907	82	711	0.134	ng	0.00
11) 2-Methylnaphthalene-d10	12.146	152	920	0.118	ng	0.00
14) 2,4,6-Tribromophenol	15.895	330	188	0.080	ng	0.01
15) 2-Fluorobiphenyl	13.019	172	1352	0.083	ng	0.00
27) Fluoranthene-d10	19.169	212	2368	0.108	ng	0.00
31) Terphenyl-d14	19.773	244	1725	0.104	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.268	88	329	0.126	ng	93
3) n-Nitrosodimethylamine	3.586	42	537	0.114	ng	# 96
6) bis(2-Chloroethyl)ether	7.183	93	819	0.140	ng	99
9) Naphthalene	10.594	128	1990	0.123	ng	# 91
10) Hexachlorobutadiene	10.883	225	454	0.089	ng	# 99
12) 2-Methylnaphthalene	12.217	142	1185	0.116	ng	96
16) Acenaphthylene	14.110	152	1733	0.098	ng	99
17) Acenaphthene	14.452	154	1194	0.098	ng	98
18) Fluorene	15.446	166	1627	0.104	ng	99
20) 4,6-Dinitro-2-methylph...	15.535	198	151	0.079	ng	# 36
21) 4-Bromophenyl-phenylether	16.342	248	518	0.089	ng	# 81
22) Hexachlorobenzene	16.453	284	651	0.085	ng	99
23) Atrazine	16.615	200	419	0.097	ng	# 85
24) Pentachlorophenol	16.801	266	298	0.089	ng	97
25) Phenanthrene	17.173	178	2633	0.105	ng	100
26) Anthracene	17.273	178	2113	0.093	ng	96
28) Fluoranthene	19.201	202	3077	0.104	ng	99
30) Pyrene	19.564	202	3179	0.099	ng	99
32) Benzo(a)anthracene	21.313	228	2523	0.088	ng	96
33) Chrysene	21.366	228	2908	0.099	ng	97
34) Bis(2-ethylhexyl)phtha...	21.241	149	1810	0.114	ng	99
36) Indeno(1,2,3-cd)pyrene	25.911	276	2385	0.075	ng	98
37) Benzo(b)fluoranthene	22.917	252	2367	0.083	ng	# 67
38) Benzo(k)fluoranthene	22.961	252	2537	0.087	ng	# 65
39) Benzo(a)pyrene	23.499	252	2200	0.089	ng	# 53
40) Dibenzo(a,h)anthracene	25.928	278	1827	0.073	ng	# 62
41) Benzo(g,h,i)perylene	26.607	276	2299	0.083	ng	# 83


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
Data File : BN036409.D
Acq On : 10 Feb 2025 12:25
Operator : RC/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Quant Time: Feb 11 00:34:59 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 00:33:05 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.753 min Scan# 64
Delta R.T. 0.000 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:152 Resp: 2370
Ion Ratio Lower Upper
152 100
150 154.7 123.7 185.5
115 65.5 52.5 78.7



#2
1,4-Dioxane
Concen: 0.126 ng
RT: 3.268 min Scan# 26
Delta R.T. 0.000 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25

Tgt Ion: 88 Resp: 329
Ion Ratio Lower Upper
88 100
43 48.9 33.7 50.5
58 81.8 68.9 103.3









#3
n-Nitrosodimethylamine
Concen: 0.114 ng
RT: 3.586 min Scan# 70
Delta R.T. 0.007 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion: 42 Resp: 537
Ion Ratio Lower Upper
42 100
74 87.0 71.8 107.6
44 5.4 7.8 11.6#



#4
2-Fluorophenol
Concen: 0.099 ng
RT: 5.348 min Scan# 314
Delta R.T. 0.000 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25

Tgt Ion: 112 Resp: 598
Ion Ratio Lower Upper
112 100
64 64.2 53.4 80.0
63 36.0 30.3 45.5









#5
Phenol-d6
Concen: 0.095 ng
RT: 6.937 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion: 99 Resp: 672
Ion Ratio Lower Upper
99 100
42 36.8 21.7 32.5#
71 42.9 32.6 49.0



#6
bis(2-Chloroethyl)ether
Concen: 0.140 ng
RT: 7.183 min Scan# 568
Delta R.T. 0.000 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25

Tgt Ion: 93 Resp: 819
Ion Ratio Lower Upper
93 100
63 82.8 66.3 99.5
95 31.5 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.552 min Scan# 911
Delta R.T. 0.011 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:	136	Resp:	5687
Ion	Ratio	Lower	Upper
136	100		
137	13.3	10.1	15.1
54	10.9	11.8	17.6#
68	7.4	7.2	10.8



#8
Nitrobenzene-d5
Concen: 0.134 ng
RT: 8.907 min Scan# 780
Delta R.T. 0.000 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25

Tgt Ion:	82	Resp:	711
Ion	Ratio	Lower	Upper
82	100		
128	47.7	31.9	47.9
54	68.5	53.1	79.7









#9
Naphthalene
Concen: 0.123 ng
RT: 10.594 min Scan# 911
Delta R.T. 0.000 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:128 Resp: 1990
Ion Ratio Lower Upper
128 100
129 15.6 9.6 14.4#
127 18.3 12.0 18.0#



#10
Hexachlorobutadiene
Concen: 0.089 ng
RT: 10.883 min Scan# 965
Delta R.T. 0.000 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25

Tgt Ion:225 Resp: 454
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.1 50.9 76.3









#11
2-Methylnaphthalene-d10
Concen: 0.118 ng
RT: 12.146 min Scan# 1149
Delta R.T. 0.005 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:152 Resp: 920
Ion Ratio Lower Upper
152 100
151 21.6 16.6 25.0



#12
2-Methylnaphthalene
Concen: 0.116 ng
RT: 12.217 min Scan# 1163
Delta R.T. 0.005 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25

Tgt Ion:142 Resp: 1185
Ion Ratio Lower Upper
142 100
141 93.1 72.8 109.2
115 50.3 35.5 53.3









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.398 min Scan# 1448
Delta R.T. 0.011 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:164 Resp: 3837
Ion Ratio Lower Upper
164 100
162 100.8 84.1 126.1
160 48.4 41.4 62.0



#14
2,4,6-Tribromophenol
Concen: 0.080 ng
RT: 15.895 min Scan# 1584
Delta R.T. 0.012 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25

Tgt Ion:330 Resp: 188
Ion Ratio Lower Upper
330 100
332 101.6 76.6 114.8
141 50.5 37.8 56.8





#15

2-Fluorobiphenyl

Concen: 0.083 ng

RT: 13.019 min Scan# 1320

Delta R.T. 0.000 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:172 Resp: 1352

Ion Ratio Lower Upper

172 100

171 39.3 29.6 44.4

170 29.6 19.8 29.6



#16

Acenaphthylene

Concen: 0.098 ng

RT: 14.110 min Scan# 1422

Delta R.T. 0.000 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25




Tgt Ion:152 Resp: 1733


Ion Ratio Lower Upper

152 100

151 19.4 15.8 23.8

153 13.3 10.2 15.2









#17
Acenaphthene
Concen: 0.098 ng
RT: 14.452 min Scan# 1454
Delta R.T. 0.000 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion	Ratio	Lower	Upper
154	100		
153	114.2	93.3	139.9
152	61.5	48.8	73.2



#18
Fluorene
Concen: 0.104 ng
RT: 15.446 min Scan# 1547
Delta R.T. 0.000 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25

Tgt Ion	Ratio	Lower	Upper
166	100		
165	100.3	79.5	119.3
167	13.0	10.4	15.6





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.136 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:188 Resp: 8539

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 13.6 9.8 14.6



#20

4,6-Dinitro-2-methylphenol

Concen: 0.079 ng

RT: 15.535 min Scan# 1555

Delta R.T. 0.012 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25




Tgt Ion:198 Resp: 151


Ion Ratio Lower Upper

198 100

51 203.2 86.6 129.8#

105 91.6 57.5 86.3#









#21
4-Bromophenyl-phenylether
Concen: 0.089 ng
RT: 16.342 min Scan# 1620
Delta R.T. 0.012 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion	Ratio	Lower	Upper
248	100		
250	103.8	76.1	114.1
141	61.7	71.7	107.5#



#22
Hexachlorobenzene
Concen: 0.085 ng
RT: 16.453 min Scan# 1629
Delta R.T. 0.000 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25

Tgt Ion	Ratio	Lower	Upper
284	100		
142	41.5	33.4	50.0
249	35.0	28.6	43.0









#23
Atrazine
Concen: 0.097 ng
RT: 16.615 min Scan# 1642
Delta R.T. 0.000 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:200 Resp: 419
Ion Ratio Lower Upper
200 100
173 40.9 23.2 34.8#
215 57.5 40.0 60.0



#24
Pentachlorophenol
Concen: 0.089 ng
RT: 16.801 min Scan# 1657
Delta R.T. 0.000 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25

Tgt Ion:266 Resp: 298
Ion Ratio Lower Upper
266 100
264 60.1 50.6 76.0
268 63.1 51.9 77.9





#25

Phenanthrene

Concen: 0.105 ng

RT: 17.173 min Scan# 1687

Delta R.T. 0.000 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:178 Resp: 2633

Ion Ratio Lower Upper

178 100

176 19.6 15.7 23.5

179 15.5 12.4 18.6



#26

Anthracene

Concen: 0.093 ng

RT: 17.273 min Scan# 1695

Delta R.T. 0.000 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25




Tgt Ion:178 Resp: 2113


Ion Ratio Lower Upper

178 100

176 20.8 14.9 22.3

179 14.6 12.4 18.6





#27

Fluoranthene-d10

Concen: 0.108 ng

RT: 19.169 min Scan# 1936

Delta R.T. 0.000 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:212 Resp: 2368

Ion Ratio Lower Upper

212 100

106 14.8 11.5 17.3

104 9.8 7.1 10.7



#28

Fluoranthene

Concen: 0.104 ng

RT: 19.201 min Scan# 1943

Delta R.T. 0.000 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25




Tgt Ion:202 Resp: 3077


Ion Ratio Lower Upper

202 100

101 12.2 9.2 13.8

203 16.6 13.4 20.0





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.331 min Scan# 21

Delta R.T. 0.009 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.1





Tgt Ion:240 Resp: 8027

Ion Ratio Lower Upper

240 100

120 11.7 13.3 19.9#

236 28.4 23.0 34.6



#30

Pyrene

Concen: 0.099 ng

RT: 19.564 min Scan# 2021

Delta R.T. 0.000 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25




Tgt Ion:202 Resp: 3179


Ion Ratio Lower Upper

202 100

200 21.4 16.9 25.3

203 17.8 13.9 20.9









#31
Terphenyl-d14
Concen: 0.104 ng
RT: 19.773 min Scan# 2066
Delta R.T. 0.005 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:244 Resp: 1725
Ion Ratio Lower Upper
244 100
212 13.3 8.1 12.1#
122 17.7 12.8 19.2



#32
Benzo(a)anthracene
Concen: 0.088 ng
RT: 21.313 min Scan# 2262
Delta R.T. 0.000 min
Lab File: BN036409.D
Acq: 10 Feb 2025 12:25

Tgt Ion:228 Resp: 2523
Ion Ratio Lower Upper
228 100
226 30.0 22.2 33.2
229 22.7 16.5 24.7





#33

Chrysene

Concen: 0.099 ng

RT: 21.366 min Scan# 21

Delta R.T. 0.009 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:228 Resp: 2908

Ion Ratio Lower Upper

228 100

226 32.9 25.5 38.3

229 23.1 16.4 24.6



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.114 ng

RT: 21.241 min Scan# 2254

Delta R.T. 0.000 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25




Tgt Ion:149 Resp: 1810


Ion Ratio Lower Upper

149 100

167 26.2 21.2 31.8

279 3.1 2.7 4.1





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.598 min Scan# 21

Delta R.T. 0.003 min

Lab File: BN036409.D


Acq: 10 Feb 2025 12:25

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.1





Tgt Ion:264 Resp: 8068

Ion Ratio Lower Upper

264 100

260 25.7 20.9 31.3

265 73.6 60.7 91.1



#36

Indeno(1,2,3-cd)pyrene


Concen: 0.075 ng

RT: 25.911 min Scan# 3583

Delta R.T. 0.012 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25





Tgt Ion:276 Resp: 2385


Ion Ratio Lower Upper

276 100

138 25.9 22.2 33.2

277 24.8 19.8 29.6





#37

Benzo(b)fluoranthene

Concen: 0.083 ng

RT: 22.917 min Scan# 2557

Delta R.T. 0.006 min

Lab File: BN036409.D


Acq: 10 Feb 2025 12:25

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:252 Resp: 2367

Ion Ratio Lower Upper

252 100

253 42.0 21.9 32.9#

125 37.0 15.0 22.6#



#38

Benzo(k)fluoranthene


Concen: 0.087 ng

RT: 22.961 min Scan# 2574

Delta R.T. 0.003 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25





Tgt Ion:252 Resp: 2537


Ion Ratio Lower Upper

252 100

253 44.3 22.2 33.4#

125 36.6 15.0 22.4#





#39

Benzo(a)pyrene

Concen: 0.089 ng

RT: 23.499 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.1





Tgt Ion:252 Resp: 2200

Ion Ratio Lower Upper

252 100

253 53.1 24.4 36.6#

125 48.8 18.2 27.2#



#40

Dibenzo(a,h)anthracene

Concen: 0.073 ng

RT: 25.928 min Scan# 3589

Delta R.T. 0.012 min

Lab File: BN036409.D

Acq: 10 Feb 2025 12:25




Tgt Ion:278 Resp: 1827


Ion Ratio Lower Upper

278 100

139 36.1 18.5 27.7#

279 56.9 24.8 37.2#





#41

Benzo(g,h,i)perylene

Concen: 0.083 ng

RT: 26.607 min Scan# 3819

Delta R.T. 0.006 min

Lab File: BN036409.D


Acq: 10 Feb 2025 12:25

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:276 Resp: 2299

Ion Ratio Lower Upper

276 100

277 34.3 20.7 31.1#

138 36.1 21.8 32.6#



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
 Data File : BN036410.D
 Acq On : 10 Feb 2025 13:01
 Operator : RC/JU
 Sample : SSTDICC0.2
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.2

Quant Time: Feb 11 00:35:26 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 00:33:05 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.753	152	2157	0.400	ng	0.00
7) Naphthalene-d8	10.551	136	4791	0.400	ng	# 0.01
13) Acenaphthene-d10	14.387	164	3185	0.400	ng	0.00
19) Phenanthrene-d10	17.136	188	7435	0.400	ng	0.00
29) Chrysene-d12	21.331	240	6531	0.400	ng	# 0.00
35) Perylene-d12	23.595	264	6918	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	1029	0.187	ng	0.00
5) Phenol-d6	6.937	99	1086	0.170	ng	0.00
8) Nitrobenzene-d5	8.907	82	869	0.195	ng	0.00
11) 2-Methylnaphthalene-d10	12.146	152	1397	0.213	ng	0.00
14) 2,4,6-Tribromophenol	15.895	330	289	0.147	ng	0.01
15) 2-Fluorobiphenyl	13.019	172	2214	0.163	ng	0.00
27) Fluoranthene-d10	19.173	212	3879	0.203	ng	0.00
31) Terphenyl-d14	19.773	244	2765	0.204	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.268	88	471	0.198	ng	97
3) n-Nitrosodimethylamine	3.579	42	840	0.196	ng	# 98
6) bis(2-Chloroethyl)ether	7.183	93	1154	0.217	ng	98
9) Naphthalene	10.594	128	2733	0.200	ng	# 95
10) Hexachlorobutadiene	10.882	225	702	0.164	ng	# 99
12) 2-Methylnaphthalene	12.217	142	1706	0.198	ng	97
16) Acenaphthylene	14.110	152	2654	0.180	ng	98
17) Acenaphthene	14.452	154	1791	0.177	ng	98
18) Fluorene	15.446	166	2596	0.200	ng	100
20) 4,6-Dinitro-2-methylph...	15.535	198	248	0.149	ng	# 85
21) 4-Bromophenyl-phenylether	16.342	248	844	0.166	ng	# 80
22) Hexachlorobenzene	16.453	284	1101	0.166	ng	97
23) Atrazine	16.615	200	705	0.188	ng	95
24) Pentachlorophenol	16.801	266	463	0.159	ng	95
25) Phenanthrene	17.173	178	4051	0.186	ng	100
26) Anthracene	17.273	178	3467	0.175	ng	98
28) Fluoranthene	19.201	202	4918	0.190	ng	100
30) Pyrene	19.564	202	5119	0.196	ng	100
32) Benzo(a)anthracene	21.313	228	4166	0.180	ng	98
33) Chrysene	21.357	228	4754	0.200	ng	99
34) Bis(2-ethylhexyl)phtha...	21.241	149	2858	0.221	ng	100
36) Indeno(1,2,3-cd)pyrene	25.905	276	4459	0.164	ng	99
37) Benzo(b)fluoranthene	22.914	252	4220	0.172	ng	# 85
38) Benzo(k)fluoranthene	22.958	252	4335	0.173	ng	# 85
39) Benzo(a)pyrene	23.499	252	3740	0.177	ng	# 80
40) Dibenzo(a,h)anthracene	25.919	278	3533	0.164	ng	# 86
41) Benzo(g,h,i)perylene	26.601	276	4192	0.177	ng	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
Data File : BN036410.D
Acq On : 10 Feb 2025 13:01
Operator : RC/JU
Sample : SSTDICC0.2
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2

Quant Time: Feb 11 00:35:26 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 00:33:05 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.753 min Scan# 64
Delta R.T. 0.000 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:152 Resp: 2157
Ion Ratio Lower Upper
152 100
150 152.0 123.7 185.5
115 65.9 52.5 78.7



#2
1,4-Dioxane
Concen: 0.198 ng
RT: 3.268 min Scan# 26
Delta R.T. -0.000 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01

Tgt Ion: 88 Resp: 471
Ion Ratio Lower Upper
88 100
43 48.2 33.7 50.5
58 85.8 68.9 103.3









#3
n-Nitrosodimethylamine
Concen: 0.196 ng
RT: 3.579 min Scan# 69
Delta R.T. -0.000 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion: 42 Resp: 840
Ion Ratio Lower Upper
42 100
74 90.6 71.8 107.6
44 14.5 7.8 11.6#



#4
2-Fluorophenol
Concen: 0.187 ng
RT: 5.348 min Scan# 314
Delta R.T. -0.000 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01

Tgt Ion: 112 Resp: 1029
Ion Ratio Lower Upper
112 100
64 64.8 53.4 80.0
63 37.2 30.3 45.5









#5
Phenol-d6
Concen: 0.170 ng
RT: 6.937 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion	Ratio	Lower	Upper
99	100		
42	26.8	21.7	32.5
71	40.7	32.6	49.0



#6
bis(2-Chloroethyl)ether
Concen: 0.217 ng
RT: 7.183 min Scan# 568
Delta R.T. -0.000 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01

Tgt Ion	Ratio	Lower	Upper
93	100		
63	81.6	66.3	99.5
95	34.1	26.2	39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.551 min Scan# 91
Delta R.T. 0.011 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:	136	Resp:	4791
Ion	Ratio	Lower	Upper
136	100		
137	13.2	10.1	15.1
54	11.1	11.8	17.6#
68	7.8	7.2	10.8



#8
Nitrobenzene-d5
Concen: 0.195 ng
RT: 8.907 min Scan# 780
Delta R.T. -0.000 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01

Tgt Ion:	82	Resp:	869
Ion	Ratio	Lower	Upper
82	100		
128	43.5	31.9	47.9
54	68.8	53.1	79.7





#9

Naphthalene

Concen: 0.200 ng

RT: 10.594 min Scan# 91

Delta R.T. -0.000 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:128 Resp: 2733

Ion Ratio Lower Upper

128 100

129 14.8 9.6 14.4#

127 16.7 12.0 18.0



#10

Hexachlorobutadiene

Concen: 0.164 ng

RT: 10.882 min Scan# 965

Delta R.T. -0.000 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01




Tgt Ion:225 Resp: 702


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 64.1 50.9 76.3









#11
2-Methylnaphthalene-d10
Concen: 0.213 ng
RT: 12.146 min Scan# 1149
Delta R.T. 0.005 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:152 Resp: 1397
Ion Ratio Lower Upper
152 100
151 21.0 16.6 25.0



#12
2-Methylnaphthalene
Concen: 0.198 ng
RT: 12.217 min Scan# 1163
Delta R.T. 0.005 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01

Tgt Ion:142 Resp: 1706
Ion Ratio Lower Upper
142 100
141 89.9 72.8 109.2
115 49.4 35.5 53.3









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.387 min Scan# 1448
Delta R.T. -0.000 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:164 Resp: 3185
Ion Ratio Lower Upper
164 100
162 106.9 84.1 126.1
160 53.5 41.4 62.0



#14
2,4,6-Tribromophenol
Concen: 0.147 ng
RT: 15.895 min Scan# 1584
Delta R.T. 0.012 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01

Tgt Ion:330 Resp: 289
Ion Ratio Lower Upper
330 100
332 99.0 76.6 114.8
141 49.8 37.8 56.8









#15
2-Fluorobiphenyl
Concen: 0.163 ng
RT: 13.019 min Scan# 1320
Delta R.T. -0.000 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion:172 Resp: 2214
Ion Ratio Lower Upper
172 100
171 38.5 29.6 44.4
170 26.7 19.8 29.6



#16
Acenaphthylene
Concen: 0.180 ng
RT: 14.110 min Scan# 1422
Delta R.T. -0.000 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01

Tgt Ion:152 Resp: 2654
Ion Ratio Lower Upper
152 100
151 20.2 15.8 23.8
153 13.9 10.2 15.2









#17
Acenaphthene
Concen: 0.177 ng
RT: 14.452 min Scan# 1454
Delta R.T. -0.000 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion	154	Resp	1791
Ion Ratio	Lower	Upper	
154	100		
153	119.4	93.3	139.9
152	61.4	48.8	73.2



#18
Fluorene
Concen: 0.200 ng
RT: 15.446 min Scan# 1547
Delta R.T. -0.000 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01

Tgt Ion	166	Resp	2596
Ion Ratio	Lower	Upper	
166	100		
165	99.4	79.5	119.3
167	13.8	10.4	15.6





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.136 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:188 Resp: 7435

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 13.6 9.8 14.6



#20

4,6-Dinitro-2-methylphenol

Concen: 0.149 ng

RT: 15.535 min Scan# 1555

Delta R.T. 0.012 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01




Tgt Ion:198 Resp: 248


Ion Ratio Lower Upper

198 100

51 133.3 86.6 129.8#

105 73.2 57.5 86.3









#21
4-Bromophenyl-phenylether
Concen: 0.166 ng
RT: 16.342 min Scan# 1620
Delta R.T. 0.012 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.2


Tgt Ion:248 Resp: 844
Ion Ratio Lower Upper
248 100
250 102.3 76.1 114.1
141 58.7 71.7 107.5#



#22
Hexachlorobenzene
Concen: 0.166 ng
RT: 16.453 min Scan# 1629
Delta R.T. -0.000 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01

Tgt Ion:284 Resp: 1101
Ion Ratio Lower Upper
284 100
142 39.9 33.4 50.0
249 33.7 28.6 43.0





#23

Atrazine

Concen: 0.188 ng

RT: 16.615 min Scan# 1642

Delta R.T. -0.000 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:200 Resp: 705

Ion Ratio Lower Upper

200 100

173 34.0 23.2 34.8

215 52.0 40.0 60.0



#24

Pentachlorophenol

Concen: 0.159 ng

RT: 16.801 min Scan# 1657

Delta R.T. -0.000 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01




Tgt Ion:266 Resp: 463


Ion Ratio Lower Upper

266 100

264 60.9 50.6 76.0

268 60.0 51.9 77.9





#25

Phenanthrene

Concen: 0.186 ng

RT: 17.173 min Scan# 1687

Delta R.T. -0.000 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:178 Resp: 4051

Ion Ratio Lower Upper

178 100

176 19.6 15.7 23.5

179 15.3 12.4 18.6



#26

Anthracene

Concen: 0.175 ng

RT: 17.273 min Scan# 1695

Delta R.T. -0.000 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01




Tgt Ion:178 Resp: 3467


Ion Ratio Lower Upper

178 100

176 19.6 14.9 22.3

179 16.5 12.4 18.6





#27

Fluoranthene-d10

Concen: 0.203 ng

RT: 19.173 min Scan# 1937

Delta R.T. 0.005 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:212 Resp: 3879

Ion Ratio Lower Upper

212 100

106 14.1 11.5 17.3

104 8.6 7.1 10.7



#28

Fluoranthene

Concen: 0.190 ng

RT: 19.201 min Scan# 1943

Delta R.T. -0.000 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01




Tgt Ion:202 Resp: 4918


Ion Ratio Lower Upper

202 100

101 11.6 9.2 13.8

203 16.7 13.4 20.0





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.331 min Scan# 21

Delta R.T. 0.009 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.2





Tgt Ion:240 Resp: 6531

Ion Ratio Lower Upper

240 100

120 12.1 13.3 19.9#

236 29.5 23.0 34.6



#30

Pyrene

Concen: 0.196 ng

RT: 19.564 min Scan# 2021

Delta R.T. -0.000 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01




Tgt Ion:202 Resp: 5119


Ion Ratio Lower Upper

202 100

200 21.0 16.9 25.3

203 17.6 13.9 20.9









#31
Terphenyl-d14
Concen: 0.204 ng
RT: 19.773 min Scan# 2066
Delta R.T. 0.005 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.2


Tgt Ion	Ratio	Lower	Upper
244	100		
212	11.8	8.1	12.1
122	15.9	12.8	19.2



#32
Benzo(a)anthracene
Concen: 0.180 ng
RT: 21.313 min Scan# 2262
Delta R.T. -0.000 min
Lab File: BN036410.D
Acq: 10 Feb 2025 13:01

Tgt Ion	Ratio	Lower	Upper
228	100		
226	28.0	22.2	33.2
229	21.9	16.5	24.7





#33

Chrysene

Concen: 0.200 ng

RT: 21.357 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:228 Resp: 4754

Ion Ratio Lower Upper

228 100

226 32.5 25.5 38.3

229 21.1 16.4 24.6



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.221 ng

RT: 21.241 min Scan# 2254

Delta R.T. -0.000 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01




Tgt Ion:149 Resp: 2858


Ion Ratio Lower Upper

149 100

167 26.6 21.2 31.8

279 2.9 2.7 4.1





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.595 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:264 Resp: 6918

Ion Ratio Lower Upper

264 100

260 25.7 20.9 31.3

265 80.8 60.7 91.1



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.164 ng

RT: 25.905 min Scan# 3581

Delta R.T. 0.006 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01




Tgt Ion:276 Resp: 4459


Ion Ratio Lower Upper

276 100

138 28.4 22.2 33.2

277 25.1 19.8 29.6





#37

Benzo(b)fluoranthene

Concen: 0.172 ng

RT: 22.914 min Scan# 2558

Delta R.T. 0.003 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2




Tgt Ion:252 Resp: 4220

Ion Ratio Lower Upper

252 100

253 34.5 21.9 32.9#

125 26.1 15.0 22.6#



#38

Benzo(k)fluoranthene

Concen: 0.173 ng

RT: 22.958 min Scan# 2573

Delta R.T. -0.000 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01




Tgt Ion:252 Resp: 4335


Ion Ratio Lower Upper

252 100

253 35.3 22.2 33.4#

125 26.0 15.0 22.4#





#39

Benzo(a)pyrene

Concen: 0.177 ng

RT: 23.499 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:252 Resp: 3740

Ion Ratio Lower Upper

252 100

253 40.4 24.4 36.6#

125 33.6 18.2 27.2#



#40

Dibenzo(a,h)anthracene

Concen: 0.164 ng

RT: 25.919 min Scan# 3586

Delta R.T. 0.003 min

Lab File: BN036410.D

Acq: 10 Feb 2025 13:01




Tgt Ion:278 Resp: 3533


Ion Ratio Lower Upper

278 100

139 28.3 18.5 27.7#

279 40.0 24.8 37.2#





#41

Benzo(g,h,i)perylene

Concen: 0.177 ng

RT: 26.601 min Scan# 3819

Delta R.T. -0.000 min

Lab File: BN036410.D


Acq: 10 Feb 2025 13:01

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.2





Tgt Ion:276 Resp: 4192

Ion Ratio Lower Upper

276 100

277 29.0 20.7 31.1

138 31.5 21.8 32.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
 Data File : BN036411.D
 Acq On : 10 Feb 2025 13:36
 Operator : RC/JU
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4

Quant Time: Feb 11 00:35:50 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 00:33:05 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.753	152	2219	0.400	ng	0.00
7) Naphthalene-d8	10.541	136	5528	0.400	ng	0.00
13) Acenaphthene-d10	14.388	164	3606	0.400	ng	0.00
19) Phenanthrene-d10	17.136	188	8328	0.400	ng	0.00
29) Chrysene-d12	21.322	240	7484	0.400	ng	0.00
35) Perylene-d12	23.595	264	7735	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	2078	0.368	ng	0.00
5) Phenol-d6	6.937	99	2290	0.348	ng	0.00
8) Nitrobenzene-d5	8.907	82	2015	0.391	ng	0.00
11) 2-Methylnaphthalene-d10	12.141	152	3329	0.440	ng	0.00
14) 2,4,6-Tribromophenol	15.882	330	670	0.302	ng	0.00
15) 2-Fluorobiphenyl	13.019	172	4965	0.323	ng	0.00
27) Fluoranthene-d10	19.169	212	8851	0.413	ng	0.00
31) Terphenyl-d14	19.768	244	6378	0.412	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.268	88	961	0.392	ng	99
3) n-Nitrosodimethylamine	3.579	42	1695	0.385	ng	100
6) bis(2-Chloroethyl)ether	7.183	93	2410	0.440	ng	100
9) Naphthalene	10.594	128	6168	0.391	ng	100
10) Hexachlorobutadiene	10.883	225	1567	0.317	ng	# 100
12) 2-Methylnaphthalene	12.212	142	4078	0.411	ng	100
16) Acenaphthylene	14.110	152	6103	0.366	ng	100
17) Acenaphthene	14.452	154	4132	0.362	ng	100
18) Fluorene	15.446	166	5991	0.407	ng	99
20) 4,6-Dinitro-2-methylph...	15.523	198	572	0.307	ng	100
21) 4-Bromophenyl-phenylether	16.329	248	1920	0.337	ng	100
22) Hexachlorobenzene	16.453	284	2369	0.318	ng	100
23) Atrazine	16.615	200	1560	0.372	ng	100
24) Pentachlorophenol	16.801	266	1018	0.313	ng	99
25) Phenanthrene	17.173	178	9119	0.374	ng	100
26) Anthracene	17.273	178	8056	0.363	ng	100
28) Fluoranthene	19.201	202	11264	0.389	ng	100
30) Pyrene	19.564	202	11479	0.384	ng	100
32) Benzo(a)anthracene	21.313	228	9677	0.364	ng	100
33) Chrysene	21.358	228	10178	0.373	ng	100
34) Bis(2-ethylhexyl)phtha...	21.241	149	5815	0.392	ng	100
36) Indeno(1,2,3-cd)pyrene	25.899	276	10661	0.351	ng	100
37) Benzo(b)fluoranthene	22.911	252	9743	0.355	ng	100
38) Benzo(k)fluoranthene	22.958	252	10543	0.377	ng	100
39) Benzo(a)pyrene	23.499	252	8521	0.362	ng	100
40) Dibenzo(a,h)anthracene	25.917	278	8314	0.344	ng	100
41) Benzo(g,h,i)perylene	26.601	276	9701	0.366	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
Data File : BN036411.D
Acq On : 10 Feb 2025 13:36
Operator : RC/JU
Sample : SSTDICCC0.4
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Quant Time: Feb 11 00:35:50 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 00:33:05 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.753 min Scan# 64
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:152 Resp: 2219
Ion Ratio Lower Upper
152 100
150 154.6 123.7 185.5
115 65.6 52.5 78.7



#2
1,4-Dioxane
Concen: 0.392 ng
RT: 3.268 min Scan# 26
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36

Tgt Ion: 88 Resp: 961
Ion Ratio Lower Upper
88 100
43 41.4 33.7 50.5
58 87.1 68.9 103.3









#3
n-Nitrosodimethylamine
Concen: 0.385 ng
RT: 3.579 min Scan# 69
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion: 42 Resp: 1695
Ion Ratio Lower Upper
42 100
74 89.7 71.8 107.6
44 9.7 7.8 11.6



#4
2-Fluorophenol
Concen: 0.368 ng
RT: 5.348 min Scan# 314
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36

Tgt Ion: 112 Resp: 2078
Ion Ratio Lower Upper
112 100
64 66.7 53.4 80.0
63 37.9 30.3 45.5









#5
Phenol-d6
Concen: 0.348 ng
RT: 6.937 min Scan# 511
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion: 99 Resp: 2290
Ion Ratio Lower Upper
99 100
42 27.1 21.7 32.5
71 40.8 32.6 49.0



#6
bis(2-Chloroethyl)ether
Concen: 0.440 ng
RT: 7.183 min Scan# 568
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36

Tgt Ion: 93 Resp: 2410
Ion Ratio Lower Upper
93 100
63 82.9 66.3 99.5
95 32.8 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.541 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	136	Resp:	5528
Ion	Ratio	Lower	Upper
136	100		
137	12.6	10.1	15.1
54	14.7	11.8	17.6
68	9.0	7.2	10.8



#8
Nitrobenzene-d5
Concen: 0.391 ng
RT: 8.907 min Scan# 780
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36

Tgt Ion:	82	Resp:	2015
Ion	Ratio	Lower	Upper
82	100		
128	39.9	31.9	47.9
54	66.4	53.1	79.7









#9
Naphthalene
Concen: 0.391 ng
RT: 10.594 min Scan# 911
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:128 Resp: 6168
Ion Ratio Lower Upper
128 100
129 12.0 9.6 14.4
127 15.0 12.0 18.0



#10
Hexachlorobutadiene
Concen: 0.317 ng
RT: 10.883 min Scan# 965
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36

Tgt Ion:225 Resp: 1567
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.6 50.9 76.3









#11
2-Methylnaphthalene-d10
Concen: 0.440 ng
RT: 12.141 min Scan# 1148
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:152 Resp: 3329
Ion Ratio Lower Upper
152 100
151 20.8 16.6 25.0



#12
2-Methylnaphthalene
Concen: 0.411 ng
RT: 12.212 min Scan# 1162
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36

Tgt Ion:142 Resp: 4078
Ion Ratio Lower Upper
142 100
141 91.0 72.8 109.2
115 44.4 35.5 53.3









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.388 min Scan# 1448
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:164 Resp: 3606
Ion Ratio Lower Upper
164 100
162 105.1 84.1 126.1
160 51.7 41.4 62.0



#14
2,4,6-Tribromophenol
Concen: 0.302 ng
RT: 15.882 min Scan# 1583
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36

Tgt Ion:330 Resp: 670
Ion Ratio Lower Upper
330 100
332 95.7 76.6 114.8
141 47.3 37.8 56.8





#15

2-Fluorobiphenyl

Concen: 0.323 ng

RT: 13.019 min Scan# 1320

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:172 Resp: 4965

Ion Ratio Lower Upper

172 100

171 37.0 29.6 44.4

170 24.7 19.8 29.6



#16

Acenaphthylene

Concen: 0.366 ng

RT: 14.110 min Scan# 1422

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36




Tgt Ion:152 Resp: 6103


Ion Ratio Lower Upper

152 100

151 19.8 15.8 23.8

153 12.7 10.2 15.2





#17

Acenaphthene

Concen: 0.362 ng

RT: 14.452 min Scan# 1454

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:154 Resp: 4132

Ion Ratio Lower Upper

154 100

153 116.6 93.3 139.9

152 61.0 48.8 73.2



#18

Fluorene

Concen: 0.407 ng

RT: 15.446 min Scan# 1547

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36




Tgt Ion:166 Resp: 5991


Ion Ratio Lower Upper

166 100

165 100.0 79.5 119.3

167 13.0 10.4 15.6





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.136 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:188 Resp: 8328

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 12.2 9.8 14.6



#20

4,6-Dinitro-2-methylphenol

Concen: 0.307 ng

RT: 15.523 min Scan# 1554

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36




Tgt Ion:198 Resp: 572


Ion Ratio Lower Upper

198 100

51 108.2 86.6 129.8

105 71.9 57.5 86.3









#21
4-Bromophenyl-phenylether
Concen: 0.337 ng
RT: 16.329 min Scan# 1619
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion: 248 Resp: 1920
Ion Ratio Lower Upper
248 100
250 95.1 76.1 114.1
141 89.6 71.7 107.5



#22
Hexachlorobenzene
Concen: 0.318 ng
RT: 16.453 min Scan# 1629
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36

Tgt Ion: 284 Resp: 2369
Ion Ratio Lower Upper
284 100
142 41.7 33.4 50.0
249 35.4 28.6 43.0









#23
Atrazine
Concen: 0.372 ng
RT: 16.615 min Scan# 1642
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion	Ratio	Lower	Upper
200	100		
173	29.0	23.2	34.8
215	50.0	40.0	60.0



#24
Pentachlorophenol
Concen: 0.313 ng
RT: 16.801 min Scan# 1657
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36

Tgt Ion	Ratio	Lower	Upper
266	100		
264	64.0	50.6	76.0
268	65.1	51.9	77.9









#25
Phenanthrene
Concen: 0.374 ng
RT: 17.173 min Scan# 1687
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:178 Resp: 9119
Ion Ratio Lower Upper
178 100
176 19.6 15.7 23.5
179 15.5 12.4 18.6



#26
Anthracene
Concen: 0.363 ng
RT: 17.273 min Scan# 1695
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36

Tgt Ion:178 Resp: 8056
Ion Ratio Lower Upper
178 100
176 18.7 14.9 22.3
179 15.3 12.4 18.6





#27

Fluoranthene-d10

Concen: 0.413 ng

RT: 19.169 min Scan# 1936

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:212 Resp: 8851

Ion Ratio Lower Upper

212 100

106 14.4 11.5 17.3

104 8.9 7.1 10.7



#28

Fluoranthene

Concen: 0.389 ng

RT: 19.201 min Scan# 1943

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36




Tgt Ion:202 Resp: 11264


Ion Ratio Lower Upper

202 100

101 11.5 9.2 13.8

203 16.7 13.4 20.0





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.322 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:240 Resp: 7484

Ion Ratio Lower Upper

240 100

120 16.6 13.3 19.9

236 28.8 23.0 34.6



#30

Pyrene

Concen: 0.384 ng

RT: 19.564 min Scan# 2021

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36




Tgt Ion:202 Resp: 11479


Ion Ratio Lower Upper

202 100

200 21.1 16.9 25.3

203 17.4 13.9 20.9









#31
Terphenyl-d14
Concen: 0.412 ng
RT: 19.768 min Scan# 2065
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion	Ratio	Lower	Upper
244	100		
212	10.1	8.1	12.1
122	16.0	12.8	19.2



#32
Benzo(a)anthracene
Concen: 0.364 ng
RT: 21.313 min Scan# 2262
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36

Tgt Ion	Ratio	Lower	Upper
228	100		
226	27.7	22.2	33.2
229	20.6	16.5	24.7









#33
Chrysene
Concen: 0.373 ng
RT: 21.358 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion	Ratio	Lower	Upper
228	100		
226	31.9	25.5	38.3
229	20.5	16.4	24.6



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.392 ng
RT: 21.241 min Scan# 2254
Delta R.T. 0.000 min
Lab File: BN036411.D
Acq: 10 Feb 2025 13:36

Tgt Ion	Ratio	Lower	Upper
149	100		
167	26.5	21.2	31.8
279	3.4	2.7	4.1





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.595 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:264 Resp: 7735

Ion Ratio Lower Upper

264 100

260 26.1 20.9 31.3

265 75.9 60.7 91.1



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.351 ng

RT: 25.899 min Scan# 3579

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36




Tgt Ion:276 Resp: 10661


Ion Ratio Lower Upper

276 100

138 27.8 22.2 33.2

277 24.8 19.8 29.6





#37

Benzo(b)fluoranthene

Concen: 0.355 ng

RT: 22.911 min Scan# 2557

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4




Tgt Ion:252 Resp: 9743

Ion Ratio Lower Upper

252 100

253 27.4 21.9 32.9

125 18.8 15.0 22.6



#38

Benzo(k)fluoranthene

Concen: 0.377 ng

RT: 22.958 min Scan# 2573

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36




Tgt Ion:252 Resp: 10543


Ion Ratio Lower Upper

252 100

253 27.8 22.2 33.4

125 18.7 15.0 22.4





#39

Benzo(a)pyrene

Concen: 0.362 ng

RT: 23.499 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:252 Resp: 8521

Ion Ratio Lower Upper

252 100

253 30.5 24.4 36.6

125 22.7 18.2 27.2



#40

Dibenzo(a,h)anthracene

Concen: 0.344 ng

RT: 25.917 min Scan# 3585

Delta R.T. 0.000 min

Lab File: BN036411.D

Acq: 10 Feb 2025 13:36




Tgt Ion:278 Resp: 8314


Ion Ratio Lower Upper

278 100

139 23.1 18.5 27.7

279 31.0 24.8 37.2





#41

Benzo(g,h,i)perylene

Concen: 0.366 ng

RT: 26.601 min Scan# 3819

Delta R.T. 0.000 min

Lab File: BN036411.D


Acq: 10 Feb 2025 13:36

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4





Tgt Ion:276 Resp: 9701

Ion Ratio Lower Upper

276 100

277 25.9 20.7 31.1

138 27.2 21.8 32.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
 Data File : BN036412.D
 Acq On : 10 Feb 2025 14:12
 Operator : RC/JU
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Quant Time: Feb 11 00:36:15 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 00:33:05 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.753	152	2190	0.400	ng	0.00
7) Naphthalene-d8	10.541	136	5510	0.400	ng	0.00
13) Acenaphthene-d10	14.388	164	3611	0.400	ng	0.00
19) Phenanthrene-d10	17.136	188	8139	0.400	ng	0.00
29) Chrysene-d12	21.322	240	7521	0.400	ng	0.00
35) Perylene-d12	23.595	264	7758	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	4030	0.723	ng	0.00
5) Phenol-d6	6.937	99	4653	0.715	ng	0.00
8) Nitrobenzene-d5	8.907	82	4082	0.795	ng	0.00
11) 2-Methylnaphthalene-d10	12.136	152	6480	0.860	ng	0.00
14) 2,4,6-Tribromophenol	15.883	330	1331	0.598	ng	0.00
15) 2-Fluorobiphenyl	13.019	172	10765	0.699	ng	0.00
27) Fluoranthene-d10	19.169	212	17242	0.823	ng	0.00
31) Terphenyl-d14	19.768	244	12467	0.800	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.268	88	1812	0.748	ng	99
3) n-Nitrosodimethylamine	3.579	42	3171	0.729	ng	# 98
6) bis(2-Chloroethyl)ether	7.176	93	4947	0.915	ng	97
9) Naphthalene	10.594	128	11992	0.762	ng	98
10) Hexachlorobutadiene	10.883	225	2996	0.608	ng	# 100
12) 2-Methylnaphthalene	12.212	142	7947	0.804	ng	97
16) Acenaphthylene	14.110	152	12153	0.727	ng	99
17) Acenaphthene	14.452	154	8143	0.712	ng	98
18) Fluorene	15.435	166	11752	0.798	ng	100
20) 4,6-Dinitro-2-methylph...	15.523	198	1199	0.658	ng	# 66
21) 4-Bromophenyl-phenylether	16.329	248	3776	0.677	ng	94
22) Hexachlorobenzene	16.441	284	4670	0.642	ng	96
23) Atrazine	16.602	200	3020	0.737	ng	94
24) Pentachlorophenol	16.801	266	1989	0.625	ng	99
25) Phenanthrene	17.173	178	18094	0.758	ng	99
26) Anthracene	17.273	178	15928	0.735	ng	99
28) Fluoranthene	19.197	202	22075	0.781	ng	99
30) Pyrene	19.564	202	22410	0.746	ng	99
32) Benzo(a)anthracene	21.304	228	18881	0.707	ng	98
33) Chrysene	21.358	228	21269	0.776	ng	97
34) Bis(2-ethylhexyl)phtha...	21.241	149	11207	0.752	ng	99
36) Indeno(1,2,3-cd)pyrene	25.893	276	21570	0.708	ng	98
37) Benzo(b)fluoranthene	22.908	252	20019	0.727	ng	# 93
38) Benzo(k)fluoranthene	22.955	252	20570	0.733	ng	# 93
39) Benzo(a)pyrene	23.493	252	17289	0.731	ng	# 91
40) Dibenzo(a,h)anthracene	25.911	278	16872	0.697	ng	93
41) Benzo(g,h,i)perylene	26.598	276	19078	0.718	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
Data File : BN036412.D
Acq On : 10 Feb 2025 14:12
Operator : RC/JU
Sample : SSTDIC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Quant Time: Feb 11 00:36:15 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 00:33:05 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.753 min Scan# 64
Delta R.T. 0.000 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:152 Resp: 2190
Ion Ratio Lower Upper
152 100
150 152.7 123.7 185.5
115 65.6 52.5 78.7



#2
1,4-Dioxane
Concen: 0.748 ng
RT: 3.268 min Scan# 26
Delta R.T. 0.000 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12

Tgt Ion: 88 Resp: 1812
Ion Ratio Lower Upper
88 100
43 41.0 33.7 50.5
58 86.3 68.9 103.3









#3
n-Nitrosodimethylamine
Concen: 0.729 ng
RT: 3.579 min Scan# 69
Delta R.T. 0.000 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion: 42 Resp: 3171
Ion Ratio Lower Upper
42 100
74 91.4 71.8 107.6
44 7.2 7.8 11.6#



#4
2-Fluorophenol
Concen: 0.723 ng
RT: 5.348 min Scan# 314
Delta R.T. 0.000 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12

Tgt Ion: 112 Resp: 4030
Ion Ratio Lower Upper
112 100
64 65.3 53.4 80.0
63 36.6 30.3 45.5









#5
Phenol-d6
Concen: 0.715 ng
RT: 6.937 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion: 99 Resp: 4653
Ion Ratio Lower Upper
99 100
42 27.6 21.7 32.5
71 40.0 32.6 49.0



#6
bis(2-Chloroethyl)ether
Concen: 0.915 ng
RT: 7.176 min Scan# 567
Delta R.T. -0.007 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12

Tgt Ion: 93 Resp: 4947
Ion Ratio Lower Upper
93 100
63 80.8 66.3 99.5
95 31.0 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.541 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	136	Resp:	5510
Ion	Ratio	Lower	Upper
136	100		
137	12.1	10.1	15.1
54	12.7	11.8	17.6
68	8.4	7.2	10.8



#8
Nitrobenzene-d5
Concen: 0.795 ng
RT: 8.907 min Scan# 780
Delta R.T. 0.000 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12

Tgt Ion:	82	Resp:	4082
Ion	Ratio	Lower	Upper
82	100		
128	40.7	31.9	47.9
54	59.3	53.1	79.7









#9
Naphthalene
Concen: 0.762 ng
RT: 10.594 min Scan# 911
Delta R.T. 0.000 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion	Ratio	Lower	Upper
128	100		
129	11.5	9.6	14.4
127	14.0	12.0	18.0



#10
Hexachlorobutadiene
Concen: 0.608 ng
RT: 10.883 min Scan# 965
Delta R.T. 0.000 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12

Tgt Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.0	50.9	76.3









#11
2-Methylnaphthalene-d10
Concen: 0.860 ng
RT: 12.136 min Scan# 1148
Delta R.T. -0.005 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:152 Resp: 6480
Ion Ratio Lower Upper
152 100
151 20.7 16.6 25.0



#12
2-Methylnaphthalene
Concen: 0.804 ng
RT: 12.212 min Scan# 1162
Delta R.T. 0.000 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12

Tgt Ion:142 Resp: 7947
Ion Ratio Lower Upper
142 100
141 88.7 72.8 109.2
115 41.6 35.5 53.3









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.388 min Scan# 1448
Delta R.T. 0.000 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:164 Resp: 3611
Ion Ratio Lower Upper
164 100
162 105.0 84.1 126.1
160 51.6 41.4 62.0



#14
2,4,6-Tribromophenol
Concen: 0.598 ng
RT: 15.883 min Scan# 1583
Delta R.T. 0.000 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12

Tgt Ion:330 Resp: 1331
Ion Ratio Lower Upper
330 100
332 95.8 76.6 114.8
141 47.5 37.8 56.8





#15

2-Fluorobiphenyl

Concen: 0.699 ng

RT: 13.019 min Scan# 1320

Delta R.T. 0.000 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.8





Tgt Ion:172 Resp: 10765

Ion Ratio Lower Upper

172 100

171 35.2 29.6 44.4

170 23.7 19.8 29.6



#16

Acenaphthylene

Concen: 0.727 ng

RT: 14.110 min Scan# 1422

Delta R.T. 0.000 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12




Tgt Ion:152 Resp: 12153


Ion Ratio Lower Upper

152 100

151 20.0 15.8 23.8

153 13.0 10.2 15.2









#17
Acenaphthene
Concen: 0.712 ng
RT: 14.452 min Scan# 1454
Delta R.T. 0.000 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion	154	Resp	8143
Ion Ratio	Lower	Upper	
154	100		
153	115.5	93.3	139.9
152	59.0	48.8	73.2



#18
Fluorene
Concen: 0.798 ng
RT: 15.435 min Scan# 1546
Delta R.T. -0.011 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12

Tgt Ion	166	Resp	11752
Ion Ratio	Lower	Upper	
166	100		
165	99.6	79.5	119.3
167	13.3	10.4	15.6





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.136 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:188 Resp: 8139

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 11.1 9.8 14.6



#20

4,6-Dinitro-2-methylphenol

Concen: 0.658 ng

RT: 15.523 min Scan# 1554

Delta R.T. 0.000 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12




Tgt Ion:198 Resp: 1199


Ion Ratio Lower Upper

198 100

51 68.0 86.6 129.8#

105 49.9 57.5 86.3#









#21
4-Bromophenyl-phenylether
Concen: 0.677 ng
RT: 16.329 min Scan# 1619
Delta R.T. 0.000 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:248 Resp: 3776
Ion Ratio Lower Upper
248 100
250 94.9 76.1 114.1
141 77.8 71.7 107.5



#22
Hexachlorobenzene
Concen: 0.642 ng
RT: 16.441 min Scan# 1628
Delta R.T. -0.012 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12

Tgt Ion:284 Resp: 4670
Ion Ratio Lower Upper
284 100
142 39.5 33.4 50.0
249 33.7 28.6 43.0









#23
Atrazine
Concen: 0.737 ng
RT: 16.602 min Scan# 1641
Delta R.T. -0.012 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:200 Resp: 3020
Ion Ratio Lower Upper
200 100
173 31.0 23.2 34.8
215 45.3 40.0 60.0



#24
Pentachlorophenol
Concen: 0.625 ng
RT: 16.801 min Scan# 1657
Delta R.T. 0.000 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12

Tgt Ion:266 Resp: 1989
Ion Ratio Lower Upper
266 100
264 64.5 50.6 76.0
268 64.4 51.9 77.9









#25
Phenanthrene
Concen: 0.758 ng
RT: 17.173 min Scan# 1687
Delta R.T. 0.000 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:178 Resp: 18094
Ion Ratio Lower Upper
178 100
176 19.8 15.7 23.5
179 15.1 12.4 18.6



#26
Anthracene
Concen: 0.735 ng
RT: 17.273 min Scan# 1695
Delta R.T. 0.000 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12

Tgt Ion:178 Resp: 15928
Ion Ratio Lower Upper
178 100
176 18.3 14.9 22.3
179 15.4 12.4 18.6





#27

Fluoranthene-d10

Concen: 0.823 ng

RT: 19.169 min Scan# 1936

Delta R.T. 0.000 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:212 Resp: 17242

Ion Ratio Lower Upper

212 100

106 14.2 11.5 17.3

104 8.4 7.1 10.7



#28

Fluoranthene

Concen: 0.781 ng

RT: 19.197 min Scan# 1942

Delta R.T. -0.005 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12




Tgt Ion:202 Resp: 22075


Ion Ratio Lower Upper

202 100

101 11.4 9.2 13.8

203 17.0 13.4 20.0





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.322 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:240 Resp: 7521

Ion Ratio Lower Upper

240 100

120 14.3 13.3 19.9

236 28.3 23.0 34.6



#30

Pyrene

Concen: 0.746 ng

RT: 19.564 min Scan# 2021

Delta R.T. 0.000 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12




Tgt Ion:202 Resp: 22410


Ion Ratio Lower Upper

202 100

200 21.2 16.9 25.3

203 17.9 13.9 20.9









#31
Terphenyl-d14
Concen: 0.800 ng
RT: 19.768 min Scan# 2065
Delta R.T. 0.000 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:244 Resp: 12467
Ion Ratio Lower Upper
244 100
212 9.6 8.1 12.1
122 14.3 12.8 19.2



#32
Benzo(a)anthracene
Concen: 0.707 ng
RT: 21.304 min Scan# 2261
Delta R.T. -0.009 min
Lab File: BN036412.D
Acq: 10 Feb 2025 14:12

Tgt Ion:228 Resp: 18881
Ion Ratio Lower Upper
228 100
226 28.5 22.2 33.2
229 19.3 16.5 24.7





#33

Chrysene

Concen: 0.776 ng

RT: 21.358 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:228 Resp: 21269

Ion Ratio Lower Upper

228 100

226 30.6 25.5 38.3

229 19.1 16.4 24.6



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.752 ng

RT: 21.241 min Scan# 2254

Delta R.T. 0.000 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12




Tgt Ion:149 Resp: 11207


Ion Ratio Lower Upper

149 100

167 27.0 21.2 31.8

279 2.9 2.7 4.1





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.595 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.8





Tgt Ion:264 Resp: 7758

Ion Ratio Lower Upper

264 100

260 25.9 20.9 31.3

265 77.2 60.7 91.1



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.708 ng

RT: 25.893 min Scan# 3577

Delta R.T. -0.006 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12




Tgt Ion:276 Resp: 21570


Ion Ratio Lower Upper

276 100

138 28.9 22.2 33.2

277 23.9 19.8 29.6





#37

Benzo(b)fluoranthene

Concen: 0.727 ng

RT: 22.908 min Scan# 2556

Delta R.T. -0.003 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8




Tgt Ion:252 Resp: 20019

Ion Ratio Lower Upper

252 100

253 24.2 21.9 32.9

125 15.0 15.0 22.6#



#38

Benzo(k)fluoranthene

Concen: 0.733 ng

RT: 22.955 min Scan# 2572

Delta R.T. -0.003 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12




Tgt Ion:252 Resp: 20570


Ion Ratio Lower Upper

252 100

253 24.9 22.2 33.4

125 14.8 15.0 22.4#





#39

Benzo(a)pyrene

Concen: 0.731 ng

RT: 23.493 min Scan# 21

Delta R.T. -0.006 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:252 Resp: 17289

Ion Ratio Lower Upper

252 100

253 26.0 24.4 36.6

125 17.9 18.2 27.2#



#40

Dibenzo(a,h)anthracene

Concen: 0.697 ng

RT: 25.911 min Scan# 3583

Delta R.T. -0.006 min

Lab File: BN036412.D

Acq: 10 Feb 2025 14:12




Tgt Ion:278 Resp: 16872


Ion Ratio Lower Upper

278 100

139 20.6 18.5 27.7

279 26.5 24.8 37.2





#41

Benzo(g,h,i)perylene

Concen: 0.718 ng

RT: 26.598 min Scan# 3818

Delta R.T. -0.003 min

Lab File: BN036412.D


Acq: 10 Feb 2025 14:12

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:276 Resp: 19078

Ion Ratio Lower Upper

276 100

277 24.6 20.7 31.1

138 26.0 21.8 32.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
 Data File : BN036413.D
 Acq On : 10 Feb 2025 14:48
 Operator : RC/JU
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Feb 11 00:36:38 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 00:33:05 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.753	152	2193	0.400	ng	0.00
7) Naphthalene-d8	10.541	136	5625	0.400	ng	0.00
13) Acenaphthene-d10	14.388	164	3645	0.400	ng	0.00
19) Phenanthrene-d10	17.136	188	8137	0.400	ng	0.00
29) Chrysene-d12	21.322	240	7773	0.400	ng	# 0.00
35) Perylene-d12	23.592	264	7967	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	8016	1.436	ng	0.00
5) Phenol-d6	6.937	99	9644	1.481	ng	0.00
8) Nitrobenzene-d5	8.907	82	8250	1.574	ng	0.00
11) 2-Methylnaphthalene-d10	12.136	152	13436	1.746	ng	0.00
14) 2,4,6-Tribromophenol	15.882	330	2844	1.266	ng	0.00
15) 2-Fluorobiphenyl	13.008	172	22810	1.467	ng	-0.01
27) Fluoranthene-d10	19.169	212	35740	1.707	ng	0.00
31) Terphenyl-d14	19.768	244	25927	1.611	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.268	88	3606	1.487	ng	98
3) n-Nitrosodimethylamine	3.572	42	6208	1.425	ng	99
6) bis(2-Chloroethyl)ether	7.175	93	9821	1.813	ng	98
9) Naphthalene	10.594	128	24185	1.505	ng	97
10) Hexachlorobutadiene	10.882	225	5933	1.178	ng	# 100
12) 2-Methylnaphthalene	12.207	142	16325	1.617	ng	99
16) Acenaphthylene	14.110	152	25289	1.499	ng	100
17) Acenaphthene	14.452	154	17130	1.483	ng	97
18) Fluorene	15.435	166	24341	1.638	ng	100
20) 4,6-Dinitro-2-methylph...	15.522	198	2725	1.497	ng	# 47
21) 4-Bromophenyl-phenylether	16.329	248	7690	1.379	ng	91
22) Hexachlorobenzene	16.453	284	9401	1.292	ng	97
23) Atrazine	16.602	200	6316	1.541	ng	95
24) Pentachlorophenol	16.789	266	4352	1.368	ng	99
25) Phenanthrene	17.173	178	37027	1.552	ng	100
26) Anthracene	17.273	178	33031	1.524	ng	99
28) Fluoranthene	19.197	202	45708	1.617	ng	99
30) Pyrene	19.559	202	46256	1.489	ng	100
32) Benzo(a)anthracene	21.304	228	40416	1.463	ng	99
33) Chrysene	21.357	228	43661	1.542	ng	97
34) Bis(2-ethylhexyl)phtha...	21.241	149	23672	1.538	ng	99
36) Indeno(1,2,3-cd)pyrene	25.890	276	46083	1.472	ng	98
37) Benzo(b)fluoranthene	22.908	252	42491	1.502	ng	# 89
38) Benzo(k)fluoranthene	22.952	252	42933	1.489	ng	# 89
39) Benzo(a)pyrene	23.493	252	36490	1.503	ng	# 86
40) Dibenzo(a,h)anthracene	25.905	278	36765	1.478	ng	# 90
41) Benzo(g,h,i)perylene	26.592	276	40441	1.482	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
Data File : BN036413.D
Acq On : 10 Feb 2025 14:48
Operator : RC/JU
Sample : SSTDIC1.6
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDIC1.6

Quant Time: Feb 11 00:36:38 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 00:33:05 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.753 min Scan# 64
Delta R.T. 0.000 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 2193
Ion Ratio Lower Upper
152 100
150 156.9 123.7 185.5
115 66.1 52.5 78.7



#2
1,4-Dioxane
Concen: 1.487 ng
RT: 3.268 min Scan# 26
Delta R.T. -0.000 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48

Tgt Ion: 88 Resp: 3606
Ion Ratio Lower Upper
88 100
43 40.7 33.7 50.5
58 84.0 68.9 103.3









#3
n-Nitrosodimethylamine
Concen: 1.425 ng
RT: 3.572 min Scan# 61
Delta R.T. -0.007 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 42 Resp: 6208
Ion Ratio Lower Upper
42 100
74 90.8 71.8 107.6
44 7.9 7.8 11.6



#4
2-Fluorophenol
Concen: 1.436 ng
RT: 5.348 min Scan# 314
Delta R.T. -0.000 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48

Tgt Ion: 112 Resp: 8016
Ion Ratio Lower Upper
112 100
64 65.9 53.4 80.0
63 37.6 30.3 45.5









#5
Phenol-d6
Concen: 1.481 ng
RT: 6.937 min Scan# 511
Delta R.T. -0.000 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion: 99 Resp: 9644
Ion Ratio Lower Upper
99 100
42 27.7 21.7 32.5
71 41.1 32.6 49.0



#6
bis(2-Chloroethyl)ether
Concen: 1.813 ng
RT: 7.175 min Scan# 567
Delta R.T. -0.007 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48

Tgt Ion: 93 Resp: 9821
Ion Ratio Lower Upper
93 100
63 81.0 66.3 99.5
95 31.5 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.541 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:	136	Resp:	5625
Ion	Ratio	Lower	Upper
136	100		
137	11.9	10.1	15.1
54	12.0	11.8	17.6
68	7.6	7.2	10.8



#8
Nitrobenzene-d5
Concen: 1.574 ng
RT: 8.907 min Scan# 780
Delta R.T. -0.000 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48

Tgt Ion:	82	Resp:	8250
Ion	Ratio	Lower	Upper
82	100		
128	40.6	31.9	47.9
54	56.8	53.1	79.7





#9
Naphthalene

Concen: 1.505 ng

RT: 10.594 min Scan# 91

Delta R.T. -0.000 min

Lab File: BN036413.D

Acq: 10 Feb 2025 14:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:128 Resp: 24185

Ion Ratio Lower Upper

128 100

129 11.0 9.6 14.4

127 13.7 12.0 18.0



#10

Hexachlorobutadiene

Concen: 1.178 ng

RT: 10.882 min Scan# 965

Delta R.T. -0.000 min

Lab File: BN036413.D

Acq: 10 Feb 2025 14:48




Tgt Ion:225 Resp: 5933


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 63.8 50.9 76.3









#11
2-Methylnaphthalene-d10
Concen: 1.746 ng
RT: 12.136 min Scan# 1147
Delta R.T. -0.005 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:152 Resp: 13436
Ion Ratio Lower Upper
152 100
151 20.8 16.6 25.0



#12
2-Methylnaphthalene
Concen: 1.617 ng
RT: 12.207 min Scan# 1161
Delta R.T. -0.005 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48

Tgt Ion:142 Resp: 16325
Ion Ratio Lower Upper
142 100
141 91.0 72.8 109.2
115 41.9 35.5 53.3









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.388 min Scan# 1448
Delta R.T. -0.000 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:164 Resp: 3645
Ion Ratio Lower Upper
164 100
162 105.5 84.1 126.1
160 52.0 41.4 62.0



#14
2,4,6-Tribromophenol
Concen: 1.266 ng
RT: 15.882 min Scan# 1583
Delta R.T. -0.000 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48

Tgt Ion:330 Resp: 2844
Ion Ratio Lower Upper
330 100
332 94.6 76.6 114.8
141 46.0 37.8 56.8









#15
2-Fluorobiphenyl
Concen: 1.467 ng
RT: 13.008 min Scan# 1319
Delta R.T. -0.011 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:172 Resp: 22810
Ion Ratio Lower Upper
172 100
171 36.9 29.6 44.4
170 25.0 19.8 29.6



#16
Acenaphthylene
Concen: 1.499 ng
RT: 14.110 min Scan# 1422
Delta R.T. -0.000 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48

Tgt Ion:152 Resp: 25289
Ion Ratio Lower Upper
152 100
151 20.0 15.8 23.8
153 12.9 10.2 15.2









#17
Acenaphthene
Concen: 1.483 ng
RT: 14.452 min Scan# 1454
Delta R.T. -0.000 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion	154	153	152
Resp:	17130	114.1	58.1
Ion Ratio	100	93.3	48.8
Lower			
Upper		139.9	73.2



#18
Fluorene
Concen: 1.638 ng
RT: 15.435 min Scan# 1546
Delta R.T. -0.011 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48

Tgt Ion	166	165	167
Resp:	24341	99.5	13.2
Ion Ratio	100	79.5	10.4
Lower			
Upper		119.3	15.6





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.136 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN036413.D

Acq: 10 Feb 2025 14:48

Instrument :

BNA_N





ClientSampleId :

SSTDICC1.6

Tgt Ion:188 Resp: 8137

Ion Ratio Lower Upper

188	100		
94	0.0	0.0	0.0
80	10.5	9.8	14.6



#20

4,6-Dinitro-2-methylphenol

Concen: 1.497 ng

RT: 15.522 min Scan# 1554

Delta R.T. -0.000 min




Lab File: BN036413.D


Acq: 10 Feb 2025 14:48

Tgt Ion:198 Resp: 2725

Ion Ratio Lower Upper

198	100		
51	46.4	86.6	129.8#
105	34.9	57.5	86.3#









#21
4-Bromophenyl-phenylether
Concen: 1.379 ng
RT: 16.329 min Scan# 1619
Delta R.T. -0.000 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:248 Resp: 7690
Ion Ratio Lower Upper
248 100
250 92.4 76.1 114.1
141 74.4 71.7 107.5



#22
Hexachlorobenzene
Concen: 1.292 ng
RT: 16.453 min Scan# 1629
Delta R.T. -0.000 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48

Tgt Ion:284 Resp: 9401
Ion Ratio Lower Upper
284 100
142 39.9 33.4 50.0
249 34.3 28.6 43.0









#23
Atrazine
Concen: 1.541 ng
RT: 16.602 min Scan# 1641
Delta R.T. -0.012 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:200 Resp: 6316
Ion Ratio Lower Upper
200 100
173 27.9 23.2 34.8
215 44.8 40.0 60.0



#24
Pentachlorophenol
Concen: 1.368 ng
RT: 16.789 min Scan# 1656
Delta R.T. -0.012 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48

Tgt Ion:266 Resp: 4352
Ion Ratio Lower Upper
266 100
264 63.3 50.6 76.0
268 63.3 51.9 77.9









#25
Phenanthrene
Concen: 1.552 ng
RT: 17.173 min Scan# 1687
Delta R.T. -0.000 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:178 Resp: 37027
Ion Ratio Lower Upper
178 100
176 19.6 15.7 23.5
179 15.1 12.4 18.6



#26
Anthracene
Concen: 1.524 ng
RT: 17.273 min Scan# 1695
Delta R.T. -0.000 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48

Tgt Ion:178 Resp: 33031
Ion Ratio Lower Upper
178 100
176 18.9 14.9 22.3
179 15.2 12.4 18.6





#27

Fluoranthene-d10

Concen: 1.707 ng

RT: 19.169 min Scan# 1936

Delta R.T. -0.000 min

Lab File: BN036413.D

Acq: 10 Feb 2025 14:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:212 Resp: 35740

Ion Ratio Lower Upper

212 100

106 14.4 11.5 17.3

104 8.5 7.1 10.7



#28

Fluoranthene

Concen: 1.617 ng

RT: 19.197 min Scan# 1942

Delta R.T. -0.005 min

Lab File: BN036413.D

Acq: 10 Feb 2025 14:48




Tgt Ion:202 Resp: 45708


Ion Ratio Lower Upper

202 100

101 11.8 9.2 13.8

203 17.0 13.4 20.0





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.322 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN036413.D

Acq: 10 Feb 2025 14:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:240 Resp: 7773

Ion Ratio Lower Upper

240 100

120 12.6 13.3 19.9#

236 28.8 23.0 34.6



#30

Pyrene

Concen: 1.489 ng

RT: 19.559 min Scan# 2020

Delta R.T. -0.005 min

Lab File: BN036413.D

Acq: 10 Feb 2025 14:48




Tgt Ion:202 Resp: 46256


Ion Ratio Lower Upper

202 100

200 21.2 16.9 25.3

203 17.7 13.9 20.9









#31
Terphenyl-d14
Concen: 1.611 ng
RT: 19.768 min Scan# 2065
Delta R.T. -0.000 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48




Instrument :
BNA_N
ClientSampleId :
SSTDICC1.6


Tgt Ion:244 Resp: 25927
Ion Ratio Lower Upper
244 100
212 9.1 8.1 12.1
122 13.8 12.8 19.2



#32
Benzo(a)anthracene
Concen: 1.463 ng
RT: 21.304 min Scan# 2261
Delta R.T. -0.009 min
Lab File: BN036413.D
Acq: 10 Feb 2025 14:48

Tgt Ion:228 Resp: 40416
Ion Ratio Lower Upper
228 100
226 27.9 22.2 33.2
229 19.5 16.5 24.7





#33

Chrysene

Concen: 1.542 ng

RT: 21.357 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN036413.D

Acq: 10 Feb 2025 14:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:228 Resp: 43661

Ion Ratio Lower Upper

228 100

226 30.3 25.5 38.3

229 19.5 16.4 24.6



#34

Bis(2-ethylhexyl)phthalate

Concen: 1.538 ng

RT: 21.241 min Scan# 2254

Delta R.T. -0.000 min

Lab File: BN036413.D

Acq: 10 Feb 2025 14:48




Tgt Ion:149 Resp: 23672


Ion Ratio Lower Upper

149 100

167 26.9 21.2 31.8

279 3.0 2.7 4.1





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.592 min Scan# 21

Delta R.T. -0.003 min

Lab File: BN036413.D


Acq: 10 Feb 2025 14:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6






Tgt Ion:264 Resp: 7967

Ion Ratio Lower Upper

264 100

260 26.1 20.9 31.3

265 76.5 60.7 91.1



#36

Indeno(1,2,3-cd)pyrene

Concen: 1.472 ng

RT: 25.890 min Scan# 3576

Delta R.T. -0.009 min

Lab File: BN036413.D

Acq: 10 Feb 2025 14:48




Tgt Ion:276 Resp: 46083


Ion Ratio Lower Upper

276 100

138 29.4 22.2 33.2

277 25.3 19.8 29.6





#37

Benzo(b)fluoranthene

Concen: 1.502 ng

RT: 22.908 min Scan# 2556

Delta R.T. -0.003 min

Lab File: BN036413.D

Acq: 10 Feb 2025 14:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:252 Resp: 42491

Ion Ratio Lower Upper

252 100

253 22.6 21.9 32.9

125 13.0 15.0 22.6#



#38

Benzo(k)fluoranthene

Concen: 1.489 ng

RT: 22.952 min Scan# 2571

Delta R.T. -0.006 min

Lab File: BN036413.D

Acq: 10 Feb 2025 14:48




Tgt Ion:252 Resp: 42933


Ion Ratio Lower Upper

252 100

253 22.8 22.2 33.4

125 13.2 15.0 22.4#





#39

Benzo(a)pyrene

Concen: 1.503 ng

RT: 23.493 min Scan# 21

Delta R.T. -0.006 min

Lab File: BN036413.D

Acq: 10 Feb 2025 14:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:252 Resp: 36490

Ion Ratio Lower Upper

252 100

253 23.7 24.4 36.6#

125 15.3 18.2 27.2#



#40

Dibenzo(a,h)anthracene

Concen: 1.478 ng

RT: 25.905 min Scan# 3581

Delta R.T. -0.012 min

Lab File: BN036413.D

Acq: 10 Feb 2025 14:48




Tgt Ion:278 Resp: 36765


Ion Ratio Lower Upper

278 100

139 19.7 18.5 27.7

279 24.3 24.8 37.2#





#41

Benzo(g,h,i)perylene

Concen: 1.482 ng

RT: 26.592 min Scan# 3816

Delta R.T. -0.009 min

Lab File: BN036413.D


Acq: 10 Feb 2025 14:48

Instrument :

BNA_N

ClientSampleId :

SSTDICC1.6





Tgt Ion:276 Resp: 40441

Ion Ratio Lower Upper

276 100

277 24.3 20.7 31.1

138 26.3 21.8 32.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
 Data File : BN036414.D
 Acq On : 10 Feb 2025 15:24
 Operator : RC/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: Feb 11 00:37:03 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 00:33:05 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.753	152	2264	0.400	ng	0.00
7) Naphthalene-d8	10.541	136	5888	0.400	ng	# 0.00
13) Acenaphthene-d10	14.388	164	3870	0.400	ng	0.00
19) Phenanthrene-d10	17.136	188	8443	0.400	ng	0.00
29) Chrysene-d12	21.322	240	8450	0.400	ng	# 0.00
35) Perylene-d12	23.592	264	8327	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	18096	3.140	ng	0.00
5) Phenol-d6	6.930	99	22949	3.414	ng	0.00
8) Nitrobenzene-d5	8.897	82	19636	3.579	ng	-0.01
11) 2-Methylnaphthalene-d10	12.131	152	31489	3.909	ng	-0.01
14) 2,4,6-Tribromophenol	15.883	330	7004	2.937	ng	0.00
15) 2-Fluorobiphenyl	13.009	172	53807	3.260	ng	-0.01
27) Fluoranthene-d10	19.169	212	84968	3.911	ng	0.00
31) Terphenyl-d14	19.768	244	61695	3.526	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.261	88	7842	3.133	ng	97
3) n-Nitrosodimethylamine	3.572	42	13931	3.098	ng	# 98
6) bis(2-Chloroethyl)ether	7.176	93	22184	3.968	ng	99
9) Naphthalene	10.594	128	55857	3.321	ng	97
10) Hexachlorobutadiene	10.883	225	13261	2.516	ng	# 100
12) 2-Methylnaphthalene	12.207	142	38458	3.640	ng	98
16) Acenaphthylene	14.110	152	60797	3.395	ng	100
17) Acenaphthene	14.452	154	39414	3.214	ng	98
18) Fluorene	15.435	166	56634	3.589	ng	100
20) 4,6-Dinitro-2-methylph...	15.510	198	7256	3.841	ng	# 74
21) 4-Bromophenyl-phenylether	16.329	248	17824	3.081	ng	# 86
22) Hexachlorobenzene	16.441	284	21378	2.832	ng	97
23) Atrazine	16.602	200	15456	3.634	ng	92
24) Pentachlorophenol	16.789	266	11492	3.481	ng	100
25) Phenanthrene	17.173	178	85950	3.473	ng	99
26) Anthracene	17.260	178	78802	3.504	ng	99
28) Fluoranthene	19.197	202	108527	3.700	ng	99
30) Pyrene	19.559	202	110153	3.262	ng	100
32) Benzo(a)anthracene	21.304	228	99445	3.312	ng	99
33) Chrysene	21.358	228	103234	3.354	ng	97
34) Bis(2-ethylhexyl)phtha...	21.241	149	58234	3.480	ng	99
36) Indeno(1,2,3-cd)pyrene	25.890	276	108596	3.320	ng	97
37) Benzo(b)fluoranthene	22.905	252	101833	3.445	ng	# 88
38) Benzo(k)fluoranthene	22.952	252	102040	3.385	ng	# 88
39) Benzo(a)pyrene	23.490	252	87191	3.437	ng	# 84
40) Dibenzo(a,h)anthracene	25.905	278	86887	3.343	ng	# 90
41) Benzo(g,h,i)perylene	26.592	276	93260	3.269	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
Data File : BN036414.D
Acq On : 10 Feb 2025 15:24
Operator : RC/JU
Sample : SSTDICC3.2
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Quant Time: Feb 11 00:37:03 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 00:33:05 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.753 min Scan# 64
Delta R.T. 0.000 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 2264
Ion Ratio Lower Upper
152 100
150 154.9 123.7 185.5
115 65.7 52.5 78.7



#2
1,4-Dioxane
Concen: 3.133 ng
RT: 3.261 min Scan# 25
Delta R.T. -0.007 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24

Tgt Ion: 88 Resp: 7842
Ion Ratio Lower Upper
88 100
43 40.6 33.7 50.5
58 83.7 68.9 103.3









#3
n-Nitrosodimethylamine
Concen: 3.098 ng
RT: 3.572 min Scan# 61
Delta R.T. -0.007 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 42 Resp: 13931
Ion Ratio Lower Upper
42 100
74 91.3 71.8 107.6
44 6.4 7.8 11.6#



#4
2-Fluorophenol
Concen: 3.140 ng
RT: 5.348 min Scan# 314
Delta R.T. 0.000 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24

Tgt Ion: 112 Resp: 18096
Ion Ratio Lower Upper
112 100
64 66.2 53.4 80.0
63 36.9 30.3 45.5









#5
Phenol-d6
Concen: 3.414 ng
RT: 6.930 min Scan# 51
Delta R.T. -0.007 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 99 Resp: 22949
Ion Ratio Lower Upper
99 100
42 28.0 21.7 32.5
71 40.5 32.6 49.0



#6
bis(2-Chloroethyl)ether
Concen: 3.968 ng
RT: 7.176 min Scan# 567
Delta R.T. -0.007 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24

Tgt Ion: 93 Resp: 22184
Ion Ratio Lower Upper
93 100
63 82.3 66.3 99.5
95 31.7 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.541 min Scan# 911
Delta R.T. 0.000 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:	136	Resp:	5888
Ion	Ratio	Lower	Upper
136	100		
137	12.2	10.1	15.1
54	11.4	11.8	17.6#
68	7.3	7.2	10.8



#8
Nitrobenzene-d5
Concen: 3.579 ng
RT: 8.897 min Scan# 779
Delta R.T. -0.011 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24

Tgt Ion:	82	Resp:	19636
Ion	Ratio	Lower	Upper
82	100		
128	32.1	31.9	47.9
54	65.7	53.1	79.7









#9
Naphthalene
Concen: 3.321 ng
RT: 10.594 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:128 Resp: 55857
Ion Ratio Lower Upper
128 100
129 11.0 9.6 14.4
127 13.3 12.0 18.0



#10
Hexachlorobutadiene
Concen: 2.516 ng
RT: 10.883 min Scan# 965
Delta R.T. 0.000 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24

Tgt Ion:225 Resp: 13261
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.6 50.9 76.3









#11
2-Methylnaphthalene-d10
Concen: 3.909 ng
RT: 12.131 min Scan# 1146
Delta R.T. -0.010 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 31489
Ion Ratio Lower Upper
152 100
151 21.0 16.6 25.0



#12
2-Methylnaphthalene
Concen: 3.640 ng
RT: 12.207 min Scan# 1161
Delta R.T. -0.005 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24

Tgt Ion:142 Resp: 38458
Ion Ratio Lower Upper
142 100
141 89.8 72.8 109.2
115 41.3 35.5 53.3





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.388 min Scan# 1448

Delta R.T. 0.000 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:164 Resp: 3870

Ion Ratio Lower Upper

164 100

162 104.3 84.1 126.1

160 49.6 41.4 62.0



#14

2,4,6-Tribromophenol

Concen: 2.937 ng

RT: 15.883 min Scan# 1583

Delta R.T. 0.000 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24




Tgt Ion:330 Resp: 7004


Ion Ratio Lower Upper

330 100

332 96.5 76.6 114.8

141 48.2 37.8 56.8









#15
2-Fluorobiphenyl
Concen: 3.260 ng
RT: 13.009 min Scan# 1319
Delta R.T. -0.011 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:172 Resp: 53807
Ion Ratio Lower Upper
172 100
171 36.3 29.6 44.4
170 24.4 19.8 29.6



#16
Acenaphthylene
Concen: 3.395 ng
RT: 14.110 min Scan# 1422
Delta R.T. 0.000 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24

Tgt Ion:152 Resp: 60797
Ion Ratio Lower Upper
152 100
151 19.8 15.8 23.8
153 12.9 10.2 15.2





#17

Acenaphthene

Concen: 3.214 ng

RT: 14.452 min Scan# 1454

Delta R.T. 0.000 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:154 Resp: 39414

Ion Ratio Lower Upper

154 100

153 115.0 93.3 139.9

152 57.9 48.8 73.2



#18

Fluorene

Concen: 3.589 ng

RT: 15.435 min Scan# 1546

Delta R.T. -0.011 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24




Tgt Ion:166 Resp: 56634


Ion Ratio Lower Upper

166 100

165 99.4 79.5 119.3

167 12.6 10.4 15.6





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.136 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:188 Resp: 8443

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 10.1 9.8 14.6



#20

4,6-Dinitro-2-methylphenol

Concen: 3.841 ng

RT: 15.510 min Scan# 1553

Delta R.T. -0.012 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24




Tgt Ion:198 Resp: 7256


Ion Ratio Lower Upper

198 100

51 74.8 86.6 129.8#

105 57.0 57.5 86.3#









#21
4-Bromophenyl-phenylether
Concen: 3.081 ng
RT: 16.329 min Scan# 1619
Delta R.T. 0.000 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:248 Resp: 17824
Ion Ratio Lower Upper
248 100
250 95.9 76.1 114.1
141 63.3 71.7 107.5#



#22
Hexachlorobenzene
Concen: 2.832 ng
RT: 16.441 min Scan# 1628
Delta R.T. -0.012 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24

Tgt Ion:284 Resp: 21378
Ion Ratio Lower Upper
284 100
142 39.4 33.4 50.0
249 34.3 28.6 43.0









#23
Atrazine
Concen: 3.634 ng
RT: 16.602 min Scan# 1641
Delta R.T. -0.012 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:200 Resp: 15456
Ion Ratio Lower Upper
200 100
173 24.8 23.2 34.8
215 44.0 40.0 60.0



#24
Pentachlorophenol
Concen: 3.481 ng
RT: 16.789 min Scan# 1656
Delta R.T. -0.012 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24

Tgt Ion:266 Resp: 11492
Ion Ratio Lower Upper
266 100
264 63.3 50.6 76.0
268 64.3 51.9 77.9





#25

Phenanthrene

Concen: 3.473 ng

RT: 17.173 min Scan# 1687

Delta R.T. 0.000 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:178 Resp: 85950

Ion Ratio Lower Upper

178 100

176 19.3 15.7 23.5

179 15.1 12.4 18.6



#26

Anthracene

Concen: 3.504 ng

RT: 17.260 min Scan# 1694

Delta R.T. -0.012 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24




Tgt Ion:178 Resp: 78802


Ion Ratio Lower Upper

178 100

176 18.9 14.9 22.3

179 15.2 12.4 18.6





#27

Fluoranthene-d10

Concen: 3.911 ng

RT: 19.169 min Scan# 1936

Delta R.T. 0.000 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:212 Resp: 84968

Ion Ratio Lower Upper

212 100

106 14.5 11.5 17.3

104 8.6 7.1 10.7



#28

Fluoranthene

Concen: 3.700 ng

RT: 19.197 min Scan# 1942

Delta R.T. -0.005 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24




Tgt Ion:202 Resp: 108527


Ion Ratio Lower Upper

202 100

101 11.9 9.2 13.8

203 17.1 13.4 20.0





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.322 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:240 Resp: 8450

Ion Ratio Lower Upper

240 100

120 12.0 13.3 19.9#

236 27.8 23.0 34.6



#30

Pyrene

Concen: 3.262 ng

RT: 19.559 min Scan# 2020

Delta R.T. -0.005 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24




Tgt Ion:202 Resp: 110153


Ion Ratio Lower Upper

202 100

200 21.2 16.9 25.3

203 17.7 13.9 20.9









#31
Terphenyl-d14
Concen: 3.526 ng
RT: 19.768 min Scan# 2065
Delta R.T. 0.000 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:244 Resp: 61695
Ion Ratio Lower Upper
244 100
212 8.9 8.1 12.1
122 12.8 12.8 19.2#



#32
Benzo(a)anthracene
Concen: 3.312 ng
RT: 21.304 min Scan# 2261
Delta R.T. -0.009 min
Lab File: BN036414.D
Acq: 10 Feb 2025 15:24

Tgt Ion:228 Resp: 99445
Ion Ratio Lower Upper
228 100
226 27.6 22.2 33.2
229 19.5 16.5 24.7





#33

Chrysene

Concen: 3.354 ng

RT: 21.358 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:228 Resp: 103234

Ion Ratio Lower Upper

228 100

226 30.1 25.5 38.3

229 19.5 16.4 24.6



#34

Bis(2-ethylhexyl)phthalate

Concen: 3.480 ng

RT: 21.241 min Scan# 2254

Delta R.T. 0.000 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24




Tgt Ion:149 Resp: 58234


Ion Ratio Lower Upper

149 100

167 27.1 21.2 31.8

279 3.0 2.7 4.1





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.592 min Scan# 21

Delta R.T. -0.003 min

Lab File: BN036414.D


Acq: 10 Feb 2025 15:24

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2






Tgt Ion:264 Resp: 8327

Ion Ratio Lower Upper

264 100

260 25.4 20.9 31.3

265 73.1 60.7 91.1



#36

Indeno(1,2,3-cd)pyrene

Concen: 3.320 ng

RT: 25.890 min Scan# 3576

Delta R.T. -0.009 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24




Tgt Ion:276 Resp: 108596


Ion Ratio Lower Upper

276 100

138 29.9 22.2 33.2

277 25.1 19.8 29.6





#37

Benzo(b)fluoranthene

Concen: 3.445 ng

RT: 22.905 min Scan# 2555

Delta R.T. -0.006 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:252 Resp: 101833

Ion Ratio Lower Upper

252 100

253 22.1 21.9 32.9

125 12.5 15.0 22.6#



#38

Benzo(k)fluoranthene

Concen: 3.385 ng

RT: 22.952 min Scan# 2571

Delta R.T. -0.006 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24




Tgt Ion:252 Resp: 102040


Ion Ratio Lower Upper

252 100

253 22.2 22.2 33.4#

125 12.2 15.0 22.4#





#39

Benzo(a)pyrene

Concen: 3.437 ng

RT: 23.490 min Scan# 21

Delta R.T. -0.009 min

Lab File: BN036414.D


Acq: 10 Feb 2025 15:24

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:252 Resp: 87191

Ion Ratio Lower Upper

252 100

253 22.2 24.4 36.6#

125 14.0 18.2 27.2#



#40

Dibenzo(a,h)anthracene

Concen: 3.343 ng

RT: 25.905 min Scan# 3581

Delta R.T. -0.012 min

Lab File: BN036414.D

Acq: 10 Feb 2025 15:24




Tgt Ion:278 Resp: 86887


Ion Ratio Lower Upper

278 100

139 20.0 18.5 27.7

279 23.7 24.8 37.2#





#41

Benzo(g,h,i)perylene

Concen: 3.269 ng

RT: 26.592 min Scan# 3816

Delta R.T. -0.009 min

Lab File: BN036414.D


Acq: 10 Feb 2025 15:24

Instrument :

BNA_N

ClientSampleId :

SSTDICC3.2





Tgt Ion:276 Resp: 93260

Ion Ratio Lower Upper

276 100

277 23.7 20.7 31.1

138 25.9 21.8 32.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
 Data File : BN036415.D
 Acq On : 10 Feb 2025 16:00
 Operator : RC/JU
 Sample : SSTDICC5.0
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Quant Time: Feb 11 00:37:30 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 00:33:05 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.753	152	2586	0.400	ng	0.00
7) Naphthalene-d8	10.541	136	6761	0.400	ng	# 0.00
13) Acenaphthene-d10	14.388	164	4542	0.400	ng	0.00
19) Phenanthrene-d10	17.136	188	9953	0.400	ng	# 0.00
29) Chrysene-d12	21.322	240	9960	0.400	ng	# 0.00
35) Perylene-d12	23.592	264	9411	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	28619	4.348	ng	0.00
5) Phenol-d6	6.930	99	37616	4.898	ng	0.00
8) Nitrobenzene-d5	8.897	82	32241	5.118	ng	-0.01
11) 2-Methylnaphthalene-d10	12.131	152	52187	5.642	ng	-0.01
14) 2,4,6-Tribromophenol	15.883	330	12444	4.447	ng	0.00
15) 2-Fluorobiphenyl	13.009	172	88478	4.567	ng	-0.01
27) Fluoranthene-d10	19.164	212	143795	5.615	ng	0.00
31) Terphenyl-d14	19.764	244	104936	5.088	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.261	88	12303	4.303	ng	97
3) n-Nitrosodimethylamine	3.572	42	21660	4.216	ng	# 99
6) bis(2-Chloroethyl)ether	7.176	93	35791	5.604	ng	99
9) Naphthalene	10.584	128	90689	4.696	ng	97
10) Hexachlorobutadiene	10.883	225	21383	3.534	ng	# 100
12) 2-Methylnaphthalene	12.207	142	63405	5.226	ng	97
16) Acenaphthylene	14.110	152	103358	4.917	ng	99
17) Acenaphthene	14.452	154	66375	4.612	ng	97
18) Fluorene	15.435	166	93442	5.045	ng	96
21) 4-Bromophenyl-phenylether	16.329	248	29550	4.334	ng	# 86
22) Hexachlorobenzene	16.441	284	35446	3.983	ng	97
23) Atrazine	16.602	200	26508	5.288	ng	92
24) Pentachlorophenol	16.789	266	20784	5.340	ng	99
25) Phenanthrene	17.173	178	143424	4.916	ng	100
26) Anthracene	17.260	178	135332	5.105	ng	100
28) Fluoranthene	19.197	202	181794	5.257	ng	99
30) Pyrene	19.559	202	185711	4.666	ng	100
32) Benzo(a)anthracene	21.304	228	169514	4.790	ng	99
33) Chrysene	21.358	228	170028	4.687	ng	97
34) Bis(2-ethylhexyl)phtha...	21.241	149	101929	5.168	ng	99
36) Indeno(1,2,3-cd)pyrene	25.890	276	172989	4.679	ng	97
37) Benzo(b)fluoranthene	22.905	252	166596	4.986	ng	# 88
38) Benzo(k)fluoranthene	22.952	252	166186	4.879	ng	# 88
39) Benzo(a)pyrene	23.490	252	141837	4.947	ng	# 83
40) Dibenzo(a,h)anthracene	25.905	278	138330	4.709	ng	# 89
41) Benzo(g,h,i)perylene	26.595	276	146932	4.557	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
Data File : BN036415.D
Acq On : 10 Feb 2025 16:00
Operator : RC/JU
Sample : SSTDICC5.0
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Quant Time: Feb 11 00:37:30 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 00:33:05 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.753 min Scan# 64
Delta R.T. 0.000 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 152 Resp: 2586
Ion Ratio Lower Upper
152 100
150 155.8 123.7 185.5
115 66.1 52.5 78.7



#2
1,4-Dioxane
Concen: 4.303 ng
RT: 3.261 min Scan# 25
Delta R.T. -0.007 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00

Tgt Ion: 88 Resp: 12303
Ion Ratio Lower Upper
88 100
43 40.4 33.7 50.5
58 82.9 68.9 103.3









#3
n-Nitrosodimethylamine
Concen: 4.216 ng
RT: 3.572 min Scan# 61
Delta R.T. -0.007 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 42 Resp: 21660
Ion Ratio Lower Upper
42 100
74 89.7 71.8 107.6
44 5.6 7.8 11.6#



#4
2-Fluorophenol
Concen: 4.348 ng
RT: 5.348 min Scan# 314
Delta R.T. 0.000 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00

Tgt Ion: 112 Resp: 28619
Ion Ratio Lower Upper
112 100
64 66.1 53.4 80.0
63 37.1 30.3 45.5









#5
Phenol-d6
Concen: 4.898 ng
RT: 6.930 min Scan# 51
Delta R.T. -0.007 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion: 99 Resp: 37616
Ion Ratio Lower Upper
99 100
42 27.6 21.7 32.5
71 40.0 32.6 49.0



#6
bis(2-Chloroethyl)ether
Concen: 5.604 ng
RT: 7.176 min Scan# 567
Delta R.T. -0.007 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00

Tgt Ion: 93 Resp: 35791
Ion Ratio Lower Upper
93 100
63 81.8 66.3 99.5
95 31.7 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.541 min Scan# 911
Delta R.T. 0.000 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:	136	Resp:	6761
Ion	Ratio	Lower	Upper
136	100		
137	11.7	10.1	15.1
54	10.5	11.8	17.6#
68	7.1	7.2	10.8#



#8
Nitrobenzene-d5
Concen: 5.118 ng
RT: 8.897 min Scan# 779
Delta R.T. -0.011 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00

Tgt Ion:	82	Resp:	32241
Ion	Ratio	Lower	Upper
82	100		
128	33.1	31.9	47.9
54	64.0	53.1	79.7









#9
Naphthalene
Concen: 4.696 ng
RT: 10.584 min Scan# 91
Delta R.T. -0.011 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:128 Resp: 90689
Ion Ratio Lower Upper
128 100
129 10.9 9.6 14.4
127 13.6 12.0 18.0



#10
Hexachlorobutadiene
Concen: 3.534 ng
RT: 10.883 min Scan# 965
Delta R.T. 0.000 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00

Tgt Ion:225 Resp: 21383
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.5 50.9 76.3









#11
2-Methylnaphthalene-d10
Concen: 5.642 ng
RT: 12.131 min Scan# 1148
Delta R.T. -0.010 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:152 Resp: 52187
Ion Ratio Lower Upper
152 100
151 21.0 16.6 25.0



#12
2-Methylnaphthalene
Concen: 5.226 ng
RT: 12.207 min Scan# 1161
Delta R.T. -0.005 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00

Tgt Ion:142 Resp: 63405
Ion Ratio Lower Upper
142 100
141 89.5 72.8 109.2
115 40.5 35.5 53.3









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.388 min Scan# 1448
Delta R.T. 0.000 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:164 Resp: 4542
Ion Ratio Lower Upper
164 100
162 102.9 84.1 126.1
160 50.4 41.4 62.0



#14
2,4,6-Tribromophenol
Concen: 4.447 ng
RT: 15.883 min Scan# 1583
Delta R.T. 0.000 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00

Tgt Ion:330 Resp: 12444
Ion Ratio Lower Upper
330 100
332 95.6 76.6 114.8
141 46.6 37.8 56.8





#15

2-Fluorobiphenyl

Concen: 4.567 ng

RT: 13.009 min Scan# 1319

Delta R.T. -0.011 min

Lab File: BN036415.D

Acq: 10 Feb 2025 16:00

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:172 Resp: 88478

Ion Ratio Lower Upper

172 100

171 36.2 29.6 44.4

170 24.0 19.8 29.6



#16

Acenaphthylene

Concen: 4.917 ng

RT: 14.110 min Scan# 1422

Delta R.T. 0.000 min

Lab File: BN036415.D

Acq: 10 Feb 2025 16:00




Tgt Ion:152 Resp: 103358


Ion Ratio Lower Upper

152 100

151 20.1 15.8 23.8

153 12.9 10.2 15.2









#17
Acenaphthene
Concen: 4.612 ng
RT: 14.452 min Scan# 1454
Delta R.T. 0.000 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:154 Resp: 66375
Ion Ratio Lower Upper
154 100
153 114.4 93.3 139.9
152 57.1 48.8 73.2



#18
Fluorene
Concen: 5.045 ng
RT: 15.435 min Scan# 1546
Delta R.T. -0.011 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00

Tgt Ion:166 Resp: 93442
Ion Ratio Lower Upper
166 100
165 95.5 79.5 119.3
167 12.8 10.4 15.6





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.136 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN036415.D

Acq: 10 Feb 2025 16:00

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:188 Resp: 9953

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 9.3 9.8 14.6#



#21

4-Bromophenyl-phenylether

Concen: 4.334 ng

RT: 16.329 min Scan# 1619

Delta R.T. 0.000 min

Lab File: BN036415.D

Acq: 10 Feb 2025 16:00




Tgt Ion:248 Resp: 29550


Ion Ratio Lower Upper

248 100

250 95.8 76.1 114.1

141 62.8 71.7 107.5#





#22

Hexachlorobenzene

Concen: 3.983 ng

RT: 16.441 min Scan# 1628

Delta R.T. -0.012 min

Lab File: BN036415.D

Acq: 10 Feb 2025 16:00

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:284 Resp: 35446

Ion Ratio Lower Upper

284 100

142 39.2 33.4 50.0

249 34.2 28.6 43.0



#23

Atrazine

Concen: 5.288 ng

RT: 16.602 min Scan# 1641

Delta R.T. -0.012 min

Lab File: BN036415.D

Acq: 10 Feb 2025 16:00




Tgt Ion:200 Resp: 26508


Ion Ratio Lower Upper

200 100

173 24.1 23.2 34.8

215 45.2 40.0 60.0





#24

Pentachlorophenol

Concen: 5.340 ng

RT: 16.789 min Scan# 1656

Delta R.T. -0.012 min

Lab File: BN036415.D

Acq: 10 Feb 2025 16:00

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:266 Resp: 20784

Ion Ratio Lower Upper

266 100

264 63.1 50.6 76.0

268 63.4 51.9 77.9



#25

Phenanthrene

Concen: 4.916 ng

RT: 17.173 min Scan# 1687

Delta R.T. 0.000 min

Lab File: BN036415.D

Acq: 10 Feb 2025 16:00




Tgt Ion:178 Resp: 143424


Ion Ratio Lower Upper

178 100

176 19.5 15.7 23.5

179 15.2 12.4 18.6









#26
Anthracene
Concen: 5.105 ng
RT: 17.260 min Scan# 1694
Delta R.T. -0.012 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:178 Resp: 135332
Ion Ratio Lower Upper
178 100
176 18.7 14.9 22.3
179 15.2 12.4 18.6



#27
Fluoranthene-d10
Concen: 5.615 ng
RT: 19.164 min Scan# 1935
Delta R.T. -0.005 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00

Tgt Ion:212 Resp: 143795
Ion Ratio Lower Upper
212 100
106 14.7 11.5 17.3
104 8.6 7.1 10.7









#28
Fluoranthene
Concen: 5.257 ng
RT: 19.197 min Scan# 1943
Delta R.T. -0.005 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:202 Resp: 181794
Ion Ratio Lower Upper
202 100
101 12.0 9.2 13.8
203 17.3 13.4 20.0



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.322 min Scan# 2263
Delta R.T. 0.000 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00

Tgt Ion:240 Resp: 9960
Ion Ratio Lower Upper
240 100
120 11.8 13.3 19.9#
236 27.8 23.0 34.6









#30
Pyrene
Concen: 4.666 ng
RT: 19.559 min Scan# 20
Delta R.T. -0.005 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:202 Resp: 185711
Ion Ratio Lower Upper
202 100
200 21.2 16.9 25.3
203 17.7 13.9 20.9



#31
Terphenyl-d14
Concen: 5.088 ng
RT: 19.764 min Scan# 2064
Delta R.T. -0.005 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00

Tgt Ion:244 Resp: 104936
Ion Ratio Lower Upper
244 100
212 8.9 8.1 12.1
122 14.8 12.8 19.2





#32

Benzo(a)anthracene

Concen: 4.790 ng

RT: 21.304 min Scan# 21

Delta R.T. -0.009 min

Lab File: BN036415.D

Acq: 10 Feb 2025 16:00

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:228 Resp: 169514

Ion Ratio Lower Upper

228 100

226 27.5 22.2 33.2

229 19.4 16.5 24.7



#33

Chrysene

Concen: 4.687 ng

RT: 21.358 min Scan# 2267

Delta R.T. 0.000 min

Lab File: BN036415.D

Acq: 10 Feb 2025 16:00




Tgt Ion:228 Resp: 170028


Ion Ratio Lower Upper

228 100

226 30.0 25.5 38.3

229 19.3 16.4 24.6









#34
Bis(2-ethylhexyl)phthalate
Concen: 5.168 ng
RT: 21.241 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0


Tgt Ion:149 Resp: 101929
Ion Ratio Lower Upper
149 100
167 26.9 21.2 31.8
279 2.9 2.7 4.1



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.592 min Scan# 2790
Delta R.T. -0.003 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00

Tgt Ion:264 Resp: 9411
Ion Ratio Lower Upper
264 100
260 25.0 20.9 31.3
265 69.8 60.7 91.1





#36

Indeno(1,2,3-cd)pyrene

Concen: 4.679 ng

RT: 25.890 min Scan# 3579

Delta R.T. -0.009 min

Lab File: BN036415.D

Acq: 10 Feb 2025 16:00

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:276 Resp: 172989

Ion Ratio Lower Upper

276 100

138 30.0 22.2 33.2

277 25.2 19.8 29.6



#37

Benzo(b)fluoranthene

Concen: 4.986 ng

RT: 22.905 min Scan# 2555

Delta R.T. -0.006 min

Lab File: BN036415.D

Acq: 10 Feb 2025 16:00




Tgt Ion:252 Resp: 166596


Ion Ratio Lower Upper

252 100

253 21.9 21.9 32.9#


125 12.2 15.0 22.6#








#38
Benzo(k)fluoranthene
Concen: 4.879 ng
RT: 22.952 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0







Tgt Ion:252 Resp: 166186
Ion Ratio Lower Upper
252 100
253 22.0 22.2 33.4#
125 12.1 15.0 22.4#



#39
Benzo(a)pyrene
Concen: 4.947 ng
RT: 23.490 min Scan# 2755
Delta R.T. -0.009 min
Lab File: BN036415.D
Acq: 10 Feb 2025 16:00

Tgt Ion:252 Resp: 141837
Ion Ratio Lower Upper
252 100
253 21.8 24.4 36.6#
125 13.5 18.2 27.2#





#40

Dibenzo(a,h)anthracene

Concen: 4.709 ng

RT: 25.905 min Scan# 3581

Delta R.T. -0.012 min

Lab File: BN036415.D

Acq: 10 Feb 2025 16:00

Instrument :

BNA_N

ClientSampleId :

SSTDICC5.0





Tgt Ion:278 Resp: 138330

Ion Ratio Lower Upper

278 100

139 19.8 18.5 27.7

279 23.5 24.8 37.2#



#41

Benzo(g,h,i)perylene

Concen: 4.557 ng

RT: 26.595 min Scan# 3817

Delta R.T. -0.006 min

Lab File: BN036415.D

Acq: 10 Feb 2025 16:00




Tgt Ion:276 Resp: 146932

Ion Ratio Lower Upper

276 100

277 23.6 20.7 31.1

138 25.5 21.8 32.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
 Data File : BN036416.D
 Acq On : 10 Feb 2025 16:36
 Operator : RC/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN021025

Quant Time: Feb 11 01:25:46 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 01:17:14 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.753	152	2722	0.400	ng	0.00
7) Naphthalene-d8	10.541	136	6737	0.400	ng	0.00
13) Acenaphthene-d10	14.388	164	4328	0.400	ng	0.00
19) Phenanthrene-d10	17.136	188	9717	0.400	ng	0.00
29) Chrysene-d12	21.322	240	8903	0.400	ng	0.00
35) Perylene-d12	23.592	264	8977	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.348	112	2386	0.371	ng	0.00
5) Phenol-d6	6.937	99	2616	0.347	ng	0.00
8) Nitrobenzene-d5	8.907	82	2386	0.359	ng	0.00
11) 2-Methylnaphthalene-d10	12.136	152	4026	0.389	ng	0.00
14) 2,4,6-Tribromophenol	15.882	330	713	0.332	ng	0.00
15) 2-Fluorobiphenyl	13.019	172	6616	0.407	ng	0.00
27) Fluoranthene-d10	19.169	212	10243	0.379	ng	0.00
31) Terphenyl-d14	19.768	244	7744	0.407	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.268	88	1221	0.410	ng	Qvalue 98
3) n-Nitrosodimethylamine	3.579	42	1993	0.385	ng	96
6) bis(2-Chloroethyl)ether	7.183	93	3230	0.409	ng	96
9) Naphthalene	10.594	128	7554	0.389	ng	99
10) Hexachlorobutadiene	10.882	225	1910	0.404	ng	# 99
12) 2-Methylnaphthalene	12.212	142	5146	0.404	ng	98
16) Acenaphthylene	14.110	152	8061	0.422	ng	100
17) Acenaphthene	14.452	154	5199	0.407	ng	99
18) Fluorene	15.446	166	7222	0.397	ng	98
20) 4,6-Dinitro-2-methylph...	15.522	198	664	0.348	ng	92
21) 4-Bromophenyl-phenylether	16.329	248	2290	0.395	ng	96
22) Hexachlorobenzene	16.453	284	2904	0.406	ng	97
23) Atrazine	16.615	200	1916	0.396	ng	98
24) Pentachlorophenol	16.801	266	1075	0.316	ng	99
25) Phenanthrene	17.173	178	11446	0.408	ng	100
26) Anthracene	17.273	178	10171	0.411	ng	99
28) Fluoranthene	19.197	202	13415	0.389	ng	100
30) Pyrene	19.559	202	13881	0.405	ng	100
32) Benzo(a)anthracene	21.304	228	11819	0.403	ng	99
33) Chrysene	21.357	228	13425	0.423	ng	98
34) Bis(2-ethylhexyl)phtha...	21.241	149	6985	0.383	ng	98
36) Indeno(1,2,3-cd)pyrene	25.896	276	13738	0.438	ng	100
37) Benzo(b)fluoranthene	22.908	252	12283	0.415	ng	97
38) Benzo(k)fluoranthene	22.955	252	13099	0.430	ng	97
39) Benzo(a)pyrene	23.493	252	11703	0.454	ng	95
40) Dibenzo(a,h)anthracene	25.908	278	10585	0.427	ng	96
41) Benzo(g,h,i)perylene	26.595	276	10921	0.389	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
Data File : BN036416.D
Acq On : 10 Feb 2025 16:36
Operator : RC/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN021025

Quant Time: Feb 11 01:25:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 01:17:14 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.753 min Scan# 64
Delta R.T. 0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36




Instrument :
BNA_N
ClientSampleId :
ICVBN021025


Tgt Ion:152 Resp: 2722
Ion Ratio Lower Upper
152 100
150 154.0 123.7 185.5
115 66.1 52.5 78.7



#2
1,4-Dioxane
Concen: 0.410 ng
RT: 3.268 min Scan# 26
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36

Tgt Ion: 88 Resp: 1221
Ion Ratio Lower Upper
88 100
43 41.8 33.7 50.5
58 89.2 68.9 103.3









#3
n-Nitrosodimethylamine
Concen: 0.385 ng
RT: 3.579 min Scan# 69
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36




Instrument :
BNA_N
ClientSampleId :
ICVBN021025


Tgt Ion: 42 Resp: 1993
Ion Ratio Lower Upper
42 100
74 93.5 71.8 107.6
44 7.9 7.8 11.6



#4
2-Fluorophenol
Concen: 0.371 ng
RT: 5.348 min Scan# 314
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36

Tgt Ion: 112 Resp: 2386
Ion Ratio Lower Upper
112 100
64 66.4 53.4 80.0
63 36.3 30.3 45.5









#5
Phenol-d6
Concen: 0.347 ng
RT: 6.937 min Scan# 51
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36




Instrument :
BNA_N
ClientSampleId :
ICVBN021025


Tgt Ion:	99	Resp:	2616
Ion	Ratio	Lower	Upper
99	100		
42	28.4	21.7	32.5
71	39.6	32.6	49.0



#6
bis(2-Chloroethyl)ether
Concen: 0.409 ng
RT: 7.183 min Scan# 568
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36

Tgt Ion:	93	Resp:	3230
Ion	Ratio	Lower	Upper
93	100		
63	78.7	66.3	99.5
95	31.0	26.2	39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.541 min Scan# 911
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36




Instrument :
BNA_N
ClientSampleId :
ICVBN021025


Tgt Ion:	136	Resp:	6737
Ion	Ratio	Lower	Upper
136	100		
137	11.9	10.1	15.1
54	12.7	11.8	17.6
68	8.6	7.2	10.8



#8
Nitrobenzene-d5
Concen: 0.359 ng
RT: 8.907 min Scan# 780
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36

Tgt Ion:	82	Resp:	2386
Ion	Ratio	Lower	Upper
82	100		
128	39.6	31.9	47.9
54	61.1	53.1	79.7



#9
Naphthalene

Concen: 0.389 ng

RT: 10.594 min Scan# 91

Delta R.T. -0.000 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36

Instrument :

BNA_N

ClientSampleId :

ICVBN021025





Tgt Ion:128 Resp: 7554

Ion Ratio Lower Upper

128 100

129 11.8 9.6 14.4

127 14.3 12.0 18.0



#10

Hexachlorobutadiene

Concen: 0.404 ng

RT: 10.882 min Scan# 965

Delta R.T. -0.000 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36




Tgt Ion:225 Resp: 1910


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 64.6 50.9 76.3









#11
2-Methylnaphthalene-d10
Concen: 0.389 ng
RT: 12.136 min Scan# 1147
Delta R.T. -0.005 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36




Instrument :
BNA_N
ClientSampleId :
ICVBN021025


Tgt Ion:152 Resp: 4026
Ion Ratio Lower Upper
152 100
151 20.9 16.6 25.0



#12
2-Methylnaphthalene
Concen: 0.404 ng
RT: 12.212 min Scan# 1162
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36

Tgt Ion:142 Resp: 5146
Ion Ratio Lower Upper
142 100
141 89.3 72.8 109.2
115 41.9 35.5 53.3





#13

Acenaphthene-d10

Concen: 0.400 ng

RT: 14.388 min Scan# 1448

Delta R.T. -0.000 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36

Instrument :

BNA_N

ClientSampleId :

ICVBN021025





Tgt Ion:164 Resp: 4328

Ion Ratio Lower Upper

164 100

162 104.9 84.1 126.1

160 51.1 41.4 62.0



#14

2,4,6-Tribromophenol

Concen: 0.332 ng

RT: 15.882 min Scan# 1583

Delta R.T. -0.000 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36




Tgt Ion:330 Resp: 713


Ion Ratio Lower Upper

330 100

332 97.6 76.6 114.8

141 48.9 37.8 56.8





#15

2-Fluorobiphenyl

Concen: 0.407 ng

RT: 13.019 min Scan# 11

Delta R.T. -0.000 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36

Instrument :

BNA_N

ClientSampleId :

ICVBN021025





Tgt Ion:172 Resp: 6616

Ion Ratio Lower Upper

172 100

171 35.4 29.6 44.4

170 23.8 19.8 29.6



#16

Acenaphthylene

Concen: 0.422 ng

RT: 14.110 min Scan# 1422

Delta R.T. -0.000 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36




Tgt Ion:152 Resp: 8061


Ion Ratio Lower Upper

152 100

151 19.8 15.8 23.8

153 12.9 10.2 15.2









#17
Acenaphthene
Concen: 0.407 ng
RT: 14.452 min Scan# 1454
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36




Instrument :
BNA_N
ClientSampleId :
ICVBN021025


Tgt Ion	154	Resp	5199
Ion Ratio	Lower	Upper	
154	100		
153	117.1	93.3	139.9
152	60.5	48.8	73.2



#18
Fluorene
Concen: 0.397 ng
RT: 15.446 min Scan# 1547
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36

Tgt Ion	166	Resp	7222
Ion Ratio	Lower	Upper	
166	100		
165	101.4	79.5	119.3
167	13.7	10.4	15.6





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.136 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36

Instrument :

BNA_N

ClientSampleId :

ICVBN021025





Tgt Ion:188 Resp: 9717

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 11.3 9.8 14.6



#20

4,6-Dinitro-2-methylphenol

Concen: 0.348 ng

RT: 15.522 min Scan# 1554

Delta R.T. -0.001 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36




Tgt Ion:198 Resp: 664


Ion Ratio Lower Upper

198 100

51 103.2 86.6 129.8

105 60.5 57.5 86.3









#21
4-Bromophenyl-phenylether
Concen: 0.395 ng
RT: 16.329 min Scan# 1619
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36




Instrument :
BNA_N
ClientSampleId :
ICVBN021025


Tgt Ion:248 Resp: 2290
Ion Ratio Lower Upper
248 100
250 91.7 76.1 114.1
141 85.4 71.7 107.5



#22
Hexachlorobenzene
Concen: 0.406 ng
RT: 16.453 min Scan# 1629
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36

Tgt Ion:284 Resp: 2904
Ion Ratio Lower Upper
284 100
142 39.4 33.4 50.0
249 34.2 28.6 43.0









#23
Atrazine
Concen: 0.396 ng
RT: 16.615 min Scan# 1642
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36




Instrument :
BNA_N
ClientSampleId :
ICVBN021025


Tgt Ion	Ratio	Lower	Upper
200	100		
173	27.7	23.2	34.8
215	49.2	40.0	60.0



#24
Pentachlorophenol
Concen: 0.316 ng
RT: 16.801 min Scan# 1657
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36

Tgt Ion	Ratio	Lower	Upper
266	100		
264	62.2	50.6	76.0
268	63.9	51.9	77.9









#25
Phenanthrene
Concen: 0.408 ng
RT: 17.173 min Scan# 1687
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36




Instrument :
BNA_N
ClientSampleId :
ICVBN021025


Tgt Ion:178 Resp: 11446
Ion Ratio Lower Upper
178 100
176 19.5 15.7 23.5
179 15.3 12.4 18.6



#26
Anthracene
Concen: 0.411 ng
RT: 17.273 min Scan# 1695
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36

Tgt Ion:178 Resp: 10171
Ion Ratio Lower Upper
178 100
176 18.8 14.9 22.3
179 15.2 12.4 18.6





#27

Fluoranthene-d10

Concen: 0.379 ng

RT: 19.169 min Scan# 1936

Delta R.T. -0.000 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36

Instrument :

BNA_N

ClientSampleId :

ICVBN021025





Tgt Ion:212 Resp: 10243

Ion Ratio Lower Upper

212 100

106 14.6 11.5 17.3

104 8.6 7.1 10.7



#28

Fluoranthene

Concen: 0.389 ng

RT: 19.197 min Scan# 1942

Delta R.T. -0.005 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36




Tgt Ion:202 Resp: 13415


Ion Ratio Lower Upper

202 100

101 11.4 9.2 13.8

203 16.9 13.4 20.0





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.322 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36

Instrument :

BNA_N

ClientSampleId :

ICVBN021025





Tgt Ion:240 Resp: 8903

Ion Ratio Lower Upper

240 100

120 13.4 13.3 19.9

236 28.2 23.0 34.6



#30

Pyrene

Concen: 0.405 ng

RT: 19.559 min Scan# 2020

Delta R.T. -0.005 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36




Tgt Ion:202 Resp: 13881


Ion Ratio Lower Upper

202 100

200 21.2 16.9 25.3

203 17.4 13.9 20.9









#31
Terphenyl-d14
Concen: 0.407 ng
RT: 19.768 min Scan# 2065
Delta R.T. -0.000 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36




Instrument :
BNA_N
ClientSampleId :
ICVBN021025


Tgt Ion	Ratio	Lower	Upper
244	100		
212	9.8	8.1	12.1
122	15.7	12.8	19.2



#32
Benzo(a)anthracene
Concen: 0.403 ng
RT: 21.304 min Scan# 2261
Delta R.T. -0.009 min
Lab File: BN036416.D
Acq: 10 Feb 2025 16:36

Tgt Ion	Ratio	Lower	Upper
228	100		
226	28.2	22.2	33.2
229	19.9	16.5	24.7





#33

Chrysene

Concen: 0.423 ng

RT: 21.357 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36

Instrument :

BNA_N

ClientSampleId :

ICVBN021025





Tgt Ion:228 Resp: 13425

Ion Ratio Lower Upper

228 100

226 30.7 25.5 38.3

229 19.7 16.4 24.6



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.383 ng

RT: 21.241 min Scan# 2254

Delta R.T. -0.000 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36




Tgt Ion:149 Resp: 6985


Ion Ratio Lower Upper

149 100

167 27.5 21.2 31.8

279 3.5 2.7 4.1





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.592 min Scan# 21

Delta R.T. -0.003 min

Lab File: BN036416.D


Acq: 10 Feb 2025 16:36

Instrument :

BNA_N

ClientSampleId :

ICVBN021025






Tgt Ion:264 Resp: 8977

Ion Ratio Lower Upper

264 100

260 25.5 20.9 31.3

265 66.9 60.7 91.1



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.438 ng

RT: 25.896 min Scan# 3578

Delta R.T. -0.003 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36




Tgt Ion:276 Resp: 13738


Ion Ratio Lower Upper

276 100

138 27.6 22.2 33.2

277 24.3 19.8 29.6





#37

Benzo(b)fluoranthene

Concen: 0.415 ng

RT: 22.908 min Scan# 21

Delta R.T. -0.003 min

Lab File: BN036416.D


Acq: 10 Feb 2025 16:36

Instrument :

BNA_N

ClientSampleId :

ICVBN021025





Tgt Ion:252 Resp: 12283

Ion Ratio Lower Upper

252 100

253 26.8 21.9 32.9

125 21.6 15.0 22.6



#38

Benzo(k)fluoranthene

Concen: 0.430 ng

RT: 22.955 min Scan# 2572

Delta R.T. -0.003 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36




Tgt Ion:252 Resp: 13099


Ion Ratio Lower Upper

252 100

253 26.5 22.2 33.4

125 17.6 15.0 22.4





#39

Benzo(a)pyrene

Concen: 0.454 ng

RT: 23.493 min Scan# 21

Delta R.T. -0.006 min

Lab File: BN036416.D


Acq: 10 Feb 2025 16:36

Instrument :

BNA_N

ClientSampleId :

ICVBN021025





Tgt Ion:252 Resp: 11703

Ion Ratio Lower Upper

252 100

253 29.0 24.4 36.6

125 26.4 18.2 27.2



#40

Dibenzo(a,h)anthracene


Concen: 0.427 ng

RT: 25.908 min Scan# 3582

Delta R.T. -0.009 min

Lab File: BN036416.D

Acq: 10 Feb 2025 16:36





Tgt Ion:278 Resp: 10585


Ion Ratio Lower Upper

278 100

139 21.9 18.5 27.7

279 28.4 24.8 37.2





#41

Benzo(g,h,i)perylene

Concen: 0.389 ng

RT: 26.595 min Scan# 3817

Delta R.T. -0.006 min

Lab File: BN036416.D


Acq: 10 Feb 2025 16:36

Instrument :

BNA_N

ClientSampleId :

ICVBN021025





Tgt Ion:276 Resp: 10921

Ion Ratio Lower Upper

276 100

277 25.1 20.7 31.1

138 28.1 21.8 32.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
 Data File : BN036416.D
 Acq On : 10 Feb 2025 16:36
 Operator : RC/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN021025

Quant Time: Feb 11 01:25:46 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 01:17:14 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	123	0.00
2	1,4-Dioxane	0.438	0.449	-2.5	127	0.00
3	n-Nitrosodimethylamine	0.760	0.732	3.7	118	0.00
4 S	2-Fluorophenol	0.945	0.877	7.2	115	0.00
5 S	Phenol-d6	1.109	0.961	13.3	114	0.00
6	bis(2-Chloroethyl)ether	1.160	1.187	-2.3	134	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	122	0.00
8 S	Nitrobenzene-d5	0.395	0.354	10.4	118	0.00
9	Naphthalene	1.154	1.121	2.9	122	0.00
10	Hexachlorobutadiene	0.281	0.284	-1.1	122	0.00
11 SURR	2-Methylnaphthalene-d10	0.615	0.598	2.8	121	0.00
12	2-Methylnaphthalene	0.757	0.764	-0.9	126	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	120	0.00
14 S	2,4,6-Tribromophenol	0.198	0.165	16.7	106	0.00
15 S	2-Fluorobiphenyl	1.504	1.529	-1.7	133	0.00
16	Acenaphthylene	1.767	1.863	-5.4	132	0.00
17	Acenaphthene	1.180	1.201	-1.8	126	0.00
18	Fluorene	1.680	1.669	0.7	121	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	117	0.00
20	4,6-Dinitro-2-methylphenol	0.078	0.068	12.8	116	0.00
21	4-Bromophenyl-phenylether	0.239	0.236	1.3	119	0.00
22	Hexachlorobenzene	0.295	0.299	-1.4	123	0.00
23	Atrazine	0.199	0.197	1.0	123	0.00
24	Pentachlorophenol	0.140	0.111	20.7	106	0.00
25	Phenanthrene	1.156	1.178	-1.9	126	0.00
26	Anthracene	1.020	1.047	-2.6	126	0.00
27 SURR	Fluoranthene-d10	1.112	1.054	5.2	116	0.00
28	Fluoranthene	1.421	1.381	2.8	119	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	119	0.00
30	Pyrene	1.541	1.559	-1.2	121	0.00
31 S	Terphenyl-d14	0.854	0.870	-1.9	121	0.00
32	Benzo(a)anthracene	1.316	1.328	-0.9	122	0.00
33	Chrysene	1.425	1.508	-5.8	132	0.00
34	Bis(2-ethylhexyl)phthalate	0.820	0.785	4.3	120	0.00
35 I	Perylene-d12	1.000	1.000	0.0	116	0.00
36	Indeno(1,2,3-cd)pyrene	1.398	1.530	-9.4	129	0.00
37	Benzo(b)fluoranthene	1.317	1.368	-3.9	126	0.00
38	Benzo(k)fluoranthene	1.356	1.459	-7.6	124	0.00
39 C	Benzo(a)pyrene	1.150	1.304	-13.4	137	0.00
40	Dibenzo(a,h)anthracene	1.103	1.179	-6.9	127	0.00
41	Benzo(g,h,i)perylene	1.250	1.217	2.6	113	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
Data File : BN036416.D
Acq On : 10 Feb 2025 16:36
Operator : RC/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN021025

Quant Time: Feb 11 01:25:46 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 01:17:14 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	123	0.00
2	1,4-Dioxane	0.400	0.410	-2.5	127	0.00
3	n-Nitrosodimethylamine	0.400	0.385	3.8	118	0.00
4 S	2-Fluorophenol	0.400	0.371	7.3	115	0.00
5 S	Phenol-d6	0.400	0.347	13.3	114	0.00
6	bis(2-Chloroethyl)ether	0.400	0.409	-2.2	134	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	122	0.00
8 S	Nitrobenzene-d5	0.400	0.359	10.3	118	0.00
9	Naphthalene	0.400	0.389	2.8	122	0.00
10	Hexachlorobutadiene	0.400	0.404	-1.0	122	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.389	2.8	121	0.00
12	2-Methylnaphthalene	0.400	0.404	-1.0	126	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	120	0.00
14 S	2,4,6-Tribromophenol	0.400	0.332	17.0	106	0.00
15 S	2-Fluorobiphenyl	0.400	0.407	-1.7	133	0.00
16	Acenaphthylene	0.400	0.422	-5.5	132	0.00
17	Acenaphthene	0.400	0.407	-1.7	126	0.00
18	Fluorene	0.400	0.397	0.8	121	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	117	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.348	13.0	116	0.00
21	4-Bromophenyl-phenylether	0.400	0.395	1.3	119	0.00
22	Hexachlorobenzene	0.400	0.406	-1.5	123	0.00
23	Atrazine	0.400	0.396	1.0	123	0.00
24	Pentachlorophenol	0.400	0.316	21.0	106	0.00
25	Phenanthrene	0.400	0.408	-2.0	126	0.00
26	Anthracene	0.400	0.411	-2.7	126	0.00
27 SURR	Fluoranthene-d10	0.400	0.379	5.3	116	0.00
28	Fluoranthene	0.400	0.389	2.8	119	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	119	0.00
30	Pyrene	0.400	0.405	-1.3	121	0.00
31 S	Terphenyl-d14	0.400	0.407	-1.7	121	0.00
32	Benzo(a)anthracene	0.400	0.403	-0.8	122	0.00
33	Chrysene	0.400	0.423	-5.7	132	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.383	4.3	120	0.00
35 I	Perylene-d12	0.400	0.400	0.0	116	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.438	-9.5	129	0.00
37	Benzo(b)fluoranthene	0.400	0.415	-3.7	126	0.00
38	Benzo(k)fluoranthene	0.400	0.430	-7.5	124	0.00
39 C	Benzo(a)pyrene	0.400	0.454	-13.5	137	0.00
40	Dibenzo(a,h)anthracene	0.400	0.427	-6.7	127	0.00
41	Benzo(g,h,i)perylene	0.400	0.389	2.8	113	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG No.: Q1380

Instrument ID: BNA_N Calibration Date/Time: 02/20/2025 10:11

Lab File ID: BN036473.D Init. Calib. Date(s): 02/10/2025 02/10/2025

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 12:25 16:00

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.615	0.585		-4.9	20.0
Fluoranthene-d10	1.112	1.062		-4.5	20.0
2-Fluorophenol	0.945	0.879		-7.0	20.0
Phenol-d6	1.109	1.046		-5.7	20.0
Nitrobenzene-d5	0.395	0.399		1.0	20.0
2-Fluorobiphenyl	1.504	1.478		-1.7	20.0
2,4,6-Tribromophenol	0.198	0.165		-16.7	20.0
Terphenyl-d14	0.854	0.795		-6.9	20.0
1,4-Dioxane	0.438	0.446		1.8	20.0

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
 Data File : BN036473.D
 Acq On : 20 Feb 2025 10:11
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4

Quant Time: Feb 21 00:08:51 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 01:17:14 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.746	152	1766	0.400	ng	0.00
7) Naphthalene-d8	10.541	136	4292	0.400	ng	# 0.00
13) Acenaphthene-d10	14.387	164	2899	0.400	ng	0.00
19) Phenanthrene-d10	17.136	188	6167	0.400	ng	0.00
29) Chrysene-d12	21.313	240	5687	0.400	ng	# 0.00
35) Perylene-d12	23.586	264	5826	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.341	112	1553	0.372	ng	0.00
5) Phenol-d6	6.923	99	1848	0.377	ng	-0.01
8) Nitrobenzene-d5	8.896	82	1714	0.405	ng	-0.01
11) 2-Methylnaphthalene-d10	12.131	152	2509	0.380	ng	-0.01
14) 2,4,6-Tribromophenol	15.882	330	478	0.333	ng	0.00
15) 2-Fluorobiphenyl	13.008	172	4284	0.393	ng	-0.01
27) Fluoranthene-d10	19.164	212	6549	0.382	ng	0.00
31) Terphenyl-d14	19.768	244	4524	0.373	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.261	88	788	0.408	ng	98
3) n-Nitrosodimethylamine	3.571	42	1397	0.416	ng	# 98
6) bis(2-Chloroethyl)ether	7.168	93	2055	0.401	ng	99
9) Naphthalene	10.583	128	4883	0.394	ng	99
10) Hexachlorobutadiene	10.872	225	1213	0.403	ng	# 100
12) 2-Methylnaphthalene	12.207	142	3149	0.388	ng	97
16) Acenaphthylene	14.099	152	4699	0.367	ng	98
17) Acenaphthene	14.452	154	3240	0.379	ng	99
18) Fluorene	15.435	166	4564	0.375	ng	98
20) 4,6-Dinitro-2-methylph...	15.522	198	394	0.326	ng	96
21) 4-Bromophenyl-phenylether	16.329	248	1446	0.393	ng	# 85
22) Hexachlorobenzene	16.441	284	1837	0.404	ng	98
23) Atrazine	16.602	200	1102	0.359	ng	94
24) Pentachlorophenol	16.788	266	753	0.349	ng	99
25) Phenanthrene	17.173	178	6947	0.390	ng	99
26) Anthracene	17.260	178	6132	0.390	ng	99
28) Fluoranthene	19.192	202	8510	0.389	ng	99
30) Pyrene	19.559	202	8707	0.398	ng	99
32) Benzo(a)anthracene	21.304	228	7305	0.390	ng	99
33) Chrysene	21.348	228	8493	0.419	ng	100
34) Bis(2-ethylhexyl)phtha...	21.232	149	3979	0.341	ng	100
36) Indeno(1,2,3-cd)pyrene	25.887	276	8094	0.397	ng	98
37) Benzo(b)fluoranthene	22.902	252	8146	0.425	ng	98
38) Benzo(k)fluoranthene	22.946	252	8381	0.424	ng	99
39) Benzo(a)pyrene	23.487	252	7017	0.419	ng	96
40) Dibenzo(a,h)anthracene	25.902	278	6114	0.380	ng	99
41) Benzo(g,h,i)perylene	26.586	276	7190	0.395	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
Data File : BN036473.D
Acq On : 20 Feb 2025 10:11
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4

Quant Time: Feb 21 00:08:51 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 01:17:14 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.746 min Scan# 64
Delta R.T. -0.007 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 1766
Ion Ratio Lower Upper
152 100
150 152.7 123.7 185.5
115 66.8 52.5 78.7



#2
1,4-Dioxane
Concen: 0.408 ng
RT: 3.261 min Scan# 25
Delta R.T. -0.007 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11

Tgt Ion: 88 Resp: 788
Ion Ratio Lower Upper
88 100
43 45.8 33.7 50.5
58 86.5 68.9 103.3









#3
n-Nitrosodimethylamine
Concen: 0.416 ng
RT: 3.571 min Scan# 61
Delta R.T. -0.007 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 42 Resp: 1397
Ion Ratio Lower Upper
42 100
74 89.1 71.8 107.6
44 6.1 7.8 11.6#



#4
2-Fluorophenol
Concen: 0.372 ng
RT: 5.341 min Scan# 313
Delta R.T. -0.007 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11

Tgt Ion: 112 Resp: 1553
Ion Ratio Lower Upper
112 100
64 66.1 53.4 80.0
63 38.9 30.3 45.5









#5
Phenol-d6
Concen: 0.377 ng
RT: 6.923 min Scan# 51
Delta R.T. -0.015 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion: 99 Resp: 1848
Ion Ratio Lower Upper
99 100
42 31.0 21.7 32.5
71 41.3 32.6 49.0



#6
bis(2-Chloroethyl)ether
Concen: 0.401 ng
RT: 7.168 min Scan# 566
Delta R.T. -0.015 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11

Tgt Ion: 93 Resp: 2055
Ion Ratio Lower Upper
93 100
63 82.0 66.3 99.5
95 31.6 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.541 min Scan# 911
Delta R.T. -0.000 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	136	Resp:	4292
Ion	Ratio	Lower	Upper
136	100		
137	13.1	10.1	15.1
54	11.0	11.8	17.6#
68	7.7	7.2	10.8



#8
Nitrobenzene-d5
Concen: 0.405 ng
RT: 8.896 min Scan# 779
Delta R.T. -0.011 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11

Tgt Ion:	82	Resp:	1714
Ion	Ratio	Lower	Upper
82	100		
128	37.5	31.9	47.9
54	67.8	53.1	79.7









#9
Naphthalene
Concen: 0.394 ng
RT: 10.583 min Scan# 91
Delta R.T. -0.011 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	128	Resp:	4883
Ion	Ratio	Lower	Upper
128	100		
129	13.0	9.6	14.4
127	15.1	12.0	18.0



#10
Hexachlorobutadiene
Concen: 0.403 ng
RT: 10.872 min Scan# 964
Delta R.T. -0.011 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11

Tgt Ion:	225	Resp:	1213
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.3	50.9	76.3









#11
2-Methylnaphthalene-d10
Concen: 0.380 ng
RT: 12.131 min Scan# 1146
Delta R.T. -0.010 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:152 Resp: 2509
Ion Ratio Lower Upper
152 100
151 20.9 16.6 25.0



#12
2-Methylnaphthalene
Concen: 0.388 ng
RT: 12.207 min Scan# 1161
Delta R.T. -0.005 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11

Tgt Ion:142 Resp: 3149
Ion Ratio Lower Upper
142 100
141 87.4 72.8 109.2
115 46.2 35.5 53.3









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.387 min Scan# 1448
Delta R.T. -0.000 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:164 Resp: 2899
Ion Ratio Lower Upper
164 100
162 100.6 84.1 126.1
160 50.2 41.4 62.0



#14
2,4,6-Tribromophenol
Concen: 0.333 ng
RT: 15.882 min Scan# 1583
Delta R.T. -0.000 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11

Tgt Ion:330 Resp: 478
Ion Ratio Lower Upper
330 100
332 94.4 76.6 114.8
141 47.3 37.8 56.8









#15
2-Fluorobiphenyl
Concen: 0.393 ng
RT: 13.008 min Scan# 1319
Delta R.T. -0.011 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:172	Resp:	4284
Ion Ratio	Lower	Upper
172	100	
171	38.3	29.6 44.4
170	26.1	19.8 29.6



#16
Acenaphthylene
Concen: 0.367 ng
RT: 14.099 min Scan# 1421
Delta R.T. -0.011 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11

Tgt Ion:152	Resp:	4699
Ion Ratio	Lower	Upper
152	100	
151	20.6	15.8 23.8
153	13.3	10.2 15.2









#17
Acenaphthene
Concen: 0.379 ng
RT: 14.452 min Scan# 1454
Delta R.T. -0.000 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:154	Resp:	3240
Ion Ratio	Lower	Upper
154	100	
153	117.2	93.3 139.9
152	62.6	48.8 73.2



#18
Fluorene
Concen: 0.375 ng
RT: 15.435 min Scan# 1546
Delta R.T. -0.011 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11

Tgt Ion:166	Resp:	4564
Ion Ratio	Lower	Upper
166	100	
165	101.9	79.5 119.3
167	12.9	10.4 15.6





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.136 min Scan# 1

Delta R.T. -0.000 min

Lab File: BN036473.D

Acq: 20 Feb 2025 10:11

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:188 Resp: 6167

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 9.8 9.8 14.6



#20

4,6-Dinitro-2-methylphenol

Concen: 0.326 ng

RT: 15.522 min Scan# 1554

Delta R.T. -0.001 min

Lab File: BN036473.D

Acq: 20 Feb 2025 10:11




Tgt Ion:198 Resp: 394


Ion Ratio Lower Upper

198 100

51 108.3 86.6 129.8

105 63.1 57.5 86.3









#21
4-Bromophenyl-phenylether
Concen: 0.393 ng
RT: 16.329 min Scan# 1619
Delta R.T. -0.000 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:248 Resp: 1446
Ion Ratio Lower Upper
248 100
250 99.6 76.1 114.1
141 64.5 71.7 107.5#



#22
Hexachlorobenzene
Concen: 0.404 ng
RT: 16.441 min Scan# 1628
Delta R.T. -0.013 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11

Tgt Ion:284 Resp: 1837
Ion Ratio Lower Upper
284 100
142 42.5 33.4 50.0
249 34.7 28.6 43.0









#23
Atrazine
Concen: 0.359 ng
RT: 16.602 min Scan# 1641
Delta R.T. -0.013 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:200 Resp: 1102
Ion Ratio Lower Upper
200 100
173 31.6 23.2 34.8
215 45.4 40.0 60.0



#24
Pentachlorophenol
Concen: 0.349 ng
RT: 16.788 min Scan# 1656
Delta R.T. -0.013 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11

Tgt Ion:266 Resp: 753
Ion Ratio Lower Upper
266 100
264 62.5 50.6 76.0
268 63.7 51.9 77.9









#25
Phenanthrene
Concen: 0.390 ng
RT: 17.173 min Scan# 1687
Delta R.T. -0.000 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:178 Resp: 6947
Ion Ratio Lower Upper
178 100
176 19.9 15.7 23.5
179 15.6 12.4 18.6



#26
Anthracene
Concen: 0.390 ng
RT: 17.260 min Scan# 1694
Delta R.T. -0.013 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11

Tgt Ion:178 Resp: 6132
Ion Ratio Lower Upper
178 100
176 19.0 14.9 22.3
179 15.3 12.4 18.6









#27
Fluoranthene-d10
Concen: 0.382 ng
RT: 19.164 min Scan# 1935
Delta R.T. -0.005 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion	Ratio	Lower	Upper
212	100		
106	15.0	11.5	17.3
104	9.1	7.1	10.7



#28
Fluoranthene
Concen: 0.389 ng
RT: 19.192 min Scan# 1941
Delta R.T. -0.009 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11

Tgt Ion	Ratio	Lower	Upper
202	100		
101	12.0	9.2	13.8
203	16.9	13.4	20.0









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.313 min Scan# 21
Delta R.T. -0.009 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:240 Resp: 5687
Ion Ratio Lower Upper
240 100
120 20.0 13.3 19.9#
236 29.3 23.0 34.6



#30
Pyrene
Concen: 0.398 ng
RT: 19.559 min Scan# 2020
Delta R.T. -0.005 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11

Tgt Ion:202 Resp: 8707
Ion Ratio Lower Upper
202 100
200 21.2 16.9 25.3
203 17.8 13.9 20.9









#31
Terphenyl-d14
Concen: 0.373 ng
RT: 19.768 min Scan# 2065
Delta R.T. -0.000 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:244 Resp: 4524
Ion Ratio Lower Upper
244 100
212 11.2 8.1 12.1
122 15.9 12.8 19.2



#32
Benzo(a)anthracene
Concen: 0.390 ng
RT: 21.304 min Scan# 2261
Delta R.T. -0.009 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11

Tgt Ion:228 Resp: 7305
Ion Ratio Lower Upper
228 100
226 28.3 22.2 33.2
229 20.2 16.5 24.7









#33
Chrysene
Concen: 0.419 ng
RT: 21.348 min Scan# 21
Delta R.T. -0.009 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:	228	Resp:	8493
Ion Ratio	Lower	Upper	
228	100		
226	31.8	25.5	38.3
229	20.0	16.4	24.6



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.341 ng
RT: 21.232 min Scan# 2253
Delta R.T. -0.009 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11

Tgt Ion:	149	Resp:	3979
Ion Ratio	Lower	Upper	
149	100		
167	26.3	21.2	31.8
279	3.1	2.7	4.1









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.586 min Scan# 21
Delta R.T. -0.009 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4


Tgt Ion:264 Resp: 5826
Ion Ratio Lower Upper
264 100
260 26.4 20.9 31.3
265 81.1 60.7 91.1



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.397 ng
RT: 25.887 min Scan# 3575
Delta R.T. -0.012 min
Lab File: BN036473.D
Acq: 20 Feb 2025 10:11

Tgt Ion:276 Resp: 8094
Ion Ratio Lower Upper
276 100
138 28.7 22.2 33.2
277 23.6 19.8 29.6





#37

Benzo(b)fluoranthene

Concen: 0.425 ng

RT: 22.902 min Scan# 21

Delta R.T. -0.009 min

Lab File: BN036473.D

Acq: 20 Feb 2025 10:11

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:252 Resp: 8146

Ion Ratio Lower Upper

252 100

253 27.5 21.9 32.9

125 21.1 15.0 22.6



#38

Benzo(k)fluoranthene

Concen: 0.424 ng

RT: 22.946 min Scan# 2569

Delta R.T. -0.012 min

Lab File: BN036473.D

Acq: 20 Feb 2025 10:11




Tgt Ion:252 Resp: 8381


Ion Ratio Lower Upper

252 100

253 27.8 22.2 33.4

125 19.4 15.0 22.4





#39

Benzo(a)pyrene

Concen: 0.419 ng

RT: 23.487 min Scan# 21

Delta R.T. -0.012 min

Lab File: BN036473.D

Acq: 20 Feb 2025 10:11

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:252 Resp: 7017

Ion Ratio Lower Upper

252 100

253 30.9 24.4 36.6

125 26.6 18.2 27.2



#40

Dibenzo(a,h)anthracene

Concen: 0.380 ng

RT: 25.902 min Scan# 3580

Delta R.T. -0.015 min

Lab File: BN036473.D

Acq: 20 Feb 2025 10:11




Tgt Ion:278 Resp: 6114


Ion Ratio Lower Upper

278 100

139 23.6 18.5 27.7

279 32.0 24.8 37.2





#41

Benzo(g,h,i)perylene

Concen: 0.395 ng

RT: 26.586 min Scan# 3814

Delta R.T. -0.015 min

Lab File: BN036473.D


Acq: 20 Feb 2025 10:11

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4





Tgt Ion:276 Resp: 7190

Ion Ratio Lower Upper

276 100

277 25.7 20.7 31.1

138 28.8 21.8 32.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
 Data File : BN036473.D
 Acq On : 20 Feb 2025 10:11
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 21 00:08:51 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 01:17:14 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	80	0.00
2	1,4-Dioxane	0.438	0.446	-1.8	82	0.00
3	n-Nitrosodimethylamine	0.760	0.791	-4.1	82	0.00
4 S	2-Fluorophenol	0.945	0.879	7.0	75	0.00
5 S	Phenol-d6	1.109	1.046	5.7	81	-0.01
6	bis(2-Chloroethyl)ether	1.160	1.164	-0.3	85	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	78	0.00
8 S	Nitrobenzene-d5	0.395	0.399	-1.0	85	-0.01
9	Naphthalene	1.154	1.138	1.4	79	-0.01
10	Hexachlorobutadiene	0.281	0.283	-0.7	77	-0.01
11 SURR	2-Methylnaphthalene-d10	0.615	0.585	4.9	75	-0.01
12	2-Methylnaphthalene	0.757	0.734	3.0	77	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	80	0.00
14 S	2,4,6-Tribromophenol	0.198	0.165	16.7	71	0.00
15 S	2-Fluorobiphenyl	1.504	1.478	1.7	86	-0.01
16	Acenaphthylene	1.767	1.621	8.3	77	-0.01
17	Acenaphthene	1.180	1.118	5.3	78	0.00
18	Fluorene	1.680	1.574	6.3	76	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	74	0.00
20	4,6-Dinitro-2-methylphenol	0.078	0.064	17.9	69	0.00
21	4-Bromophenyl-phenylether	0.239	0.234	2.1	75	0.00
22	Hexachlorobenzene	0.295	0.298	-1.0	78	-0.01
23	Atrazine	0.199	0.179	10.1	71	-0.01
24	Pentachlorophenol	0.140	0.122	12.9	74	-0.01
25	Phenanthrene	1.156	1.126	2.6	76	0.00
26	Anthracene	1.020	0.994	2.5	76	-0.01
27 SURR	Fluoranthene-d10	1.112	1.062	4.5	74	0.00
28	Fluoranthene	1.421	1.380	2.9	76	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	76	0.00
30	Pyrene	1.541	1.531	0.6	76	0.00
31 S	Terphenyl-d14	0.854	0.795	6.9	71	0.00
32	Benzo(a)anthracene	1.316	1.285	2.4	75	0.00
33	Chrysene	1.425	1.493	-4.8	83	0.00
34	Bis(2-ethylhexyl)phthalate	0.820	0.700	14.6	68	0.00
35 I	Perylene-d12	1.000	1.000	0.0	75	0.00
36	Indeno(1,2,3-cd)pyrene	1.398	1.389	0.6	76	-0.01
37	Benzo(b)fluoranthene	1.317	1.398	-6.2	84	0.00
38	Benzo(k)fluoranthene	1.356	1.439	-6.1	79	-0.01
39 C	Benzo(a)pyrene	1.150	1.204	-4.7	82	-0.01
40	Dibenzo(a,h)anthracene	1.103	1.049	4.9	74	-0.01
41	Benzo(g,h,i)perylene	1.250	1.234	1.3	74	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
Data File : BN036473.D
Acq On : 20 Feb 2025 10:11
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Feb 21 00:08:51 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 01:17:14 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	80	0.00
2	1,4-Dioxane	0.400	0.408	-2.0	82	0.00
3	n-Nitrosodimethylamine	0.400	0.416	-4.0	82	0.00
4 S	2-Fluorophenol	0.400	0.372	7.0	75	0.00
5 S	Phenol-d6	0.400	0.377	5.8	81	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.401	-0.3	85	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	78	0.00
8 S	Nitrobenzene-d5	0.400	0.405	-1.3	85	-0.01
9	Naphthalene	0.400	0.394	1.5	79	-0.01
10	Hexachlorobutadiene	0.400	0.403	-0.8	77	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.380	5.0	75	-0.01
12	2-Methylnaphthalene	0.400	0.388	3.0	77	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	80	0.00
14 S	2,4,6-Tribromophenol	0.400	0.333	16.8	71	0.00
15 S	2-Fluorobiphenyl	0.400	0.393	1.8	86	-0.01
16	Acenaphthylene	0.400	0.367	8.3	77	-0.01
17	Acenaphthene	0.400	0.379	5.3	78	0.00
18	Fluorene	0.400	0.375	6.3	76	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	74	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.326	18.5	69	0.00
21	4-Bromophenyl-phenylether	0.400	0.393	1.8	75	0.00
22	Hexachlorobenzene	0.400	0.404	-1.0	78	-0.01
23	Atrazine	0.400	0.359	10.3	71	-0.01
24	Pentachlorophenol	0.400	0.349	12.8	74	-0.01
25	Phenanthrene	0.400	0.390	2.5	76	0.00
26	Anthracene	0.400	0.390	2.5	76	-0.01
27 SURR	Fluoranthene-d10	0.400	0.382	4.5	74	0.00
28	Fluoranthene	0.400	0.389	2.8	76	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	76	0.00
30	Pyrene	0.400	0.398	0.5	76	0.00
31 S	Terphenyl-d14	0.400	0.373	6.8	71	0.00
32	Benzo(a)anthracene	0.400	0.390	2.5	75	0.00
33	Chrysene	0.400	0.419	-4.7	83	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.341	14.8	68	0.00
35 I	Perylene-d12	0.400	0.400	0.0	75	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.397	0.8	76	-0.01
37	Benzo(b)fluoranthene	0.400	0.425	-6.2	84	0.00
38	Benzo(k)fluoranthene	0.400	0.424	-6.0	79	-0.01
39 C	Benzo(a)pyrene	0.400	0.419	-4.7	82	-0.01
40	Dibenzo(a,h)anthracene	0.400	0.380	5.0	74	-0.01
41	Benzo(g,h,i)perylene	0.400	0.395	1.3	74	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG No.: Q1380

Instrument ID: BNA_N Calibration Date/Time: 02/20/2025 16:08

Lab File ID: BN036482.D Init. Calib. Date(s): 02/10/2025 02/10/2025

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 12:25 16:00

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.615	0.590		-4.1	50.0
Fluoranthene-d10	1.112	1.050		-5.6	50.0
2-Fluorophenol	0.945	0.903		-4.4	50.0
Phenol-d6	1.109	1.067		-3.8	50.0
Nitrobenzene-d5	0.395	0.392		-0.8	50.0
2-Fluorobiphenyl	1.504	1.565		4.1	50.0
2,4,6-Tribromophenol	0.198	0.163		-17.7	50.0
Terphenyl-d14	0.854	0.801		-6.2	50.0
1,4-Dioxane	0.438	0.416		-5.0	50.0

All other compounds must meet a minimum RRF of 0.010.

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
 Data File : BN036482.D
 Acq On : 20 Feb 2025 16:08
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 02/21/2025
 Supervised By :Jagrut Upadhyay 02/21/2025

Quant Time: Feb 21 00:12:25 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 01:17:14 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.746	152	2017	0.400	ng	0.00
7) Naphthalene-d8	10.541	136	4822	0.400	ng	# 0.00
13) Acenaphthene-d10	14.388	164	3262	0.400	ng	0.00
19) Phenanthrene-d10	17.124	188	6914	0.400	ng	#-0.01
29) Chrysene-d12	21.322	240	6182	0.400	ng	# 0.00
35) Perylene-d12	23.587	264	6375	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.334	112	1821	0.382	ng	-0.01
5) Phenol-d6	6.923	99	2152	0.385	ng	-0.01
8) Nitrobenzene-d5	8.897	82	1892	0.398	ng	-0.01
11) 2-Methylnaphthalene-d10	12.131	152	2847	0.384	ng	-0.01
14) 2,4,6-Tribromophenol	15.883	330	533	0.330	ng	0.00
15) 2-Fluorobiphenyl	13.009	172	5105	0.416	ng	-0.01
27) Fluoranthene-d10	19.164	212	7260	0.378	ng	0.00
31) Terphenyl-d14	19.768	244	4950	0.375	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.261	88	839	0.380	ng	97
3) n-Nitrosodimethylamine	3.579	42	1556	0.406	ng	96
6) bis(2-Chloroethyl)ether	7.168	93	2310	0.395	ng	99
9) Naphthalene	10.584	128	5486	0.394	ng	99
10) Hexachlorobutadiene	10.872	225	1352	0.399	ng	# 99
12) 2-Methylnaphthalene	12.207	142	3576	0.392	ng	98
16) Acenaphthylene	14.099	152	5339	0.371	ng	100
17) Acenaphthene	14.441	154	3586	0.373	ng	99
18) Fluorene	15.435	166	5179	0.378	ng	99
20) 4,6-Dinitro-2-methylph...	15.523	198	443	0.327	ng	93
21) 4-Bromophenyl-phenylether	16.329	248	1615	0.392	ng	# 84
22) Hexachlorobenzene	16.441	284	2053	0.403	ng	98
23) Atrazine	16.602	200	1254	0.364	ng	96
24) Pentachlorophenol	16.789	266	845	0.349	ng	95
25) Phenanthrene	17.173	178	7909	0.396	ng	100
26) Anthracene	17.260	178	6847	0.388	ng	100
28) Fluoranthene	19.192	202	9570	0.390	ng	99
30) Pyrene	19.559	202	9761	0.410	ng	99
32) Benzo(a)anthracene	21.304	228	7952	0.391	ng	99
33) Chrysene	21.358	228	9374	0.426	ng	99
34) Bis(2-ethylhexyl)phtha...	21.232	149	4422	0.349	ng	99
36) Indeno(1,2,3-cd)pyrene	25.885	276	8878	0.398	ng	99
37) Benzo(b)fluoranthene	22.902	252	8760	0.417	ng	98
38) Benzo(k)fluoranthene	22.949	252	9091m	0.421	ng	
39) Benzo(a)pyrene	23.490	252	7583	0.414	ng	96
40) Dibenzo(a,h)anthracene	25.905	278	6854	0.390	ng	99
41) Benzo(g,h,i)perylene	26.586	276	7966	0.400	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
 Data File : BN036482.D
 Acq On : 20 Feb 2025 16:08
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :

BNA_N

ClientSampleId :


SSTDCCC0.4EC


Manual Integrations**APPROVED**

Reviewed By :Rahul Chavli 02/21/2025

Supervised By :Jagrut Upadhyay 02/21/2025


Quant Time: Feb 21 00:12:25 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 01:17:14 2025
 Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.746 min Scan# 64
Delta R.T. -0.007 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion: 152 Resp: 201
Ion Ratio Lower Upper
152 100
150 154.9 123.7 185.5
115 65.0 52.5 78.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#2
1,4-Dioxane
Concen: 0.380 ng
RT: 3.261 min Scan# 25
Delta R.T. -0.007 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08

Tgt Ion: 88 Resp: 839
Ion Ratio Lower Upper
88 100
43 45.4 33.7 50.5
58 88.0 68.9 103.3









#3
n-Nitrosodimethylamine
Concen: 0.406 ng
RT: 3.579 min Scan# 69
Delta R.T. 0.000 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion: 42 Resp: 1550
Ion Ratio Lower Upper
42 100
74 85.9 71.8 107.6
44 9.6 7.8 11.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#4
2-Fluorophenol
Concen: 0.382 ng
RT: 5.334 min Scan# 312
Delta R.T. -0.014 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08

Tgt Ion:112 Resp: 1821
Ion Ratio Lower Upper
112 100
64 66.1 53.4 80.0
63 40.9 30.3 45.5









#5
Phenol-d6
Concen: 0.385 ng
RT: 6.923 min Scan# 51
Delta R.T. -0.014 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion: 99 Resp: 215
Ion Ratio Lower Upper
99 100
42 30.4 21.7 32.5
71 41.6 32.6 49.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#6
bis(2-Chloroethyl)ether
Concen: 0.395 ng
RT: 7.168 min Scan# 566
Delta R.T. -0.014 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08


Tgt Ion: 93 Resp: 2310
Ion Ratio Lower Upper
93 100
63 82.3 66.3 99.5
95 31.2 26.2 39.4





#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.541 min Scan# 911
Delta R.T. 0.000 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08




Instrument :
BNA_N
ClientSampleId :
SSTDC00.4EC



Tgt Ion: 136 Resp: 482
Ion Ratio Lower Upper
136 100
137 12.2 10.1 15.1
54 10.6 11.8 17.6
68 7.0 7.2 10.8#




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#8
Nitrobenzene-d5
Concen: 0.398 ng
RT: 8.897 min Scan# 779
Delta R.T. -0.011 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08

Tgt Ion: 82 Resp: 1892
Ion Ratio Lower Upper
82 100
128 37.4 31.9 47.9
54 64.4 53.1 79.7









#9
Naphthalene
Concen: 0.394 ng
RT: 10.584 min Scan# 91
Delta R.T. -0.011 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC

Tgt Ion:128 Resp: 548
Ion Ratio Lower Upper
128 100
129 12.7 9.6 14.4
127 14.9 12.0 18.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#10
Hexachlorobutadiene
Concen: 0.399 ng
RT: 10.872 min Scan# 964
Delta R.T. -0.011 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08


Tgt Ion:225 Resp: 1352
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.6 50.9 76.3





#11
2-Methylnaphthalene-d10
Concen: 0.384 ng
RT: 12.131 min Scan# 1146
Delta R.T. -0.010 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:152 Resp: 284
Ion Ratio Lower Upper
152 100
151 21.8 16.6 25.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#12
2-Methylnaphthalene
Concen: 0.392 ng
RT: 12.207 min Scan# 1161
Delta R.T. -0.005 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08


Tgt Ion:142 Resp: 3576
Ion Ratio Lower Upper
142 100
141 88.9 72.8 109.2
115 44.6 35.5 53.3





#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.388 min Scan# 1448
Delta R.T. 0.000 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:164 Resp: 326
Ion Ratio Lower Upper
164 100
162 102.1 84.1 126.1
160 50.9 41.4 62.0

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025




#14
2,4,6-Tribromophenol
Concen: 0.330 ng
RT: 15.883 min Scan# 1583
Delta R.T. 0.000 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08




Tgt Ion:330 Resp: 533
Ion Ratio Lower Upper
330 100
332 96.6 76.6 114.8
141 49.9 37.8 56.8





#15
2-Fluorobiphenyl
Concen: 0.416 ng
RT: 13.009 min Scan# 1319
Delta R.T. -0.011 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:172 Resp: 510
Ion Ratio Lower Upper
172 100
171 36.7 29.6 44.4
170 24.4 19.8 29.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#16
Acenaphthylene
Concen: 0.371 ng
RT: 14.099 min Scan# 1421
Delta R.T. -0.011 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08


Tgt Ion:152 Resp: 5339
Ion Ratio Lower Upper
152 100
151 19.6 15.8 23.8
153 12.6 10.2 15.2





#17
Acenaphthene
Concen: 0.373 ng
RT: 14.441 min Scan# 1453
Delta R.T. -0.011 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:154 Resp: 3580
Ion Ratio Lower Upper
154 100
153 118.5 93.3 139.9
152 61.0 48.8 73.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#18
Fluorene
Concen: 0.378 ng
RT: 15.435 min Scan# 1546
Delta R.T. -0.011 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08

Tgt Ion:166 Resp: 5179
Ion Ratio Lower Upper
166 100
165 99.0 79.5 119.3
167 13.8 10.4 15.6





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.124 min Scan# 1

Delta R.T. -0.012 min

Lab File: BN036482.D

Acq: 20 Feb 2025 16:08

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:	188	Resp:	6914
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	15.8	9.8	14.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 02/21/2025

Supervised By :Jagrut Upadhyay 02/21/2025



#20

4,6-Dinitro-2-methylphenol

Concen: 0.327 ng




RT: 15.523 min Scan# 1554


Delta R.T. -0.000 min

Lab File: BN036482.D

Acq: 20 Feb 2025 16:08


Tgt Ion:	198	Resp:	443
Ion Ratio	Lower	Upper	
198	100		
51	105.2	86.6	129.8
105	61.0	57.5	86.3





#21
4-Bromophenyl-phenylether
Concen: 0.392 ng
RT: 16.329 min Scan# 1619
Delta R.T. 0.000 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:248 Resp: 1619
Ion Ratio Lower Upper
248 100
250 97.5 76.1 114.1
141 60.3 71.7 107.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#22
Hexachlorobenzene
Concen: 0.403 ng
RT: 16.441 min Scan# 1628
Delta R.T. -0.012 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08

Tgt Ion:284 Resp: 2053
Ion Ratio Lower Upper
284 100
142 43.4 33.4 50.0
249 34.5 28.6 43.0









#23
Atrazine
Concen: 0.364 ng
RT: 16.602 min Scan# 1641
Delta R.T. -0.012 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08

Instrument :
BNA_N
ClientSampleId :
SSTDC00.4EC

Tgt Ion:200 Resp: 1254
Ion Ratio Lower Upper
200 100
173 32.5 23.2 34.8
215 48.5 40.0 60.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#24
Pentachlorophenol
Concen: 0.349 ng
RT: 16.789 min Scan# 1656
Delta R.T. -0.012 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08


Tgt Ion:266 Resp: 845
Ion Ratio Lower Upper
266 100
264 59.2 50.6 76.0
268 60.7 51.9 77.9





#25
Phenanthrene
Concen: 0.396 ng
RT: 17.173 min Scan# 1687
Delta R.T. 0.000 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:178 Resp: 7905
Ion Ratio Lower Upper
178 100
176 19.6 15.7 23.5
179 15.5 12.4 18.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#26
Anthracene
Concen: 0.388 ng
RT: 17.260 min Scan# 1694
Delta R.T. -0.012 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08

Tgt Ion:178 Resp: 6847
Ion Ratio Lower Upper
178 100
176 19.0 14.9 22.3
179 15.5 12.4 18.6





#27

Fluoranthene-d10

Concen: 0.378 ng

RT: 19.164 min Scan# 1935

Delta R.T. -0.005 min

Lab File: BN036482.D

Acq: 20 Feb 2025 16:08

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC

Tgt Ion:212 Resp: 7260

Ion Ratio Lower Upper

212 100

106 14.5 11.5 17.3





104 8.8 7.1 10.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 02/21/2025

Supervised By :Jagrut Upadhyay 02/21/2025



#28

Fluoranthene

Concen: 0.390 ng

RT: 19.192 min Scan# 1941

Delta R.T. -0.009 min

Lab File: BN036482.D

Acq: 20 Feb 2025 16:08




Tgt Ion:202 Resp: 9570


Ion Ratio Lower Upper

202 100

101 11.3 9.2 13.8


203 17.0 13.4 20.0





#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.322 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC




Tgt Ion:240 Resp: 6182
Ion Ratio Lower Upper
240 100
120 13.1 13.3 19.9
236 29.0 23.0 34.6

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025




#30
Pyrene
Concen: 0.410 ng
RT: 19.559 min Scan# 2020
Delta R.T. -0.005 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08



Tgt Ion:202 Resp: 9761
Ion Ratio Lower Upper
202 100
200 21.1 16.9 25.3
203 18.0 13.9 20.9






#31
Terphenyl-d14
Concen: 0.375 ng
RT: 19.768 min Scan# 2065
Delta R.T. 0.000 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:244 Resp: 4950
Ion Ratio Lower Upper
244 100
212 10.7 8.1 12.1
122 15.5 12.8 19.2

Manual Integrations

APPROVED




Reviewed By :Rahul Chavli 02/21/2025


Supervised By :Jagrut Upadhyay 02/21/2025



#32
Benzo(a)anthracene
Concen: 0.391 ng
RT: 21.304 min Scan# 2261
Delta R.T. -0.009 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08

Tgt Ion:228 Resp: 7952
Ion Ratio Lower Upper
228 100
226 28.3 22.2 33.2
229 19.9 16.5 24.7





#33

Chrysene

Concen: 0.426 ng

RT: 21.358 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036482.D


Acq: 20 Feb 2025 16:08

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC



Tgt Ion:228 Resp: 9374

Ion Ratio Lower Upper

228 100

226 31.0 25.5 38.3




229 20.2 16.4 24.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 02/21/2025

Supervised By :Jagrut Upadhyay 02/21/2025



#34

Bis(2-ethylhexyl)phthalate


Concen: 0.349 ng

RT: 21.232 min Scan# 2253

Delta R.T. -0.009 min

Lab File: BN036482.D

Acq: 20 Feb 2025 16:08





Tgt Ion:149 Resp: 4422


Ion Ratio Lower Upper

149 100

167 26.1 21.2 31.8


279 4.0 2.7 4.1





#35
Perylene-d12
Concen: 0.400 ng
RT: 23.587 min Scan# 21
Delta R.T. -0.009 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC



Tgt Ion:264 Resp: 6375
Ion Ratio Lower Upper
264 100
260 26.5 20.9 31.3
265 78.5 60.7 91.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.398 ng
RT: 25.885 min Scan# 3574
Delta R.T. -0.015 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08

Tgt Ion:276 Resp: 8878
Ion Ratio Lower Upper
276 100
138 28.2 22.2 33.2
277 24.6 19.8 29.6





#37

Benzo(b)fluoranthene

Concen: 0.417 ng

RT: 22.902 min Scan# 21

Delta R.T. -0.009 min

Lab File: BN036482.D


Acq: 20 Feb 2025 16:08

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC



Tgt Ion:252 Resp: 8760

Ion Ratio Lower Upper

252 100

253 27.4 21.9 32.9




125 20.9 15.0 22.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 02/21/2025

Supervised By :Jagrut Upadhyay 02/21/2025



#38

Benzo(k)fluoranthene


Concen: 0.421 ng m

RT: 22.949 min Scan# 2570

Delta R.T. -0.009 min

Lab File: BN036482.D

Acq: 20 Feb 2025 16:08





Tgt Ion:252 Resp: 9091


Ion Ratio Lower Upper

252 100

253 27.3 22.2 33.4

125 19.1 15.0 22.4





#39

Benzo(a)pyrene

Concen: 0.414 ng

RT: 23.490 min Scan# 21

Delta R.T. -0.009 min

Lab File: BN036482.D

Acq: 20 Feb 2025 16:08

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:	252	Resp:	758
Ion Ratio	Lower	Upper	
252	100		
253	29.9	24.4	36.6
125	27.1	18.2	27.2

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 02/21/2025

Supervised By :Jagrut Upadhyay 02/21/2025



#40

Dibenzo(a,h)anthracene

Concen: 0.390 ng




RT: 25.905 min Scan# 3581


Delta R.T. -0.012 min

Lab File: BN036482.D

Acq: 20 Feb 2025 16:08

Tgt Ion:	278	Resp:	6854
Ion Ratio	Lower	Upper	
278	100		
139	22.5	18.5	27.7
279	30.8	24.8	37.2





#41

Benzo(g,h,i)perylene

Concen: 0.400 ng

RT: 26.586 min Scan# 3814

Delta R.T. -0.015 min

Lab File: BN036482.D


Acq: 20 Feb 2025 16:08

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC



Tgt Ion: 276 Resp: 796

Ion Ratio Lower Upper

276 100

277 26.2 20.7 31.1



138 29.0 21.8 32.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 02/21/2025

Supervised By :Jagrut Upadhyay 02/21/2025



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
 Data File : BN036482.D
 Acq On : 20 Feb 2025 16:08
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 21 00:12:25 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 01:17:14 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	91	0.00
2	1,4-Dioxane	0.438	0.416	5.0	87	0.00
3	n-Nitrosodimethylamine	0.760	0.771	-1.4	92	0.00
4 S	2-Fluorophenol	0.945	0.903	4.4	88	-0.01
5 S	Phenol-d6	1.109	1.067	3.8	94	-0.01
6	bis(2-Chloroethyl)ether	1.160	1.145	1.3	96	-0.01
7 I	Naphthalene-d8	1.000	1.000	0.0	87	0.00
8 S	Nitrobenzene-d5	0.395	0.392	0.8	94	-0.01
9	Naphthalene	1.154	1.138	1.4	89	-0.01
10	Hexachlorobutadiene	0.281	0.280	0.4	86	-0.01
11 SURR	2-Methylnaphthalene-d10	0.615	0.590	4.1	86	-0.01
12	2-Methylnaphthalene	0.757	0.742	2.0	88	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	90	0.00
14 S	2,4,6-Tribromophenol	0.198	0.163	17.7	80	0.00
15 S	2-Fluorobiphenyl	1.504	1.565	-4.1	103	-0.01
16	Acenaphthylene	1.767	1.637	7.4	87	-0.01
17	Acenaphthene	1.180	1.099	6.9	87	-0.01
18	Fluorene	1.680	1.588	5.5	86	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	83	-0.01
20	4,6-Dinitro-2-methylphenol	0.078	0.064	17.9	77	0.00
21	4-Bromophenyl-phenylether	0.239	0.234	2.1	84	0.00
22	Hexachlorobenzene	0.295	0.297	-0.7	87	-0.01
23	Atrazine	0.199	0.181	9.0	80	-0.01
24	Pentachlorophenol	0.140	0.122	12.9	83	-0.01
25	Phenanthrene	1.156	1.144	1.0	87	0.00
26	Anthracene	1.020	0.990	2.9	85	-0.01
27 SURR	Fluoranthene-d10	1.112	1.050	5.6	82	0.00
28	Fluoranthene	1.421	1.384	2.6	85	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	83	0.00
30	Pyrene	1.541	1.579	-2.5	85	0.00
31 S	Terphenyl-d14	0.854	0.801	6.2	78	0.00
32	Benzo(a)anthracene	1.316	1.286	2.3	82	0.00
33	Chrysene	1.425	1.516	-6.4	92	0.00
34	Bis(2-ethylhexyl)phthalate	0.820	0.715	12.8	76	0.00
35 I	Perylene-d12	1.000	1.000	0.0	82	0.00
36	Indeno(1,2,3-cd)pyrene	1.398	1.393	0.4	83	-0.01
37	Benzo(b)fluoranthene	1.317	1.374	-4.3	90	0.00
38	Benzo(k)fluoranthene	1.356	1.426	-5.2	86	0.00
39 C	Benzo(a)pyrene	1.150	1.189	-3.4	89	0.00
40	Dibenzo(a,h)anthracene	1.103	1.075	2.5	82	-0.01
41	Benzo(g,h,i)perylene	1.250	1.250	0.0	82	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
 Data File : BN036482.D
 Acq On : 20 Feb 2025 16:08
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Feb 21 00:12:25 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 01:17:14 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	91	0.00
2	1,4-Dioxane	0.400	0.380	5.0	87	0.00
3	n-Nitrosodimethylamine	0.400	0.406	-1.5	92	0.00
4 S	2-Fluorophenol	0.400	0.382	4.5	88	-0.01
5 S	Phenol-d6	0.400	0.385	3.8	94	-0.01
6	bis(2-Chloroethyl)ether	0.400	0.395	1.3	96	-0.01
7 I	Naphthalene-d8	0.400	0.400	0.0	87	0.00
8 S	Nitrobenzene-d5	0.400	0.398	0.5	94	-0.01
9	Naphthalene	0.400	0.394	1.5	89	-0.01
10	Hexachlorobutadiene	0.400	0.399	0.3	86	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.384	4.0	86	-0.01
12	2-Methylnaphthalene	0.400	0.392	2.0	88	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	90	0.00
14 S	2,4,6-Tribromophenol	0.400	0.330	17.5	80	0.00
15 S	2-Fluorobiphenyl	0.400	0.416	-4.0	103	-0.01
16	Acenaphthylene	0.400	0.371	7.3	87	-0.01
17	Acenaphthene	0.400	0.373	6.8	87	-0.01
18	Fluorene	0.400	0.378	5.5	86	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	83	-0.01
20	4,6-Dinitro-2-methylphenol	0.400	0.327	18.3	77	0.00
21	4-Bromophenyl-phenylether	0.400	0.392	2.0	84	0.00
22	Hexachlorobenzene	0.400	0.403	-0.8	87	-0.01
23	Atrazine	0.400	0.364	9.0	80	-0.01
24	Pentachlorophenol	0.400	0.349	12.8	83	-0.01
25	Phenanthrene	0.400	0.396	1.0	87	0.00
26	Anthracene	0.400	0.388	3.0	85	-0.01
27 SURR	Fluoranthene-d10	0.400	0.378	5.5	82	0.00
28	Fluoranthene	0.400	0.390	2.5	85	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	83	0.00
30	Pyrene	0.400	0.410	-2.5	85	0.00
31 S	Terphenyl-d14	0.400	0.375	6.3	78	0.00
32	Benzo(a)anthracene	0.400	0.391	2.3	82	0.00
33	Chrysene	0.400	0.426	-6.5	92	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.349	12.8	76	0.00
35 I	Perylene-d12	0.400	0.400	0.0	82	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.398	0.5	83	-0.01
37	Benzo(b)fluoranthene	0.400	0.417	-4.2	90	0.00
38	Benzo(k)fluoranthene	0.400	0.421	-5.2	86	0.00
39 C	Benzo(a)pyrene	0.400	0.414	-3.5	89	0.00
40	Dibenzo(a,h)anthracene	0.400	0.390	2.5	82	-0.01
41	Benzo(g,h,i)perylene	0.400	0.400	0.0	82	-0.01

(#) = Out of Range

SPCC's out = 0 CCC's out = 0



QC SAMPLE DATA



- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18
- 19

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
Data File : BN036408.D
Acq On : 10 Feb 2025 11:46
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Feb 11 01:17:14 2025



AutoFind: Scans 2464, 2465, 2466; Background Corrected with Scan 2458

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	51.4	106809	PASS
68	69	0.00	2	0.7	724	PASS
69	198	0.00	100	47.7	99211	PASS
70	69	0.00	2	0.6	601	PASS
127	198	10	80	48.3	100501	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	207936	PASS
199	198	5	9	7.0	14493	PASS
275	198	10	60	24.7	51437	PASS
365	198	1	100	3.3	6829	PASS
441	198	0.01	100	7.1	14816	PASS
442	442	50	100	100.0	108824	PASS
443	442	15	24	20.1	21911	PASS

Instrument :
BNA_N
ClientSampleId :
DFTPP




DDT Breakdown

Date	Instrument Name	DFTPP Data File
2/10/2025	BNA_N	<u>BN036408.D</u>
Compound Name	Response	Retention Time
DDT	740190	20.628
DDD	8470	20.239
DDE	0	19.728
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
8470	748660	1.13

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
Data File : BN036408.D
Acq On : 10 Feb 2025 11:46
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Feb 11 01:55:02 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 25 04:23:53 2024
Response via : Initial Calibration



TIC: BN036408.D\data.ms

(70) Pentachlorophenol (C)

16.792min (-0.024) 22697.76 ng




response 239758

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	60.20
264.00	61.60	59.58
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN021025\
Data File : BN036408.D
Acq On : 10 Feb 2025 11:46
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Feb 11 01:55:02 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 25 04:23:53 2024
Response via : Initial Calibration



TIC: BN036408.D\data.ms

(77) Benzidine

19.392min (-0.012) 0.00 ng

response 1390167



Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.06
183.00	13.20	12.43
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
Data File : BN036472.D
Acq On : 20 Feb 2025 09:32
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Tue Feb 11 01:17:14 2025



AutoFind: Scans 2463, 2464, 2465; Background Corrected with Scan 2457

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	57.4	80973	PASS
68	69	0.00	2	1.5	1126	PASS
69	198	0.00	100	52.1	73475	PASS
70	69	0.00	2	0.5	370	PASS
127	198	10	80	50.3	70957	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	141053	PASS
199	198	5	9	6.9	9758	PASS
275	198	10	60	24.3	34267	PASS
365	198	1	100	3.5	4912	PASS
441	198	0.01	100	8.4	11880	PASS
442	442	50	100	100.0	70216	PASS
443	442	15	24	19.3	13522	PASS

Instrument :
BNA_N
ClientSampleId :
DFTPP




DDT Breakdown

Date	Instrument Name	DFTPP Data File
2/20/2025	BNA_N	<u>BN036472.D</u>
Compound Name	Response	Retention Time
DDT	465194	20.621
DDD	5607	20.227
DDE	141	19.674
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
5748	470942	1.22

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
Data File : BN036472.D
Acq On : 20 Feb 2025 09:32
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Feb 21 02:03:38 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 25 04:23:53 2024
Response via : Initial Calibration



TIC: BN036472.D\data.ms




(70) Pentachlorophenol (C)**16.786min (-0.030) 19225.56 ng****response 150123**

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	62.22
264.00	61.60	63.77
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
Data File : BN036472.D
Acq On : 20 Feb 2025 09:32
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Feb 21 02:03:38 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Dec 25 04:23:53 2024
Response via : Initial Calibration



TIC: BN036472.D\data.ms

(77) Benzidine

19.380min (-0.024) 0.00 ng

response 787833

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.45
183.00	13.20	12.84
0.00	0.00	0.00

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005	Date Received:	
Client Sample ID:	PB166758BL	SDG No.:	Q1380
Lab Sample ID:	PB166758BL	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036474.D	1	02/18/25 11:40	02/20/25 10:47	PB166758

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.35		30 - 150		87%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.42		30 - 150		104%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.37		55 - 111		93%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.35		53 - 106		88%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.42		58 - 132		104%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1870	7.746				
1146-65-2	Naphthalene-d8	3910	10.551				
15067-26-2	Acenaphthene-d10	2310	14.398				
1517-22-2	Phenanthrene-d10	4890	17.148				
1719-03-5	Chrysene-d12	4490	21.331				
1520-96-3	Perylene-d12	3910	23.595				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
Data File : BN036474.D
Acq On : 20 Feb 2025 10:47
Operator : RC/JU
Sample : PB166758BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB166758BL

Quant Time: Feb 21 00:09:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 01:17:14 2025
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.746	152	1874	0.400	ng	0.00
7) Naphthalene-d8	10.551	136	3905	0.400	ng	# 0.01
13) Acenaphthene-d10	14.398	164	2305	0.400	ng	0.01
19) Phenanthrene-d10	17.148	188	4889	0.400	ng	# 0.01
29) Chrysene-d12	21.331	240	4485	0.400	ng	0.00
35) Perylene-d12	23.595	264	3907	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.341	112	1773	0.400	ng	0.00
5) Phenol-d6	6.930	99	1793	0.345	ng	0.00
8) Nitrobenzene-d5	8.918	82	1432	0.372	ng	0.01
11) 2-Methylnaphthalene-d10	12.156	152	2086	0.348	ng	0.02
14) 2,4,6-Tribromophenol	15.907	330	220	0.193	ng	0.02
15) 2-Fluorobiphenyl	13.040	172	3041	0.351	ng	0.02
27) Fluoranthene-d10	19.169	212	5645	0.415	ng	0.00
31) Terphenyl-d14	19.768	244	3980	0.416	ng	0.00


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
Data File : BN036474.D
Acq On : 20 Feb 2025 10:47
Operator : RC/JU
Sample : PB166758BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB166758BL

Quant Time: Feb 21 00:09:23 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 01:17:14 2025
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.746 min Scan# 64
Delta R.T. -0.007 min
Lab File: BN036474.D
Acq: 20 Feb 2025 10:47




Instrument :
BNA_N
ClientSampleId :
PB166758BL


Tgt Ion:152 Resp: 1874
Ion Ratio Lower Upper
152 100
150 155.4 123.7 185.5
115 66.8 52.5 78.7



#4
2-Fluorophenol
Concen: 0.400 ng
RT: 5.341 min Scan# 313
Delta R.T. -0.007 min
Lab File: BN036474.D
Acq: 20 Feb 2025 10:47

Tgt Ion:112 Resp: 1773
Ion Ratio Lower Upper
112 100
64 68.7 53.4 80.0
63 39.0 30.3 45.5









#5
Phenol-d6
Concen: 0.345 ng
RT: 6.930 min Scan# 511
Delta R.T. -0.007 min
Lab File: BN036474.D
Acq: 20 Feb 2025 10:47




Instrument :
BNA_N
ClientSampleId :
PB166758BL


Tgt Ion:	99	Resp:	1793
Ion	Ratio	Lower	Upper
99	100		
42	27.4	21.7	32.5
71	41.9	32.6	49.0



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.551 min Scan# 934
Delta R.T. 0.011 min
Lab File: BN036474.D
Acq: 20 Feb 2025 10:47

Tgt Ion:	136	Resp:	3905
Ion	Ratio	Lower	Upper
136	100		
137	14.5	10.1	15.1
54	16.5	11.8	17.6
68	11.2	7.2	10.8









#8
Nitrobenzene-d5
Concen: 0.372 ng
RT: 8.918 min Scan# 711
Delta R.T. 0.011 min
Lab File: BN036474.D
Acq: 20 Feb 2025 10:47




Instrument :
BNA_N
ClientSampleId :
PB166758BL


Tgt Ion:	82	Resp:	1432
Ion Ratio	Lower	Upper	
82	100		
128	45.0	31.9	47.9
54	70.3	53.1	79.7



#11
2-Methylnaphthalene-d10
Concen: 0.348 ng
RT: 12.156 min Scan# 1151
Delta R.T. 0.015 min
Lab File: BN036474.D
Acq: 20 Feb 2025 10:47

Tgt Ion:	152	Resp:	2086
Ion Ratio	Lower	Upper	
152	100		
151	20.8	16.6	25.0









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.398 min Scan# 1448
Delta R.T. 0.011 min
Lab File: BN036474.D
Acq: 20 Feb 2025 10:47




Instrument :
BNA_N
ClientSampleId :
PB166758BL


Tgt Ion:164 Resp: 2305
Ion Ratio Lower Upper
164 100
162 103.6 84.1 126.1
160 53.8 41.4 62.0



#14
2,4,6-Tribromophenol
Concen: 0.193 ng
RT: 15.907 min Scan# 1585
Delta R.T. 0.025 min
Lab File: BN036474.D
Acq: 20 Feb 2025 10:47

Tgt Ion:330 Resp: 220
Ion Ratio Lower Upper
330 100
332 93.2 76.6 114.8
141 53.2 37.8 56.8









#15
2-Fluorobiphenyl
Concen: 0.351 ng
RT: 13.040 min Scan# 11
Delta R.T. 0.021 min
Lab File: BN036474.D
Acq: 20 Feb 2025 10:47




Instrument :
BNA_N
ClientSampleId :
PB166758BL


Tgt Ion:	172	Resp:	3041
Ion Ratio	Lower	Upper	
172	100		
171	39.2	29.6	44.4
170	27.1	19.8	29.6



#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.148 min Scan# 1685
Delta R.T. 0.012 min
Lab File: BN036474.D
Acq: 20 Feb 2025 10:47

Tgt Ion:	188	Resp:	4889
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	15.3	9.8	14.6#









#27
Fluoranthene-d10
Concen: 0.415 ng
RT: 19.169 min Scan# 1936
Delta R.T. -0.000 min
Lab File: BN036474.D
Acq: 20 Feb 2025 10:47




Instrument :
BNA_N
ClientSampleId :
PB166758BL


Tgt Ion	Ratio	Lower	Upper
212	100		
106	14.0	11.5	17.3
104	9.2	7.1	10.7



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.331 min Scan# 2264
Delta R.T. 0.009 min
Lab File: BN036474.D
Acq: 20 Feb 2025 10:47

Tgt Ion	Ratio	Lower	Upper
240	100		
120	17.0	13.3	19.9
236	31.0	23.0	34.6









#31
Terphenyl-d14
Concen: 0.416 ng
RT: 19.768 min Scan# 2065
Delta R.T. -0.000 min
Lab File: BN036474.D
Acq: 20 Feb 2025 10:47




Instrument :
BNA_N
ClientSampleId :
PB166758BL

Tgt Ion	Ratio	Lower	Upper
244	100		
212	12.0	8.1	12.1
122	17.0	12.8	19.2



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.595 min Scan# 2791
Delta R.T. -0.000 min
Lab File: BN036474.D
Acq: 20 Feb 2025 10:47

Tgt Ion	Ratio	Lower	Upper
264	100		
260	27.1	20.9	31.3
265	121.1	60.7	91.1#



Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005	Date Received:	
Client Sample ID:	PB166758BS	SDG No.:	Q1380
Lab Sample ID:	PB166758BS	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	1000 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036480.D	1	02/18/25 11:40	02/20/25 14:24	PB166758

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.35		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.52		30 - 150		129%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.41		30 - 150		101%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.39		55 - 111		97%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.47	*	53 - 106		118%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.43		58 - 132		108%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1830	7.746				
1146-65-2	Naphthalene-d8	4130	10.541				
15067-26-2	Acenaphthene-d10	2360	14.388				
1517-22-2	Phenanthrene-d10	5260	17.136				
1719-03-5	Chrysene-d12	4820	21.322				
1520-96-3	Perylene-d12	4360	23.589				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
 Data File : BN036480.D
 Acq On : 20 Feb 2025 14:24
 Operator : RC/JU
 Sample : PB166758BS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB166758BS

Manual Integrations
 APPROVED

Reviewed By :Rahul Chavli 02/21/2025
 Supervised By :Jagrut Upadhyay 02/21/2025

Quant Time: Feb 21 00:11:38 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 01:17:14 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.746	152	1827	0.400	ng	0.00
7) Naphthalene-d8	10.541	136	4127	0.400	ng	# 0.00
13) Acenaphthene-d10	14.388	164	2357	0.400	ng	0.00
19) Phenanthrene-d10	17.136	188	5256	0.400	ng	0.00
29) Chrysene-d12	21.322	240	4823	0.400	ng	0.00
35) Perylene-d12	23.589	264	4362	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.341	112	1752	0.406	ng	0.00
5) Phenol-d6	6.923	99	1995	0.394	ng	-0.01
8) Nitrobenzene-d5	8.897	82	1572	0.386	ng	-0.01
11) 2-Methylnaphthalene-d10	12.131	152	3269	0.515	ng	-0.01
14) 2,4,6-Tribromophenol	15.882	330	432	0.370	ng	0.00
15) 2-Fluorobiphenyl	13.008	172	4173	0.471	ng	-0.01
27) Fluoranthene-d10	19.164	212	5915	0.405	ng	0.00
31) Terphenyl-d14	19.764	244	4423	0.430	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.261	88	699	0.350	ng	# 51
3) n-Nitrosodimethylamine	3.572	42	1437	0.414	ng	# 97
6) bis(2-Chloroethyl)ether	7.168	93	2058	0.388	ng	98
9) Naphthalene	10.584	128	4763	0.400	ng	99
10) Hexachlorobutadiene	10.872	225	1195	0.412	ng	# 100
12) 2-Methylnaphthalene	12.207	142	3110	0.398	ng	98
16) Acenaphthylene	14.099	152	4781	0.459	ng	100
17) Acenaphthene	14.452	154	3067	0.441	ng	99
18) Fluorene	15.435	166	4315	0.436	ng	99
20) 4,6-Dinitro-2-methylph...	15.523	198	393	0.381	ng	92
21) 4-Bromophenyl-phenylether	16.329	248	1336	0.426	ng	# 88
22) Hexachlorobenzene	16.441	284	1675	0.433	ng	98
23) Atrazine	16.602	200	1144	0.437	ng	95
24) Pentachlorophenol	16.789	266	915	0.498	ng	99
25) Phenanthrene	17.173	178	6880	0.453	ng	99
26) Anthracene	17.260	178	6286	0.469	ng	99
28) Fluoranthene	19.192	202	8127	0.435	ng	99
30) Pyrene	19.559	202	8425	0.454	ng	100
32) Benzo(a)anthracene	21.304	228	7065	0.445	ng	99
33) Chrysene	21.358	228	8131	0.473	ng	99
34) Bis(2-ethylhexyl)phtha...	21.241	149	3779	0.382	ng	100
36) Indeno(1,2,3-cd)pyrene	25.890	276	7355	0.482	ng	99
37) Benzo(b)fluoranthene	22.902	252	6950	0.484	ng	97
38) Benzo(k)fluoranthene	22.949	252	7857m	0.531	ng	
39) Benzo(a)pyrene	23.490	252	6505	0.519	ng	97
40) Dibenzo(a,h)anthracene	25.908	278	5603m	0.466	ng	
41) Benzo(g,h,i)perylene	26.589	276	5921	0.434	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
Data File : BN036480.D
Acq On : 20 Feb 2025 14:24
Operator : RC/JU
Sample : PB166758BS
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :

BNA_N

ClientSampleId :

PB166758BS

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 02/21/2025

Supervised By :Jagrut Upadhyay 02/21/2025


Quant Time: Feb 21 00:11:38 2025


Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M

Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION

QLast Update : Tue Feb 11 01:17:14 2025

Response via : Initial Calibration






#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.746 min Scan# 64
Delta R.T. -0.007 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24

Instrument :

BNA_N

ClientSampleId :




PB166758BS




Tgt Ion: 152 Resp: 182
Ion Ratio Lower Upper
152 100
150 158.6 123.7 185.5
115 70.3 52.5 78.7

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025




#2
1,4-Dioxane
Concen: 0.350 ng
RT: 3.261 min Scan# 25
Delta R.T. -0.007 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24




Tgt Ion: 88 Resp: 699
Ion Ratio Lower Upper
88 100
43 125.0 33.7 50.5#
58 94.6 68.9 103.3





#3
n-Nitrosodimethylamine
Concen: 0.414 ng
RT: 3.572 min Scan# 61
Delta R.T. -0.007 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24




Instrument :
BNA_N
ClientSampleId :
PB166758BS



Tgt Ion: 42 Resp: 143
Ion Ratio Lower Upper
42 100
74 88.9 71.8 107.6
44 17.9 7.8 11.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#4
2-Fluorophenol
Concen: 0.406 ng
RT: 5.341 min Scan# 313
Delta R.T. -0.007 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24


Tgt Ion:112 Resp: 1752
Ion Ratio Lower Upper
112 100
64 67.2 53.4 80.0
63 37.0 30.3 45.5





#5
Phenol-d6
Concen: 0.394 ng
RT: 6.923 min Scan# 51
Delta R.T. -0.014 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24




Instrument :
BNA_N
ClientSampleId :
PB166758BS



Tgt Ion: 99 Resp: 199
Ion Ratio Lower Upper
99 100
42 31.6 21.7 32.5
71 42.0 32.6 49.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#6
bis(2-Chloroethyl)ether
Concen: 0.388 ng
RT: 7.168 min Scan# 566
Delta R.T. -0.014 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24


Tgt Ion: 93 Resp: 2058
Ion Ratio Lower Upper
93 100
63 84.9 66.3 99.5
95 32.0 26.2 39.4





#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.541 min Scan# 911
Delta R.T. 0.000 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24




Instrument :
BNA_N
ClientSampleId :
PB166758BS



Tgt Ion:136 Resp: 412
Ion Ratio Lower Upper
136 100
137 13.1 10.1 15.1
54 11.3 11.8 17.6
68 7.4 7.2 10.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#8
Nitrobenzene-d5
Concen: 0.386 ng
RT: 8.897 min Scan# 779
Delta R.T. -0.011 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24

Tgt Ion: 82 Resp: 1572
Ion Ratio Lower Upper
82 100
128 38.5 31.9 47.9
54 67.2 53.1 79.7









#9
Naphthalene
Concen: 0.400 ng
RT: 10.584 min Scan# 91
Delta R.T. -0.011 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24

Instrument :
BNA_N
ClientSampleId :
PB166758BS

Tgt Ion:128 Resp: 476
Ion Ratio Lower Upper
128 100
129 12.5 9.6 14.4
127 15.2 12.0 18.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#10
Hexachlorobutadiene
Concen: 0.412 ng
RT: 10.872 min Scan# 964
Delta R.T. -0.011 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24


Tgt Ion:225 Resp: 1195
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.4 50.9 76.3





#11
2-Methylnaphthalene-d10
Concen: 0.515 ng
RT: 12.131 min Scan# 1146
Delta R.T. -0.010 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24




Instrument :
BNA_N
ClientSampleId :
PB166758BS




Tgt Ion:152 Resp: 3269
Ion Ratio Lower Upper
152 100
151 15.2 16.6 25.0

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025




#12
2-Methylnaphthalene
Concen: 0.398 ng
RT: 12.207 min Scan# 1161
Delta R.T. -0.005 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24




Tgt Ion:142 Resp: 3110
Ion Ratio Lower Upper
142 100
141 90.1 72.8 109.2
115 46.4 35.5 53.3





#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.388 min Scan# 1448
Delta R.T. 0.000 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24




Instrument :
BNA_N
ClientSampleId :
PB166758BS




Tgt Ion:164 Resp: 235
Ion Ratio Lower Upper
164 100
162 104.0 84.1 126.1
160 52.6 41.4 62.0

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025




#14
2,4,6-Tribromophenol
Concen: 0.370 ng
RT: 15.882 min Scan# 1583
Delta R.T. 0.000 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24




Tgt Ion:330 Resp: 432
Ion Ratio Lower Upper
330 100
332 93.8 76.6 114.8
141 49.5 37.8 56.8





#15
2-Fluorobiphenyl
Concen: 0.471 ng
RT: 13.008 min Scan# 11
Delta R.T. -0.011 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24




Instrument :
BNA_N
ClientSampleId :
PB166758BS



Tgt Ion:172 Resp: 417
Ion Ratio Lower Upper
172 100
171 38.0 29.6 44.4
170 25.5 19.8 29.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#16
Acenaphthylene
Concen: 0.459 ng
RT: 14.099 min Scan# 1421
Delta R.T. -0.011 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24


Tgt Ion:152 Resp: 4781
Ion Ratio Lower Upper
152 100
151 19.7 15.8 23.8
153 12.9 10.2 15.2





#17
Acenaphthene
Concen: 0.441 ng
RT: 14.452 min Scan# 1454
Delta R.T. 0.000 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24




Instrument :
BNA_N
ClientSampleId :
PB166758BS



Tgt Ion:154 Resp: 306
Ion Ratio Lower Upper
154 100
153 117.2 93.3 139.9
152 61.5 48.8 73.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#18
Fluorene
Concen: 0.436 ng
RT: 15.435 min Scan# 1546
Delta R.T. -0.011 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24


Tgt Ion:166 Resp: 4315
Ion Ratio Lower Upper
166 100
165 100.1 79.5 119.3
167 13.5 10.4 15.6





#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.136 min Scan# 11
Delta R.T. 0.000 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24




Instrument :
BNA_N
ClientSampleId :
PB166758BS



Tgt Ion:188 Resp: 525
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 10.7 9.8 14.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#20
4,6-Dinitro-2-methylphenol
Concen: 0.381 ng
RT: 15.523 min Scan# 1554
Delta R.T. -0.000 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24


Tgt Ion:198 Resp: 393
Ion Ratio Lower Upper
198 100
51 102.4 86.6 129.8
105 62.7 57.5 86.3





#21
4-Bromophenyl-phenylether
Concen: 0.426 ng
RT: 16.329 min Scan# 1619
Delta R.T. 0.000 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24




Instrument :
BNA_N
ClientSampleId :
PB166758BS



Tgt Ion:248 Resp: 1330
Ion Ratio Lower Upper
248 100
250 94.7 76.1 114.1
141 65.9 71.7 107.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#22
Hexachlorobenzene
Concen: 0.433 ng
RT: 16.441 min Scan# 1628
Delta R.T. -0.012 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24

Tgt Ion:284 Resp: 1675
Ion Ratio Lower Upper
284 100
142 42.7 33.4 50.0
249 34.7 28.6 43.0





#23

Atrazine

Concen: 0.437 ng

RT: 16.602 min Scan# 1144

Delta R.T. -0.012 min

Lab File: BN036480.D

Acq: 20 Feb 2025 14:24

Instrument :

BNA_N

ClientSampleId :





PB166758BS

Tgt Ion:	200	Resp:	1144
Ion	Ratio	Lower	Upper
200	100		
173	32.2	23.2	34.8
215	47.2	40.0	60.0

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 02/21/2025

Supervised By :Jagrut Upadhyay 02/21/2025



#24

Pentachlorophenol

Concen: 0.498 ng




RT: 16.789 min Scan# 1656


Delta R.T. -0.012 min

Lab File: BN036480.D

Acq: 20 Feb 2025 14:24


Tgt Ion:	266	Resp:	915
Ion	Ratio	Lower	Upper
266	100		
264	63.0	50.6	76.0
268	66.7	51.9	77.9





#25
Phenanthrene
Concen: 0.453 ng
RT: 17.173 min Scan# 1687
Delta R.T. 0.000 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24




Instrument :
BNA_N
ClientSampleId :
PB166758BS



Tgt Ion:178 Resp: 6880
Ion Ratio Lower Upper
178 100
176 19.8 15.7 23.5
179 15.8 12.4 18.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#26
Anthracene
Concen: 0.469 ng
RT: 17.260 min Scan# 1694
Delta R.T. -0.012 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24

Tgt Ion:178 Resp: 6286
Ion Ratio Lower Upper
178 100
176 18.8 14.9 22.3
179 15.0 12.4 18.6





#27

Fluoranthene-d10

Concen: 0.405 ng

RT: 19.164 min Scan# 1935

Delta R.T. -0.005 min

Lab File: BN036480.D

Acq: 20 Feb 2025 14:24

Instrument :

BNA_N

ClientSampleId :

PB166758BS





Tgt Ion:	212	Resp:	591
Ion Ratio	Lower	Upper	
212	100		
106	14.7	11.5	17.3
104	9.4	7.1	10.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 02/21/2025

Supervised By :Jagrut Upadhyay 02/21/2025



#28

Fluoranthene

Concen: 0.435 ng




RT: 19.192 min Scan# 1941


Delta R.T. -0.009 min

Lab File: BN036480.D

Acq: 20 Feb 2025 14:24


Tgt Ion:	202	Resp:	8127
Ion Ratio	Lower	Upper	
202	100		
101	11.2	9.2	13.8
203	17.0	13.4	20.0





#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.322 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24




Instrument :
BNA_N
ClientSampleId :
PB166758BS




Tgt Ion:240 Resp: 482
Ion Ratio Lower Upper
240 100
120 14.4 13.3 19.9
236 28.9 23.0 34.6

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025




#30
Pyrene
Concen: 0.454 ng
RT: 19.559 min Scan# 2020
Delta R.T. -0.005 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24



Tgt Ion:202 Resp: 8425
Ion Ratio Lower Upper
202 100
200 21.3 16.9 25.3
203 17.7 13.9 20.9






#31
Terphenyl-d14
Concen: 0.430 ng
RT: 19.764 min Scan# 2064
Delta R.T. -0.005 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24

Instrument :

BNA_N

ClientSampleId :




PB166758BS




Tgt Ion:244 Resp: 442
Ion Ratio Lower Upper
244 100
212 11.0 8.1 12.1
122 17.2 12.8 19.2

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025




#32
Benzo(a)anthracene
Concen: 0.445 ng
RT: 21.304 min Scan# 2261
Delta R.T. -0.009 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24




Tgt Ion:228 Resp: 7065
Ion Ratio Lower Upper
228 100
226 28.8 22.2 33.2
229 20.7 16.5 24.7





#33
Chrysene
Concen: 0.473 ng
RT: 21.358 min Scan# 2135
Delta R.T. 0.000 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24




Instrument :
BNA_N
ClientSampleId :
PB166758BS



Tgt Ion:228 Resp: 813
Ion Ratio Lower Upper
228 100
226 31.0 25.5 38.3
229 20.8 16.4 24.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.382 ng
RT: 21.241 min Scan# 2254
Delta R.T. 0.000 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24

Tgt Ion:149 Resp: 3779
Ion Ratio Lower Upper
149 100
167 26.6 21.2 31.8
279 3.3 2.7 4.1





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.589 min Scan# 21

Delta R.T. -0.006 min

Lab File: BN036480.D


Acq: 20 Feb 2025 14:24

Instrument :

BNA_N

ClientSampleId :

PB166758BS



Tgt Ion:264 Resp: 4362

Ion Ratio Lower Upper

264 100

260 27.8 20.9 31.3



265 108.5 60.7 91.1

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 02/21/2025

Supervised By :Jagrut Upadhyay 02/21/2025



#36

Indeno(1,2,3-cd)pyrene


Concen: 0.482 ng

RT: 25.890 min Scan# 3576

Delta R.T. -0.009 min

Lab File: BN036480.D

Acq: 20 Feb 2025 14:24





Tgt Ion:276 Resp: 7355


Ion Ratio Lower Upper

276 100

138 28.6 22.2 33.2

277 24.1 19.8 29.6





#37

Benzo(b)fluoranthene

Concen: 0.484 ng

RT: 22.902 min Scan# 2554

Delta R.T. -0.009 min

Lab File: BN036480.D


Acq: 20 Feb 2025 14:24

Instrument :

BNA_N

ClientSampleId :

PB166758BS



Tgt Ion:252 Resp: 6950

Ion Ratio Lower Upper

252 100

253 28.8 21.9 32.9




125 20.1 15.0 22.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 02/21/2025

Supervised By :Jagrut Upadhyay 02/21/2025



#38

Benzo(k)fluoranthene


Concen: 0.531 ng m

RT: 22.949 min Scan# 2570

Delta R.T. -0.009 min

Lab File: BN036480.D

Acq: 20 Feb 2025 14:24





Tgt Ion:252 Resp: 7857


Ion Ratio Lower Upper

252 100

253 29.0 22.2 33.4

125 20.3 15.0 22.4





#39

Benzo(a)pyrene

Concen: 0.519 ng

RT: 23.490 min Scan# 21

Delta R.T. -0.009 min

Lab File: BN036480.D

Acq: 20 Feb 2025 14:24

Instrument :

BNA_N

ClientSampleId :

PB166758BS





Tgt Ion:	252	Resp:	650
Ion Ratio	Lower	Upper	
252	100		
253	31.1	24.4	36.6
125	25.0	18.2	27.2

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 02/21/2025

Supervised By :Jagrut Upadhyay 02/21/2025



#40

Dibenzo(a,h)anthracene

Concen: 0.466 ng m




RT: 25.908 min Scan# 3582


Delta R.T. -0.009 min

Lab File: BN036480.D

Acq: 20 Feb 2025 14:24

Tgt Ion:	278	Resp:	5603
Ion Ratio	Lower	Upper	
278	100		
139	23.1	18.5	27.7
279	32.0	24.8	37.2





#41

Benzo(g,h,i)perylene

Concen: 0.434 ng

RT: 26.589 min Scan# 3815

Delta R.T. -0.012 min

Lab File: BN036480.D


Acq: 20 Feb 2025 14:24

Instrument :

BNA_N

ClientSampleId :

PB166758BS



Tgt Ion:276 Resp: 592

Ion Ratio Lower Upper

276 100

277 26.5 20.7 31.1



138 29.6 21.8 32.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 02/21/2025

Supervised By :Jagrut Upadhyay 02/21/2025



Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	02/13/25
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005	Date Received:	02/17/25
Client Sample ID:	BP-VPB-192-GW-780-782MS	SDG No.:	Q1380
Lab Sample ID:	Q1380-10MS	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	320 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036477.D	1	02/18/25 11:40	02/20/25 12:36	PB166758

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.83		0.21	0.63	0.63	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.32		30 - 150		80%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.45		30 - 150		112%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.32		55 - 111		80%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.44	*	53 - 106		110%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.54	*	58 - 132		135%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1710		7.746			
1146-65-2	Naphthalene-d8	4000		10.541			
15067-26-2	Acenaphthene-d10	2500		14.387			
1517-22-2	Phenanthrene-d10	5570		17.124			
1719-03-5	Chrysene-d12	5520		21.313			
1520-96-3	Perylene-d12	5300		23.586			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
 Data File : BN036477.D
 Acq On : 20 Feb 2025 12:36
 Operator : RC/JU
 Sample : Q1380-10MS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-192-GW-780-782MS

Manual Integrations APPROVED

Reviewed By :Rahul Chavli 02/21/2025
 Supervised By :Jagrut Upadhyay 02/21/2025

Quant Time: Feb 21 00:10:19 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 01:17:14 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.746	152	1709	0.400	ng	0.00
7) Naphthalene-d8	10.541	136	3995	0.400	ng	# 0.00
13) Acenaphthene-d10	14.387	164	2495	0.400	ng	0.00
19) Phenanthrene-d10	17.124	188	5565	0.400	ng	#-0.01
29) Chrysene-d12	21.313	240	5523	0.400	ng	0.00
35) Perylene-d12	23.586	264	5300	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.341	112	974	0.241	ng	0.00
5) Phenol-d6	6.923	99	1025	0.216	ng	-0.01
8) Nitrobenzene-d5	8.897	82	1261	0.320	ng	-0.01
11) 2-Methylnaphthalene-d10	12.126	152	1968	0.321	ng	-0.02
14) 2,4,6-Tribromophenol	15.882	330	500	0.404	ng	0.00
15) 2-Fluorobiphenyl	13.008	172	4139	0.441	ng	-0.01
27) Fluoranthene-d10	19.160	212	6898	0.446	ng	0.00
31) Terphenyl-d14	19.763	244	6357	0.539	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.261	88	493	0.264	ng	# 50
3) n-Nitrosodimethylamine	3.572	42	905	0.279	ng	# 93
6) bis(2-Chloroethyl)ether	7.168	93	1675	0.338	ng	98
9) Naphthalene	10.583	128	3881	0.337	ng	98
10) Hexachlorobutadiene	10.872	225	805	0.287	ng	# 98
12) 2-Methylnaphthalene	12.202	142	2609	0.345	ng	97
16) Acenaphthylene	14.099	152	4341	0.394	ng	99
17) Acenaphthene	14.441	154	2787	0.379	ng	98
18) Fluorene	15.435	166	4131	0.394	ng	97
20) 4,6-Dinitro-2-methylph...	15.522	198	376	0.344	ng	# 81
21) 4-Bromophenyl-phenylether	16.329	248	1426	0.430	ng	# 74
22) Hexachlorobenzene	16.441	284	1658	0.404	ng	99
23) Atrazine	16.590	200	1195	0.431	ng	# 92
24) Pentachlorophenol	16.788	266	1439	0.739	ng	97
25) Phenanthrene	17.173	178	7646	0.475	ng	99
26) Anthracene	17.260	178	6840	0.482	ng	99
28) Fluoranthene	19.192	202	9601	0.486	ng	98
30) Pyrene	19.554	202	10137	0.477	ng	100
32) Benzo(a)anthracene	21.304	228	8882	0.489	ng	99
33) Chrysene	21.348	228	9509	0.483	ng	99
34) Bis(2-ethylhexyl)phtha...	21.232	149	5246	0.463	ng	99
36) Indeno(1,2,3-cd)pyrene	25.884	276	8360	0.451	ng	97
37) Benzo(b)fluoranthene	22.899	252	8647	0.495	ng	99
38) Benzo(k)fluoranthene	22.946	252	8912m	0.496	ng	
39) Benzo(a)pyrene	23.484	252	7527	0.494	ng	97
40) Dibenzo(a,h)anthracene	25.899	278	6349	0.434	ng	99
41) Benzo(g,h,i)perylene	26.586	276	6415	0.387	ng	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
Data File : BN036477.D
Acq On : 20 Feb 2025 12:36
Operator : RC/JU
Sample : Q1380-10MS
Misc :
ALS Vial : 6 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MS

Quant Time: Feb 21 00:10:19 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 01:17:14 2025
Response via : Initial Calibration

Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025





#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.746 min Scan# 64
Delta R.T. -0.007 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MS



Tgt Ion: 152 Resp: 1709
Ion Ratio Lower Upper
152 100
150 155.3 123.7 185.5
115 67.7 52.5 78.7




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#2
1,4-Dioxane
Concen: 0.264 ng
RT: 3.261 min Scan# 25
Delta R.T. -0.007 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36

Tgt Ion: 88 Resp: 493
Ion Ratio Lower Upper
88 100
43 128.0 33.7 50.5#
58 93.9 68.9 103.3









#3
n-Nitrosodimethylamine
Concen: 0.279 ng
RT: 3.572 min Scan# 61
Delta R.T. -0.007 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36

Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MS

Tgt Ion: 42 Resp: 90
Ion Ratio Lower Upper
42 100
74 94.6 71.8 107.6
44 18.8 7.8 11.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#4
2-Fluorophenol
Concen: 0.241 ng
RT: 5.341 min Scan# 313
Delta R.T. -0.007 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36

Tgt Ion:112 Resp: 974
Ion Ratio Lower Upper
112 100
64 69.7 53.4 80.0
63 40.5 30.3 45.5









#5
Phenol-d6
Concen: 0.216 ng
RT: 6.923 min Scan# 51
Delta R.T. -0.015 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36

Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MS

Tgt Ion: 99 Resp: 102
Ion Ratio Lower Upper
99 100
42 39.0 21.7 32.5
71 43.8 32.6 49.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#6
bis(2-Chloroethyl)ether
Concen: 0.338 ng
RT: 7.168 min Scan# 566
Delta R.T. -0.015 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36


Tgt Ion: 93 Resp: 1675
Ion Ratio Lower Upper
93 100
63 84.7 66.3 99.5
95 34.0 26.2 39.4





#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.541 min Scan# 911
Delta R.T. -0.000 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MS



Tgt Ion: 136 Resp: 399
Ion Ratio Lower Upper
136 100
137 13.1 10.1 15.1
54 10.6 11.8 17.6
68 7.2 7.2 10.8




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#8
Nitrobenzene-d5
Concen: 0.320 ng
RT: 8.897 min Scan# 779
Delta R.T. -0.011 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36

Tgt Ion: 82 Resp: 1261
Ion Ratio Lower Upper
82 100
128 40.4 31.9 47.9
54 68.9 53.1 79.7





#9

Naphthalene

Concen: 0.337 ng

RT: 10.583 min Scan# 91

Delta R.T. -0.011 min

Lab File: BN036477.D

Acq: 20 Feb 2025 12:36

Instrument :

BNA_N

ClientSampleId :





BP-VPB-192-GW-780-782MS

Tgt Ion:	128	Resp:	388
Ion	Ratio	Lower	Upper
128	100		
129	12.7	9.6	14.4
127	15.6	12.0	18.0

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 02/21/2025

Supervised By :Jagrut Upadhyay 02/21/2025



#10

Hexachlorobutadiene

Concen: 0.287 ng




RT: 10.872 min Scan# 964


Delta R.T. -0.011 min

Lab File: BN036477.D

Acq: 20 Feb 2025 12:36


Tgt Ion:	225	Resp:	805
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	62.0	50.9	76.3





#11
2-Methylnaphthalene-d10
Concen: 0.321 ng
RT: 12.126 min Scan# 1145
Delta R.T. -0.015 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MS



Tgt Ion:152 Resp: 1968
Ion Ratio Lower Upper
152 100
151 20.7 16.6 25.0




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#12
2-Methylnaphthalene
Concen: 0.345 ng
RT: 12.202 min Scan# 1160
Delta R.T. -0.010 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36


Tgt Ion:142 Resp: 2609
Ion Ratio Lower Upper
142 100
141 94.2 72.8 109.2
115 46.7 35.5 53.3





#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.387 min Scan# 1448
Delta R.T. -0.000 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MS




Tgt Ion:164 Resp: 249
Ion Ratio Lower Upper
164 100
162 101.2 84.1 126.1
160 50.7 41.4 62.0

Manual Integrations
APPROVED



Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025




#14
2,4,6-Tribromophenol
Concen: 0.404 ng
RT: 15.882 min Scan# 1583
Delta R.T. -0.000 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36




Tgt Ion:330 Resp: 500
Ion Ratio Lower Upper
330 100
332 94.0 76.6 114.8
141 48.0 37.8 56.8





#15
2-Fluorobiphenyl
Concen: 0.441 ng
RT: 13.008 min Scan# 1319
Delta R.T. -0.011 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MS



Tgt Ion:172 Resp: 4139
Ion Ratio Lower Upper
172 100
171 37.6 29.6 44.4
170 25.1 19.8 29.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#16
Acenaphthylene
Concen: 0.394 ng
RT: 14.099 min Scan# 1421
Delta R.T. -0.011 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36


Tgt Ion:152 Resp: 4341
Ion Ratio Lower Upper
152 100
151 20.3 15.8 23.8
153 13.3 10.2 15.2





#17
Acenaphthene
Concen: 0.379 ng
RT: 14.441 min Scan# 1453
Delta R.T. -0.011 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MS



Tgt Ion:154 Resp: 278
Ion Ratio Lower Upper
154 100
153 119.2 93.3 139.9
152 61.4 48.8 73.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#18
Fluorene
Concen: 0.394 ng
RT: 15.435 min Scan# 1546
Delta R.T. -0.011 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36

Tgt Ion:166 Resp: 4131
Ion Ratio Lower Upper
166 100
165 102.8 79.5 119.3
167 13.5 10.4 15.6





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.124 min Scan# 1683

Delta R.T. -0.012 min

Lab File: BN036477.D

Acq: 20 Feb 2025 12:36

Instrument :

BNA_N

ClientSampleId :





BP-VPB-192-GW-780-782MS

Tgt Ion:188	Resp:	556
Ion Ratio	Lower	Upper
188	100	
94	0.0	0.0
80	16.0	9.8

Manual Integrations
APPROVED

Reviewed By :Rahul Chavli 02/21/2025

Supervised By :Jagrut Upadhyay 02/21/2025



#20

4,6-Dinitro-2-methylphenol

Concen: 0.344 ng




RT: 15.522 min Scan# 1554


Delta R.T. -0.001 min

Lab File: BN036477.D

Acq: 20 Feb 2025 12:36


Tgt Ion:198	Resp:	376
Ion Ratio	Lower	Upper
198	100	
51	90.2	86.6
105	54.5	57.5





#21
4-Bromophenyl-phenylether
Concen: 0.430 ng
RT: 16.329 min Scan# 1619
Delta R.T. -0.000 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MS



Tgt Ion:248 Resp: 1420
Ion Ratio Lower Upper
248 100
250 103.2 76.1 114.1
141 46.7 71.7 107.5




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#22
Hexachlorobenzene
Concen: 0.404 ng
RT: 16.441 min Scan# 1628
Delta R.T. -0.012 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36

Tgt Ion:284 Resp: 1658
Ion Ratio Lower Upper
284 100
142 42.2 33.4 50.0
249 34.9 28.6 43.0





#23

Atrazine

Concen: 0.431 ng

RT: 16.590 min Scan# 1640

Delta R.T. -0.025 min

Lab File: BN036477.D

Acq: 20 Feb 2025 12:36

Instrument :

BNA_N

ClientSampleId :

BP-VPB-192-GW-780-782MS





Tgt Ion:	200	Resp:	119
Ion Ratio	Lower	Upper	
200	100		
173	35.5	23.2	34.8
215	46.2	40.0	60.0

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 02/21/2025

Supervised By :Jagrut Upadhyay 02/21/2025



#24

Pentachlorophenol

Concen: 0.739 ng




RT: 16.788 min Scan# 1656


Delta R.T. -0.012 min

Lab File: BN036477.D

Acq: 20 Feb 2025 12:36


Tgt Ion:	266	Resp:	1439
Ion Ratio	Lower	Upper	
266	100		
264	61.4	50.6	76.0
268	62.2	51.9	77.9





#25
Phenanthrene
Concen: 0.475 ng
RT: 17.173 min Scan# 1687
Delta R.T. -0.000 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MS



Tgt Ion:178 Resp: 7640
Ion Ratio Lower Upper
178 100
176 20.2 15.7 23.5
179 15.6 12.4 18.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#26
Anthracene
Concen: 0.482 ng
RT: 17.260 min Scan# 1694
Delta R.T. -0.012 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36

Tgt Ion:178 Resp: 6840
Ion Ratio Lower Upper
178 100
176 19.0 14.9 22.3
179 14.9 12.4 18.6





#27

Fluoranthene-d10

Concen: 0.446 ng

RT: 19.160 min Scan# 1936

Delta R.T. -0.009 min

Lab File: BN036477.D

Acq: 20 Feb 2025 12:36

Instrument :

BNA_N

ClientSampleId :

BP-VPB-192-GW-780-782MS





Tgt Ion:	212	Resp:	6898
Ion Ratio	Lower	Upper	
212	100		
106	15.1	11.5	17.3
104	9.8	7.1	10.7

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 02/21/2025

Supervised By :Jagrut Upadhyay 02/21/2025



#28

Fluoranthene

Concen: 0.486 ng




RT: 19.192 min Scan# 1941


Delta R.T. -0.009 min

Lab File: BN036477.D

Acq: 20 Feb 2025 12:36


Tgt Ion:	202	Resp:	9601
Ion Ratio	Lower	Upper	
202	100		
101	12.7	9.2	13.8
203	17.3	13.4	20.0





#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.313 min Scan# 21
Delta R.T. -0.009 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MS



Tgt Ion:240 Resp: 552
Ion Ratio Lower Upper
240 100
120 19.3 13.3 19.9
236 29.6 23.0 34.6




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#30
Pyrene
Concen: 0.477 ng
RT: 19.554 min Scan# 2019
Delta R.T. -0.009 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36


Tgt Ion:202 Resp: 10137
Ion Ratio Lower Upper
202 100
200 21.2 16.9 25.3
203 17.6 13.9 20.9





#31
Terphenyl-d14
Concen: 0.539 ng
RT: 19.763 min Scan# 2064
Delta R.T. -0.005 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MS



Tgt Ion:244 Resp: 635
Ion Ratio Lower Upper
244 100
212 10.7 8.1 12.1
122 15.1 12.8 19.2




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#32
Benzo(a)anthracene
Concen: 0.489 ng
RT: 21.304 min Scan# 2261
Delta R.T. -0.009 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36

Tgt Ion:228 Resp: 8882
Ion Ratio Lower Upper
228 100
226 28.0 22.2 33.2
229 20.7 16.5 24.7





#33

Chrysene

Concen: 0.483 ng

RT: 21.348 min Scan# 21

Delta R.T. -0.009 min

Lab File: BN036477.D

Acq: 20 Feb 2025 12:36

Instrument :

BNA_N

ClientSampleId :

BP-VPB-192-GW-780-782MS





Tgt Ion: 228	Resp: 950
Ion Ratio	Lower Upper
228 100	
226 31.6	25.5 38.3
229 20.0	16.4 24.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 02/21/2025

Supervised By :Jagrut Upadhyay 02/21/2025



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.463 ng




RT: 21.232 min Scan# 2253


Delta R.T. -0.009 min

Lab File: BN036477.D

Acq: 20 Feb 2025 12:36


Tgt Ion: 149	Resp: 5246
Ion Ratio	Lower Upper
149 100	
167 26.2	21.2 31.8
279 2.9	2.7 4.1





#35
Perylene-d12
Concen: 0.400 ng
RT: 23.586 min Scan# 21
Delta R.T. -0.009 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MS



Tgt Ion:264 Resp: 5300
Ion Ratio Lower Upper
264 100
260 26.0 20.9 31.3
265 91.1 60.7 91.1




Manual Integrations
APPROVED


Reviewed By :Rahul Chavli 02/21/2025
Supervised By :Jagrut Upadhyay 02/21/2025



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.451 ng
RT: 25.884 min Scan# 3574
Delta R.T. -0.015 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36

Tgt Ion:276 Resp: 8360
Ion Ratio Lower Upper
276 100
138 30.2 22.2 33.2
277 24.7 19.8 29.6





#37

Benzo(b)fluoranthene

Concen: 0.495 ng

RT: 22.899 min Scan# 2553

Delta R.T. -0.012 min

Lab File: BN036477.D


Acq: 20 Feb 2025 12:36

Instrument :

BNA_N

ClientSampleId :

BP-VPB-192-GW-780-782MS






Tgt Ion	Ratio	Lower	Upper
252	100		
253	27.4	21.9	32.9
125	20.3	15.0	22.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 02/21/2025

Supervised By :Jagrut Upadhyay 02/21/2025



#38

Benzo(k)fluoranthene


Concen: 0.496 ng m

RT: 22.946 min Scan# 2569



Delta R.T. -0.012 min


Lab File: BN036477.D

Acq: 20 Feb 2025 12:36



Tgt Ion	Ratio	Lower	Upper
252	100		
253	27.5	22.2	33.4
125	19.6	15.0	22.4





#39

Benzo(a)pyrene

Concen: 0.494 ng

RT: 23.484 min Scan# 21

Delta R.T. -0.015 min

Lab File: BN036477.D


Acq: 20 Feb 2025 12:36

Instrument :

BNA_N

ClientSampleId :

BP-VPB-192-GW-780-782MS



Tgt Ion:252 Resp: 752

Ion Ratio Lower Upper

252 100

253 30.6 24.4 36.6




125 25.6 18.2 27.2

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 02/21/2025

Supervised By :Jagrut Upadhyay 02/21/2025



#40

Dibenzo(a,h)anthracene


Concen: 0.434 ng

RT: 25.899 min Scan# 3579

Delta R.T. -0.018 min

Lab File: BN036477.D

Acq: 20 Feb 2025 12:36





Tgt Ion:278 Resp: 6349


Ion Ratio Lower Upper

278 100

139 23.3 18.5 27.7

279 30.5 24.8 37.2





#41

Benzo(g,h,i)perylene

Concen: 0.387 ng

RT: 26.586 min Scan# 3814

Delta R.T. -0.015 min

Lab File: BN036477.D


Acq: 20 Feb 2025 12:36

Instrument :

BNA_N

ClientSampleId :

BP-VPB-192-GW-780-782MS



Tgt Ion:276 Resp: 641

Ion Ratio Lower Upper

276 100

277 26.3 20.7 31.1



138 29.4 21.8 32.6

Manual Integrations

APPROVED

Reviewed By :Rahul Chavli 02/21/2025

Supervised By :Jagrut Upadhyay 02/21/2025



Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	02/13/25
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005	Date Received:	02/17/25
Client Sample ID:	BP-VPB-192-GW-780-782MSD	SDG No.:	Q1380
Lab Sample ID:	Q1380-11MSD	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	300 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036478.D	1	02/18/25 11:40	02/20/25 13:12	PB166758

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.92		0.23	0.67	0.67	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.32		30 - 150		80%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.47		30 - 150		116%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.32		55 - 111		80%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.45	*	53 - 106		113%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.55	*	58 - 132		138%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1790		7.746			
1146-65-2	Naphthalene-d8	4250		10.53			
15067-26-2	Acenaphthene-d10	2640		14.387			
1517-22-2	Phenanthrene-d10	5810		17.124			
1719-03-5	Chrysene-d12	5870		21.313			
1520-96-3	Perylene-d12	5480		23.584			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
 Data File : BN036478.D
 Acq On : 20 Feb 2025 13:12
 Operator : RC/JU
 Sample : Q1380-11MSD
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-192-GW-780-782MSD

Quant Time: Feb 21 00:10:48 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 01:17:14 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.746	152	1787	0.400	ng	0.00
7) Naphthalene-d8	10.530	136	4250	0.400	ng	-0.01
13) Acenaphthene-d10	14.387	164	2636	0.400	ng	0.00
19) Phenanthrene-d10	17.124	188	5810	0.400	ng	#-0.01
29) Chrysene-d12	21.313	240	5865	0.400	ng	0.00
35) Perylene-d12	23.584	264	5479	0.400	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.341	112	1045	0.247	ng	0.00
5) Phenol-d6	6.923	99	1140	0.230	ng	-0.01
8) Nitrobenzene-d5	8.897	82	1342	0.320	ng	-0.01
11) 2-Methylnaphthalene-d10	12.126	152	2100	0.321	ng	-0.02
14) 2,4,6-Tribromophenol	15.882	330	527	0.403	ng	0.00
15) 2-Fluorobiphenyl	13.008	172	4503	0.454	ng	-0.01
27) Fluoranthene-d10	19.164	212	7507	0.465	ng	0.00
31) Terphenyl-d14	19.763	244	6922	0.553	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.261	88	542	0.277	ng	# 40
3) n-Nitrosodimethylamine	3.572	42	979	0.288	ng	# 97
6) bis(2-Chloroethyl)ether	7.168	93	1774	0.342	ng	99
9) Naphthalene	10.583	128	4190	0.342	ng	99
10) Hexachlorobutadiene	10.872	225	844	0.283	ng	# 99
12) 2-Methylnaphthalene	12.202	142	2829	0.352	ng	99
16) Acenaphthylene	14.099	152	4650	0.399	ng	99
17) Acenaphthene	14.441	154	3020	0.388	ng	98
18) Fluorene	15.435	166	4515	0.408	ng	95
20) 4,6-Dinitro-2-methylph...	15.510	198	396	0.347	ng	97
21) 4-Bromophenyl-phenylether	16.329	248	1544	0.446	ng	# 75
22) Hexachlorobenzene	16.441	284	1773	0.414	ng	97
23) Atrazine	16.602	200	1299	0.449	ng	96
24) Pentachlorophenol	16.788	266	1498	0.737	ng	99
25) Phenanthrene	17.173	178	8339	0.497	ng	100
26) Anthracene	17.260	178	7296	0.493	ng	99
28) Fluoranthene	19.192	202	10396	0.504	ng	98
30) Pyrene	19.554	202	10879	0.482	ng	99
32) Benzo(a)anthracene	21.304	228	9605	0.498	ng	99
33) Chrysene	21.348	228	10250	0.491	ng	99
34) Bis(2-ethylhexyl)phtha...	21.232	149	5656	0.470	ng	100
36) Indeno(1,2,3-cd)pyrene	25.884	276	8831	0.461	ng	97
37) Benzo(b)fluoranthene	22.902	252	9474	0.525	ng	99
38) Benzo(k)fluoranthene	22.946	252	9481	0.510	ng	99
39) Benzo(a)pyrene	23.487	252	7896	0.501	ng	98
40) Dibenzo(a,h)anthracene	25.902	278	6691	0.443	ng	99
41) Benzo(g,h,i)perylene	26.586	276	6735	0.393	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
Data File : BN036478.D
Acq On : 20 Feb 2025 13:12
Operator : RC/JU
Sample : Q1380-11MSD
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MSD


Quant Time: Feb 21 00:10:48 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 01:17:14 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.746 min Scan# 64
Delta R.T. -0.007 min
Lab File: BN036478.D
Acq: 20 Feb 2025 13:12

Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MSD







Tgt Ion: 152 Resp: 1787
Ion Ratio Lower Upper
152 100
150 152.1 123.7 185.5
115 67.2 52.5 78.7



#2
1,4-Dioxane
Concen: 0.277 ng
RT: 3.261 min Scan# 25
Delta R.T. -0.007 min
Lab File: BN036478.D
Acq: 20 Feb 2025 13:12

Tgt Ion: 88 Resp: 542
Ion Ratio Lower Upper
88 100
43 153.0 33.7 50.5#
58 89.1 68.9 103.3









#3
n-Nitrosodimethylamine
Concen: 0.288 ng
RT: 3.572 min Scan# 61
Delta R.T. -0.007 min
Lab File: BN036478.D
Acq: 20 Feb 2025 13:12




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MSD


Tgt Ion: 42 Resp: 979
Ion Ratio Lower Upper
42 100
74 90.6 71.8 107.6
44 16.6 7.8 11.6#



#4
2-Fluorophenol
Concen: 0.247 ng
RT: 5.341 min Scan# 313
Delta R.T. -0.007 min
Lab File: BN036478.D
Acq: 20 Feb 2025 13:12

Tgt Ion: 112 Resp: 1045
Ion Ratio Lower Upper
112 100
64 69.5 53.4 80.0
63 39.5 30.3 45.5









#5
Phenol-d6
Concen: 0.230 ng
RT: 6.923 min Scan# 511
Delta R.T. -0.015 min
Lab File: BN036478.D
Acq: 20 Feb 2025 13:12




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MSD


Tgt Ion: 99 Resp: 1140
Ion Ratio Lower Upper
99 100
42 39.7 21.7 32.5#
71 46.6 32.6 49.0



#6
bis(2-Chloroethyl)ether
Concen: 0.342 ng
RT: 7.168 min Scan# 566
Delta R.T. -0.015 min
Lab File: BN036478.D
Acq: 20 Feb 2025 13:12

Tgt Ion: 93 Resp: 1774
Ion Ratio Lower Upper
93 100
63 84.4 66.3 99.5
95 32.9 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.530 min Scan# 911
Delta R.T. -0.011 min
Lab File: BN036478.D
Acq: 20 Feb 2025 13:12




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MSD


Tgt Ion:	136	Resp:	4250
Ion	Ratio	Lower	Upper
136	100		
137	13.0	10.1	15.1
54	15.0	11.8	17.6
68	9.8	7.2	10.8



#8
Nitrobenzene-d5
Concen: 0.320 ng
RT: 8.897 min Scan# 779
Delta R.T. -0.011 min
Lab File: BN036478.D
Acq: 20 Feb 2025 13:12

Tgt Ion:	82	Resp:	1342
Ion	Ratio	Lower	Upper
82	100		
128	40.0	31.9	47.9
54	67.6	53.1	79.7









#9
Naphthalene
Concen: 0.342 ng
RT: 10.583 min Scan# 91
Delta R.T. -0.011 min
Lab File: BN036478.D
Acq: 20 Feb 2025 13:12




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MSD


Tgt Ion:	128	Resp:	4190
Ion	Ratio	Lower	Upper
128	100		
129	12.7	9.6	14.4
127	14.8	12.0	18.0



#10
Hexachlorobutadiene
Concen: 0.283 ng
RT: 10.872 min Scan# 964
Delta R.T. -0.011 min
Lab File: BN036478.D
Acq: 20 Feb 2025 13:12

Tgt Ion:	225	Resp:	844
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	62.8	50.9	76.3









#11
2-Methylnaphthalene-d10
Concen: 0.321 ng
RT: 12.126 min Scan# 1145
Delta R.T. -0.015 min
Lab File: BN036478.D
Acq: 20 Feb 2025 13:12




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MSD


Tgt Ion:152 Resp: 2100
Ion Ratio Lower Upper
152 100
151 21.5 16.6 25.0



#12
2-Methylnaphthalene
Concen: 0.352 ng
RT: 12.202 min Scan# 1160
Delta R.T. -0.010 min
Lab File: BN036478.D
Acq: 20 Feb 2025 13:12

Tgt Ion:142 Resp: 2829
Ion Ratio Lower Upper
142 100
141 90.3 72.8 109.2
115 45.7 35.5 53.3









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.387 min Scan# 1448
Delta R.T. -0.000 min
Lab File: BN036478.D
Acq: 20 Feb 2025 13:12




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MSD


Tgt Ion:	164	Resp:	2636
Ion	Ratio	Lower	Upper
164	100		
162	102.7	84.1	126.1
160	50.7	41.4	62.0



#14
2,4,6-Tribromophenol
Concen: 0.403 ng
RT: 15.882 min Scan# 1583
Delta R.T. -0.000 min
Lab File: BN036478.D
Acq: 20 Feb 2025 13:12

Tgt Ion:	330	Resp:	527
Ion	Ratio	Lower	Upper
330	100		
332	96.0	76.6	114.8
141	48.4	37.8	56.8









#15
2-Fluorobiphenyl
Concen: 0.454 ng
RT: 13.008 min Scan# 1319
Delta R.T. -0.011 min
Lab File: BN036478.D
Acq: 20 Feb 2025 13:12




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MSD


Tgt Ion:	172	Resp:	4503
Ion Ratio	Lower	Upper	
172	100		
171	36.4	29.6	44.4
170	24.0	19.8	29.6



#16
Acenaphthylene
Concen: 0.399 ng
RT: 14.099 min Scan# 1421
Delta R.T. -0.011 min
Lab File: BN036478.D
Acq: 20 Feb 2025 13:12

Tgt Ion:	152	Resp:	4650
Ion Ratio	Lower	Upper	
152	100		
151	19.9	15.8	23.8
153	13.4	10.2	15.2





#17

Acenaphthene

Concen: 0.388 ng

RT: 14.441 min Scan# 1453

Delta R.T. -0.011 min

Lab File: BN036478.D

Acq: 20 Feb 2025 13:12

Instrument :

BNA_N

ClientSampleId :

BP-VPB-192-GW-780-782MSD





Tgt Ion:154 Resp: 3020

Ion Ratio Lower Upper

154 100

153 119.4 93.3 139.9

152 61.4 48.8 73.2



#18

Fluorene

Concen: 0.408 ng

RT: 15.435 min Scan# 1546

Delta R.T. -0.011 min

Lab File: BN036478.D

Acq: 20 Feb 2025 13:12




Tgt Ion:166 Resp: 4515


Ion Ratio Lower Upper

166 100

165 104.8 79.5 119.3

167 13.5 10.4 15.6









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.124 min Scan# 1683
Delta R.T. -0.013 min
Lab File: BN036478.D
Acq: 20 Feb 2025 13:12




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MSD


Tgt Ion	188	94	80
Ratio	100	0.0	17.6
Lower		0.0	9.8
Upper		0.0	14.6
Resp:	5810		



#20
4,6-Dinitro-2-methylphenol
Concen: 0.347 ng
RT: 15.510 min Scan# 1553
Delta R.T. -0.013 min
Lab File: BN036478.D
Acq: 20 Feb 2025 13:12

Tgt Ion	198	51	105
Ratio	100	113.7	71.8
Lower		86.6	57.5
Upper		129.8	86.3
Resp:	396		









#21
4-Bromophenyl-phenylether
Concen: 0.446 ng
RT: 16.329 min Scan# 1619
Delta R.T. -0.000 min
Lab File: BN036478.D
Acq: 20 Feb 2025 13:12




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MSD


Tgt Ion:248 Resp: 1544
Ion Ratio Lower Upper
248 100
250 102.0 76.1 114.1
141 47.4 71.7 107.5#



#22
Hexachlorobenzene
Concen: 0.414 ng
RT: 16.441 min Scan# 1628
Delta R.T. -0.013 min
Lab File: BN036478.D
Acq: 20 Feb 2025 13:12

Tgt Ion:284 Resp: 1773
Ion Ratio Lower Upper
284 100
142 42.9 33.4 50.0
249 33.6 28.6 43.0









#23
Atrazine
Concen: 0.449 ng
RT: 16.602 min Scan# 1641
Delta R.T. -0.013 min
Lab File: BN036478.D
Acq: 20 Feb 2025 13:12




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MSD


Tgt Ion	Ratio	Lower	Upper
200	100		
173	26.2	23.2	34.8
215	47.7	40.0	60.0



#24
Pentachlorophenol
Concen: 0.737 ng
RT: 16.788 min Scan# 1656
Delta R.T. -0.013 min
Lab File: BN036478.D
Acq: 20 Feb 2025 13:12

Tgt Ion	Ratio	Lower	Upper
266	100		
264	63.4	50.6	76.0
268	66.0	51.9	77.9









#25
Phenanthrene
Concen: 0.497 ng
RT: 17.173 min Scan# 1687
Delta R.T. -0.000 min
Lab File: BN036478.D
Acq: 20 Feb 2025 13:12




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MSD


Tgt Ion:178 Resp: 8339
Ion Ratio Lower Upper
178 100
176 19.9 15.7 23.5
179 15.6 12.4 18.6



#26
Anthracene
Concen: 0.493 ng
RT: 17.260 min Scan# 1694
Delta R.T. -0.013 min
Lab File: BN036478.D
Acq: 20 Feb 2025 13:12

Tgt Ion:178 Resp: 7296
Ion Ratio Lower Upper
178 100
176 19.2 14.9 22.3
179 14.9 12.4 18.6





#27

Fluoranthene-d10

Concen: 0.465 ng

RT: 19.164 min Scan# 1935

Delta R.T. -0.005 min

Lab File: BN036478.D

Acq: 20 Feb 2025 13:12

Instrument :

BNA_N

ClientSampleId :

BP-VPB-192-GW-780-782MSD





Tgt Ion:212 Resp: 7507

Ion Ratio Lower Upper

212 100

106 15.1 11.5 17.3

104 9.7 7.1 10.7



#28

Fluoranthene

Concen: 0.504 ng

RT: 19.192 min Scan# 1941

Delta R.T. -0.009 min

Lab File: BN036478.D

Acq: 20 Feb 2025 13:12




Tgt Ion:202 Resp: 10396


Ion Ratio Lower Upper

202 100

101 13.1 9.2 13.8

203 17.0 13.4 20.0









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.313 min Scan# 21
Delta R.T. -0.009 min
Lab File: BN036478.D
Acq: 20 Feb 2025 13:12




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MSD


Tgt Ion	Ratio	Lower	Upper
240	100		
120	19.7	13.3	19.9
236	29.6	23.0	34.6



#30
Pyrene
Concen: 0.482 ng
RT: 19.554 min Scan# 2019
Delta R.T. -0.009 min
Lab File: BN036478.D
Acq: 20 Feb 2025 13:12

Tgt Ion	Ratio	Lower	Upper
202	100		
200	21.4	16.9	25.3
203	17.7	13.9	20.9









#31
Terphenyl-d14
Concen: 0.553 ng
RT: 19.763 min Scan# 2064
Delta R.T. -0.005 min
Lab File: BN036478.D
Acq: 20 Feb 2025 13:12




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MSD


Tgt Ion	Ratio	Lower	Upper
244	100		
212	10.3	8.1	12.1
122	15.3	12.8	19.2



#32
Benzo(a)anthracene
Concen: 0.498 ng
RT: 21.304 min Scan# 2261
Delta R.T. -0.009 min
Lab File: BN036478.D
Acq: 20 Feb 2025 13:12

Tgt Ion	Ratio	Lower	Upper
228	100		
226	27.4	22.2	33.2
229	20.0	16.5	24.7





#33

Chrysene

Concen: 0.491 ng

RT: 21.348 min Scan# 21

Delta R.T. -0.009 min

Lab File: BN036478.D

Acq: 20 Feb 2025 13:12

Instrument :

BNA_N

ClientSampleId :

BP-VPB-192-GW-780-782MSD





Tgt Ion:228 Resp: 10250

Ion Ratio Lower Upper

228 100

226 31.7 25.5 38.3

229 19.9 16.4 24.6



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.470 ng

RT: 21.232 min Scan# 2253

Delta R.T. -0.009 min

Lab File: BN036478.D

Acq: 20 Feb 2025 13:12




Tgt Ion:149 Resp: 5656


Ion Ratio Lower Upper

149 100

167 26.5 21.2 31.8

279 3.3 2.7 4.1









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.584 min Scan# 21
Delta R.T. -0.012 min
Lab File: BN036478.D
Acq: 20 Feb 2025 13:12




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MSD


Tgt Ion:	264	Resp:	5479
Ion Ratio	Lower	Upper	
264	100		
260	26.9	20.9	31.3
265	90.7	60.7	91.1



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.461 ng
RT: 25.884 min Scan# 3574
Delta R.T. -0.015 min
Lab File: BN036478.D
Acq: 20 Feb 2025 13:12

Tgt Ion:	276	Resp:	8831
Ion Ratio	Lower	Upper	
276	100		
138	30.5	22.2	33.2
277	24.8	19.8	29.6





#37

Benzo(b)fluoranthene

Concen: 0.525 ng

RT: 22.902 min Scan# 21

Delta R.T. -0.009 min

Lab File: BN036478.D

Acq: 20 Feb 2025 13:12

Instrument :

BNA_N

ClientSampleId :

BP-VPB-192-GW-780-782MSD





Tgt Ion:252 Resp: 9474

Ion Ratio Lower Upper

252 100

253 27.2 21.9 32.9

125 19.7 15.0 22.6



#38

Benzo(k)fluoranthene

Concen: 0.510 ng

RT: 22.946 min Scan# 2569

Delta R.T. -0.012 min

Lab File: BN036478.D

Acq: 20 Feb 2025 13:12




Tgt Ion:252 Resp: 9481


Ion Ratio Lower Upper

252 100

253 27.3 22.2 33.4

125 19.8 15.0 22.4





#39

Benzo(a)pyrene

Concen: 0.501 ng

RT: 23.487 min Scan# 21

Delta R.T. -0.012 min

Lab File: BN036478.D


Acq: 20 Feb 2025 13:12

Instrument :



BNA_N

ClientSampleId :

BP-VPB-192-GW-780-782MSD



Tgt Ion:	252	Resp:	7896
Ion Ratio	Lower	Upper	
252	100		
253	29.7	24.4	36.6
125	23.7	18.2	27.2



#40

Dibenzo(a,h)anthracene


Concen: 0.443 ng

RT: 25.902 min Scan# 3580



Delta R.T. -0.015 min


Lab File: BN036478.D

Acq: 20 Feb 2025 13:12



Tgt Ion:	278	Resp:	6691
Ion Ratio	Lower	Upper	
278	100		
139	23.6	18.5	27.7
279	30.9	24.8	37.2





#41

Benzo(g,h,i)perylene

Concen: 0.393 ng

RT: 26.586 min Scan# 31

Delta R.T. -0.015 min

Lab File: BN036478.D


Acq: 20 Feb 2025 13:12

Instrument :

BNA_N

ClientSampleId :

BP-VPB-192-GW-780-782MSD





Tgt Ion:276 Resp: 6735

Ion Ratio Lower Upper

276 100

277 26.3 20.7 31.1

138 28.9 21.8 32.6





284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900, Fax : 908 789 8922

Manual Integration Report

Sequence:	BN021025	Instrument	BNA_n

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason

Manual Integration Report

Sequence:	bn022025	Instrument	BNA_n

Sample ID	File ID	Parameter	Review By	Review On	Supervised By	Supervised On	Reason
Q1380-10MS	BN036477.D	Benzo(k)fluoranthene	Rahul	2/21/2025 10:56:50 AM	Jagrut	2/21/2025 11:16:44 AM	Peak Integrated by Software
PB166758BS	BN036480.D	Benzo(k)fluoranthene	Rahul	2/21/2025 10:56:55 AM	Jagrut	2/21/2025 11:16:47 AM	Peak Integrated by Software
PB166758BS	BN036480.D	Dibenzo(a,h)anthracene	Rahul	2/21/2025 10:56:55 AM	Jagrut	2/21/2025 11:16:47 AM	Peak Integrated by Software
SSTDCCC0.4	BN036482.D	Benzo(k)fluoranthene	Rahul	2/21/2025 10:57:01 AM	Jagrut	2/21/2025 11:16:50 AM	Peak Integrated by Software

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN021025

Review By	Rahul	Review On	2/11/2025 4:21:36 PM
Supervise By	Jagrut	Supervise On	2/11/2025 6:22:45 PM
SubDirectory	BN021025	HP Acquire Method	BNA_N, 8270_SiM HP Processing Method bn021025
STD. NAME	STD REF.#		
Tune/Reschk	SP6717		
Initial Calibration Stds	SP6738,SP6736,SP6735,SP6734,SP6733,SP6732,SP6731		
CCC	SP6735		
Internal Standard/PEM	SP6682,1ul/100ul sample		
ICV/I.BLK	SP6684		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BN036408.D	10 Feb 2025 11:46	RC/JU	Ok
2	SSTDICC0.1	BN036409.D	10 Feb 2025 12:25	RC/JU	Ok
3	SSTDICC0.2	BN036410.D	10 Feb 2025 13:01	RC/JU	Ok
4	SSTDICCC0.4	BN036411.D	10 Feb 2025 13:36	RC/JU	Ok
5	SSTDICC0.8	BN036412.D	10 Feb 2025 14:12	RC/JU	Ok
6	SSTDICC1.6	BN036413.D	10 Feb 2025 14:48	RC/JU	Ok
7	SSTDICC3.2	BN036414.D	10 Feb 2025 15:24	RC/JU	Ok
8	SSTDICC5.0	BN036415.D	10 Feb 2025 16:00	RC/JU	Ok
9	SSTDICV0.4	BN036416.D	10 Feb 2025 16:36	RC/JU	Ok
10	PB166297BL	BN036417.D	10 Feb 2025 17:48	RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QCBatch ID # BN022025

Review By	Rahul	Review On	2/21/2025 10:57:17 AM
Supervise By	Jagrut	Supervise On	2/21/2025 11:17:01 AM
SubDirectory	BN022025	HP Acquire Method	BNA_N, 8270_SiM HP Processing Method bn021025
STD. NAME	STD REF.#		
Tune/Reschk	SP6717		
Initial Calibration Stds	SP6738,SP6736,SP6735,SP6734,SP6733,SP6732,SP6731		
CCC	SP6735		
Internal Standard/PEM	SP6682,1ul/100ul sample		
ICV/I.BLK	SP6684		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BN036472.D	20 Feb 2025 09:32	RC/JU	Ok
2	SSTDCCC0.4	BN036473.D	20 Feb 2025 10:11	RC/JU	Ok
3	PB166758BL	BN036474.D	20 Feb 2025 10:47	RC/JU	Ok
4	Q1380-01	BN036475.D	20 Feb 2025 11:24	RC/JU	Ok
5	Q1380-09	BN036476.D	20 Feb 2025 12:00	RC/JU	Ok
6	Q1380-10MS	BN036477.D	20 Feb 2025 12:36	RC/JU	Ok,M
7	Q1380-11MSD	BN036478.D	20 Feb 2025 13:12	RC/JU	Ok
8	Q1380-12	BN036479.D	20 Feb 2025 13:48	RC/JU	Ok
9	PB166758BS	BN036480.D	20 Feb 2025 14:24	RC/JU	Ok,M
10	SP6741	BN036481.D	20 Feb 2025 15:32	RC/JU	Ok
11	SSTDCCC0.4	BN036482.D	20 Feb 2025 16:08	RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN021025

Review By	Rahul	Review On	2/11/2025 4:21:36 PM
Supervise By	Jagrut	Supervise On	2/11/2025 6:22:45 PM
SubDirectory	BN021025	HP Acquire Method	BNA_N, 8270_HP Processing Method bn021025

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6738,SP6736,SP6735,SP6734,SP6733,SP6732,SP6731
CCC	SP6735
Internal Standard/PEM	SP6682,1ul/100ul sample
ICV/I.BLK	SP6684
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN036408.D	10 Feb 2025 11:46		RC/JU	Ok
2	SSTDICC0.1	SSTDICC0.1	BN036409.D	10 Feb 2025 12:25		RC/JU	Ok
3	SSTDICC0.2	SSTDICC0.2	BN036410.D	10 Feb 2025 13:01		RC/JU	Ok
4	SSTDICCC0.4	SSTDICCC0.4	BN036411.D	10 Feb 2025 13:36	The Calibration is Good For DOD	RC/JU	Ok
5	SSTDICC0.8	SSTDICC0.8	BN036412.D	10 Feb 2025 14:12		RC/JU	Ok
6	SSTDICC1.6	SSTDICC1.6	BN036413.D	10 Feb 2025 14:48		RC/JU	Ok
7	SSTDICC3.2	SSTDICC3.2	BN036414.D	10 Feb 2025 15:24		RC/JU	Ok
8	SSTDICC5.0	SSTDICC5.0	BN036415.D	10 Feb 2025 16:00	Compound #20 removed from 5.0ppm	RC/JU	Ok
9	SSTDICV0.4	ICVBN021025	BN036416.D	10 Feb 2025 16:36		RC/JU	Ok
10	PB166297BL	PB166297BL	BN036417.D	10 Feb 2025 17:48	Analyzed for contamination check	RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN022025

Review By	Rahul	Review On	2/21/2025 10:57:17 AM
Supervise By	Jagrut	Supervise On	2/21/2025 11:17:01 AM
SubDirectory	BN022025	HP Acquire Method	BNA_N, 8270_HP Processing Method bn021025

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6738,SP6736,SP6735,SP6734,SP6733,SP6732,SP6731
CCC	SP6735
Internal Standard/PEM	SP6682,1ul/100ul sample
ICV/I.BLK	SP6684
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN036472.D	20 Feb 2025 09:32		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN036473.D	20 Feb 2025 10:11		RC/JU	Ok
3	PB166758BL	PB166758BL	BN036474.D	20 Feb 2025 10:47		RC/JU	Ok
4	Q1380-01	BP-VPB-192-EB-20250	BN036475.D	20 Feb 2025 11:24		RC/JU	Ok
5	Q1380-09	BP-VPB-192-GW-780-7	BN036476.D	20 Feb 2025 12:00		RC/JU	Ok
6	Q1380-10MS	BP-VPB-192-GW-780-7	BN036477.D	20 Feb 2025 12:36		RC/JU	Ok,M
7	Q1380-11MSD	BP-VPB-192-GW-780-7	BN036478.D	20 Feb 2025 13:12		RC/JU	Ok
8	Q1380-12	VPB192-HYD-2025021	BN036479.D	20 Feb 2025 13:48	Surrogate Fail	RC/JU	Ok
9	PB166758BS	PB166758BS	BN036480.D	20 Feb 2025 14:24		RC/JU	Ok,M
10	SP6741	SP6741	BN036481.D	20 Feb 2025 15:32	8270 Sim Surrogate	RC/JU	Ok
11	SSTDCCC0.4	SSTDCCC0.4EC	BN036482.D	20 Feb 2025 16:08		RC/JU	Ok,M

M : Manual Integration

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 2/19/2025

OVENTEMP IN Celsius(°C): 107
Time IN: 16:40
In Date: 02/18/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:11
Out Date: 02/19/2025
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB134739

Lab ID	Client SampleID	Dish #	Dish Wt (g) (A)	Sample Wt (g)	Dish + Sample Wt (g) (B)	Dish+Dry Sample Wt (g) (C)	% Solid	Comments
Q1380-06	BP-VPB-192-GW-725-727	1	1.15	8.48	9.63	1.78	7.4	SLUDGE SAMPLE
Q1380-07	Q1380-06MS	2	1.15	8.48	9.63	1.78	7.4	SLUDGE SAMPLE
Q1380-08	Q1380-06MSD	3	1.15	8.48	9.63	1.78	7.4	SLUDGE SAMPLE
Q1382-01	OU4-PCS-TC-01-021725	4	1.15	8.56	9.71	9.02	91.9	
Q1382-03	OU4-PCS-TC-02-021725	5	1.15	8.59	9.74	9.03	91.7	
Q1382-05	OU4-PCS-TC-03-021725	6	1.15	8.56	9.71	9.06	92.4	
Q1382-07	OU4-PCS-TC-04-021725	7	1.16	8.80	9.96	9.3	92.5	
Q1382-09	OU4-PCS-TC-05-021725	8	1.16	8.65	9.81	9.12	92.0	
Q1382-11	OU4-PCS-TC-06-021725	9	1.12	8.76	9.88	9.2	92.2	
Q1382-13	OU4-PCS-TC-07-021725	10	1.11	8.67	9.78	9.1	92.2	
Q1382-15	OU4-PCS-TC-08-021725	11	1.19	8.61	9.8	9.16	92.6	
Q1382-17	OU4-PCS-TC-09-021725	12	1.14	8.66	9.8	9.16	92.6	
Q1382-19	OU4-PCS-TC-10-021725	13	1.14	8.84	9.98	9.4	93.4	
Q1383-01	OU4-PCS-TC-11-021725	14	1.17	8.74	9.91	9.26	92.6	
Q1383-03	OU4-PCS-TC-12-021725	15	1.15	8.73	9.88	9.2	92.2	
Q1383-05	OU4-PCS-TC-13-021725	16	1.12	8.76	9.88	9.22	92.5	
Q1383-07	OU4-PCS-TC-14-021725	17	1.18	8.58	9.76	9.08	92.1	
Q1383-09	OU4-PCS-TC-15-021725	18	1.13	8.77	9.9	9.24	92.5	
Q1383-11	OU4-PCS-TC-16-021725	19	1.16	8.61	9.77	9.09	92.1	
Q1383-13	OU4-PCS-TC-17-021725	20	1.19	8.58	9.77	9.03	91.4	
Q1383-15	OU4-PCS-TC-18-021725	21	1.15	8.60	9.75	9.05	91.9	
Q1383-17	OU4-PCS-TC-19-021725	22	1.15	8.84	9.99	9.29	92.1	
Q1383-19	OU4-PCS-TC-20-021725	23	1.13	8.67	9.8	9.12	92.2	
Q1383-21	OU4-CF-15-021725	24	1.16	8.50	9.66	9.02	92.5	
Q1384-01	OR-03-021823025	25	1.15	8.81	9.96	8.96	88.6	
Q1384-02	OR-03-021823025-E2	26	1.17	8.82	9.99	9.04	89.2	
Q1385-01	ST-1	27	1.00	1.00	2.00	2.00	100.0	CONCRETE sample
Q1385-02	ST-1-E2	28	1.00	1.00	2.00	2.00	100.0	CONCRETE sample

WORKLIST(Hardcopy Internal Chain)

WorkList Name : %1-021825 WorkList ID : 187742 Department : Wet-Chemistry Date : 02-18-2025 08:35:18

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1380-06	BP-VPB-192-GW-725-727	Solid	Percent Solids	Cool 4 deg C	TETR06	H31	02/10/2025	Chemtech -SO
Q1380-07	Q1380-06MS	Solid	Percent Solids	Cool 4 deg C	TETR06	H31	02/10/2025	Chemtech -SO
Q1380-08	Q1380-06MSD	Solid	Percent Solids	Cool 4 deg C	TETR06	H31	02/10/2025	Chemtech -SO
Q1382-01	OU4-PCS-TC-01-021725	Solid	Percent Solids	Cool 4 deg C	NOBI03	H31	02/17/2025	Chemtech -SO
Q1382-03	OU4-PCS-TC-02-021725	Solid	Percent Solids	Cool 4 deg C	NOBI03	H31	02/17/2025	Chemtech -SO
Q1382-05	OU4-PCS-TC-03-021725	Solid	Percent Solids	Cool 4 deg C	NOBI03	H31	02/17/2025	Chemtech -SO
Q1382-07	OU4-PCS-TC-04-021725	Solid	Percent Solids	Cool 4 deg C	NOBI03	H31	02/17/2025	Chemtech -SO
Q1382-09	OU4-PCS-TC-05-021725	Solid	Percent Solids	Cool 4 deg C	NOBI03	H31	02/17/2025	Chemtech -SO
Q1382-11	OU4-PCS-TC-06-021725	Solid	Percent Solids	Cool 4 deg C	NOBI03	H31	02/17/2025	Chemtech -SO
Q1382-13	OU4-PCS-TC-07-021725	Solid	Percent Solids	Cool 4 deg C	NOBI03	H31	02/17/2025	Chemtech -SO
Q1382-15	OU4-PCS-TC-08-021725	Solid	Percent Solids	Cool 4 deg C	NOBI03	H31	02/17/2025	Chemtech -SO
Q1382-17	OU4-PCS-TC-09-021725	Solid	Percent Solids	Cool 4 deg C	NOBI03	H31	02/17/2025	Chemtech -SO
Q1382-19	OU4-PCS-TC-10-021725	Solid	Percent Solids	Cool 4 deg C	NOBI03	H31	02/17/2025	Chemtech -SO
Q1383-01	OU4-PCS-TC-11-021725	Solid	Percent Solids	Cool 4 deg C	NOBI03	H31	02/17/2025	Chemtech -SO
Q1383-03	OU4-PCS-TC-12-021725	Solid	Percent Solids	Cool 4 deg C	NOBI03	H31	02/17/2025	Chemtech -SO
Q1383-05	OU4-PCS-TC-13-021725	Solid	Percent Solids	Cool 4 deg C	NOBI03	H31	02/17/2025	Chemtech -SO
Q1383-07	OU4-PCS-TC-14-021725	Solid	Percent Solids	Cool 4 deg C	NOBI03	H31	02/17/2025	Chemtech -SO
Q1383-09	OU4-PCS-TC-15-021725	Solid	Percent Solids	Cool 4 deg C	NOBI03	H31	02/17/2025	Chemtech -SO
Q1383-11	OU4-PCS-TC-16-021725	Solid	Percent Solids	Cool 4 deg C	NOBI03	H31	02/17/2025	Chemtech -SO
Q1383-13	OU4-PCS-TC-17-021725	Solid	Percent Solids	Cool 4 deg C	NOBI03	H31	02/17/2025	Chemtech -SO
Q1383-15	OU4-PCS-TC-18-021725	Solid	Percent Solids	Cool 4 deg C	NOBI03	H31	02/17/2025	Chemtech -SO

Date/Time: 02/18/25 15:00
 Raw Sample Received by: *TH Weller*
 Raw Sample Relinquished by: *OP*
 Date/Time: 02/18/25 15:00
 Raw Sample Received by: *OP*
 Raw Sample Relinquished by: *OP*

WORKLIST(Hardcopy Internal Chain)

8134439

WorkList Name : %1-021825 WorkList ID : 187742 Department : Wet-Chemistry Date : 02-18-2025 08:35:18

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1383-17	OU4-PCS-TC-19-021725	Solid	Percent Solids	Cool 4 deg C	NOBI03	H31	02/17/2025	Chemtech -SO
Q1383-19	OU4-PCS-TC-20-021725	Solid	Percent Solids	Cool 4 deg C	NOBI03	H31	02/17/2025	Chemtech -SO
Q1383-21	OU4-CF-15-021725	Solid	Percent Solids	Cool 4 deg C	NOBI03	H31	02/17/2025	Chemtech -SO
Q1384-01	OR-03-021823025	Solid	Percent Solids	Cool 4 deg C	PSEG05	N41	02/18/2025	Chemtech -SO
Q1384-02	OR-03-021823025-E2	Solid	Percent Solids	Cool 4 deg C	PSEG05	N41	02/18/2025	Chemtech -SO
Q1385-01	ST-1	Solid	Percent Solids	Cool 4 deg C	PSEG03	H21	02/18/2025	Chemtech -SO
Q1385-02	ST-1-E2	Solid	Percent Solids	Cool 4 deg C	PSEG03	H21	02/18/2025	Chemtech -SO

Date/Time 02/18/25 15:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

Date/Time 02/18/25 17:00
 Raw Sample Received by: [Signature]
 Raw Sample Relinquished by: [Signature]

SOP ID: M3510C,3580A-Extraction SVOC-20

Clean Up SOP #: N/A

Matrix : Water

Welgh By: N/A **Extraction By:** RS

Balance check: N/A **Filter By:** RS

Balance ID: N/A **pH Meter ID:** N/A

pH Strip Lot#: E3880 **Hood ID:** 4,5,6,7

Extraction Method: ☒ Seperatory Funnel ☐ Continious Liquid/Liquid ☐ Sonication ☐ Waste Dilution ☐ Soxhlet

Extraction Start Date : 02/18/2025

Extraction Start Time : 11:40

Extraction End Date : 02/18/2025

Extraction End Time : 16:35

Concentration By: EH

Supervisor By : rajesh

Standard Name	MLS USED	Concentration ug/mL	STD REF. # FROM LOG
Spike Sol 1	1.0ML	0.4 PPM	SP6739
Surrogate	1.0ML	0.4 PPM	SP6718
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A
N/A	N/A	N/A	N/A

Chemical Used	ML/SAMPLE USED	Lot Number
Methylene Chloride	N/A	E3878
Baked Na2SO4	N/A	EP2589
10N NaoH	N/A	EP2559
H2SO4 1:1	N/A	EP2565
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A
N/A	N/A	N/A

Extraction Conformance/Non-Conformance Comments:

1.5 ML Vial lot# 2210673. pH Adjusted<2 with 1:1 H2SO4 &>11 with 10 N NaOH. Q1380-09,10,11 Limited volume Recd.

KD Bath ID: Water bath -01,02 **Envap ID:** NEVAP-02

KD Bath Temperature: 60 °C **Envap Temperature:** 40 °C

Date / Time	Prepped Sample Relinquished By/Location	Received By/Location
02/18/25	RP (Ext-2.4)	RLSVOC
16:40	Preparation Group	Analysis Group

Analytical Method: M3510C,3580A-Extraction SVOC-20

Concentration Date: 02/18/2025

Sample ID	Client Sample ID	Test	g / mL	PH	Surr/Spike By:		Final Vol. (mL)	JarID	Comments	Prep Pos
					AddedBy	VerifiedBy				
PB166758BL	SBLK758	SVOC-SIMGrou p1	1000	6	RUPESH	rajesh	1			SEP-01
PB166758BS	SLCS758	SVOC-SIMGrou p1	1000	6	RUPESH	rajesh	1			2
Q1380-01	BP-VPB-192-EB-2025021 2	SVOC-SIMGrou p1	890	6	RUPESH	rajesh	1	C		3
Q1380-09	BP-VPB-192-GW-780- ~~787~~ 782	SVOC-SIMGrou p1	380	6	RUPESH	rajesh	1	C	Muddy	4
Q1380-10	Q1380-09MS	SVOC-SIMGrou p1	320	6	RUPESH	rajesh	1		Muddy	5
Q1380-11	Q1380-09MSD	SVOC-SIMGrou p1	300	6	RUPESH	rajesh	1		Muddy	6
Q1380-12	VPB192-HYD-20250214	SVOC-SIMGrou p1	940	6	RUPESH	rajesh	1	C		7

* Extracts relinquished on the same date as received.

WORKLIST(Hardcopy Internal Chain)

Worklist Name : Q1380 Worklist ID : 187746 Department : Extraction Date : 02-18-2025 11:35:43

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
Q1380-01	BP-VPB-192-EB-20250212	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	H31	02/12/2025	8270-Modified
Q1380-09	BP-VPB-192-GW-780-782	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	H31	02/13/2025	8270-Modified
Q1380-10	Q1380-09MS	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	H31	02/13/2025	8270-Modified
Q1380-11	Q1380-09MSD	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	H31	02/13/2025	8270-Modified
Q1380-12	VPB192-HYD-20250214	Water	SVOC-SIMGroup1	Cool 4 deg C	TETR06	H31	02/14/2025	8270-Modified

Date/Time 02/18/25 11:37
Raw Sample Received by: RS (Ext lab)
Raw Sample Relinquished by: [Signature]

Date/Time 02/18/25 12:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: RS (Ext lab)

Prep Standard - Chemical Standard Summary

Order ID : Q1380
Test : SVOC-SIMGroup1
Prepbatch ID : PB166758,
Sequence ID/Qc Batch ID: BN022025,

Standard ID :
EP2559,EP2565,EP2589,SP6682,SP6683,SP6684,SP6717,SP6718,SP6730,SP6731,SP6732,SP6733,SP6734,SP6735,SP6736,SP6738,SP6739,

Chemical ID :
1ul/100ul
sample,E3551,E3657,E3828,E3846,E3871,E3873,E3874,E3878,M5173,S10104,S10246,S11074,S11495,S11650,S11785,S11831,S11832,S12114,S12142,S12189,S12208,S12270,S12328,S12469,S12478,S12517,S12525,S12791,S12966,W3112,

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
1874	10 N SODIUM HYDROXIDE SOLN	EP2559	11/14/2024	05/14/2025	Rajesh Parikh	Extraction_SC ALE_2 (EX-SC-2)	None	RUPESHKUMAR SHAH 11/14/2024

FROM 1000.00000ml of W3112 + 400.00000gram of E3657 = Final Quantity: 1000.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
314	1.1 H2SO4 SOLN	EP2565	11/20/2024	05/20/2025	Rajesh Parikh	None	None	RUPESHKUMAR SHAH 11/20/2024

FROM 1000.00000ml of M5173 + 1000.00000ml of W3112 = Final Quantity: 2000.000 ml

Extractions STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3923	Baked Sodium Sulfate	EP2589	02/14/2025	07/01/2025	Rajesh Parikh	Extraction_SC ALE_2 (EX-SC-2)	None	RUPESHKUMAR SHAH 02/14/2025

FROM 4000.00000gram of E3551 = Final Quantity: 4000.000 gram

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3493	Internal Standard 0.4 PPM	SP6682	11/15/2024	05/09/2025	Jagrut Upadhyay	None	None	Yogesh Patel 12/03/2024

FROM 0.10000ml of S12328 + 4.90000ml of E3828 = Final Quantity: 5.000 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3355	8270-SIM MDL-3.2PPM CALIBRATION STOCK SOL- 2ND	SP6683	11/15/2024	04/10/2025	Jagrut Upadhyay	None	None	Yogesh Patel
SOURCE FROM 0.00630ml of S12189 + 0.01280ml of S12208 + 0.03200ml of S11074 + 0.03200ml of S11831 + 0.06400ml of S12142 + 0.06400ml of S12469 + 0.06400ml of S12517 + 19.72490ml of E3828 = Final Quantity: 20.000 ml								

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3356	8270-SIM MDL-0.4PPM CALIBRATION SOL ICV-2ND	SP6684	11/15/2024	04/10/2025	Jagrut Upadhyay	None	None	Yogesh Patel
SOURCE FROM 0.87500ml of E3828 + 0.01000ml of SP6682 + 0.12500ml of SP6683 = Final Quantity: 1.010 ml								

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3895	50 ug/ml DFTPP 8270E	SP6717	01/15/2025	03/31/2025	Rahul Chavli	None	None	Yogesh Patel
								01/16/2025

FROM 1.00000ml of S10246 + 19.00000ml of E3871 = Final Quantity: 20.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3491	8270-SIM-Surrogate 0.4 PPM	SP6718	01/17/2025	04/10/2025	Rahul Chavli	None	None	Shreena Patel
								02/07/2025

FROM 0.00400ml of S12189 + 0.00800ml of S12208 + 0.02000ml of S11831 + 99.96800ml of E3846 = Final Quantity: 100.000 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3339	8270 sim calibration stock 10ppm (CPI)	SP6730	02/04/2025	05/12/2025	Jagrut Upadhyay	None	None	Shreena Patel
								02/07/2025

FROM 0.03350ml of S10104 + 0.05000ml of S11495 + 0.12500ml of S11832 + 0.12500ml of S12114 + 0.25000ml of S12270 + 0.25000ml of S12791 + 24.16650ml of E3874 = Final Quantity: 25.000 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3361	8270-SIM MDL-5PPM CALIBRATION SOLUTION	SP6731	02/04/2025	05/09/2025	Jagrut Upadhyay	None	None	Shreena Patel
								02/07/2025

FROM 0.50000ml of E3874 + 0.01000ml of SP6682 + 0.50000ml of SP6730 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3341	8270-SIM MDL-3.2PPM CALIBRATION SOLUTION	SP6732	02/04/2025	05/09/2025	Jagrut Upadhyay	None	None	Shreena Patel
								02/07/2025

FROM 0.68000ml of E3874 + 0.01000ml of SP6682 + 0.32000ml of SP6730 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3344	8270-SIM MDL-1.6PPM CALIBRATION SOLUTION	SP6733	02/04/2025	05/09/2025	Jagrut Upadhyay	None	None	Shreena Patel
								02/07/2025

FROM 0.84000ml of E3874 + 0.01000ml of SP6682 + 0.16000ml of SP6730 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3342	8270-SIM MDL-0.8PPM CALIBRATION SOLUTION	SP6734	02/04/2025	05/09/2025	Jagrut Upadhyay	None	None	Shreena Patel
								02/07/2025

FROM 0.92000ml of E3874 + 0.01000ml of SP6682 + 0.08000ml of SP6730 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3343	8270-SIM MDL-0.4PPM CALIBRATION SOLUTION	SP6735	02/04/2025	05/09/2025	Jagrut Upadhyay	None	None	Shreena Patel
								02/07/2025

FROM 0.96000ml of E3874 + 0.01000ml of SP6682 + 0.04000ml of SP6730 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3345	8270-SIM MDL-0.2PPM CALIBRATION SOLUTION	SP6736	02/04/2025	05/09/2025	Jagrut Upadhyay	None	None	Shreena Patel
								02/07/2025

FROM 0.50000ml of E3874 + 0.01000ml of SP6682 + 0.50000ml of SP6735 = Final Quantity: 1.010 ml

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3346	8270-SIM MDL-0.1PPM CALIBRATION SOLUTION	SP6738	02/04/2025	05/09/2025	Jagrut Upadhyay	None	None	Shreena Patel
								02/07/2025

FROM 0.75000ml of E3874 + 0.01000ml of SP6682 + 0.25000ml of SP6735 = Final Quantity: 1.010 ml

SVOC STANDARD PREPARATION LOG

<u>Recipe ID</u>	<u>NAME</u>	<u>NO.</u>	<u>Prep Date</u>	<u>Expiration Date</u>	<u>Prepared By</u>	<u>ScaleID</u>	<u>PipetteID</u>	<u>Supervised By</u>
3492	8270-SIM-Spike 0.4 PPM	SP6739	02/05/2025	07/29/2025	Jagrut Upadhyay	None	None	Shreena Patel
FROM 0.00080ml of S11650 + 0.01000ml of S11785 + 0.02000ml of S12478 + 0.02000ml of S12525 + 0.02000ml of S12966 + 49.92920ml of E3873 = Final Quantity: 50.000 ml								

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19631-100 / SODIUM SULFATE, ANHYDROUS, PEST GRADE, 1	313201	07/01/2025	01/03/2024 / Rajesh	07/20/2023 / Rajesh	E3551

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
PCI Scientific Supply, Inc.	PC19510-5 / Sodium Hydroxide Pellets 2.5 Kg, Pk of 4	23B1556310	12/31/2025	12/04/2023 / Rajesh	12/01/2023 / Rajesh	E3657

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24G0862003	05/09/2025	11/09/2024 / Rajesh	11/04/2024 / Rajesh	E3828

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	06/26/2025	12/26/2024 / Rajesh	12/13/2024 / Rajesh	E3846

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24K1762005	07/14/2025	01/14/2025 / Rajesh	12/27/2024 / Rajesh	E3871

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9254-03 / Acetone, Ultra Resi (cs/4x4L)	24H2762008	07/29/2025	01/29/2025 / Rajesh	01/29/2025 / Rajesh	E3873

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	25A0262002	07/30/2025	01/30/2025 / Rajesh	01/20/2025 / Rajesh	E3874

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9644-A4 / Methylene Chloride,U-Resi, Cycle-Tainer (215L)	24K1762005	08/14/2025	02/14/2025 / Rajesh	12/27/2024 / Rajesh	E3878

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	BA-9673-33 / Sulfuric Acid, Instra-Analyzed (cs/6c2.5L)	0000281827	06/02/2025	06/01/2022 /	04/05/2022 / william	M5173

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-112090-04 / CLP Acid Surrogate Solution, 7500 mg/L, 1ml	440246	07/30/2025	01/30/2025 / anahy	12/09/2021 / Christian	S10104

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31615 / SV Mixture, GC/MS Tuning Mixture, CH2Cl2, 1mL,	A0182667	03/31/2025	01/15/2025 / Rahul	03/18/2022 / Christian	S10246

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0187043	05/15/2025	11/15/2024 / Jagrut	02/06/2023 / Christian	S11074

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110094-02 / CLP Base/Neutral Surrogate Solution, 5000 mg/L, 1ml	506889	05/12/2025	11/12/2024 / Jagrut	08/11/2023 / Yogesh	S11495

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555872 / Custom Standard, pentachlorophenol Std [CS 5328-5]	A0201728	07/29/2025	01/29/2025 / anahy	11/09/2023 / Yogesh	S11650

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31853 / 1,4-Dioxane, 2000 ug/ml , Solvent: Methylene Chloride	A0196453	07/29/2025	01/29/2025 / anahy	11/21/2023 / Rahul	S11785

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	33913 / SOM01.0 SIM Analysis Standard (Surrogate), 2000 PPM	A0201976	04/11/2025	10/11/2024 / Jagrut	11/21/2023 / rahul	S11831

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	33913 / SOM01.0 SIM Analysis Standard (Surrogate), 2000 PPM	A0201976	07/24/2025	01/24/2025 / anahy	11/21/2023 / rahul	S11832

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	z-010223-01 / 1,4-Dioxane Solution, 2,000mg/L, 1ml	454157	05/12/2025	11/12/2024 / Jagrut	03/08/2024 / Rahul	S12114

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH ₂ Cl ₂ [New Solvent 100% CH ₂ Cl ₂]	A0203726	04/30/2025	11/14/2024 / anahy	03/15/2024 / Rahul	S12142

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31087 / Acid Surrogate 10,000ug/ml, methanol, 5ml/ ampul	A0206206	04/10/2025	10/10/2024 / anahy	03/15/2024 / Rahul	S12189

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31086 / Base Neutral Surrogate 5000ug/ml, CH ₂ Cl ₂ , 5ml	A0206381	05/15/2025	11/15/2024 / Jagrut	03/15/2024 / Rahul	S12208

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	z-110381-01 / 8270 Calibration Solution, 76-1, 500 & 1,000 mg/L, 1ml	520963	07/30/2025	01/30/2025 / anahy	05/24/2024 / Rahul	S12270

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31206 / SV Mix, CLP method, Internal Std, 2000ug/mL, CH ₂ Cl ₂ , 1mL	A0206540	05/13/2025	11/13/2024 / anahy	05/30/2024 / Rahul	S12328

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0214021	05/14/2025	11/14/2024 / anahy	07/23/2024 / RAHUL	S12469

[CS 4978-1]

CHEMICAL RECEIPT LOG BOOK

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555223 / Custom 8270 Plus Std #1 [2nd lot at \$100 per ampul if requested - contact ARM with Request]	A0214021	07/29/2025	01/29/2025 / anahy	07/23/2024 / RAHUL	S12478

[CS 4978-1]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0214017	05/14/2025	11/14/2024 / anahy	07/23/2024 / RAHUL	S12517

[CS 4978-2]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	555224 / Custom 8270 Plus Std #2 [2nd lot at \$85 per ampul if requested - contact ARM with Request]	A0214017	07/29/2025	01/29/2025 / anahy	07/23/2024 / RAHUL	S12525

[CS 4978-2]

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
CPI International	Z-110816-01 / Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL, (Maximum Expiration: 180 Days)	414127	06/21/2025	01/30/2025 / anahy	05/24/2024 / Rahul	S12791

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Restek	31850 / 8270 SV Mix, 8270 Mega Mix 1mL, 1000ug/mL, CH2Cl2 [New Solvent 100% CH2Cl2]	A0219438	07/29/2025	01/29/2025 / anahy	12/11/2024 / anahy	S12966

Supplier	ItemCode / ItemName	Lot #	Expiration Date	Date Opened / Opened By	Received Date / Received By	Chemtech Lot #
Seidler Chemical	DIW / DI Water	Daily Lab-Certified	07/03/2029	07/03/2024 / lwona	07/03/2024 / lwona	W3112



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.: Z-112090-04 **Lot No.:** 440246 **Storage:** $\leq -10^{\circ}\text{C}$ **Solvent:** Methylene Chloride **Exp. Date:** 2/16/2026 **Description:** CLP Acid Surrogate Solution, 7,500 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
2-chlorophenol-d ₄	93951-73-6	99.3	248.12.7P	7487 $\pm$ 17.2
2-fluorophenol	367-12-4	99.8	10.7.3.3P	7513 $\pm$ 17.26
phenol-d ₆	13127-88-3	99.9	949.120.8P	7481 $\pm$ 17.19
2,4,6-tribromophenol	118-79-6	99.8	12.1.6P	7469 $\pm$ 17.17

Received on

02/25/21

by
CG

S9236
to

S9240

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.

Certified By: Erica Castiglione
Erica Castiglione
Chemist



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received on
03/10/22
by
CG
S10242
to
S10247

Catalog No. : 31615 **Lot No.:** A0182667
Description : GC/MS Tuning Mixture
GC/MS Tuning Mixture 1,000µg/mL, Methylene Chloride, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : March 31, 2025 **Storage:** 10°C or colder
Handling: Contains carcinogen/reproductive toxin. **Ship:** Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)	
1	Pentachlorophenol CAS # 87-86-5 (Lot 211229RSR) Purity 99%	1,003.6 µg/mL	+/- 5.8897 µg/mL +/- 45.7132 µg/mL +/- 66.0037 µg/mL	Gravimetric Unstressed Stressed
2	DFTPP (Decafluorotriphenylphosphine) CAS # 5074-71-5 (Lot Q117-147) Purity 95%	1,006.6 µg/mL	+/- 5.9074 µg/mL +/- 45.8508 µg/mL +/- 66.2023 µg/mL	Gravimetric Unstressed Stressed
3	Benzidine CAS # 92-87-5 (Lot 211228JLM) Purity 99%	1,008.4 µg/mL	+/- 5.9179 µg/mL +/- 45.9318 µg/mL +/- 66.3193 µg/mL	Gravimetric Unstressed Stressed
4	4,4'-DDT CAS # 50-29-3 (Lot 210916JLM) Purity 99%	1,007.6 µg/mL	+/- 5.9132 µg/mL +/- 45.8954 µg/mL +/- 66.2667 µg/mL	Gravimetric Unstressed Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:


250°C

Det. Temp:

330°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Morgan Craighead - Mix Technician

Date Mixed: 08-Mar-2022

Balance: B345965662

Marlene Cowan - Operations Tech I

Date Passed: 10-Mar-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: (800)356-1688
Fax: (814)353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Received on
02/08/23

b1
CG

S 11071
to
S 11075

Catalog No. : 31853 Lot No.: A0187043
Description : 1,4-dioxane
1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul
Container Size : 2 mL Pkg Amt: > 1 mL
Expiration Date : July 31, 2027 Storage: 0°C or colder
Ship: Ambient

CERTIFIED VALUES

Elution Order	Compound	Grav. Conc. (weight/volume)	Expanded Uncertainty (95% C.L.; K=2)
1	1,4-Dioxane CAS # 123-91-1 Purity 99% (Lot SHBN5929)	2,019.0 µg/mL	+/- 11.8486 µg/mL Gravimetric +/- 43.2570 µg/mL Unstressed +/- 44.5129 µg/mL Stressed

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Column:

105m x 0.53mm x 3.0µm
Rtx-502.2 (cat.#10910)

Carrier Gas:

hydrogen-constant pressure 11.0 psi.

Temp. Program:

40°C (hold 2 min.) to 240°C
@ 8°C/min. (hold 5 min.)

Inj. Temp:


200°C

Det. Temp:

250°C

Det. Type:

FID



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Brittany Federinko - Operations Tech I

Date Mixed: 07-Jul-2022

Balance: 1128360905

Marlina Cowan - Operations Tech II ARM QC

Date Passed: 12-Jul-2022

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



**PRODUCTOS
QUÍMICOS
MONTERREY, S.A. DE C.V.**

MIRADOR 201, COL. MIRADOR
MONTERREY, N.L. MEXICO
CP 64070
TEL +52 81 13 52 57 57
www.pqm.com.mx

CERTIFICATE OF ANALYSIS

PRODUCT :	SODIUM SULFATE CRYSTALS ANHYDROUS		
QUALITY :	ACS (CODE RMB3375)	FORMULA :	Na ₂ SO ₄
SPECIFICATION NUMBER :	6399	RELEASE DATE:	ABR/21/2023
LOT NUMBER :	313201		

TEST	SPECIFICATIONS	LOT VALUES
Assay (Na ₂ SO ₄)	Min. 99.0%	99.7 %
pH of a 5% solution at 25°C	5.2 - 9.2	6.1
Insoluble matter	Max. 0.01%	0.005 %
Loss on ignition	Max. 0.5%	0.1 %
Chloride (Cl)	Max. 0.001%	<0.001 %
Nitrogen compounds (as N)	Max. 5 ppm	<5 ppm
Phosphate (PO ₄)	Max. 0.001%	<0.001 %
Heavy metals (as Pb)	Max. 5 ppm	<5 ppm
Iron (Fe)	Max. 0.001%	<0.001 %
Calcium (Ca)	Max. 0.01%	0.002 %
Magnesium (Mg)	Max. 0.005%	0.001 %
Potassium (K)	Max. 0.008%	0.003 %
Extraction-concentration suitability	Passes test	Passes test
Appearance	Passes test	Passes test
Identification	Passes test	Passes test
Solubility and foreign matter	Passes test	Passes test
Retained on US Standard No. 10 sieve	Max. 1%	0.1 %
Retained on US Standard No. 60 sieve	Min. 94%	97.3 %
Through US Standard No. 60 sieve	Max. 5%	2.5 %
Through US Standard No. 100 sieve	Max. 10%	0.1 %
COMMENTS		
QC: PhC Irma Belmares		

If you need further details, please call our factory or contact our local distributor.

Recd. by R3 on 7/24/23 E 3551

RC-02-01, Ed. 3



Certificate of Analysis

Sodium Hydroxide (Pellets)

Material: 0583
Grade: ACS GRADE
Batch Number: 23B1556310

Chemical Formula: NaOH
Molecular Weight: 40
CAS #: 1310-73-2
Appearance:

Manufacture Date: 12/14/2022
Expiration Date: 12/31/2025

Storage: Room Temperature

Pellets

TEST	SPECIFICATION	ANALYSIS	DISPOSITION
Calcium	$\leq 0.005 \%$	$< 0.005 \%$	PASS
Chloride	$\leq 0.005 \%$	0.002 %	PASS
Heavy Metals	$\leq 0.002 \%$	$< 0.002 \%$	PASS
Iron	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Magnesium	$\leq 0.002 \%$	$< 0.002 \%$	PASS
Mercury	$\leq 0.1 \text{ ppm}$	$< 0.1 \text{ ppm}$	PASS
Nickel	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Nitrogen Compounds	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Phosphate	$\leq 0.001 \%$	$< 0.001 \%$	PASS
Potassium	$\leq 0.02 \%$	$< 0.02 \%$	PASS
Purity	$\geq 97.0 \%$	99.2 %	PASS
Sodium Carbonate	$\leq 1.0 \%$	0.5 %	PASS
Sulfate	$\leq 0.003 \%$	$< 0.003 \%$	PASS

Internal ID #: 710

Signature

We certify that this batch conforms to the specifications listed.


This document has been electronically produced and is valid without a signature.

Leona Edwardson, Quality Control Sr. Manager - Solon
VWR Chemicals, LLC.
28600 Fountain Parkway, Solon OH 44139 USA

Additional Information

Analysis may have been rounded to significant digits in specification limits.

Product meets analytical specifications of the grades listed.



Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)

avantor™



Material No.: 9266-A4

Batch No.: 24J0862003

Manufactured Date: 2024-09-12

Expiration Date: 2025-12-12

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	2
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	1
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8\%$	100.0 %
Color (APHA)	≤ 10	5
Residue after Evaporation	≤ 1.0 ppm	0.2 ppm
Titration Acid ($\mu\text{eq/g}$)	≤ 0.3	< 0.1
Chloride (Cl)	≤ 10 ppm	< 5 ppm
Water (by KF, coulometric)	$\leq 0.02\%$	$< 0.01\%$

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

E 3828

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Acetone
BAKER RESI-ANALYZED® Reagent
For Organic Residue Analysis

avantor™



Material No.: 9254-03
Batch No.: 24H2762008
Manufactured Date: 2024-04-18
Expiration Date: 2027-04-18
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.0 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (µeq/g)	<= 0.3	0.2
Titration Base (µeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	<0.1 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

Rec'd by RP On 12/13/24

E 3846

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

12/27/24

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)



Material No.: 9266-A4
Batch No.: 24K1762005
Manufactured Date: 2024-10-08
Expiration Date: 2026-01-07
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	<= 10	2
Assay (CH ₂ Cl ₂) (by GC, exclusive of preservative, corrected for water)	>= 99.8 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.5 ppm
Titration Acid (µeq/g)	<= 0.3	0.0
Chloride (Cl)	<= 10 ppm	<5 ppm
Water (by KF, coulometric)	<= 0.02 %	0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E 3871


Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA, 19087, U.S.A. Phone 610.386.1700

Acetone

BAKER RESI-ANALYZED® Reagent

For Organic Residue Analysis

avantor™



Material No.: 9254-03

Batch No.: 24H2762008

Manufactured Date: 2024-04-18

Expiration Date: 2027-04-18

Revision No.: 0

Certificate of Analysis

Test	Specification	Result
Assay ((CH ₃) ₂ CO) (by GC, corrected for water)	>= 99.4 %	100.0 %
Color (APHA)	<= 10	5
Residue after Evaporation	<= 1.0 ppm	0.0 ppm
Substances Reducing Permanganate	Passes Test	Passes Test
Titration Acid (μeq/g)	<= 0.3	0.2
Titration Base (μeq/g)	<= 0.6	<0.1
Water (H ₂ O)	<= 0.5 %	<0.1 %
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	<= 5	1
ECD Sensitive Impurities (as Heptachlor Epoxide) Single Peak (pg/mL)	<= 10	1

For Laboratory, Research, or Manufacturing Use

MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States

Packaging Site: Phillipsburg Mfg Ctr & DC

Recd. by RP on 1/29/25

E 3873

Jamie Croak
Director Quality Operations, Bioscience Production

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)

Avantor

Material No.: 9266-A4
Batch No.: 25A0262002
Manufactured Date: 2024-11-21
Expiration Date: 2026-02-20
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	4
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8\%$	99.9%
Color (APHA)	≤ 10	10
Residue after Evaporation	≤ 1.0 ppm	0.8 ppm
Titration Acid ($\mu\text{eq/g}$)	≤ 0.3	< 0.1
Chloride (Cl)	≤ 10 ppm	< 5 ppm
Water (by KF, coulometric)	$\leq 0.02\%$	$< 0.01\%$

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E 3874



Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700

Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA, 19087, U.S.A. Phone 610.386.1700

Methylene Chloride
ULTRA RESI-ANALYZED
For Organic Residue Analysis
(dichloromethane)



Material No.: 9266-A4
Batch No.: 24K1762005
Manufactured Date: 2024-10-08
Expiration Date: 2026-01-07
Revision No.: 0

Certificate of Analysis

Test	Specification	Result
FID-Sensitive Impurities (as 2-Octanol) Single Impurity Peak (ng/mL)	≤ 5	1
ECD Sensitive Impurities (as HeptachlorEpoxide) Single Peak (pg/mL)	≤ 10	2
Assay (CH_2Cl_2) (by GC, exclusive of preservative, corrected for water)	$\geq 99.8 \%$	100.0 %
Color (APHA)	≤ 10	5
Residue after Evaporation	$\leq 1.0 \text{ ppm}$	0.5 ppm
Titration Acid ($\mu\text{eq/g}$)	≤ 0.3	0.0
Chloride (Cl)	$\leq 10 \text{ ppm}$	$< 5 \text{ ppm}$
Water (by KF, coulometric)	$\leq 0.02 \%$	0.01 %

For Laboratory, Research, or Manufacturing Use
MEETS SPECIFICATIONS WITHIN THE EXPIRATION PERIOD

Country of Origin: United States
Packaging Site: Phillipsburg Mfg Ctr & DC

E 3878

Jamie Croak
Director Quality Operations, Bioscience Production

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC

100 Matsonford Rd, Suite 200, Radnor, PA, 19087 U.S.A. Phone 610.386.1700

Hydrochloric Acid, 36.5–38.0%
 BAKER INSTRA-ANALYZED® Reagent
 For Trace Metal Analysis



Material No.: 9530-33
 Batch No.: 0000281827
 Manufactured Date: 2021/03/30
 Retest Date: 2026/03/29
 Revision No: 1

Certificate of Analysis

Test	Specification	Result
ACS – Assay (as HCl) (by acid–base titrn)	36.5 – 38.0 %	37.6
ACS – Color (APHA)	<= 10	5
ACS – Residue after Ignition	<= 3 ppm	1
ACS – Specific Gravity at 60°/60°F	1.185 – 1.192	1.189
ACS – Bromide (Br)	<= 0.005 %	< 0.005
ACS – Extractable Organic Substances	<= 5 ppm	< 1
ACS – Free Chlorine (as Cl ₂)	<= 0.5 ppm	< 0.5
Phosphate (PO ₄)	<= 0.05 ppm	< 0.03
Sulfate (SO ₄)	<= 0.5 ppm	< 0.3
Sulfite (SO ₃)	<= 0.8 ppm	0.3
Ammonium (NH ₄)	<= 3 ppm	< 1
Trace Impurities – Arsenic (As)	<= 0.010 ppm	< 0.003
Trace Impurities – Aluminum (Al)	<= 10.0 ppb	0.5
Arsenic and Antimony (as As)	<= 5 ppb	< 3
Trace Impurities – Barium (Ba)	<= 1.0 ppb	< 0.2
Trace Impurities – Beryllium (Be)	<= 1.0 ppb	< 0.2
Trace Impurities – Bismuth (Bi)	<= 10.0 ppb	< 1.0
Trace Impurities – Boron (B)	<= 20.0 ppb	< 5.0
Trace Impurities – Cadmium (Cd)	<= 1.0 ppb	< 0.3
Trace Impurities – Calcium (Ca)	<= 50.0 ppb	15.0
Trace Impurities – Chromium (Cr)	<= 1.0 ppb	< 0.4
Trace Impurities – Cobalt (Co)	<= 1.0 ppb	< 0.3
Trace Impurities – Copper (Cu)	<= 1.0 ppb	< 0.1
Trace Impurities – Gallium (Ga)	<= 1.0 ppb	< 0.2

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
 Avantor Performance Materials, LLC
 100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700

Test	Specification	Result
Trace Impurities – Germanium (Ge)	<= 3.0 ppb	< 2.0
Trace Impurities – Gold (Au)	<= 4.0 ppb	3.0
Heavy Metals (as Pb)	<= 100 ppb	< 50
Trace Impurities – Iron (Fe)	<= 15.0 ppb	1.0
Trace Impurities – Lead (Pb)	<= 1.0 ppb	< 0.5
Trace Impurities – Lithium (Li)	<= 1.0 ppb	< 0.2
Trace Impurities – Magnesium (Mg)	<= 10.0 ppb	< 0.4
Trace Impurities – Manganese (Mn)	<= 1.0 ppb	< 0.4
Trace Impurities – Mercury (Hg)	<= 0.5 ppb	0.2
Trace Impurities – Molybdenum (Mo)	<= 10.0 ppb	< 5.0
Trace Impurities – Nickel (Ni)	<= 4.0 ppb	< 0.3
Trace Impurities – Niobium (Nb)	<= 1.0 ppb	< 0.2
Trace Impurities – Potassium (K)	<= 9.0 ppb	< 2.0
Trace Impurities – Selenium (Se), For Information Only	ppb	1.0
Trace Impurities – Silicon (Si)	<= 100.0 ppb	18.0
Trace Impurities – Silver (Ag)	<= 1.0 ppb	< 0.3
Trace Impurities – Sodium (Na)	<= 100.0 ppb	< 5.0
Trace Impurities – Strontium (Sr)	<= 1.0 ppb	< 0.2
Trace Impurities – Tantalum (Ta)	<= 1.0 ppb	< 0.9
Trace Impurities – Thallium (Tl)	<= 5.0 ppb	< 2.0
Trace Impurities – Tin (Sn)	<= 5.0 ppb	< 0.8
Trace Impurities – Titanium (Ti)	<= 1.0 ppb	< 0.2
Trace Impurities – Vanadium (V)	<= 1.0 ppb	< 0.2
Trace Impurities – Zinc (Zn)	<= 5.0 ppb	0.4
Trace Impurities – Zirconium (Zr)	<= 1.0 ppb	< 0.1

For Laboratory, Research or Manufacturing Use

Product Information (not specifications):

Appearance (clear, fuming liquid)

Meets ACS Specifications

Country of Origin: US

Packaging Site: Phillipsburg Mfg Ctr & DC


Jamie Ethier
Vice President Global Quality

For questions on this Certificate of Analysis please contact Technical Services at 855.282.6867 or +1.610.386.1700
Avantor Performance Materials, LLC
100 Matsonford Rd, Suite 200, Radnor, PA 19087. U.S.A. Phone: 610.386.1700



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110094-02	506889	≤ -10 °C	Methylene Chloride	7/25/2028	CLP Base/Neutral Surrogate Solution, 5,000 mg/L, 1 ml

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,2-dichlorobenzene-d ₄	2199-69-1	99.7	247.29.3P	5035 ± 28.02
2-fluorobiphenyl	321-60-8	99.69	8.286.1.1P	4999 ± 103.66
nitrobenzene-d5	4165-60-0	99.67	7.9.3P	4988 ± 27.32
p-terphenyl-d14	1718-51-0	99.3	9.120.8P	5005 ± 27.85

511494 } Y.P.
↓
511498 } 08/11/2023

*Not a certified value

Certified By: Shane C. Tipton
Clint Tipton
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555872 **Lot No.:** A0201728

Description : Custom Pentachlorophenol Standard
Custom Pentachlorophenol Standard 25,000µg/mL, Methanol, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2026 **Storage:** 10°C or colder

Ship: Ambient

511649
↓
511658
Y.P.
11/13/23

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pentachlorophenol	87-86-5	RP230530RSR	99%	25,000.0 µg/mL	+/- 777.0837

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Josh McCloskey - Operations Technician I

Date Mixed: 05-Sep-2023

Balance: B251644995

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31853 **Lot No.:** A0196453
Description : 1,4-dioxane
1,4-Dioxane 2,000µg/mL, Methylene Chloride, 1mL/ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : March 31, 2028 **Storage:** 0°C or colder
Ship: Ambient

S11749
↓
S11794 } RC / 11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dioxane	123-91-1	SHBN3770	99%	2,013.0 µg/mL	+/- 25.0521

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant flow 1.8 mL/min.

Temp. Program:

80°C (hold 0.1 min.) to 330°C
@ 9.6°C/min. (hold 2.86 min.)

Inj. Temp:

250°C

Det. Temp:

340°C

Det. Type:


FID

Split Vent:

100 mL/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Sam Moodler
Sam Moodler - Operations Tech I

Date Mixed: 30-Mar-2023

Balance Serial # B707717271

Jennifer Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 31-Mar-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 33913 **Lot No.:** A0201976
Description : SOM01.0 SIM Analysis Standard
SOM01.0 SIM Analysis Standard 2000µg/mL, Methylene chloride, 1mL /ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2029 **Storage:** 10°C or colder
Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

S11828
↓
S11832 } RC/
11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2-Methylnaphthalene-d10	7297-45-2	EF-135	98%	2,015.9 µg/mL	+/- 90.8098
2	Fluoranthene-d10	93951-69-0	PR-32557	99%	2,020.0 µg/mL	+/- 90.9963

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 13-Sep-2023

Balance Serial # B442140311


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 28-Sep-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 33913 **Lot No.:** A0201976
Description : SOM01.0 SIM Analysis Standard
SOM01.0 SIM Analysis Standard 2000µg/mL, Methylene chloride, 1mL /ampul
Container Size : 2 mL **Pkg Amt:** > 1 mL
Expiration Date : August 31, 2029 **Storage:** 10°C or colder
Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

S11828
↓
S11832 } RC/
11/30/23

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2-Methylnaphthalene-d10	7297-45-2	EF-135	98%	2,015.9 µg/mL	+/- 90.8098
2	Fluoranthene-d10	93951-69-0	PR-32557	99%	2,020.0 µg/mL	+/- 90.9963

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

10 ml/min.

Inj. Vol

1µl




This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.


Dakota Parson - Operations Technician I

Date Mixed: 13-Sep-2023

Balance Serial # B442140311


Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 28-Sep-2023

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:	
Z-020223-01	454157	≤ -10 °C	P/T Methanol	6/10/2026	1,4-Dioxane Solution, 2000 mg/L, 1 mL	
Compound			CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,4-dioxane			123-91-1	100	223.1.3P	1997 ± 57.08

512112 } RC/
↓
912116 } 03/08/24

*Not a certified value

Certified By: _____

Melissa Workoff
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0203726

Description : 8270 MegaMix®
8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : April 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

512117 } RC/
↓ 03/18/24
512146 }

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,001.6 µg/mL	+/- 36.4412
2	N-Nitrosodimethylamine	62-75-9	230209JLM	99%	1,005.9 µg/mL	+/- 36.5968
3	Phenol	108-95-2	MKCK1120	99%	1,003.3 µg/mL	+/- 36.5038
4	Aniline	62-53-3	X22F726	99%	1,005.8 µg/mL	+/- 36.5928
5	Bis(2-chloroethyl)ether	111-44-4	SHBL6942	99%	1,008.1 µg/mL	+/- 36.6776
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,001.8 µg/mL	+/- 36.4492
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,002.3 µg/mL	+/- 36.4654
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,003.7 µg/mL	+/- 36.5159
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,008.7 µg/mL	+/- 36.6979
10	1,2-Dichlorobenzene	95-50-1	SHBN3835	99%	1,000.3 µg/mL	+/- 36.3926
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,003.5 µg/mL	+/- 36.5099
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,007.3 µg/mL	+/- 36.6493
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	504.3 µg/mL	+/- 18.3500
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.6 µg/mL	+/- 18.3237
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,008.3 µg/mL	+/- 36.6857
16	Hexachloroethane	67-72-1	QTORH	99%	1,007.5 µg/mL	+/- 36.6554
17	Nitrobenzene	98-95-3	10224044	99%	1,008.6 µg/mL	+/- 36.6938

18	Isophorone	78-59-1	MKCC9506	99%	1,005.9	µg/mL	+/- 36.5988	
19	2-Nitrophenol	88-75-5	RP230710	99%	1,003.2	µg/mL	+/- 36.4998	
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,003.8	µg/mL	+/- 36.5200	
21	Bis(2-chloroethoxy)methane	111-91-1	13670200	99%	1,002.1	µg/mL	+/- 36.4573	1
22	2,4-Dichlorophenol	120-83-2	BCBZ6787	99%	1,003.7	µg/mL	+/- 36.5180	2
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,007.6	µg/mL	+/- 36.6574	3
24	Naphthalene	91-20-3	STBL1057	99%	1,008.3	µg/mL	+/- 36.6837	4
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,001.3	µg/mL	+/- 36.4290	5
26	Hexachlorobutadiene	87-68-3	RP230823RSR	98%	1,008.3	µg/mL	+/- 36.6829	6
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,003.1	µg/mL	+/- 36.4937	7
28	2-Methylnaphthalene	91-57-6	STBK0259	96%	1,001.9	µg/mL	+/- 36.4505	8
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	1,000.0	µg/mL	+/- 36.3838	9
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,008.5	µg/mL	+/- 36.6909	10
31	2,4,6-Trichlorophenol	88-06-2	STBJ5914	99%	1,004.4	µg/mL	+/- 36.5442	11
32	2,4,5-Trichlorophenol	95-95-4	FHN01	98%	1,001.9	µg/mL	+/- 36.4512	12
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,001.1	µg/mL	+/- 36.4230	13
34	2-Nitroaniline	88-74-4	RP230531	99%	1,002.9	µg/mL	+/- 36.4876	14
35	1,4-Dinitrobenzene	100-25-4	RP230816	99%	1,005.7	µg/mL	+/- 36.5887	15
36	Acenaphthylene	208-96-8	p06V	98%	1,009.5	µg/mL	+/- 36.7265	16
37	1,3-Dinitrobenzene	99-65-0	1-DXX-24-1	99%	1,004.4	µg/mL	+/- 36.5422	17
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,005.9	µg/mL	+/- 36.5968	18
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,003.2	µg/mL	+/- 36.4998	19
40	1,2-Dinitrobenzene	528-29-0	RP230428	99%	1,002.2	µg/mL	+/- 36.4634	
41	Acenaphthene	83-32-9	MKCR7169	99%	1,009.3	µg/mL	+/- 36.7221	
42	3-Nitroaniline	99-09-2	RP230822RSR	99%	1,003.9	µg/mL	+/- 36.5240	
43	2,4-Dinitrophenol	51-28-5	DR230417RSR	99%	1,002.0	µg/mL	+/- 36.4553	
44	Dibenzofuran	132-64-9	MKCD9952	99%	1,006.7	µg/mL	+/- 36.6251	
45	2,4-Dinitrotoluene	121-14-2	MKAA0690V	99%	1,003.8	µg/mL	+/- 36.5220	
46	4-Nitrophenol	100-02-7	RP230627	99%	1,002.3	µg/mL	+/- 36.4674	
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-30126	99%	1,008.7	µg/mL	+/- 36.6979	
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP230919	99%	1,006.3	µg/mL	+/- 36.6130	
49	Fluorene	86-73-7	10241100	99%	1,008.3	µg/mL	+/- 36.6857	
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,003.8	µg/mL	+/- 36.5220	
51	Diethylphthalate	84-66-2	MKCD2547	99%	1,008.6	µg/mL	+/- 36.6958	
52	4-Nitroaniline	100-01-6	RP230111	99%	1,001.1	µg/mL	+/- 36.4230	
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	230718JLM	99%	1,002.0	µg/mL	+/- 36.4553	

54	Diphenylamine	122-39-4	MKCH1042	99%	1,002.3	µg/mL	+/- 36.4674
55	Azobenzene	103-33-3	BCKK0887	99%	1,005.8	µg/mL	+/- 36.5928
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,003.0	µg/mL	+/- 36.4917
57	Hexachlorobenzene	118-74-1	14821700	99%	1,007.5	µg/mL	+/- 36.6554
58	Pentachlorophenol	87-86-5	RP230530RSR	99%	1,008.8	µg/mL	+/- 36.7019
59	Phenanthrene	85-01-8	MKCQ8876	99%	1,008.4	µg/mL	+/- 36.6877
60	Anthracene	120-12-7	MKCR0570	99%	1,009.0	µg/mL	+/- 36.7100
61	Carbazole	86-74-8	14351100	99%	1,000.9	µg/mL	+/- 36.4149
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,007.6	µg/mL	+/- 36.6595
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,009.6	µg/mL	+/- 36.7302
64	Pyrene	129-00-0	BCCG8479	98%	1,007.2	µg/mL	+/- 36.6453
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,002.1	µg/mL	+/- 36.4573
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.2	µg/mL	+/- 36.5705
67	Benz(a)anthracene	56-55-3	I220012022BAA	99%	1,002.2	µg/mL	+/- 36.4614
68	Chrysene	218-01-9	RP230601	99%	1,008.3	µg/mL	+/- 36.6837
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCQ3468	99%	1,001.8	µg/mL	+/- 36.4472
70	Di-n-octyl phthalate	117-84-0	14382700	99%	1,006.0	µg/mL	+/- 36.6008
71	Benzo(b)fluoranthene	205-99-2	012013B	99%	1,002.8	µg/mL	+/- 36.4836
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,003.0	µg/mL	+/- 36.4917
73	Benzo(a)pyrene	50-32-8	P54915-0703	99%	1,002.3	µg/mL	+/- 36.4674
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,009.4	µg/mL	+/- 36.7243
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,007.6	µg/mL	+/- 36.6595
76	Benzo(g,h,i)perylene	191-24-2	RP231003RSR	99%	1,002.9	µg/mL	+/- 36.4876

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31087 **Lot No.:** A0206206
Description : Acid Surrogate Mix (4/89 SOW)
Acid Surrogate 10, 000µg/mL, Methanol, 5mL/ampul
Container Size : 5 mL **Pkg Amt:** > 5 mL
Expiration Date : January 31, 2032 **Storage:** 10°C or colder
Ship: Ambient

512187 } RC/
↓
512206 } 03/18/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	2-Fluorophenol	367-12-4	STBK1705	99%	10,005.3 µg/mL	+/- 302.5390
2	Phenol-d6	13127-88-3	PR-33287A	99%	10,005.5 µg/mL	+/- 302.5475
3	2,4,6-Tribromophenol	118-79-6	RP230831RSR	99%	10,006.6 µg/mL	+/- 302.5783

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methanol
CAS # 67-56-1
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

2 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Penelope A. Riglin

Penelope Riglin - Operations Tech I

Date Mixed: 04-Jan-2024

Balance Serial # 1128360905

Christie Mills

Christie Mills - Operations Lead Tech - ARM QC

Date Passed: 08-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31086 **Lot No.:** A0206381
Description : B/N Surrogate Mix (4/89 SOW)
Base Neutral Surrogate 5000µg/mL, Methylene Chloride, 5mL/ampul
Container Size : 5 mL **Pkg Amt:** > 5 mL
Expiration Date : December 31, 2029 **Storage:** 10°C or colder
Handling: Sonicate prior to use. **Ship:** Ambient

S12207 } RC/
↓
S12221 } 03/18/24

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Nitrobenzene-d5	4165-60-0	I-25158	99%	5,029.3 µg/mL	+/- 226.5204
2	2-Fluorobiphenyl	321-60-8	00021384	99%	5,030.9 µg/mL	+/- 226.5936
3	p-Terphenyl-d14	1718-51-0	PR-32599	99%	5,026.4 µg/mL	+/- 226.3909

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

Due to the limited solubility of p-terphenyl-d14 in methanol, we do not recommend that this mixture be diluted in methanol.

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-S (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

40°C (hold 2 min.) to 330°C
@ 10°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

2 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Jess Hoy - Operations Tech I

Date Mixed: 09-Jan-2024

Balance Serial # 1128360905

Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 11-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 4

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110381-01	520963	≤ -10 °C	Methylene Chloride	10/10/2028	Method 8270 Calibration Solution, 76-1, 500 & 1,000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
acenaphthene	83-32-9	99.9	13.1.5P	1010 ± 9.89
acenaphthylene	208-96-8	97.6	14.290.1P	1014 ± 9.93
aniline	62-53-3	99.97	64.1.4P	1001 ± 9.8
anthracene	120-12-7	99.5	15.7.1P	999.6 ± 9.79
azobenzene	103-33-3	98.1	252.7.2P	999.1 ± 9.8
benzo[a]anthracene	56-55-3	100	16.7.3P	1007 ± 9.86
benzo[b]fluoranthene	205-99-2	99.8	17.421.3P	1011 ± 14.11
benzo[k]fluoranthene	207-08-9	98.9	18.421.4P	1001 ± 10.96
benzo[ghi]perylene	191-24-2	93	19.286.4P	999.6 ± 13.95
benzo[a]pyrene	50-32-8	97	20.286.2P	999.9 ± 22.24
benzyl alcohol	100-51-6	99.9	65.18.1P	1001 ± 9.82
bis(2-chloroethoxy)methane	111-91-1	99.1	31.3.15P	1000 ± 14.69
bis(2-chloroethyl)ether	111-44-4	99.8	32.7.1P	1003 ± 13.89
bis(2-chloro-1-methylethyl) ether	108-60-1	99.5	34.3.15P	999.4 ± 14.68
bis(2-ethylhexyl)adipate	103-23-1	99.5	874.7.1P	999.5 ± 9.8
bis(2-ethylhexyl)phthalate	117-81-7	99.4	33.29.1P	998.8 ± 17.03
4-bromophenyl phenyl ether	101-55-3	99.4	35.7.1.1P	1000 ± 13.85
butyl benzyl phthalate	85-68-7	98.4	36.1.6P	984.7 ± 16.79
carbazole	86-74-8	99.4	239.7.2P	1000 ± 9.8

*Not a certified value

512270 } RC/
↓
512274 } 05/24/24

Kerry Kane

Certified By: _____

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.

Certificate of Analysis

Page 2 of 4

Catalog No.: Z-110381-01

Lot No.: 520963

Expiration Date: 10/10/2028

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
4-chloroaniline	106-47-8	100	66.7.1P	1000 ± 9.79
4-chlorophenylphenyl ether	7005-72-3	98	37.158.2P	1001 ± 17.07
4-chloro-3-methylphenol	59-50-7	99	102.1.2P	1006 ± 17.16
2-chloronaphthalene	91-58-7	99.9	42.7.6P	1000 ± 9.79
2-chlorophenol	95-57-8	99.8	103.7.1P	1007 ± 13.96
chrysene	218-01-9	96	21.286.2P	998.4 ± 12.85
dibenz[a,h]anthracene	53-70-3	99.44	22.286.3P	1000 ± 9.74
dibenzofuran	132-64-9	100	67.7.2.1P	1002 ± 9.77
di-n-butyl phthalate	84-74-2	99.84	40.286.1P	1007 ± 24.48
1,2-dichlorobenzene	95-50-1	99.8	43.7.1P	1000 ± 9.79
1,3-dichlorobenzene	541-73-1	99.5	44.1.3P	999.4 ± 9.79
1,4-dichlorobenzene	106-46-7	99.9	45.29.2P	1000 ± 9.79
2,4-dichlorophenol	120-83-2	99.6	104.7.1.1P	1005 ± 13.93
diethyl phthalate	84-66-2	99.8	38.7.1P	1011 ± 14
2,4-dimethylphenol	105-67-9	99.6	105.7.1.1P	1009 ± 13.98
dimethyl phthalate	131-11-3	99.9	39.9.2P	996.5 ± 13.8
1,2-dinitrobenzene	528-29-0	99.86	86.7.3.1P	999.5 ± 9.75
1,3-dinitrobenzene	99-65-0	100	313.7.2P	998 ± 9.79
1,4-dinitrobenzene	100-25-4	100	907.7.1P	999.5 ± 9.8
2,4-dinitrophenol	51-28-5	99.9	106.1.6DP	1002 ± 13.89
2,4-dinitrotoluene	121-14-2	100	87.7.3P	999.8 ± 13.85
2,6-dinitrotoluene	606-20-2	99.4	88.7.2.1P	999.6 ± 13.85
di-n-octyl phthalate	117-84-0	99.1	41.7.5P	991.6 ± 13.74
diphenylamine	122-39-4	100	78.1.6P	998 ± 13.79
2,3,5,6-tetrachlorophenol	935-95-5	97	1112.286.1P	1004 ± 14.02
fluoranthene	206-44-0	98.6	23.7.4P	999.6 ± 9.79
fluorene	86-73-7	98.4	24.7.1P	999.7 ± 9.79

*Not a certified value

Kerry Kane

Certified By:

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.

Certificate of Analysis

Page 3 of 4

Catalog No.: Z-110381-01

Lot No.: 520963

Expiration Date: 10/10/2028

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
hexachlorobenzene	118-74-1	99	46.158.4P	999.9 ± 13.96
hexachlorobutadiene	87-68-3	97.4	47.1.4P	1000 ± 9.79
hexachlorocyclopentadiene	77-47-4	99.2	48.2.2P	1001 ± 9.8
hexachloroethane	67-72-1	99.9	49.1.4P	1003 ± 9.82
indeno[1,2,3-cd]pyrene	193-39-5	98	25.286.4P	999.4 ± 22.23
isophorone	78-59-1	98.9	90.1.4P	999.9 ± 13.85
2-methyl-4,6-dinitrophenol	534-52-1	99.6	107.421.2DP	991 ± 24.09
1-methylnaphthalene	90-12-0	97.1	249.7.5P	999.2 ± 13.95
2-methylnaphthalene	91-57-6	97.4	68.7.2P	1006 ± 22.38
2-methylphenol	95-48-7	99.6	114.7.3P	1001 ± 13.87
3-methylphenol	108-39-4	99.1	115.7.4P	499.7 ± 6.92
4-methylphenol	106-44-5	99.5	116.7.1P	501.2 ± 6.94
naphthalene	91-20-3	99.8	26.9.1P	1018 ± 9.97
2-nitroaniline	88-74-4	99.7	69.29.1P	999.6 ± 9.79
3-nitroaniline	99-09-2	100	70.7.3P	1000 ± 9.74
4-nitroaniline	100-01-6	99.7	71.29.1P	1001 ± 9.8
nitrobenzene	98-95-3	100	94.7.1P	1000 ± 13.85
2-nitrophenol	88-75-5	99.1	108.29.1P	996.5 ± 13.81
4-nitrophenol	100-02-7	100	109.7.1P	1000 ± 13.82
N-nitrosodimethylamine	62-75-9	99.5	57.3.19P	998.5 ± 14.67
N-nitrosodi-n-propylamine	621-64-7	99.8	59.286.1P	996.8 ± 17
pentachlorophenol	87-86-5	99	110.1.7P	1004 ± 13.92
phenanthrene	85-01-8	99.7	27.1.5P	999 ± 12.87
phenol	108-95-2	100	112.7.1P	998.5 ± 13.8
pyrene	129-00-0	99.2	28.9.2P	998.9 ± 9.78
pyridine	110-86-1	100	101.24.1P	999 ± 9.73
2,3,4,6-Tetrachlorophenol	58-90-2	91.8	120.421.1P	996.5 ± 13.92

*Not a certified value

Kerry Kane

Certified By:

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.

Certificate of Analysis

Page 4 of 4

Catalog No.: Z-110381-01

Lot No.: 520963

Expiration Date: 10/10/2028

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
1,2,4-trichlorobenzene	120-82-1	99.6	54.29.1P	999.6 ± 9.79
2,4,5-trichlorophenol	95-95-4	96.5	121.7.1.1P	999.5 ± 13.85
2,4,6-trichlorophenol	88-06-2	99.6	113.7.1P	996 ± 13.8

*Not a certified value



Certified By:

Kerry Kane
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values
listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31206 **Lot No.:** A0206540

Description : SV Internal Standard Mix 2mg/ml
SV Internal Standard Mix 2mg/ml 2000 µg/ml, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : December 31, 2029 **Storage:** 10°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

S12312 } RC/
↓ 05/30/24
S12331

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,4-Dichlorobenzene-d4	3855-82-1	PR-30447	99%	2,007.1 µg/mL	+/- 90.4025
2	Naphthalene-d8	1146-65-2	M-2180	99%	2,005.9 µg/mL	+/- 90.3454
3	Acenaphthene-d10	15067-26-2	PR-33507	99%	2,007.9 µg/mL	+/- 90.4385
4	Phenanthrene-d10	1517-22-2	PR-32303	99%	2,006.7 µg/mL	+/- 90.3845
5	Chrysene-d12	1719-03-5	PR-32210	99%	2,015.5 µg/mL	+/- 90.7778
6	Perylene-d12	1520-96-3	PR-33205	99%	2,014.7 µg/mL	+/- 90.7448

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Quality Confirmation Test

Column:

30m x 0.25mm x 0.25µm
Rtx-5 (cat.#10223)

Carrier Gas:

hydrogen-constant pressure 10 psi.

Temp. Program:

75°C (hold 1 min.) to 330°C
@ 20°C/min. (hold 10 min.)

Inj. Temp:

250°C

Det. Temp:

330°C

Det. Type:


FID

Split Vent:

10 ml/min.

Inj. Vol

1µl



This chromatogram represents a general set of testing conditions chosen for product acceptance. For optimal results in your lab, conditions should be adjusted for your specific instrument, method, and application.

Malina Homan
Malina Homan - Operations Technician I

Date Mixed: 12-Jan-2024

Balance Serial # 1128360905

Jennifer J. Pollino
Jennifer Pollino - Operations Tech III - ARM QC

Date Passed: 16-Jan-2024

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1
Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 } RC/
↓ 7/24/24
S12508 }

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024

Balance: 1128353505

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555223 **Lot No.:** A0214021

Description : Custom 8270 Plus Standard #1
Custom 8270 Plus Standard #1 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Handling: This product is photosensitive. **Ship:** Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	3,3'-Dichlorobenzidine	91-94-1	S240326RSR	99%	1,004.0 µg/mL	+/- 23.0487
2	Atrazine	1912-24-9	5FYWL	99%	1,005.0 µg/mL	+/- 23.0717
3	Benzidine	92-87-5	S240430RSR	99%	1,006.0 µg/mL	+/- 23.0947
4	epsilon-Caprolactam	105-60-2	Y16H012	99%	1,000.0 µg/mL	+/- 22.9569

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12449 } RC/
↓ 7/24/24
S12508 }

Rebecca Gingerich - Operations Tech II

Date Mixed: 18-Jul-2024

Balance: 1128353505

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/ μ ECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0214017

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,005.0 µg/mL	+/- 29.541899
2	Acetophenone	98-86-2	STBH8205	99%	1,005.0 µg/mL	+/- 29.541899
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,008.0 µg/mL	+/- 29.630084
4	Benzoic acid	65-85-0	MKCR2694	99%	1,010.0 µg/mL	+/- 29.688874
5	Biphenyl	92-52-4	MKCS5928	99%	1,008.0 µg/mL	+/- 29.630084

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12509 } RC/
↓
S12568 } 7/24/24

Jess Hoy - Operations Tech I

Date Mixed: 18-Jul-2024

Balance: 1128360905

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

gravimetric



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 555224 **Lot No.:** A0214017

Description : Custom 8270 Plus Standard #2

Custom 8270 Plus Standard #2 1,000µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : July 31, 2026 **Storage:** 10°C or colder

Ship: Ambient

CERTIFIED VALUES

Component #	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	1,2,4,5-Tetrachlorobenzene	95-94-3	MKCT9480	99%	1,005.0 µg/mL	+/- 29.541899
2	Acetophenone	98-86-2	STBH8205	99%	1,005.0 µg/mL	+/- 29.541899
3	Benzaldehyde	100-52-7	RD231129RSRA	99%	1,008.0 µg/mL	+/- 29.630084
4	Benzoic acid	65-85-0	MKCR2694	99%	1,010.0 µg/mL	+/- 29.688874
5	Biphenyl	92-52-4	MKCS5928	99%	1,008.0 µg/mL	+/- 29.630084

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

S12509 } RC/
↓
S12568 } 7/24/24

Jess Hoy - Operations Tech I

Date Mixed: 18-Jul-2024

Balance: 1128360905

Manufactured under Restek's ISO 9001:2015
Registered Quality System
Certificate #FM 80397

General Certified Reference Material Notes

Expiration Notes:

- Expiration date valid for unopened ampul stored in compliance with the recommended conditions.
- Uncertainty, concentration, and expiration of the CRM are based on the unopened product being stored according to the recommended condition found in the storage field.

Purity Notes:

- Purity and/or chemical identity are determined by one or more of the following techniques: GC/FID, HPLC, GC/μECD, GC/MS, LC/MS, RI, and/or melting point.
- Compounds with a listed purity of less than 99% have been weight corrected to compensate for impurities and/or salts. A correction factor is used to calculate the amount of compound necessary to achieve the desired concentration of the parent compound in solution.
- Purity of isomeric compounds is reported as the sum of the isomers.
- Purity values are rounded to the nearest whole number.

Certified Uncertainty Value Notes:

- The uncertainties are determined in accordance with ISO 17034 and Guide 35. The certified expanded uncertainty value includes gravimetric uncertainty, homogeneity between-ampul uncertainty, storage stability uncertainty and shipping stability uncertainty and were combined using the following formula:

$$U_{combined\ uncertainty} = k \sqrt{u_{gravimetric}^2 + u_{homogeneity}^2 + u_{storage\ stability}^2 + u_{shipping\ stability}^2}$$

k is a coverage factor of 2, which gives a level of confidence of approximately 95%.

- The packaged amount is the minimum sample size for which uncertainty is valid. The ampuls are over-filled to ensure that the minimum packaged amount can be sufficiently transferred.

Manufacturing Notes:

- Concentration is based upon gravimetric preparation using either a balance whose calibration has been verified daily using NIST traceable weights, and/or dilutions with Class A glassware.

Handling Notes:

- Stability of the unopened product, when stored in compliance with the recommended conditions, is guaranteed through the expiration displayed on the product label and certificate. Contact Restek for additional opened product stability information, with the knowledge/understanding that open product stability is subject to the specific handling and environmental conditions to which the product is exposed. For your convenience Restek supplies deactivated vials with most standards packed in 2mL ampuls. Larger volume deactivated vials are available through Restek as a custom ordered item. Additionally, Restek sells DMDCS for the purpose of glassware deactivation as catalog number 31861, which includes complete instructions.
- If any undissolved material is visible inside the ampul, sonicate the unopened ampul until the material is completely dissolved.



5580 Skylane Blvd
Santa Rosa, CA 95403

(707)525-5788
(800)878-7654 Toll Free
(707)545-7901 Fax

Manufacturer's Quality System
Audited & Registered
by TUV USA to ISO 9001:2015

Date Received: _____

Certificate of Analysis

Rev 0

Page 1 of 1

Catalog No.:	Lot No.:	Storage:	Solvent:	Exp. Date:	Description:
Z-110816-01	414127	≤ -10 °C	Methylene Chloride	6/21/2025	Custom 8270 Mix, 4-79, 1000 mg/L, 1 mL

Compound	CAS No.	Purity (%)	Compound Lot No.	Concentration, mg/L
atrazine	1912-24-9	99.5	337.7.3P	997 ± 5.81
benzidine	92-87-5	99.9	124.18.6.2P	991.8 ± 5.77
caprolactam	105-60-2	99.9	271.1.6P	999 ± 5.82

~~512280~~ } RCL
↓
~~512284~~ } 05/24/24

New Numbers Generated.

512790 } RCL
↓
512794 } 11/12/24

*Not a certified value

Manufactured by o2si smart solutions, Accredited to ISO 9001:2008 by NSF and ISO/IEC 17025:2005 (Certification No. 3031.01) and ISO Guide 34:2009 (Certification No. 3031.02) by A2LA

Certified By: _____

Shane Overcash
Chemist

All weights are traceable through N. I. S. T. Test No. 822/264157-00.
Concentration (correct for purity) and uncertainty (95% confidence) values listed are determined gravimetrically.



110 Benner Circle
Bellefonte, PA 16823-8812
Tel: 1-814-353-1300
Fax: 1-814-353-1309

www.restek.com

CERTIFIED REFERENCE MATERIAL

Certificate of Analysis

chromatographic plus



FOR LABORATORY USE ONLY-READ SDS PRIOR TO USE.

This Reference Material is intended for Laboratory Use Only as a standard for the qualitative and/or quantitative determination of the analyte(s) listed.

Catalog No. : 31850 **Lot No.:** A0219438

Description : 8270 MegaMix®
8270 MegaMix® 500-1000 µg/mL, Methylene Chloride, 1mL/ampul

Container Size : 2 mL **Pkg Amt:** > 1 mL

Expiration Date : September 30, 2025 **Storage:** 0°C or colder

Handling: Sonication required. Mix is photosensitive. **Ship:** Ambient

S12963
↓
S12992 } AC
12/17/20

CERTIFIED VALUES

Elution Order	Compound	CAS #	Lot #	Purity	Grav. Conc. (weight/volume)	Expanded Uncertainty * (95% C.L.; K=2)
1	Pyridine	110-86-1	SHBP6240	99%	1,008.3 µg/mL	+/- 36.6849
2	N-Nitrosodimethylamine	62-75-9	S240313RSR	99%	1,008.6 µg/mL	+/- 36.6985
3	Phenol	108-95-2	MKCK1120	99%	1,003.5 µg/mL	+/- 36.5120
4	Aniline	62-53-3	X22F726	99%	1,002.9 µg/mL	+/- 36.4893
5	Bis(2-chloroethyl)ether	111-44-4	002891T24M	99%	1,003.0 µg/mL	+/- 36.4938
6	2-Chlorophenol	95-57-8	STBJ3909	99%	1,005.6 µg/mL	+/- 36.5894
7	1,3-Dichlorobenzene	541-73-1	BCCD5315	99%	1,004.1 µg/mL	+/- 36.5348
8	1,4-Dichlorobenzene	106-46-7	MKBS7929V	99%	1,002.1 µg/mL	+/- 36.4620
9	Benzyl alcohol	100-51-6	SHBK5469	99%	1,003.5 µg/mL	+/- 36.5120
10	1,2-Dichlorobenzene	95-50-1	SHBL6287	99%	1,005.3 µg/mL	+/- 36.5757
11	2-Methylphenol (o-cresol)	95-48-7	SHBN7598	99%	1,008.4 µg/mL	+/- 36.6894
12	2,2'-oxybis(1-chloropropane)	108-60-1	29-MAR-45-5	99%	1,004.6 µg/mL	+/- 36.5530
13	3-Methylphenol (m-cresol)	108-39-4	STBJ0710	99%	502.1 µg/mL	+/- 18.2697
14	4-Methylphenol (p-cresol)	106-44-5	SHBN3411	99%	503.8 µg/mL	+/- 18.3288
15	N-Nitroso-di-n-propylamine	621-64-7	N63MG	99%	1,006.5 µg/mL	+/- 36.6212
16	Hexachloroethane	67-72-1	DAXRI	99%	1,004.5 µg/mL	+/- 36.5484
17	Nitrobenzene	98-95-3	10224044	99%	1,002.5 µg/mL	+/- 36.4757

18	Isophorone	78-59-1	MKCR3249	99%	1,003.4	µg/mL	+/- 36.5075
19	2-Nitrophenol	88-75-5	RP230710	99%	1,002.5	µg/mL	+/- 36.4757
20	2,4-Dimethylphenol	105-67-9	XW5GK	99%	1,006.5	µg/mL	+/- 36.6212
21	Bis(2-chloroethoxy)methane	111-91-1	15705100	99%	1,006.6	µg/mL	+/- 36.6257
22	2,4-Dichlorophenol	120-83-2	BCKK6969	99%	1,001.5	µg/mL	+/- 36.4393
23	1,2,4-Trichlorobenzene	120-82-1	SHBP5900	99%	1,006.4	µg/mL	+/- 36.6166
24	Naphthalene	91-20-3	STBL1057	99%	1,002.1	µg/mL	+/- 36.4620
25	4-Chloroaniline	106-47-8	BCCJ3217	99%	1,004.4	µg/mL	+/- 36.5439
26	Hexachlorobutadiene	87-68-3	X05J	98%	1,002.5	µg/mL	+/- 36.4771
27	4-Chloro-3-methylphenol	59-50-7	BCCD4461	99%	1,004.5	µg/mL	+/- 36.5484
28	2-Methylnaphthalene	91-57-6	STBL3028	99%	1,000.0	µg/mL	+/- 36.3847
29	1-Methylnaphthalene	90-12-0	5234.00-8	98%	990.2	µg/mL	+/- 36.0269
30	Hexachlorocyclopentadiene	77-47-4	099063I14L	98%	1,001.3	µg/mL	+/- 36.4325
31	2,4,6-Trichlorophenol	88-06-2	STBK8870	99%	1,006.4	µg/mL	+/- 36.6166
32	2,4,5-Trichlorophenol	95-95-4	3YFRE	97%	1,004.6	µg/mL	+/- 36.5505
33	2-Chloronaphthalene	91-58-7	RPN7O	99%	1,004.3	µg/mL	+/- 36.5393
34	2-Nitroaniline	88-74-4	RP240715RSR	99%	1,004.4	µg/mL	+/- 36.5439
35	1,4-Dinitrobenzene	100-25-4	RP240703RSR	99%	1,002.8	µg/mL	+/- 36.4847
36	Acenaphthylene	208-96-8	RP241029RSR	98%	1,000.0	µg/mL	+/- 36.3835
37	1,3-Dinitrobenzene	99-65-0	TRC3-1075941-2-1	99%	1,006.3	µg/mL	+/- 36.6121
38	Dimethylphthalate	131-11-3	358221L17K	99%	1,008.9	µg/mL	+/- 36.7076
39	2,6-Dinitrotoluene	606-20-2	BCCG1833	99%	1,006.6	µg/mL	+/- 36.6257
40	1,2-Dinitrobenzene	528-29-0	RP240701RSR	99%	1,002.5	µg/mL	+/- 36.4757
41	Acenaphthene	83-32-9	MKCR7169	99%	1,000.0	µg/mL	+/- 36.3847
42	3-Nitroaniline	99-09-2	RP240708RSR	99%	1,004.6	µg/mL	+/- 36.5530
43	2,4-Dinitrophenol	51-28-5	D240927RSR	----%	1,005.6	µg/mL	+/- 36.5894
44	Dibenzofuran	132-64-9	MKCN1772	99%	1,003.5	µg/mL	+/- 36.5120
45	2,4-Dinitrotoluene	121-14-2	102869V26E	99%	1,008.3	µg/mL	+/- 36.6849
46	4-Nitrophenol	100-02-7	20241029-2-AN	99%	1,004.8	µg/mL	+/- 36.5575
47	2,3,4,6-Tetrachlorophenol	58-90-2	PR-34476	99%	1,005.8	µg/mL	+/- 36.5939
48	2,3,5,6-Tetrachlorophenol	935-95-5	RP231219RSR	99%	1,006.4	µg/mL	+/- 36.6166
49	Fluorene	86-73-7	10246250	98%	1,000.7	µg/mL	+/- 36.4102
50	4-Chlorophenyl phenyl ether	7005-72-3	MKCT7248	99%	1,004.9	µg/mL	+/- 36.5621
51	Diethylphthalate	84-66-2	BCCJ6241	99%	1,003.9	µg/mL	+/- 36.5257
52	4-Nitroaniline	100-01-6	RP230111	99%	1,006.6	µg/mL	+/- 36.6257
53	4,6-Dinitro-2-methylphenol (Dinitro-o-cresol)	534-52-1	S241008RSR	99%	1,001.3	µg/mL	+/- 36.4302

54	Diphenylamine	122-39-4	MKCT1512	99%	1,003.0	µg/mL	+/- 36.4938
55	Azobenzene	103-33-3	BCKK0887	99%	1,002.4	µg/mL	+/- 36.4711
56	4-Bromophenyl phenyl ether	101-55-3	STBH6361	99%	1,008.8	µg/mL	+/- 36.7031
57	Hexachlorobenzene	118-74-1	15458400	99%	1,005.1	µg/mL	+/- 36.5712
58	Pentachlorophenol	87-86-5	RP240517RSR	99%	1,005.9	µg/mL	+/- 36.5984
59	Phenanthrene	85-01-8	MKCT3391	99%	1,004.9	µg/mL	+/- 36.5621
60	Anthracene	120-12-7	101492T18R	99%	1,005.1	µg/mL	+/- 36.5712
61	Carbazole	86-74-8	15276700	99%	1,005.4	µg/mL	+/- 36.5803
62	Di-n-butylphthalate	84-74-2	MKCN4337	99%	1,006.3	µg/mL	+/- 36.6121
63	Fluoranthene	206-44-0	MKCQ4728	99%	1,003.5	µg/mL	+/- 36.5120
64	Pyrene	129-00-0	BCKK2592	99%	1,002.0	µg/mL	+/- 36.4575
65	Benzyl butyl phthalate	85-68-7	X12I018	99%	1,007.5	µg/mL	+/- 36.6576
66	Bis(2-ethylhexyl)adipate	103-23-1	MKCM1988	99%	1,005.9	µg/mL	+/- 36.5984
67	Benz(a)anthracene	56-55-3	I70012022BAA	99%	1,005.5	µg/mL	+/- 36.5848
68	Chrysene	218-01-9	RP241007RSR	99%	1,005.3	µg/mL	+/- 36.5757
69	Bis(2-ethylhexyl)phthalate	117-81-7	MKCS8065	99%	1,007.5	µg/mL	+/- 36.6576
70	Di-n-octyl phthalate	117-84-0	15566400	99%	1,002.3	µg/mL	+/- 36.4666
71	Benzo(b)fluoranthene	205-99-2	052013B	99%	1,004.1	µg/mL	+/- 36.5348
72	Benzo(k)fluoranthene	207-08-9	012022K	99%	1,002.8	µg/mL	+/- 36.4847
73	Benzo(a)pyrene	50-32-8	NQLXA	98%	1,006.2	µg/mL	+/- 36.6108
74	Indeno(1,2,3-cd)pyrene	193-39-5	12-JKL-118-9	97%	1,001.8	µg/mL	+/- 36.4490
75	Dibenz(a,h)anthracene	53-70-3	2-ASA-59-1	99%	1,003.3	µg/mL	+/- 36.5029
76	Benzo(g,h,i)perylene	191-24-2	RP241014RSR	98%	1,003.8	µg/mL	+/- 36.5217

* Expanded Uncertainty displayed in same units as Grav. Conc.

Solvent: Methylene chloride
CAS # 75-09-2
Purity 99%

Tech Tips:

N-Nitrosodiphenylamine (86-30-6) is prone to breakdown in the injection port and will be converted to Diphenylamine (122-39-4). When comparing the response of Diphenylamine to mixtures manufactured using N-Nitrosodiphenylamine, a difference in response will be observed. The ratio of the MW can be used to calculate the theoretical concentration of the N-Nitrosodiphenylamine.



SHIPPING DOCUMENTS

- 1
- 2
- 3
- 4
- 5
- 6
- 7
- 8
- 9
- 10
- 11
- 12
- 13
- 14
- 15
- 16
- 17
- 18
- 19

CHEMTECH CHAIN OF CUSTODY RECORD		284 Sheffield Street, Mountainside, NJ 07092 (908) 789-8900 Fax: (908) 78-8922 www.chemtech.net		Chemtech Project Number: Q1380										
CLIENT INFORMATION		PROJECT INFORMATION		BILLING INFORMATION										
COMPANY: Tetra Tech		PROJECT NAME: NWIRP Bethpage		BILL TO: SEE CONTRACT PO#										
ADDRESS: 4433 Corporation Lane Suite 300		PROJECT #: 112G08005-WE13 LOCATION: VPB-192		ADDRESS:										
CITY: Virginia Beach STATE: VA ZIP: 23462		PROJECT MANAGER: Ernie Wu		CITY: STATE: ZIP:										
ATTENTION: Ernie Wu		E-MAIL: ernie.wu@tetrattech.com		ATTENTION: PHONE:										
PHONE: 757-466-4901 FAX: 757-461-4148		PHONE: 757-466-4901 FAX: 757-461-4148												
DATA TURNAROUND INFORMATION		DATA DELIVERABLE INFORMATION		ANALYSIS										
FAX: _____ 10 _____ DAYS* HARD COPY: _____ 10 _____ DAYS* EDD _____ 10 _____ DAYS* * TO BE APPROVED BY CHEMTECH STANDARD TURNAROUND TIME IS 10 BUSINESS DAYS		☐ RESULTS ONLY ☐ USEPA CLP ☐ RESULTS + QC ☐ New York State ASP "B" ☐ New Jersey REDUCED ☐ New York State ASP "A" ☐ New Jersey CLP ☐ Other _____ ☐ EDD Format _____		VOC(SW846-8260B) 1,4 Dioxane (8270 SIM) <table border="1" style="border-collapse: collapse; text-align: center;"> <tr> <td style="width: 20px;">1</td> <td style="width: 20px;">2</td> <td style="width: 20px;">3</td> <td style="width: 20px;">4</td> <td style="width: 20px;">5</td> <td style="width: 20px;">6</td> <td style="width: 20px;">7</td> <td style="width: 20px;">8</td> <td style="width: 20px;">9</td> </tr> </table>		1	2	3	4	5	6	7	8	9
1	2	3	4	5	6	7	8	9						
				PRESERVATIVES										
CHEMTECH SAMPLE ID	PROJECT SAMPLE IDENTIFICATION	SAMPLE MATRIX	SAMPLE TYPE COMP GRAB	SAMPLE COLLECTION DATE TIME	# of Bottles									
1.	BP-VPB-192-EB-20250212	QA	X	2/12/25 15:00	3									
2.	BP-VPB-192-TB-20250210	QA	X	2/10/25 9:00	2									
3.	BP-VPB-192-GW-725-727	AQ	X	2/10/25 15:00	6									
4.	BP-VPB-192-GW-780-782	AQ	X	2/13/25 10:05	5									
5.	VPB-192-HYD-20250214	AQ	X	2/14/25 13:30	3									
6.														
7.														
8.														
9.														
10.														

SAMPLE CUSTODY MUST BE DOCUMENTED BELOW EACH TIME SAMPLES CHANGE PROSESSION INCLUDING COURIER DELIVERY					
RELINQUISHED BY SAMPLER	DATE/TIME	RECEIVED BY	Conditions of bottles or coolers at receipt: q Compliant q Non Compliant q Cooler Temp 24C MeOH extraction requires an additional 4oz. Jar for percent solid Comments: Standard TAT q Ice in Cooler?: _____		
1.	2/17/25				
RELINQUISHED BY	DATE/TIME	RECEIVED BY	SHIPPED VIA: CLIENT: ☐ Hand Delivered ☐ Overnight CHEMTECH: ☐ Picked Up ☐ Overnight		
2.					
RELINQUISHED BY	DATE/TIME	RECEIVED FOR LAB BY	<u>Shipment Complete</u> ☐ YES ☐ NO		
3.	2-17-25				

Page 1 of 1

WHITE - CHEMTECH COPY FOR RETURN TO CLIENT YELLOW - CHEMTECH COPY PINK - SAMPLER COPY

Laboratory Certification

Certified By	License No.
CAS EPA CLP Contract	68HERH20D0011
Connecticut	PH-0830
DOD ELAP (ANAB)	L2219
Maine	2024021
Maryland	296
New Hampshire	255424 Rev 1
New Jersey	20012
New York	11376
Pennsylvania	68-00548
Soil Permit	525-24-234-08441
Texas	T104704488

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1380	TETR06	Order Date : 2/17/2025 3:59:00 PM	Project Mgr :
Client Name : Tetra Tech NUS, Inc.		NWIRP Bethpage CTO WE13 - VPB-192 #112G08005	Report Type : Level 4
Client Contact : Ernie Wu		Receive DateTime : 2/17/2025 6:01:00 PM	EDD Type : ADAPT
Invoice Name : Tetra Tech NUS, Inc.		Purchase Order :	Hard Copy Date :
Invoice Contact : Ernie Wu			Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES
Q1380-01	BP-VPB-192-EB-20250212	Water	02/12/2025	15:00					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q1380-02	BP-VPB-192-TB-20250210	Water	02/10/2025	09:00					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q1380-06	BP-VPB-192-GW-725-727	Solid	02/10/2025	15:00					
					VOCMS Group1		8260D	10 Bus. Days	
Q1380-07	Q1380-06MS	Solid	02/10/2025	15:00					
					VOCMS Group1		8260D	10 Bus. Days	
Q1380-08	Q1380-06MSD	Solid	02/10/2025	15:00					
					VOCMS Group1		8260D	10 Bus. Days	
Q1380-09	~~BP-VPB-192-GW-780-727~~ BP-VPB-192-GW-780-782	Water	02/13/2025	10:05					
					VOCMS Group1		8260-Low	10 Bus. Days	
Q1380-12	VPB192-HYD-20250214	Water	02/14/2025	13:30					
					VOCMS Group1		8260-Low	10 Bus. Days	

LOGIN REPORT/SAMPLE TRANSFER

Order ID : Q1380	TETR06	Order Date : 2/17/2025 3:59:00 PM	Project Mgr :
Client Name : Tetra Tech NUS, Inc.		NWIRP Bethpage CTO WE13 - VPB-192 #112G08005	Report Type : Level 4
Client Contact : Ernie Wu		Project Name : CTO WE13	EDD Type : ADAPT
Invoice Name : Tetra Tech NUS, Inc.		Receive DateTime : 2/18/2025 6:01:00 PM	Hard Copy Date :
Invoice Contact : Ernie Wu		Purchase Order :	Date Signoff :

LAB ID	CLIENT ID	MATRIX	SAMPLE DATE	SAMPLE TIME	TEST	TEST GROUP	METHOD	FAX DATE	DUE DATES

Relinquished By :

Date / Time : 2-18-25 1110

Received By :

Date / Time : 02/18/25 11:00

Storage Area : VOA Refridgerator Room

Reg # 6
F22
Reg # 4

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
 Data File : BN036477.D
 Acq On : 20 Feb 2025 12:36
 Operator : RC/JU
 Sample : Q1380-10MS
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-192-GW-780-782MS

Quant Time: Feb 21 00:10:19 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 01:17:14 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.746	152	1709	0.400	ng	0.00
7) Naphthalene-d8	10.541	136	3995	0.400	ng	# 0.00
13) Acenaphthene-d10	14.387	164	2495	0.400	ng	0.00
19) Phenanthrene-d10	17.124	188	5565	0.400	ng	#-0.01
29) Chrysene-d12	21.313	240	5523	0.400	ng	0.00
35) Perylene-d12	23.586	264	5300	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.341	112	974	0.241	ng	0.00
5) Phenol-d6	6.923	99	1025	0.216	ng	-0.01
8) Nitrobenzene-d5	8.897	82	1261	0.320	ng	-0.01
11) 2-Methylnaphthalene-d10	12.126	152	1968	0.321	ng	-0.02
14) 2,4,6-Tribromophenol	15.882	330	500	0.404	ng	0.00
15) 2-Fluorobiphenyl	13.008	172	4139	0.441	ng	-0.01
27) Fluoranthene-d10	19.160	212	6898	0.446	ng	0.00
31) Terphenyl-d14	19.763	244	6357	0.539	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.261	88	493	0.264	ng	# 50
3) n-Nitrosodimethylamine	3.572	42	905	0.279	ng	# 93
6) bis(2-Chloroethyl)ether	7.168	93	1675	0.338	ng	98
9) Naphthalene	10.583	128	3881	0.337	ng	98
10) Hexachlorobutadiene	10.872	225	805	0.287	ng	# 98
12) 2-Methylnaphthalene	12.202	142	2609	0.345	ng	97
16) Acenaphthylene	14.099	152	4341	0.394	ng	99
17) Acenaphthene	14.441	154	2787	0.379	ng	98
18) Fluorene	15.435	166	4131	0.394	ng	97
20) 4,6-Dinitro-2-methylph...	15.522	198	376	0.344	ng	# 81
21) 4-Bromophenyl-phenylether	16.329	248	1426	0.430	ng	# 74
22) Hexachlorobenzene	16.441	284	1658	0.404	ng	99
23) Atrazine	16.590	200	1195	0.431	ng	# 92
24) Pentachlorophenol	16.788	266	1439	0.739	ng	97
25) Phenanthrene	17.173	178	7646	0.475	ng	99
26) Anthracene	17.260	178	6840	0.482	ng	99
28) Fluoranthene	19.192	202	9601	0.486	ng	98
30) Pyrene	19.554	202	10137	0.477	ng	100
32) Benzo(a)anthracene	21.304	228	8882	0.489	ng	99
33) Chrysene	21.348	228	9509	0.483	ng	99
34) Bis(2-ethylhexyl)phtha...	21.232	149	5246	0.463	ng	99
36) Indeno(1,2,3-cd)pyrene	25.884	276	8360	0.451	ng	97
37) Benzo(b)fluoranthene	22.899	252	8647	0.495	ng	99
38) Benzo(k)fluoranthene	22.946	252	8871	0.494	ng	99
39) Benzo(a)pyrene	23.484	252	7527	0.494	ng	97
40) Dibenzo(a,h)anthracene	25.899	278	6349	0.434	ng	99
41) Benzo(g,h,i)perylene	26.586	276	6415	0.387	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
Data File : BN036477.D
Acq On : 20 Feb 2025 12:36
Operator : RC/JU
Sample : Q1380-10MS
Misc :
ALS Vial : 6 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MS


Quant Time: Feb 21 00:10:19 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 01:17:14 2025
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.746 min Scan# 64
 Delta R.T. -0.007 min
 Lab File: BN036477.D
 Acq: 20 Feb 2025 12:36


Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-192-GW-780-782MS





Tgt Ion:152 Resp: 1709
 Ion Ratio Lower Upper
 152 100
 150 155.3 123.7 185.5
 115 67.7 52.5 78.7




#2
 1,4-Dioxane
 Concen: 0.264 ng
 RT: 3.261 min Scan# 25
 Delta R.T. -0.007 min
 Lab File: BN036477.D
 Acq: 20 Feb 2025 12:36



Tgt Ion: 88 Resp: 493
 Ion Ratio Lower Upper
 88 100
 43 128.0 33.7 50.5#
 58 93.9 68.9 103.3









#3
 n-Nitrosodimethylamine
 Concen: 0.279 ng
 RT: 3.572 min Scan# 61
 Delta R.T. -0.007 min
 Lab File: BN036477.D
 Acq: 20 Feb 2025 12:36




Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-192-GW-780-782MS


Tgt Ion: 42 Resp: 905
 Ion Ratio Lower Upper
 42 100
 74 94.6 71.8 107.6
 44 18.8 7.8 11.6#



#4
 2-Fluorophenol
 Concen: 0.241 ng
 RT: 5.341 min Scan# 313
 Delta R.T. -0.007 min
 Lab File: BN036477.D
 Acq: 20 Feb 2025 12:36

Tgt Ion: 112 Resp: 974
 Ion Ratio Lower Upper
 112 100
 64 69.7 53.4 80.0
 63 40.5 30.3 45.5









#5
Phenol-d6
Concen: 0.216 ng
RT: 6.923 min Scan# 51
Delta R.T. -0.015 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MS


Tgt Ion: 99 Resp: 1025
Ion Ratio Lower Upper
99 100
42 39.0 21.7 32.5#
71 43.8 32.6 49.0



#6
bis(2-Chloroethyl)ether
Concen: 0.338 ng
RT: 7.168 min Scan# 566
Delta R.T. -0.015 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36

Tgt Ion: 93 Resp: 1675
Ion Ratio Lower Upper
93 100
63 84.7 66.3 99.5
95 34.0 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.541 min Scan# 911
Delta R.T. -0.000 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MS


Tgt Ion:	136	Resp:	3995
Ion	Ratio	Lower	Upper
136	100		
137	13.1	10.1	15.1
54	10.6	11.8	17.6#
68	7.2	7.2	10.8



#8
Nitrobenzene-d5
Concen: 0.320 ng
RT: 8.897 min Scan# 779
Delta R.T. -0.011 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36

Tgt Ion:	82	Resp:	1261
Ion	Ratio	Lower	Upper
82	100		
128	40.4	31.9	47.9
54	68.9	53.1	79.7





#9

Naphthalene

Concen: 0.337 ng

RT: 10.583 min Scan# 91

Delta R.T. -0.011 min

Lab File: BN036477.D

Acq: 20 Feb 2025 12:36





Instrument :

BNA_N

ClientSampleId :

BP-VPB-192-GW-780-782MS

Tgt Ion:	128	Resp:	3881
Ion	Ratio	Lower	Upper
128	100		
129	12.7	9.6	14.4
127	15.6	12.0	18.0



#10

Hexachlorobutadiene

Concen: 0.287 ng




RT: 10.872 min Scan# 964


Delta R.T. -0.011 min

Lab File: BN036477.D

Acq: 20 Feb 2025 12:36

Tgt Ion:	225	Resp:	805
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	62.0	50.9	76.3









#11
2-Methylnaphthalene-d10
Concen: 0.321 ng
RT: 12.126 min Scan# 1145
Delta R.T. -0.015 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MS


Tgt Ion:152 Resp: 1968
Ion Ratio Lower Upper
152 100
151 20.7 16.6 25.0



#12
2-Methylnaphthalene
Concen: 0.345 ng
RT: 12.202 min Scan# 1160
Delta R.T. -0.010 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36

Tgt Ion:142 Resp: 2609
Ion Ratio Lower Upper
142 100
141 94.2 72.8 109.2
115 46.7 35.5 53.3









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.387 min Scan# 1448
Delta R.T. -0.000 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MS


Tgt Ion:	164	Resp:	2495
Ion	Ratio	Lower	Upper
164	100		
162	101.2	84.1	126.1
160	50.7	41.4	62.0



#14
2,4,6-Tribromophenol
Concen: 0.404 ng
RT: 15.882 min Scan# 1583
Delta R.T. -0.000 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36

Tgt Ion:	330	Resp:	500
Ion	Ratio	Lower	Upper
330	100		
332	94.0	76.6	114.8
141	48.0	37.8	56.8









#15
2-Fluorobiphenyl
Concen: 0.441 ng
RT: 13.008 min Scan# 1319
Delta R.T. -0.011 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MS


Tgt Ion:172	Resp:	4139
Ion Ratio	Lower	Upper
172	100	
171	37.6	29.6 44.4
170	25.1	19.8 29.6



#16
Acenaphthylene
Concen: 0.394 ng
RT: 14.099 min Scan# 1421
Delta R.T. -0.011 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36

Tgt Ion:152	Resp:	4341
Ion Ratio	Lower	Upper
152	100	
151	20.3	15.8 23.8
153	13.3	10.2 15.2









#17
Acenaphthene
Concen: 0.379 ng
RT: 14.441 min Scan# 1453
Delta R.T. -0.011 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MS


Tgt Ion:154 Resp: 2787
Ion Ratio Lower Upper
154 100
153 119.2 93.3 139.9
152 61.4 48.8 73.2



#18
Fluorene
Concen: 0.394 ng
RT: 15.435 min Scan# 1546
Delta R.T. -0.011 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36

Tgt Ion:166 Resp: 4131
Ion Ratio Lower Upper
166 100
165 102.8 79.5 119.3
167 13.5 10.4 15.6









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.124 min Scan# 1683
Delta R.T. -0.012 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MS


Tgt Ion:188 Resp: 5565
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 16.0 9.8 14.6#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.344 ng
RT: 15.522 min Scan# 1554
Delta R.T. -0.001 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36

Tgt Ion:198 Resp: 376
Ion Ratio Lower Upper
198 100
51 90.2 86.6 129.8
105 54.5 57.5 86.3#









#21
4-Bromophenyl-phenylether
Concen: 0.430 ng
RT: 16.329 min Scan# 1619
Delta R.T. -0.000 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MS


Tgt Ion:248 Resp: 1426
Ion Ratio Lower Upper
248 100
250 103.2 76.1 114.1
141 46.7 71.7 107.5#



#22
Hexachlorobenzene
Concen: 0.404 ng
RT: 16.441 min Scan# 1628
Delta R.T. -0.012 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36

Tgt Ion:284 Resp: 1658
Ion Ratio Lower Upper
284 100
142 42.2 33.4 50.0
249 34.9 28.6 43.0









#23
Atrazine
Concen: 0.431 ng
RT: 16.590 min Scan# 1640
Delta R.T. -0.025 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MS


Tgt Ion:200 Resp: 1195
Ion Ratio Lower Upper
200 100
173 35.5 23.2 34.8#
215 46.2 40.0 60.0



#24
Pentachlorophenol
Concen: 0.739 ng
RT: 16.788 min Scan# 1656
Delta R.T. -0.012 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36

Tgt Ion:266 Resp: 1439
Ion Ratio Lower Upper
266 100
264 61.4 50.6 76.0
268 62.2 51.9 77.9









#25
Phenanthrene
Concen: 0.475 ng
RT: 17.173 min Scan# 1687
Delta R.T. -0.000 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MS


Tgt Ion:178 Resp: 7646
Ion Ratio Lower Upper
178 100
176 20.2 15.7 23.5
179 15.6 12.4 18.6



#26
Anthracene
Concen: 0.482 ng
RT: 17.260 min Scan# 1694
Delta R.T. -0.012 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36

Tgt Ion:178 Resp: 6840
Ion Ratio Lower Upper
178 100
176 19.0 14.9 22.3
179 14.9 12.4 18.6









#27
Fluoranthene-d10
Concen: 0.446 ng
RT: 19.160 min Scan# 1936
Delta R.T. -0.009 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MS


Tgt Ion:212 Resp: 6898
Ion Ratio Lower Upper
212 100
106 15.1 11.5 17.3
104 9.8 7.1 10.7



#28
Fluoranthene
Concen: 0.486 ng
RT: 19.192 min Scan# 1941
Delta R.T. -0.009 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36

Tgt Ion:202 Resp: 9601
Ion Ratio Lower Upper
202 100
101 12.7 9.2 13.8
203 17.3 13.4 20.0









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.313 min Scan# 21
Delta R.T. -0.009 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36




Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MS


Tgt Ion:240 Resp: 5523
Ion Ratio Lower Upper
240 100
120 19.3 13.3 19.9
236 29.6 23.0 34.6



#30
Pyrene
Concen: 0.477 ng
RT: 19.554 min Scan# 2019
Delta R.T. -0.009 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36

Tgt Ion:202 Resp: 10137
Ion Ratio Lower Upper
202 100
200 21.2 16.9 25.3
203 17.6 13.9 20.9









#31
 Terphenyl-d14
 Concen: 0.539 ng
 RT: 19.763 min Scan# 2064
 Delta R.T. -0.005 min
 Lab File: BN036477.D
 Acq: 20 Feb 2025 12:36




Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-192-GW-780-782MS


Tgt Ion:244 Resp: 6357
 Ion Ratio Lower Upper
 244 100
 212 10.7 8.1 12.1
 122 15.1 12.8 19.2



#32
 Benzo(a)anthracene
 Concen: 0.489 ng
 RT: 21.304 min Scan# 2261
 Delta R.T. -0.009 min
 Lab File: BN036477.D
 Acq: 20 Feb 2025 12:36

Tgt Ion:228 Resp: 8882
 Ion Ratio Lower Upper
 228 100
 226 28.0 22.2 33.2
 229 20.7 16.5 24.7









#33
 Chrysene
 Concen: 0.483 ng
 RT: 21.348 min Scan# 21
 Delta R.T. -0.009 min
 Lab File: BN036477.D
 Acq: 20 Feb 2025 12:36




Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-192-GW-780-782MS


Tgt Ion:	228	Resp:	9509
Ion Ratio	Lower	Upper	
228	100		
226	31.6	25.5	38.3
229	20.0	16.4	24.6



#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.463 ng
 RT: 21.232 min Scan# 2253
 Delta R.T. -0.009 min
 Lab File: BN036477.D
 Acq: 20 Feb 2025 12:36


Tgt Ion:	149	Resp:	5246
Ion Ratio	Lower	Upper	
149	100		
167	26.2	21.2	31.8
279	2.9	2.7	4.1








#35
Perylene-d12
Concen: 0.400 ng
RT: 23.586 min Scan# 21
Delta R.T. -0.009 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36


Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MS





Tgt Ion:264 Resp: 5300
Ion Ratio Lower Upper
264 100
260 26.0 20.9 31.3
265 91.1 60.7 91.1#




#36
Indeno(1,2,3-cd)pyrene
Concen: 0.451 ng
RT: 25.884 min Scan# 3574
Delta R.T. -0.015 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36




Tgt Ion:276 Resp: 8360
Ion Ratio Lower Upper
276 100
138 30.2 22.2 33.2
277 24.7 19.8 29.6







#37
Benzo(b)fluoranthene
Concen: 0.495 ng
RT: 22.899 min Scan# 21
Delta R.T. -0.012 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36

Instrument :
BNA_N
ClientSampleId :
BP-VPB-192-GW-780-782MS







Tgt Ion:252 Resp: 8647
Ion Ratio Lower Upper
252 100
253 27.4 21.9 32.9
125 20.3 15.0 22.6



#38
Benzo(k)fluoranthene
Concen: 0.494 ng
RT: 22.946 min Scan# 2569
Delta R.T. -0.012 min
Lab File: BN036477.D
Acq: 20 Feb 2025 12:36

Tgt Ion:252 Resp: 8871
Ion Ratio Lower Upper
252 100
253 27.5 22.2 33.4
125 19.6 15.0 22.4





#39

Benzo(a)pyrene

Concen: 0.494 ng

RT: 23.484 min Scan# 21

Delta R.T. -0.015 min

Lab File: BN036477.D


Acq: 20 Feb 2025 12:36

Instrument :

BNA_N

ClientSampleId :

BP-VPB-192-GW-780-782MS






Tgt Ion:252 Resp: 7527

Ion Ratio Lower Upper

252 100

253 30.6 24.4 36.6

125 25.6 18.2 27.2



#40

Dibenzo(a,h)anthracene

Concen: 0.434 ng

RT: 25.899 min Scan# 3579

Delta R.T. -0.018 min

Lab File: BN036477.D

Acq: 20 Feb 2025 12:36




Tgt Ion:278 Resp: 6349


Ion Ratio Lower Upper

278 100

139 23.3 18.5 27.7


279 30.5 24.8 37.2







#41
 Benzo(g,h,i)perylene
 Concen: 0.387 ng
 RT: 26.586 min Scan# 31
 Delta R.T. -0.015 min
 Lab File: BN036477.D
 Acq: 20 Feb 2025 12:36

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB-192-GW-780-782MS



Tgt Ion:	276	Resp:	6415
Ion	Ratio	Lower	Upper
276	100		
277	26.3	20.7	31.1
138	29.4	21.8	32.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
 Data File : BN036480.D
 Acq On : 20 Feb 2025 14:24
 Operator : RC/JU
 Sample : PB166758BS
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB166758BS

Quant Time: Feb 21 00:11:38 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 01:17:14 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.746	152	1827	0.400	ng	0.00
7) Naphthalene-d8	10.541	136	4127	0.400	ng	# 0.00
13) Acenaphthene-d10	14.388	164	2357	0.400	ng	0.00
19) Phenanthrene-d10	17.136	188	5256	0.400	ng	0.00
29) Chrysene-d12	21.322	240	4823	0.400	ng	0.00
35) Perylene-d12	23.589	264	4362	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.341	112	1752	0.406	ng	0.00
5) Phenol-d6	6.923	99	1995	0.394	ng	-0.01
8) Nitrobenzene-d5	8.897	82	1572	0.386	ng	-0.01
11) 2-Methylnaphthalene-d10	12.131	152	3269	0.515	ng	-0.01
14) 2,4,6-Tribromophenol	15.882	330	432	0.370	ng	0.00
15) 2-Fluorobiphenyl	13.008	172	4173	0.471	ng	-0.01
27) Fluoranthene-d10	19.164	212	5915	0.405	ng	0.00
31) Terphenyl-d14	19.764	244	4423	0.430	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.261	88	699	0.350	ng	# 51
3) n-Nitrosodimethylamine	3.572	42	1437	0.414	ng	# 97
6) bis(2-Chloroethyl)ether	7.168	93	2058	0.388	ng	98
9) Naphthalene	10.584	128	4763	0.400	ng	99
10) Hexachlorobutadiene	10.872	225	1195	0.412	ng	# 100
12) 2-Methylnaphthalene	12.207	142	3110	0.398	ng	98
16) Acenaphthylene	14.099	152	4781	0.459	ng	100
17) Acenaphthene	14.452	154	3067	0.441	ng	99
18) Fluorene	15.435	166	4315	0.436	ng	99
20) 4,6-Dinitro-2-methylph...	15.523	198	393	0.381	ng	92
21) 4-Bromophenyl-phenylether	16.329	248	1336	0.426	ng	# 88
22) Hexachlorobenzene	16.441	284	1675	0.433	ng	98
23) Atrazine	16.602	200	1144	0.437	ng	95
24) Pentachlorophenol	16.789	266	915	0.498	ng	99
25) Phenanthrene	17.173	178	6880	0.453	ng	99
26) Anthracene	17.260	178	6286	0.469	ng	99
28) Fluoranthene	19.192	202	8127	0.435	ng	99
30) Pyrene	19.559	202	8425	0.454	ng	100
32) Benzo(a)anthracene	21.304	228	7065	0.445	ng	99
33) Chrysene	21.358	228	8131	0.473	ng	99
34) Bis(2-ethylhexyl)phtha...	21.241	149	3779	0.382	ng	100
36) Indeno(1,2,3-cd)pyrene	25.890	276	7355	0.482	ng	99
37) Benzo(b)fluoranthene	22.902	252	6950	0.484	ng	97
38) Benzo(k)fluoranthene	22.949	252	7711	0.521	ng	97
39) Benzo(a)pyrene	23.490	252	6505	0.519	ng	97
40) Dibenzo(a,h)anthracene	25.908	278	5331	0.443	ng	99
41) Benzo(g,h,i)perylene	26.589	276	5921	0.434	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
Data File : BN036480.D
Acq On : 20 Feb 2025 14:24
Operator : RC/JU
Sample : PB166758BS
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB166758BS


Quant Time: Feb 21 00:11:38 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 01:17:14 2025
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.746 min Scan# 64
 Delta R.T. -0.007 min
 Lab File: BN036480.D
 Acq: 20 Feb 2025 14:24


Instrument :
 BNA_N
 ClientSampleId :
 PB166758BS





Tgt Ion:152 Resp: 1827
 Ion Ratio Lower Upper
 152 100
 150 158.6 123.7 185.5
 115 70.3 52.5 78.7




#2
 1,4-Dioxane
 Concen: 0.350 ng
 RT: 3.261 min Scan# 25
 Delta R.T. -0.007 min
 Lab File: BN036480.D
 Acq: 20 Feb 2025 14:24



Tgt Ion: 88 Resp: 699
 Ion Ratio Lower Upper
 88 100
 43 125.0 33.7 50.5#
 58 94.6 68.9 103.3









#3
n-Nitrosodimethylamine
Concen: 0.414 ng
RT: 3.572 min Scan# 61
Delta R.T. -0.007 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24




Instrument :
BNA_N
ClientSampleId :
PB166758BS


Tgt Ion: 42 Resp: 1437
Ion Ratio Lower Upper
42 100
74 88.9 71.8 107.6
44 17.9 7.8 11.6#



#4
2-Fluorophenol
Concen: 0.406 ng
RT: 5.341 min Scan# 313
Delta R.T. -0.007 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24

Tgt Ion: 112 Resp: 1752
Ion Ratio Lower Upper
112 100
64 67.2 53.4 80.0
63 37.0 30.3 45.5









#5
Phenol-d6
Concen: 0.394 ng
RT: 6.923 min Scan# 51
Delta R.T. -0.014 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24




Instrument :
BNA_N
ClientSampleId :
PB166758BS


Tgt Ion: 99 Resp: 1995
Ion Ratio Lower Upper
99 100
42 31.6 21.7 32.5
71 42.0 32.6 49.0



#6
bis(2-Chloroethyl)ether
Concen: 0.388 ng
RT: 7.168 min Scan# 566
Delta R.T. -0.014 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24


Tgt Ion: 93 Resp: 2058
Ion Ratio Lower Upper
93 100
63 84.9 66.3 99.5
95 32.0 26.2 39.4








#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.541 min Scan# 911
Delta R.T. 0.000 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24


Instrument :
BNA_N
ClientSampleId :
PB166758BS





Tgt Ion:136	Resp:	4127
Ion	Ratio	Lower Upper
136	100	
137	13.1	10.1 15.1
54	11.3	11.8 17.6#
68	7.4	7.2 10.8




#8
Nitrobenzene-d5
Concen: 0.386 ng
RT: 8.897 min Scan# 779
Delta R.T. -0.011 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24



Tgt Ion: 82	Resp:	1572
Ion	Ratio	Lower Upper
82	100	
128	38.5	31.9 47.9
54	67.2	53.1 79.7









#9
Naphthalene
Concen: 0.400 ng
RT: 10.584 min Scan# 91
Delta R.T. -0.011 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24




Instrument :
BNA_N
ClientSampleId :
PB166758BS


Tgt Ion:	128	Resp:	4763
Ion	Ratio	Lower	Upper
128	100		
129	12.5	9.6	14.4
127	15.2	12.0	18.0



#10
Hexachlorobutadiene
Concen: 0.412 ng
RT: 10.872 min Scan# 964
Delta R.T. -0.011 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24

Tgt Ion:	225	Resp:	1195
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	63.4	50.9	76.3









#11
2-Methylnaphthalene-d10
Concen: 0.515 ng
RT: 12.131 min Scan# 1146
Delta R.T. -0.010 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24




Instrument :
BNA_N
ClientSampleId :
PB166758BS


Tgt Ion:152 Resp: 3269
Ion Ratio Lower Upper
152 100
151 15.2 16.6 25.0#



#12
2-Methylnaphthalene
Concen: 0.398 ng
RT: 12.207 min Scan# 1161
Delta R.T. -0.005 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24

Tgt Ion:142 Resp: 3110
Ion Ratio Lower Upper
142 100
141 90.1 72.8 109.2
115 46.4 35.5 53.3









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.388 min Scan# 1448
Delta R.T. 0.000 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24




Instrument :
BNA_N
ClientSampleId :
PB166758BS


Tgt Ion:164 Resp: 2357
Ion Ratio Lower Upper
164 100
162 104.0 84.1 126.1
160 52.6 41.4 62.0



#14
2,4,6-Tribromophenol
Concen: 0.370 ng
RT: 15.882 min Scan# 1583
Delta R.T. 0.000 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24

Tgt Ion:330 Resp: 432
Ion Ratio Lower Upper
330 100
332 93.8 76.6 114.8
141 49.5 37.8 56.8









#15
2-Fluorobiphenyl
Concen: 0.471 ng
RT: 13.008 min Scan# 11
Delta R.T. -0.011 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24




Instrument :
BNA_N
ClientSampleId :
PB166758BS


Tgt Ion:	172	Resp:	4173
Ion Ratio	Lower	Upper	
172	100		
171	38.0	29.6	44.4
170	25.5	19.8	29.6



#16
Acenaphthylene
Concen: 0.459 ng
RT: 14.099 min Scan# 1421
Delta R.T. -0.011 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24

Tgt Ion:	152	Resp:	4781
Ion Ratio	Lower	Upper	
152	100		
151	19.7	15.8	23.8
153	12.9	10.2	15.2









#17
Acenaphthene
Concen: 0.441 ng
RT: 14.452 min Scan# 1454
Delta R.T. 0.000 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24




Instrument :
BNA_N
ClientSampleId :
PB166758BS


Tgt Ion:	154	Resp:	3067
Ion Ratio	Lower	Upper	
154	100		
153	117.2	93.3	139.9
152	61.5	48.8	73.2



#18
Fluorene
Concen: 0.436 ng
RT: 15.435 min Scan# 1546
Delta R.T. -0.011 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24

Tgt Ion:	166	Resp:	4315
Ion Ratio	Lower	Upper	
166	100		
165	100.1	79.5	119.3
167	13.5	10.4	15.6





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.136 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN036480.D

Acq: 20 Feb 2025 14:24

Instrument :

BNA_N

ClientSampleId :

PB166758BS





Tgt Ion:188 Resp: 5256

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 10.7 9.8 14.6



#20

4,6-Dinitro-2-methylphenol

Concen: 0.381 ng

RT: 15.523 min Scan# 1554

Delta R.T. -0.000 min

Lab File: BN036480.D

Acq: 20 Feb 2025 14:24




Tgt Ion:198 Resp: 393


Ion Ratio Lower Upper

198 100

51 102.4 86.6 129.8

105 62.7 57.5 86.3









#21
4-Bromophenyl-phenylether
Concen: 0.426 ng
RT: 16.329 min Scan# 1619
Delta R.T. 0.000 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24




Instrument :
BNA_N
ClientSampleId :
PB166758BS


Tgt Ion:248 Resp: 1336
Ion Ratio Lower Upper
248 100
250 94.7 76.1 114.1
141 65.9 71.7 107.5#



#22
Hexachlorobenzene
Concen: 0.433 ng
RT: 16.441 min Scan# 1628
Delta R.T. -0.012 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24

Tgt Ion:284 Resp: 1675
Ion Ratio Lower Upper
284 100
142 42.7 33.4 50.0
249 34.7 28.6 43.0





#23

Atrazine

Concen: 0.437 ng

RT: 16.602 min Scan# 1

Delta R.T. -0.012 min

Lab File: BN036480.D

Acq: 20 Feb 2025 14:24

Instrument :

BNA_N

ClientSampleId :

PB166758BS





Tgt Ion:200 Resp: 1144

Ion Ratio Lower Upper

200 100

173 32.2 23.2 34.8

215 47.2 40.0 60.0



#24

Pentachlorophenol

Concen: 0.498 ng

RT: 16.789 min Scan# 1656

Delta R.T. -0.012 min

Lab File: BN036480.D

Acq: 20 Feb 2025 14:24




Tgt Ion:266 Resp: 915


Ion Ratio Lower Upper

266 100

264 63.0 50.6 76.0

268 66.7 51.9 77.9









#25
Phenanthrene
Concen: 0.453 ng
RT: 17.173 min Scan# 1687
Delta R.T. 0.000 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24




Instrument :
BNA_N
ClientSampleId :
PB166758BS


Tgt Ion:178 Resp: 6880
Ion Ratio Lower Upper
178 100
176 19.8 15.7 23.5
179 15.8 12.4 18.6



#26
Anthracene
Concen: 0.469 ng
RT: 17.260 min Scan# 1694
Delta R.T. -0.012 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24

Tgt Ion:178 Resp: 6286
Ion Ratio Lower Upper
178 100
176 18.8 14.9 22.3
179 15.0 12.4 18.6









#27
Fluoranthene-d10
Concen: 0.405 ng
RT: 19.164 min Scan# 1935
Delta R.T. -0.005 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24




Instrument :
BNA_N
ClientSampleId :
PB166758BS


Tgt Ion:212 Resp: 5915
Ion Ratio Lower Upper
212 100
106 14.7 11.5 17.3
104 9.4 7.1 10.7



#28
Fluoranthene
Concen: 0.435 ng
RT: 19.192 min Scan# 1941
Delta R.T. -0.009 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24

Tgt Ion:202 Resp: 8127
Ion Ratio Lower Upper
202 100
101 11.2 9.2 13.8
203 17.0 13.4 20.0





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.322 min Scan# 21

Delta R.T. 0.000 min

Lab File: BN036480.D

Acq: 20 Feb 2025 14:24

Instrument :

BNA_N

ClientSampleId :

PB166758BS





Tgt Ion:240 Resp: 4823

Ion Ratio Lower Upper

240 100

120 14.4 13.3 19.9

236 28.9 23.0 34.6



#30

Pyrene

Concen: 0.454 ng

RT: 19.559 min Scan# 2020

Delta R.T. -0.005 min

Lab File: BN036480.D

Acq: 20 Feb 2025 14:24




Tgt Ion:202 Resp: 8425


Ion Ratio Lower Upper

202 100

200 21.3 16.9 25.3


203 17.7 13.9 20.9








#31
 Terphenyl-d14
 Concen: 0.430 ng
 RT: 19.764 min Scan# 2064
 Delta R.T. -0.005 min
 Lab File: BN036480.D
 Acq: 20 Feb 2025 14:24

Instrument :
 BNA_N
 ClientSampleId :
 PB166758BS







Tgt Ion:244 Resp: 4423
 Ion Ratio Lower Upper
 244 100
 212 11.0 8.1 12.1
 122 17.2 12.8 19.2



#32
 Benzo(a)anthracene
 Concen: 0.445 ng
 RT: 21.304 min Scan# 2261
 Delta R.T. -0.009 min
 Lab File: BN036480.D
 Acq: 20 Feb 2025 14:24

Tgt Ion:228 Resp: 7065
 Ion Ratio Lower Upper
 228 100
 226 28.8 22.2 33.2
 229 20.7 16.5 24.7









#33
Chrysene
Concen: 0.473 ng
RT: 21.358 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24




Instrument :
BNA_N
ClientSampleId :
PB166758BS


Tgt Ion:228 Resp: 8131
Ion Ratio Lower Upper
228 100
226 31.0 25.5 38.3
229 20.8 16.4 24.6



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.382 ng
RT: 21.241 min Scan# 2254
Delta R.T. 0.000 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24

Tgt Ion:149 Resp: 3779
Ion Ratio Lower Upper
149 100
167 26.6 21.2 31.8
279 3.3 2.7 4.1









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.589 min Scan# 21
Delta R.T. -0.006 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24




Instrument :
BNA_N
ClientSampleId :
PB166758BS


Tgt Ion:264 Resp: 4362
Ion Ratio Lower Upper
264 100
260 27.8 20.9 31.3
265 108.5 60.7 91.1#



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.482 ng
RT: 25.890 min Scan# 3576
Delta R.T. -0.009 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24

Tgt Ion:276 Resp: 7355
Ion Ratio Lower Upper
276 100
138 28.6 22.2 33.2
277 24.1 19.8 29.6









#37
Benzo(b)fluoranthene
Concen: 0.484 ng
RT: 22.902 min Scan# 21
Delta R.T. -0.009 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24




Instrument :
BNA_N
ClientSampleId :
PB166758BS


Tgt Ion:252 Resp: 6950
Ion Ratio Lower Upper
252 100
253 28.8 21.9 32.9
125 20.1 15.0 22.6



#38
Benzo(k)fluoranthene
Concen: 0.521 ng
RT: 22.949 min Scan# 2570
Delta R.T. -0.009 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24


Tgt Ion:252 Resp: 7711
Ion Ratio Lower Upper
252 100
253 29.0 22.2 33.4
125 20.3 15.0 22.4








#39
Benzo(a)pyrene
Concen: 0.519 ng
RT: 23.490 min Scan# 21
Delta R.T. -0.009 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24


Instrument :
BNA_N
ClientSampleId :
PB166758BS





Tgt Ion:252 Resp: 6505
Ion Ratio Lower Upper
252 100
253 31.1 24.4 36.6
125 25.0 18.2 27.2




#40
Dibenzo(a,h)anthracene
Concen: 0.443 ng
RT: 25.908 min Scan# 3582
Delta R.T. -0.009 min
Lab File: BN036480.D
Acq: 20 Feb 2025 14:24



Tgt Ion:278 Resp: 5331
Ion Ratio Lower Upper
278 100
139 23.1 18.5 27.7
279 32.0 24.8 37.2





#41

Benzo(g,h,i)perylene

Concen: 0.434 ng

RT: 26.589 min Scan# 3815

Delta R.T. -0.012 min

Lab File: BN036480.D


Acq: 20 Feb 2025 14:24

Instrument :

BNA_N

ClientSampleId :

PB166758BS





Tgt Ion:276 Resp: 5921

Ion Ratio Lower Upper

276 100

277 26.5 20.7 31.1

138 29.6 21.8 32.6



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
 Data File : BN036482.D
 Acq On : 20 Feb 2025 16:08
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC

Quant Time: Feb 21 00:12:25 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Feb 11 01:17:14 2025
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.746	152	2017	0.400	ng	0.00
7) Naphthalene-d8	10.541	136	4822	0.400	ng	# 0.00
13) Acenaphthene-d10	14.388	164	3262	0.400	ng	0.00
19) Phenanthrene-d10	17.124	188	6914	0.400	ng	#-0.01
29) Chrysene-d12	21.322	240	6182	0.400	ng	# 0.00
35) Perylene-d12	23.587	264	6375	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.334	112	1821	0.382	ng	-0.01
5) Phenol-d6	6.923	99	2152	0.385	ng	-0.01
8) Nitrobenzene-d5	8.897	82	1892	0.398	ng	-0.01
11) 2-Methylnaphthalene-d10	12.131	152	2847	0.384	ng	-0.01
14) 2,4,6-Tribromophenol	15.883	330	533	0.330	ng	0.00
15) 2-Fluorobiphenyl	13.009	172	5105	0.416	ng	-0.01
27) Fluoranthene-d10	19.164	212	7260	0.378	ng	0.00
31) Terphenyl-d14	19.768	244	4950	0.375	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.261	88	839	0.380	ng	Qvalue 97
3) n-Nitrosodimethylamine	3.579	42	1556	0.406	ng	96
6) bis(2-Chloroethyl)ether	7.168	93	2310	0.395	ng	99
9) Naphthalene	10.584	128	5486	0.394	ng	99
10) Hexachlorobutadiene	10.872	225	1352	0.399	ng	# 99
12) 2-Methylnaphthalene	12.207	142	3576	0.392	ng	98
16) Acenaphthylene	14.099	152	5339	0.371	ng	100
17) Acenaphthene	14.441	154	3586	0.373	ng	99
18) Fluorene	15.435	166	5179	0.378	ng	99
20) 4,6-Dinitro-2-methylph...	15.523	198	443	0.327	ng	93
21) 4-Bromophenyl-phenylether	16.329	248	1615	0.392	ng	# 84
22) Hexachlorobenzene	16.441	284	2053	0.403	ng	98
23) Atrazine	16.602	200	1254	0.364	ng	96
24) Pentachlorophenol	16.789	266	845	0.349	ng	95
25) Phenanthrene	17.173	178	7909	0.396	ng	100
26) Anthracene	17.260	178	6847	0.388	ng	100
28) Fluoranthene	19.192	202	9570	0.390	ng	99
30) Pyrene	19.559	202	9761	0.410	ng	99
32) Benzo(a)anthracene	21.304	228	7952	0.391	ng	99
33) Chrysene	21.358	228	9374	0.426	ng	99
34) Bis(2-ethylhexyl)phtha...	21.232	149	4422	0.349	ng	99
36) Indeno(1,2,3-cd)pyrene	25.885	276	8878	0.398	ng	99
37) Benzo(b)fluoranthene	22.902	252	8760	0.417	ng	98
38) Benzo(k)fluoranthene	22.949	252	8815	0.408	ng	99
39) Benzo(a)pyrene	23.490	252	7583	0.414	ng	96
40) Dibenzo(a,h)anthracene	25.905	278	6854	0.390	ng	99
41) Benzo(g,h,i)perylene	26.586	276	7966	0.400	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN022025\
Data File : BN036482.D
Acq On : 20 Feb 2025 16:08
Operator : RC/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Quant Time: Feb 21 00:12:25 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN021025.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Feb 11 01:17:14 2025
Response via : Initial Calibration








#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.746 min Scan# 64
Delta R.T. -0.007 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08


Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC





Tgt Ion:152 Resp: 2017
Ion Ratio Lower Upper
152 100
150 154.9 123.7 185.5
115 65.0 52.5 78.7




#2
1,4-Dioxane
Concen: 0.380 ng
RT: 3.261 min Scan# 25
Delta R.T. -0.007 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08



Tgt Ion: 88 Resp: 839
Ion Ratio Lower Upper
88 100
43 45.4 33.7 50.5
58 88.0 68.9 103.3









#3
 n-Nitrosodimethylamine
 Concen: 0.406 ng
 RT: 3.579 min Scan# 69
 Delta R.T. 0.000 min
 Lab File: BN036482.D
 Acq: 20 Feb 2025 16:08




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion: 42 Resp: 1556
 Ion Ratio Lower Upper
 42 100
 74 85.9 71.8 107.6
 44 9.6 7.8 11.6



#4
 2-Fluorophenol
 Concen: 0.382 ng
 RT: 5.334 min Scan# 312
 Delta R.T. -0.014 min
 Lab File: BN036482.D
 Acq: 20 Feb 2025 16:08

Tgt Ion: 112 Resp: 1821
 Ion Ratio Lower Upper
 112 100
 64 66.1 53.4 80.0
 63 40.9 30.3 45.5









#5
Phenol-d6
Concen: 0.385 ng
RT: 6.923 min Scan# 51
Delta R.T. -0.014 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion: 99 Resp: 2152
Ion Ratio Lower Upper
99 100
42 30.4 21.7 32.5
71 41.6 32.6 49.0



#6
bis(2-Chloroethyl)ether
Concen: 0.395 ng
RT: 7.168 min Scan# 566
Delta R.T. -0.014 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08

Tgt Ion: 93 Resp: 2310
Ion Ratio Lower Upper
93 100
63 82.3 66.3 99.5
95 31.2 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.541 min Scan# 91
Delta R.T. 0.000 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:	136	Resp:	4822
Ion	Ratio	Lower	Upper
136	100		
137	12.2	10.1	15.1
54	10.6	11.8	17.6#
68	7.0	7.2	10.8#



#8
Nitrobenzene-d5
Concen: 0.398 ng
RT: 8.897 min Scan# 779
Delta R.T. -0.011 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08

Tgt Ion:	82	Resp:	1892
Ion	Ratio	Lower	Upper
82	100		
128	37.4	31.9	47.9
54	64.4	53.1	79.7









#9
Naphthalene
Concen: 0.394 ng
RT: 10.584 min Scan# 91
Delta R.T. -0.011 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:128	Resp:	5486
Ion Ratio	Lower	Upper
128	100	
129	12.7	9.6 14.4
127	14.9	12.0 18.0



#10
Hexachlorobutadiene
Concen: 0.399 ng
RT: 10.872 min Scan# 964
Delta R.T. -0.011 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08

Tgt Ion:225	Resp:	1352
Ion Ratio	Lower	Upper
225	100	
223	0.0	0.0 0.0
227	64.6	50.9 76.3









#11
2-Methylnaphthalene-d10
Concen: 0.384 ng
RT: 12.131 min Scan# 1146
Delta R.T. -0.010 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:152 Resp: 2847
Ion Ratio Lower Upper
152 100
151 21.8 16.6 25.0



#12
2-Methylnaphthalene
Concen: 0.392 ng
RT: 12.207 min Scan# 1161
Delta R.T. -0.005 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08

Tgt Ion:142 Resp: 3576
Ion Ratio Lower Upper
142 100
141 88.9 72.8 109.2
115 44.6 35.5 53.3









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.388 min Scan# 1448
Delta R.T. 0.000 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:164 Resp: 3262
Ion Ratio Lower Upper
164 100
162 102.1 84.1 126.1
160 50.9 41.4 62.0



#14
2,4,6-Tribromophenol
Concen: 0.330 ng
RT: 15.883 min Scan# 1583
Delta R.T. 0.000 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08

Tgt Ion:330 Resp: 533
Ion Ratio Lower Upper
330 100
332 96.6 76.6 114.8
141 49.9 37.8 56.8









#15
2-Fluorobiphenyl
Concen: 0.416 ng
RT: 13.009 min Scan# 11
Delta R.T. -0.011 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:172	Resp:	5105
Ion Ratio	Lower	Upper
172	100	
171	36.7	29.6 44.4
170	24.4	19.8 29.6



#16
Acenaphthylene
Concen: 0.371 ng
RT: 14.099 min Scan# 1421
Delta R.T. -0.011 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08

Tgt Ion:152	Resp:	5339
Ion Ratio	Lower	Upper
152	100	
151	19.6	15.8 23.8
153	12.6	10.2 15.2









#17
 Acenaphthene
 Concen: 0.373 ng
 RT: 14.441 min Scan# 14
 Delta R.T. -0.011 min
 Lab File: BN036482.D
 Acq: 20 Feb 2025 16:08




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:	154	Resp:	3586
Ion Ratio	Lower	Upper	
154	100		
153	118.5	93.3	139.9
152	61.0	48.8	73.2



#18
 Fluorene
 Concen: 0.378 ng
 RT: 15.435 min Scan# 1546
 Delta R.T. -0.011 min
 Lab File: BN036482.D
 Acq: 20 Feb 2025 16:08

Tgt Ion:	166	Resp:	5179
Ion Ratio	Lower	Upper	
166	100		
165	99.0	79.5	119.3
167	13.8	10.4	15.6









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.124 min Scan# 1683
Delta R.T. -0.012 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:188 Resp: 6914
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 15.8 9.8 14.6#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.327 ng
RT: 15.523 min Scan# 1554
Delta R.T. -0.000 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08

Tgt Ion:198 Resp: 443
Ion Ratio Lower Upper
198 100
51 105.2 86.6 129.8
105 61.0 57.5 86.3









#21
4-Bromophenyl-phenylether
Concen: 0.392 ng
RT: 16.329 min Scan# 1619
Delta R.T. 0.000 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:248 Resp: 1615
Ion Ratio Lower Upper
248 100
250 97.5 76.1 114.1
141 60.3 71.7 107.5#



#22
Hexachlorobenzene
Concen: 0.403 ng
RT: 16.441 min Scan# 1628
Delta R.T. -0.012 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08

Tgt Ion:284 Resp: 2053
Ion Ratio Lower Upper
284 100
142 43.4 33.4 50.0
249 34.5 28.6 43.0









#23
Atrazine
Concen: 0.364 ng
RT: 16.602 min Scan# 1641
Delta R.T. -0.012 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:200 Resp: 1254
Ion Ratio Lower Upper
200 100
173 32.5 23.2 34.8
215 48.5 40.0 60.0



#24
Pentachlorophenol
Concen: 0.349 ng
RT: 16.789 min Scan# 1656
Delta R.T. -0.012 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08

Tgt Ion:266 Resp: 845
Ion Ratio Lower Upper
266 100
264 59.2 50.6 76.0
268 60.7 51.9 77.9









#25
Phenanthrene
Concen: 0.396 ng
RT: 17.173 min Scan# 1687
Delta R.T. 0.000 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:178 Resp: 7909
Ion Ratio Lower Upper
178 100
176 19.6 15.7 23.5
179 15.5 12.4 18.6



#26
Anthracene
Concen: 0.388 ng
RT: 17.260 min Scan# 1694
Delta R.T. -0.012 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08

Tgt Ion:178 Resp: 6847
Ion Ratio Lower Upper
178 100
176 19.0 14.9 22.3
179 15.5 12.4 18.6









#27
Fluoranthene-d10
Concen: 0.378 ng
RT: 19.164 min Scan# 1935
Delta R.T. -0.005 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:212 Resp: 7260
Ion Ratio Lower Upper
212 100
106 14.5 11.5 17.3
104 8.8 7.1 10.7



#28
Fluoranthene
Concen: 0.390 ng
RT: 19.192 min Scan# 1941
Delta R.T. -0.009 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08

Tgt Ion:202 Resp: 9570
Ion Ratio Lower Upper
202 100
101 11.3 9.2 13.8
203 17.0 13.4 20.0









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.322 min Scan# 21
Delta R.T. 0.000 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:240 Resp: 6182
Ion Ratio Lower Upper
240 100
120 13.1 13.3 19.9#
236 29.0 23.0 34.6



#30
Pyrene
Concen: 0.410 ng
RT: 19.559 min Scan# 2020
Delta R.T. -0.005 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08

Tgt Ion:202 Resp: 9761
Ion Ratio Lower Upper
202 100
200 21.1 16.9 25.3
203 18.0 13.9 20.9









#31
 Terphenyl-d14
 Concen: 0.375 ng
 RT: 19.768 min Scan# 2065
 Delta R.T. 0.000 min
 Lab File: BN036482.D
 Acq: 20 Feb 2025 16:08




Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC


Tgt Ion:244 Resp: 4950
 Ion Ratio Lower Upper
 244 100
 212 10.7 8.1 12.1
 122 15.5 12.8 19.2



#32
 Benzo(a)anthracene
 Concen: 0.391 ng
 RT: 21.304 min Scan# 2261
 Delta R.T. -0.009 min
 Lab File: BN036482.D
 Acq: 20 Feb 2025 16:08

Tgt Ion:228 Resp: 7952
 Ion Ratio Lower Upper
 228 100
 226 28.3 22.2 33.2
 229 19.9 16.5 24.7









#33
Chrysene
Concen: 0.426 ng
RT: 21.358 min Scan# 2135
Delta R.T. 0.000 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:228 Resp: 9374
Ion Ratio Lower Upper
228 100
226 31.0 25.5 38.3
229 20.2 16.4 24.6



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.349 ng
RT: 21.232 min Scan# 2253
Delta R.T. -0.009 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08

Tgt Ion:149 Resp: 4422
Ion Ratio Lower Upper
149 100
167 26.1 21.2 31.8
279 4.0 2.7 4.1









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.587 min Scan# 21
Delta R.T. -0.009 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08




Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC


Tgt Ion:264 Resp: 6375
Ion Ratio Lower Upper
264 100
260 26.5 20.9 31.3
265 78.5 60.7 91.1



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.398 ng
RT: 25.885 min Scan# 3574
Delta R.T. -0.015 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08


Tgt Ion:276 Resp: 8878
Ion Ratio Lower Upper
276 100
138 28.2 22.2 33.2
277 24.6 19.8 29.6







#37
Benzo(b)fluoranthene
Concen: 0.417 ng
RT: 22.902 min Scan# 21
Delta R.T. -0.009 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08

Instrument :
BNA_N
ClientSampleId :
SSTDCCC0.4EC







Tgt Ion:252 Resp: 8760
Ion Ratio Lower Upper
252 100
253 27.4 21.9 32.9
125 20.9 15.0 22.6



#38
Benzo(k)fluoranthene
Concen: 0.408 ng
RT: 22.949 min Scan# 2570
Delta R.T. -0.009 min
Lab File: BN036482.D
Acq: 20 Feb 2025 16:08

Tgt Ion:252 Resp: 8815
Ion Ratio Lower Upper
252 100
253 27.3 22.2 33.4
125 19.1 15.0 22.4





#39

Benzo(a)pyrene

Concen: 0.414 ng

RT: 23.490 min Scan# 21

Delta R.T. -0.009 min

Lab File: BN036482.D

Acq: 20 Feb 2025 16:08

Instrument :

BNA_N

ClientSampleId :

SSTDCCC0.4EC





Tgt Ion:252 Resp: 7583

Ion Ratio Lower Upper

252 100

253 29.9 24.4 36.6

125 27.1 18.2 27.2



#40

Dibenzo(a,h)anthracene

Concen: 0.390 ng

RT: 25.905 min Scan# 3581

Delta R.T. -0.012 min

Lab File: BN036482.D

Acq: 20 Feb 2025 16:08




Tgt Ion:278 Resp: 6854


Ion Ratio Lower Upper

278 100

139 22.5 18.5 27.7


279 30.8 24.8 37.2





#41
 Benzo(g,h,i)perylene
 Concen: 0.400 ng
 RT: 26.586 min Scan# 3814
 Delta R.T. -0.015 min
 Lab File: BN036482.D
 Acq: 20 Feb 2025 16:08

Instrument :
 BNA_N
 ClientSampleId :
 SSTDCCC0.4EC



Tgt Ion:	276	Resp:	7966
Ion	Ratio	Lower	Upper
276	100		
277	26.2	20.7	31.1
138	29.0	21.8	32.6

