

LAB CHRONICLE

OrderID:	Q1380	OrderDate:	2/17/2025 3:59:00 PM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage CTO WE13 - VPB-192
Contact:	Ernie Wu	Location:	#12 VOC Ref. #2 Soil, VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1380-01	BP-VPB-192-EB-2025 0212	Water			02/12/25			02/17/25
			VOCMS Group1	8260-Low			02/19/25	
Q1380-02	BP-VPB-192-TB-2025 0210	Water			02/10/25			02/17/25
			VOCMS Group1	8260-Low			02/19/25	
Q1380-06	BP-VPB-192-GW-725- 727	SOIL			02/10/25			02/17/25
			VOCMS Group1	8260D			02/19/25	
Q1380-09	BP-VPB-192-GW-780- 782	Water			02/13/25			02/17/25
			VOCMS Group1	8260-Low			02/19/25	
Q1380-12	VPB192-HYD-202502 14	Water			02/14/25			02/17/25
			VOCMS Group1	8260-Low			02/19/25	

Hit Summary Sheet
SW-846

SDG No.: Q1380
Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID:	BP-VPB-192-EB-20250212								
Q1380-01	BP-VPB-192-EB-20 Water	Acetone		4.70	J	1.40	3.80	5.00	ug/L
		Total Voc :		4.70					
		Total Concentration:		4.70					
Client ID:	BP-VPB-192-TB-20250210								
Q1380-02	BP-VPB-192-TB-20 Water	Acetone		1.40	J	1.40	3.80	5.00	ug/L
		Total Voc :		1.40					
		Total Concentration:		1.40					
Client ID:	BP-VPB-192-GW-725-727								
Q1380-06	BP-VPB-192-GW-7 SOIL	Acetone		130	J	83.5	270	330	ug/Kg
		Total Voc :		130					
		Total Concentration:		130					
Client ID:	BP-VPB-192-GW-780-782								
Q1380-09	BP-VPB-192-GW-7 Water	Acetone		9.50		1.40	3.80	5.00	ug/L
Q1380-09	BP-VPB-192-GW-7 Water	2-Butanone		1.90	J	1.30	2.50	5.00	ug/L
		Total Voc :		11.4					
		Total Concentration:		11.4					
Client ID:	VPB192-HYD-20250214								
Q1380-12	VPB192-HYD-2025 Water	Acetone		1.50	J	1.40	3.80	5.00	ug/L
Q1380-12	VPB192-HYD-2025 Water	Chloroform		0.49	J	0.26	0.50	1.00	ug/L
Q1380-12	VPB192-HYD-2025 Water	Bromodichloromethane		1.50		0.24	0.50	1.00	ug/L
Q1380-12	VPB192-HYD-2025 Water	4-Methyl-2-Pentanone		1.90	J	0.75	2.50	5.00	ug/L
Q1380-12	VPB192-HYD-2025 Water	Dibromochloromethane		2.70		0.18	0.50	1.00	ug/L
Q1380-12	VPB192-HYD-2025 Water	Bromoform		1.60		0.21	0.50	1.00	ug/L
		Total Voc :		9.69					
		Total Concentration:		9.69					



SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	02/12/25	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	02/17/25	
Client Sample ID:	BP-VPB-192-EB-20250212		SDG No.:	Q1380	
Lab Sample ID:	Q1380-01		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085788.D	1		02/19/25 14:21	VN021925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	4.70	J	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	02/12/25	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	02/17/25	
Client Sample ID:	BP-VPB-192-EB-20250212		SDG No.:	Q1380	
Lab Sample ID:	Q1380-01		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085788.D	1		02/19/25 14:21	VN021925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	55.6		81 - 118		111%	SPK: 50
1868-53-7	Dibromofluoromethane	52.0		80 - 119		104%	SPK: 50
2037-26-5	Toluene-d8	51.0		89 - 112		102%	SPK: 50
460-00-4	4-Bromofluorobenzene	49.5		85 - 114		99%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	292000	8.218				
540-36-3	1,4-Difluorobenzene	538000	9.1				
3114-55-4	Chlorobenzene-d5	479000	11.859				
3855-82-1	1,4-Dichlorobenzene-d4	200000	13.788				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	02/12/25	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	02/17/25	
Client Sample ID:	BP-VPB-192-EB-20250212		SDG No.:	Q1380	
Lab Sample ID:	Q1380-01		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085788.D	1		02/19/25 14:21	VN021925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	02/10/25	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	02/17/25	
Client Sample ID:	BP-VPB-192-TB-20250210		SDG No.:	Q1380	
Lab Sample ID:	Q1380-02		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085789.D	1		02/19/25 14:45	VN021925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	1.40	J	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	02/10/25	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	02/17/25	
Client Sample ID:	BP-VPB-192-TB-20250210		SDG No.:	Q1380	
Lab Sample ID:	Q1380-02		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085789.D	1		02/19/25 14:45	VN021925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	57.5		81 - 118		115%	SPK: 50
1868-53-7	Dibromofluoromethane	52.9		80 - 119		106%	SPK: 50
2037-26-5	Toluene-d8	50.7		89 - 112		101%	SPK: 50
460-00-4	4-Bromofluorobenzene	47.8		85 - 114		96%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	265000	8.224				
540-36-3	1,4-Difluorobenzene	488000	9.1				
3114-55-4	Chlorobenzene-d5	434000	11.859				
3855-82-1	1,4-Dichlorobenzene-d4	175000	13.788				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	02/10/25	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	02/17/25	
Client Sample ID:	BP-VPB-192-TB-20250210		SDG No.:	Q1380	
Lab Sample ID:	Q1380-02		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085789.D	1		02/19/25 14:45	VN021925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	02/10/25	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	02/17/25	
Client Sample ID:	BP-VPB-192-GW-725-727		SDG No.:	Q1380	
Lab Sample ID:	Q1380-06		Matrix:	SOIL	
Analytical Method:	SW8260		% Solid:	7.4	
Sample Wt/Vol:	5.05	Units: g	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY021241.D	1		02/19/25 16:59	VY021925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units(Dry Weight)
TARGETS							
74-87-3	Chloromethane	33.4	UQ	15.5	33.4	66.9	ug/Kg
75-01-4	Vinyl Chloride	33.4	UQ	10.3	33.4	66.9	ug/Kg
74-83-9	Bromomethane	53.5	U	13.8	53.5	66.9	ug/Kg
75-00-3	Chloroethane	33.4	UQ	13.5	33.4	66.9	ug/Kg
75-69-4	Trichlorofluoromethane	53.5	U	12.2	53.5	66.9	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	33.4	U	14.3	33.4	66.9	ug/Kg
75-35-4	1,1-Dichloroethene	33.4	U	10.4	33.4	66.9	ug/Kg
67-64-1	Acetone	130	J	83.5	270	330	ug/Kg
75-15-0	Carbon Disulfide	53.5	U	17.1	53.5	66.9	ug/Kg
1634-04-4	Methyl tert-butyl Ether	33.4	UM	9.00	33.4	66.9	ug/Kg
75-09-2	Methylene Chloride	110	U	45.6	110	130	ug/Kg
156-60-5	trans-1,2-Dichloroethene	33.4	U	11.2	33.4	66.9	ug/Kg
75-34-3	1,1-Dichloroethane	33.4	U	8.40	33.4	66.9	ug/Kg
78-93-3	2-Butanone	270	U	76.0	270	330	ug/Kg
56-23-5	Carbon Tetrachloride	33.4	U	11.6	33.4	66.9	ug/Kg
156-59-2	cis-1,2-Dichloroethene	33.4	U	8.20	33.4	66.9	ug/Kg
67-66-3	Chloroform	53.5	U	9.00	53.5	66.9	ug/Kg
71-55-6	1,1,1-Trichloroethane	33.4	U	10.4	33.4	66.9	ug/Kg
108-87-2	Methylcyclohexane	33.4	UM	11.6	33.4	66.9	ug/Kg
71-43-2	Benzene	33.4	U	9.60	33.4	66.9	ug/Kg
107-06-2	1,2-Dichloroethane	33.4	U	8.20	33.4	66.9	ug/Kg
79-01-6	Trichloroethene	33.4	UM	10.0	33.4	66.9	ug/Kg
78-87-5	1,2-Dichloropropane	33.4	U	8.80	33.4	66.9	ug/Kg
75-27-4	Bromodichloromethane	33.4	U	7.50	33.4	66.9	ug/Kg
108-10-1	4-Methyl-2-Pentanone	170	U	58.2	170	330	ug/Kg
108-88-3	Toluene	33.4	U	9.00	33.4	66.9	ug/Kg
10061-02-6	t-1,3-Dichloropropene	33.4	U	8.00	33.4	66.9	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	33.4	U	7.60	33.4	66.9	ug/Kg
79-00-5	1,1,2-Trichloroethane	33.4	U	11.2	33.4	66.9	ug/Kg
591-78-6	2-Hexanone	170	U	64.1	170	330	ug/Kg

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	02/10/25	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	02/17/25	
Client Sample ID:	BP-VPB-192-GW-725-727		SDG No.:	Q1380	
Lab Sample ID:	Q1380-06		Matrix:	SOIL	
Analytical Method:	SW8260		% Solid:	7.4	
Sample Wt/Vol:	5.05	Units: g	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY021241.D	1		02/19/25 16:59	VY021925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units(Dry Weight)
124-48-1	Dibromochloromethane	33.4	U	8.70	33.4	66.9	ug/Kg
127-18-4	Tetrachloroethene	33.4	UM	11.9	33.4	66.9	ug/Kg
108-90-7	Chlorobenzene	33.4	UM	9.90	33.4	66.9	ug/Kg
100-41-4	Ethyl Benzene	33.4	UM	8.30	33.4	66.9	ug/Kg
179601-23-1	m/p-Xylenes	66.9	UM	18.1	66.9	130	ug/Kg
95-47-6	o-Xylene	33.4	UM	9.40	33.4	66.9	ug/Kg
100-42-5	Styrene	33.4	UM	8.00	33.4	66.9	ug/Kg
75-25-2	Bromoform	33.4	U	10.8	33.4	66.9	ug/Kg
98-82-8	Isopropylbenzene	33.4	U	9.00	33.4	66.9	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	33.4	UM	14.7	33.4	66.9	ug/Kg
541-73-1	1,3-Dichlorobenzene	33.4	UM	9.90	33.4	66.9	ug/Kg
106-46-7	1,4-Dichlorobenzene	33.4	UM	10.7	33.4	66.9	ug/Kg
95-50-1	1,2-Dichlorobenzene	33.4	UM	7.90	33.4	66.9	ug/Kg
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	66.6		71 - 136		133%	SPK: 50
1868-53-7	Dibromofluoromethane	53.3		78 - 119		107%	SPK: 50
2037-26-5	Toluene-d8	50.3		85 - 116		101%	SPK: 50
460-00-4	4-Bromofluorobenzene	53.7		79 - 119		107%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	224000	7.713				
540-36-3	1,4-Difluorobenzene	425000	8.622				
3114-55-4	Chlorobenzene-d5	405000	11.42				
3855-82-1	1,4-Dichlorobenzene-d4	161000	13.353				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	02/10/25	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	02/17/25	
Client Sample ID:	BP-VPB-192-GW-725-727		SDG No.:	Q1380	
Lab Sample ID:	Q1380-06		Matrix:	SOIL	
Analytical Method:	SW8260		% Solid:	7.4	
Sample Wt/Vol:	5.05	Units: g	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY021241.D	1		02/19/25 16:59	VY021925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units(Dry Weight)
------------	-----------	-------	-----------	-----	-----	------------	-------------------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	02/13/25	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	02/17/25	
Client Sample ID:	BP-VPB-192-GW-780-782		SDG No.:	Q1380	
Lab Sample ID:	Q1380-09		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085790.D	1		02/19/25 15:08	VN021925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	9.50		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	1.90	J	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	02/13/25	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	02/17/25	
Client Sample ID:	BP-VPB-192-GW-780-782		SDG No.:	Q1380	
Lab Sample ID:	Q1380-09		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085790.D	1		02/19/25 15:08	VN021925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	57.5		81 - 118		115%	SPK: 50
1868-53-7	Dibromofluoromethane	53.5		80 - 119		107%	SPK: 50
2037-26-5	Toluene-d8	50.6		89 - 112		101%	SPK: 50
460-00-4	4-Bromofluorobenzene	45.8		85 - 114		92%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	260000	8.218				
540-36-3	1,4-Difluorobenzene	471000	9.1				
3114-55-4	Chlorobenzene-d5	404000	11.865				
3855-82-1	1,4-Dichlorobenzene-d4	157000	13.788				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	02/13/25	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	02/17/25	
Client Sample ID:	BP-VPB-192-GW-780-782		SDG No.:	Q1380	
Lab Sample ID:	Q1380-09		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085790.D	1		02/19/25 15:08	VN021925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected
 LOQ = Limit of Quantitation
 MDL = Method Detection Limit
 LOD = Limit of Detection
 E = Value Exceeds Calibration Range
 Q = indicates LCS control criteria did not meet requirements
 M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value
 B = Analyte Found in Associated Method Blank
 N = Presumptive Evidence of a Compound
 * = Values outside of QC limits
 D = Dilution
 () = Laboratory InHouse Limit
 A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	02/14/25	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	02/17/25	
Client Sample ID:	VPB192-HYD-20250214		SDG No.:	Q1380	
Lab Sample ID:	Q1380-12		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085787.D	1		02/19/25 13:57	VN021925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	1.50	J	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
67-66-3	Chloroform	0.49	J	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	1.50		0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	1.90	J	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	02/14/25	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	02/17/25	
Client Sample ID:	VPB192-HYD-20250214		SDG No.:	Q1380	
Lab Sample ID:	Q1380-12		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085787.D	1		02/19/25 13:57	VN021925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	2.70		0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	1.60		0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	55.0		81 - 118		110%	SPK: 50
1868-53-7	Dibromofluoromethane	52.8		80 - 119		106%	SPK: 50
2037-26-5	Toluene-d8	50.4		89 - 112		101%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.2		85 - 114		96%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	298000	8.218				
540-36-3	1,4-Difluorobenzene	540000	9.094				
3114-55-4	Chlorobenzene-d5	473000	11.859				
3855-82-1	1,4-Dichlorobenzene-d4	194000	13.788				
TENTATIVE IDENTIFIED COMPOUNDS							
75-43-4	Dichlorofluoromethane	N.D					

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	02/14/25	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	02/17/25	
Client Sample ID:	VPB192-HYD-20250214		SDG No.:	Q1380	
Lab Sample ID:	Q1380-12		Matrix:	Water	
Analytical Method:	SW8260		% Solid:	0	
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085787.D	1		02/19/25 13:57	VN021925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
------------	-----------	-------	-----------	-----	-----	------------	-------

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SDG No.: **Q1380**

Client: **Tetra Tech NUS, Inc.**

Analytical Method: **SW8260D**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1380-06	BP-VPB-192-GW-725-727	1,2-Dichloroethane-d4	50	66.6	133	71	136
		Dibromofluoromethane	50	53.3	107	78	119
		Toluene-d8	50	50.3	101	85	116
		4-Bromofluorobenzene	50	53.7	107	79	119
Q1380-07MS	BP-VPB-192-GW-725-727MS	1,2-Dichloroethane-d4	50	40.9	82	71	136
		Dibromofluoromethane	50	39.9	80	78	119
		Toluene-d8	50	36.4	73 *	85	116
		4-Bromofluorobenzene	50	31.4	63 *	79	119
Q1380-08MSD	BP-VPB-192-GW-725-727MSD	1,2-Dichloroethane-d4	50	56.3	113	71	136
		Dibromofluoromethane	50	55.0	110	78	119
		Toluene-d8	50	50.0	100	85	116
		4-Bromofluorobenzene	50	42.4	85	79	119
VY0219SBL01	VY0219SBL01	1,2-Dichloroethane-d4	50	54.2	108	71	136
		Dibromofluoromethane	50	50.2	100	78	119
		Toluene-d8	50	49.6	99	85	116
		4-Bromofluorobenzene	50	47.8	96	79	119
VY0219SBS01	VY0219SBS01	1,2-Dichloroethane-d4	50	50.3	101	71	136
		Dibromofluoromethane	50	48.6	97	78	119
		Toluene-d8	50	47.3	95	85	116
		4-Bromofluorobenzene	50	50.7	101	79	119
VY0226SBL01	VY0226SBL01	1,2-Dichloroethane-d4	50	53.1	106	71	136
		Dibromofluoromethane	50	49.5	99	78	119
		Toluene-d8	50	49.1	98	85	116
		4-Bromofluorobenzene	50	50.1	100	79	119
VY0226SBS01	VY0226SBS01	1,2-Dichloroethane-d4	50	46.7	93	71	136
		Dibromofluoromethane	50	47.4	95	78	119
		Toluene-d8	50	47.6	95	85	116
		4-Bromofluorobenzene	50	47.3	95	79	119

Surrogate Summary

SDG No.: **Q1380**

Client: **Tetra Tech NUS, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
Q1380-01	BP-VPB-192-EB-20250212	1,2-Dichloroethane-d4	50	55.6	111	81	118
		Dibromofluoromethane	50	52.0	104	80	119
		Toluene-d8	50	51.0	102	89	112
		4-Bromofluorobenzene	50	49.5	99	85	114
Q1380-02	BP-VPB-192-TB-20250210	1,2-Dichloroethane-d4	50	57.5	115	81	118
		Dibromofluoromethane	50	52.9	106	80	119
		Toluene-d8	50	50.7	101	89	112
		4-Bromofluorobenzene	50	47.8	96	85	114
Q1380-09	BP-VPB-192-GW-780-782	1,2-Dichloroethane-d4	50	57.5	115	81	118
		Dibromofluoromethane	50	53.5	107	80	119
		Toluene-d8	50	50.5	101	89	112
		4-Bromofluorobenzene	50	45.8	92	85	114
Q1380-12	VPB192-HYD-20250214	1,2-Dichloroethane-d4	50	55.0	110	81	118
		Dibromofluoromethane	50	52.8	106	80	119
		Toluene-d8	50	50.4	101	89	112
		4-Bromofluorobenzene	50	48.2	96	85	114
VN0219WBL01	VN0219WBL01	1,2-Dichloroethane-d4	50	56.1	112	81	118
		Dibromofluoromethane	50	54.6	109	80	119
		Toluene-d8	50	51.9	104	89	112
		4-Bromofluorobenzene	50	48.1	96	85	114
VN0219WBS01	VN0219WBS01	1,2-Dichloroethane-d4	50	52.7	105	81	118
		Dibromofluoromethane	50	52.5	105	80	119
		Toluene-d8	50	54.8	110	89	112
		4-Bromofluorobenzene	50	56.0	112	85	114
VN0219WBSD0	VN0219WBSD01	1,2-Dichloroethane-d4	50	55.5	111	81	118
		Dibromofluoromethane	50	54.7	109	80	119
		Toluene-d8	50	55.1	110	89	112
		4-Bromofluorobenzene	50	56.6	113	85	114

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1380

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260D

Parameter	Spike	Sample	Result	Units	Rec		RPD	RPD	Limits		RPD
		Result			Qual	Qual		Low	High		
Lab Sample ID :	Q1380-07MS	Client Sample ID :	BP-VPB-192-GW-725-727MS					Datafile :	VY021342.D		
Chloromethane	670	0	520	ug/Kg	78				50	136	
Vinyl chloride	670	0	550	ug/Kg	82				56	135	
Bromomethane	670	0	550	ug/Kg	82				53	143	
Chloroethane	670	0	580	ug/Kg	87				59	139	
Trichlorofluoromethane	670	0	570	ug/Kg	85				62	140	
1,1,2-Trichlorotrifluoroethane	670	0	530	ug/Kg	79				66	136	
1,1-Dichloroethene	670	0	560	ug/Kg	84				70	131	
Acetone	3300	130	2600	ug/Kg	75				36	164	
Carbon disulfide	670	0	570	ug/Kg	85				63	132	
Methyl tert-butyl Ether	670	0	570	ug/Kg	85				73	125	
Methylene Chloride	670	0	590	ug/Kg	88				70	128	
trans-1,2-Dichloroethene	670	0	560	ug/Kg	84				74	125	
1,1-Dichloroethane	670	0	580	ug/Kg	87				76	125	
2-Butanone	3300	0	2500	ug/Kg	76				51	148	
Carbon Tetrachloride	670	0	520	ug/Kg	78				70	135	
cis-1,2-Dichloroethene	670	0	580	ug/Kg	87				77	123	
Chloroform	670	0	580	ug/Kg	87				78	123	
1,1,1-Trichloroethane	670	0	570	ug/Kg	85				73	130	
Methylcyclohexane	670	0	310	ug/Kg	46	*			66	133	
Benzene	670	0	540	ug/Kg	81				77	121	
1,2-Dichloroethane	670	0	550	ug/Kg	82				73	128	
Trichloroethene	670	0	510	ug/Kg	76	*			77	123	
1,2-Dichloropropane	670	0	540	ug/Kg	81				76	123	
Bromodichloromethane	670	0	560	ug/Kg	84				75	127	
4-Methyl-2-Pentanone	3300	0	2400	ug/Kg	73				65	135	
Toluene	670	0	520	ug/Kg	78				77	121	
t-1,3-Dichloropropene	670	0	520	ug/Kg	78				71	130	
cis-1,3-Dichloropropene	670	0	520	ug/Kg	78				74	126	
1,1,2-Trichloroethane	670	0	540	ug/Kg	81				78	121	
2-Hexanone	3300	0	2400	ug/Kg	73				53	145	
Dibromochloromethane	670	0	540	ug/Kg	81				74	126	
Tetrachloroethene	670	0	480	ug/Kg	72	*			73	128	
Chlorobenzene	670	0	510	ug/Kg	76	*			79	120	
Ethyl Benzene	670	0	500	ug/Kg	75	*			76	122	
m/p-Xylenes	1300	0	960	ug/Kg	74	*			77	124	
o-Xylene	670	0	500	ug/Kg	75	*			77	123	
Styrene	670	0	340	ug/Kg	51	*			76	124	
Bromoform	670	0	560	ug/Kg	84				67	132	
Isopropylbenzene	670	0	590	ug/Kg	88				68	134	
1,1,2,2-Tetrachloroethane	670	0	750	ug/Kg	112				70	124	
1,3-Dichlorobenzene	670	0	450	ug/Kg	67	*			77	121	
1,4-Dichlorobenzene	670	0	460	ug/Kg	69	*			75	120	
1,2-Dichlorobenzene	670	0	480	ug/Kg	72	*			78	121	

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1380

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260D

Limits											
Parameter	Spike	Sample	Result	Units	Rec		RPD		Low	High	RPD
		Result			Qual	RPD	Qual				
Lab Sample ID :	Q1380-08MSD	Client Sample ID :	BP-VPB-192-GW-725-727MSD					Datafile :	VY021343.D		
Chloromethane	670	0	750	ug/Kg	112		36	*	50	136	20
Vinyl chloride	670	0	800	ug/Kg	119		37	*	56	135	20
Bromomethane	670	0	770	ug/Kg	115		33	*	53	143	20
Chloroethane	670	0	790	ug/Kg	118		31	*	59	139	20
Trichlorofluoromethane	670	0	810	ug/Kg	121		35	*	62	140	20
1,1,2-Trichlorotrifluoroethane	670	0	770	ug/Kg	115		37	*	66	136	20
1,1-Dichloroethene	670	0	800	ug/Kg	119		35	*	70	131	20
Acetone	3400	130	3900	ug/Kg	111		39	*	36	164	20
Carbon disulfide	670	0	790	ug/Kg	118		32	*	63	132	20
Methyl tert-butyl Ether	670	0	850	ug/Kg	127	*	39	*	73	125	20
Methylene Chloride	670	0	820	ug/Kg	122		33	*	70	128	20
trans-1,2-Dichloroethene	670	0	790	ug/Kg	118		34	*	74	125	20
1,1-Dichloroethane	670	0	800	ug/Kg	119		32	*	76	125	20
2-Butanone	3400	0	3900	ug/Kg	115		41	*	51	148	20
Carbon Tetrachloride	670	0	740	ug/Kg	110		35	*	70	135	20
cis-1,2-Dichloroethene	670	0	810	ug/Kg	121		33	*	77	123	20
Chloroform	670	0	810	ug/Kg	121		33	*	78	123	20
1,1,1-Trichloroethane	670	0	800	ug/Kg	119		34	*	73	130	20
Methylcyclohexane	670	0	480	ug/Kg	72		43	*	66	133	20
Benzene	670	0	770	ug/Kg	115		35	*	77	121	20
1,2-Dichloroethane	670	0	810	ug/Kg	121		38	*	73	128	20
Trichloroethene	670	0	730	ug/Kg	109		35	*	77	123	20
1,2-Dichloropropane	670	0	770	ug/Kg	115		35	*	76	123	20
Bromodichloromethane	670	0	790	ug/Kg	118		34	*	75	127	20
4-Methyl-2-Pentanone	3400	0	3900	ug/Kg	115		45	*	65	135	20
Toluene	670	0	740	ug/Kg	110		35	*	77	121	20
t-1,3-Dichloropropene	670	0	750	ug/Kg	112		36	*	71	130	20
cis-1,3-Dichloropropene	670	0	740	ug/Kg	110		35	*	74	126	20
1,1,2-Trichloroethane	670	0	810	ug/Kg	121		40	*	78	121	20
2-Hexanone	3400	0	3900	ug/Kg	115		45	*	53	145	20
Dibromochloromethane	670	0	790	ug/Kg	118		38	*	74	126	20
Tetrachloroethene	670	0	680	ug/Kg	101		34	*	73	128	20
Chlorobenzene	670	0	720	ug/Kg	107		34	*	79	120	20
Ethyl Benzene	670	0	710	ug/Kg	106		35	*	76	122	20
m/p-Xylenes	1300	0	1400	ug/Kg	108		37	*	77	124	20
o-Xylene	670	0	710	ug/Kg	106		35	*	77	123	20
Styrene	670	0	460	ug/Kg	69	*	30	*	76	124	20
Bromoform	670	0	820	ug/Kg	122		38	*	67	132	20
Isopropylbenzene	670	0	860	ug/Kg	128		37	*	68	134	20
1,1,2,2-Tetrachloroethane	670	0	1100	ug/Kg	164	*	38	*	70	124	20
1,3-Dichlorobenzene	670	0	660	ug/Kg	99		38	*	77	121	20
1,4-Dichlorobenzene	670	0	690	ug/Kg	103		40	*	75	120	20
1,2-Dichlorobenzene	670	0	720	ug/Kg	107		40	*	78	121	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1380

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Datafile : VN085785.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0219WBS01	Chloromethane	20	17.2	ug/L	86			50	139	
	Vinyl chloride	20	17.2	ug/L	86			58	137	
	Bromomethane	20	17.2	ug/L	86			53	141	
	Chloroethane	20	17.0	ug/L	85			60	138	
	Trichlorofluoromethane	20	17.1	ug/L	86			65	141	
	1,1,2-Trichlorotrifluoroethane	20	17.7	ug/L	89			70	136	
	1,1-Dichloroethene	20	17.4	ug/L	87			71	131	
	Acetone	100	93.8	ug/L	94			39	160	
	Carbon disulfide	20	16.8	ug/L	84			64	133	
	Methyl tert-butyl Ether	20	17.9	ug/L	90			71	124	
	Methylene Chloride	20	16.8	ug/L	84			74	124	
	trans-1,2-Dichloroethene	20	17.0	ug/L	85			75	124	
	1,1-Dichloroethane	20	17.7	ug/L	89			77	125	
	2-Butanone	100	89.6	ug/L	90			56	143	
	Carbon Tetrachloride	20	17.5	ug/L	88			72	136	
	cis-1,2-Dichloroethene	20	17.6	ug/L	88			78	123	
	Chloroform	20	17.5	ug/L	88			79	124	
	1,1,1-Trichloroethane	20	17.5	ug/L	88			74	131	
	Methylcyclohexane	20	18.7	ug/L	94			72	132	
	Benzene	20	18.1	ug/L	91			79	120	
	1,2-Dichloroethane	20	17.7	ug/L	89			73	128	
	Trichloroethene	20	17.3	ug/L	86			79	123	
	1,2-Dichloropropane	20	17.8	ug/L	89			78	122	
	Bromodichloromethane	20	17.6	ug/L	88			79	125	
	4-Methyl-2-Pentanone	100	89.2	ug/L	89			67	130	
	Toluene	20	19.0	ug/L	95			80	121	
	t-1,3-Dichloropropene	20	18.6	ug/L	93			73	127	
	cis-1,3-Dichloropropene	20	18.5	ug/L	93			75	124	
	1,1,2-Trichloroethane	20	17.8	ug/L	89			80	119	
	2-Hexanone	100	94.1	ug/L	94			57	139	
	Dibromochloromethane	20	18.2	ug/L	91			74	126	
	Tetrachloroethene	20	18.0	ug/L	90			74	129	
	Chlorobenzene	20	17.9	ug/L	90			82	118	
	Ethyl Benzene	20	18.8	ug/L	94			79	121	
	m/p-Xylenes	40	38.9	ug/L	97			80	121	
	o-Xylene	20	19.1	ug/L	96			78	122	
	Styrene	20	18.7	ug/L	94			78	123	
	Bromoform	20	18.3	ug/L	92			66	130	
	Isopropylbenzene	20	18.8	ug/L	94			72	131	
	1,1,2,2-Tetrachloroethane	20	16.6	ug/L	83			71	121	
	1,3-Dichlorobenzene	20	17.8	ug/L	89			80	119	
	1,4-Dichlorobenzene	20	17.5	ug/L	88			79	118	
	1,2-Dichlorobenzene	20	17.5	ug/L	88			80	119	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1380

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260-Low

Datafile : VN085786.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VN0219WBSD01	Chloromethane	20	16.5	ug/L	83	4		50	139	20
	Vinyl chloride	20	16.8	ug/L	84	2		58	137	20
	Bromomethane	20	17.1	ug/L	86	0		53	141	20
	Chloroethane	20	16.2	ug/L	81	5		60	138	20
	Trichlorofluoromethane	20	16.7	ug/L	84	2		65	141	20
	1,1,2-Trichlorotrifluoroethane	20	17.1	ug/L	86	3		70	136	20
	1,1-Dichloroethene	20	17.2	ug/L	86	1		71	131	20
	Acetone	100	90.3	ug/L	90	4		39	160	20
	Carbon disulfide	20	15.9	ug/L	79	6		64	133	20
	Methyl tert-butyl Ether	20	19.2	ug/L	96	6		71	124	20
	Methylene Chloride	20	17.2	ug/L	86	2		74	124	20
	trans-1,2-Dichloroethene	20	16.9	ug/L	85	0		75	124	20
	1,1-Dichloroethane	20	17.3	ug/L	86	3		77	125	20
	2-Butanone	100	91.4	ug/L	91	1		56	143	20
	Carbon Tetrachloride	20	17.0	ug/L	85	3		72	136	20
	cis-1,2-Dichloroethene	20	17.6	ug/L	88	0		78	123	20
	Chloroform	20	17.5	ug/L	88	0		79	124	20
	1,1,1-Trichloroethane	20	17.2	ug/L	86	2		74	131	20
	Methylcyclohexane	20	17.3	ug/L	86	9		72	132	20
	Benzene	20	17.8	ug/L	89	2		79	120	20
	1,2-Dichloroethane	20	18.2	ug/L	91	2		73	128	20
	Trichloroethene	20	16.7	ug/L	84	2		79	123	20
	1,2-Dichloropropane	20	18.1	ug/L	91	2		78	122	20
	Bromodichloromethane	20	18.0	ug/L	90	2		79	125	20
	4-Methyl-2-Pentanone	100	96.3	ug/L	96	8		67	130	20
	Toluene	20	18.8	ug/L	94	1		80	121	20
	t-1,3-Dichloropropene	20	18.7	ug/L	94	1		73	127	20
	cis-1,3-Dichloropropene	20	18.8	ug/L	94	1		75	124	20
	1,1,2-Trichloroethane	20	17.9	ug/L	90	1		80	119	20
	2-Hexanone	100	97.6	ug/L	98	4		57	139	20
	Dibromochloromethane	20	18.9	ug/L	95	4		74	126	20
	Tetrachloroethene	20	17.5	ug/L	88	2		74	129	20
	Chlorobenzene	20	17.5	ug/L	88	2		82	118	20
	Ethyl Benzene	20	17.9	ug/L	90	4		79	121	20
	m/p-Xylenes	40	37.6	ug/L	94	3		80	121	20
	o-Xylene	20	18.5	ug/L	93	3		78	122	20
	Styrene	20	18.1	ug/L	91	3		78	123	20
	Bromoform	20	18.5	ug/L	93	1		66	130	20
	Isopropylbenzene	20	18.1	ug/L	91	3		72	131	20
	1,1,2,2-Tetrachloroethane	20	17.6	ug/L	88	6		71	121	20
	1,3-Dichlorobenzene	20	17.4	ug/L	87	2		80	119	20
	1,4-Dichlorobenzene	20	17.1	ug/L	86	2		79	118	20
	1,2-Dichlorobenzene	20	17.1	ug/L	86	2		80	119	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1380

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260D

Datafile : VY021238.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0219SBS01	Chloromethane	20	30.0	ug/Kg	150	*		50	136	
	Vinyl chloride	20	29.0	ug/Kg	145	*		56	135	
	Bromomethane	20	28.3	ug/Kg	142			53	143	
	Chloroethane	20	29.1	ug/Kg	146	*		59	139	
	Trichlorofluoromethane	20	24.4	ug/Kg	122			62	140	
	1,1,2-Trichlorotrifluoroethane	20	23.3	ug/Kg	117			66	136	
	1,1-Dichloroethene	20	23.7	ug/Kg	119			70	131	
	Acetone	100	130	ug/Kg	130			36	164	
	Carbon disulfide	20	24.6	ug/Kg	123			63	132	
	Methyl tert-butyl Ether	20	23.4	ug/Kg	117			73	125	
	Methylene Chloride	20	24.3	ug/Kg	121			70	128	
	trans-1,2-Dichloroethene	20	23.4	ug/Kg	117			74	125	
	1,1-Dichloroethane	20	23.4	ug/Kg	117			76	125	
	2-Butanone	100	120	ug/Kg	120			51	148	
	Carbon Tetrachloride	20	21.4	ug/Kg	107			70	135	
	cis-1,2-Dichloroethene	20	23.0	ug/Kg	115			77	123	
	Chloroform	20	23.3	ug/Kg	117			78	123	
	1,1,1-Trichloroethane	20	22.7	ug/Kg	114			73	130	
	Methylcyclohexane	20	22.7	ug/Kg	114			66	133	
	Benzene	20	22.6	ug/Kg	113			77	121	
	1,2-Dichloroethane	20	22.7	ug/Kg	114			73	128	
	Trichloroethene	20	22.0	ug/Kg	110			77	123	
	1,2-Dichloropropane	20	22.6	ug/Kg	113			76	123	
	Bromodichloromethane	20	22.1	ug/Kg	111			75	127	
	4-Methyl-2-Pentanone	100	110	ug/Kg	110			65	135	
	Toluene	20	22.6	ug/Kg	113			77	121	
	t-1,3-Dichloropropene	20	21.9	ug/Kg	110			71	130	
	cis-1,3-Dichloropropene	20	22.2	ug/Kg	111			74	126	
	1,1,2-Trichloroethane	20	21.7	ug/Kg	109			78	121	
	2-Hexanone	100	110	ug/Kg	110			53	145	
	Dibromochloromethane	20	21.4	ug/Kg	107			74	126	
	Tetrachloroethene	20	21.8	ug/Kg	109			73	128	
	Chlorobenzene	20	21.9	ug/Kg	110			79	120	
	Ethyl Benzene	20	21.9	ug/Kg	110			76	122	
	m/p-Xylenes	40	45.1	ug/Kg	113			77	124	
	o-Xylene	20	22.4	ug/Kg	112			77	123	
	Styrene	20	22.2	ug/Kg	111			76	124	
	Bromoform	20	21.4	ug/Kg	107			67	132	
	Isopropylbenzene	20	21.3	ug/Kg	106			68	134	
	1,1,2,2-Tetrachloroethane	20	20.7	ug/Kg	104			70	124	
	1,3-Dichlorobenzene	20	21.1	ug/Kg	106			77	121	
	1,4-Dichlorobenzene	20	21.1	ug/Kg	106			75	120	
	1,2-Dichlorobenzene	20	20.8	ug/Kg	104			78	121	

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1380

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8260D

Datafile : VY021322.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	Low	Limits High	RPD
VY0226SBS01	Chloromethane	20	18.9	ug/Kg	95			50	136	
	Vinyl chloride	20	19.3	ug/Kg	97			56	135	
	Bromomethane	20	19.2	ug/Kg	96			53	143	
	Chloroethane	20	19.0	ug/Kg	95			59	139	
	Trichlorofluoromethane	20	19.0	ug/Kg	95			62	140	
	1,1,2-Trichlorotrifluoroethane	20	19.8	ug/Kg	99			66	136	
	1,1-Dichloroethene	20	19.0	ug/Kg	95			70	131	
	Acetone	100	94.9	ug/Kg	95			36	164	
	Carbon disulfide	20	18.6	ug/Kg	93			63	132	
	Methyl tert-butyl Ether	20	19.1	ug/Kg	96			73	125	
	Methylene Chloride	20	18.5	ug/Kg	93			70	128	
	trans-1,2-Dichloroethene	20	19.0	ug/Kg	95			74	125	
	1,1-Dichloroethane	20	19.1	ug/Kg	96			76	125	
	2-Butanone	100	92.9	ug/Kg	93			51	148	
	Carbon Tetrachloride	20	19.0	ug/Kg	95			70	135	
	cis-1,2-Dichloroethene	20	19.2	ug/Kg	96			77	123	
	Chloroform	20	19.3	ug/Kg	97			78	123	
	1,1,1-Trichloroethane	20	19.2	ug/Kg	96			73	130	
	Methylcyclohexane	20	18.9	ug/Kg	95			66	133	
	Benzene	20	19.3	ug/Kg	97			77	121	
	1,2-Dichloroethane	20	19.2	ug/Kg	96			73	128	
	Trichloroethene	20	19.2	ug/Kg	96			77	123	
	1,2-Dichloropropane	20	19.1	ug/Kg	96			76	123	
	Bromodichloromethane	20	19.1	ug/Kg	96			75	127	
	4-Methyl-2-Pentanone	100	93.3	ug/Kg	93			65	135	
	Toluene	20	19.3	ug/Kg	97			77	121	
	t-1,3-Dichloropropene	20	19.0	ug/Kg	95			71	130	
	cis-1,3-Dichloropropene	20	19.3	ug/Kg	97			74	126	
	1,1,2-Trichloroethane	20	19.1	ug/Kg	96			78	121	
	2-Hexanone	100	93.8	ug/Kg	94			53	145	
	Dibromochloromethane	20	19.3	ug/Kg	97			74	126	
	Tetrachloroethene	20	19.2	ug/Kg	96			73	128	
	Chlorobenzene	20	19.0	ug/Kg	95			79	120	
	Ethyl Benzene	20	19.0	ug/Kg	95			76	122	
	m/p-Xylenes	40	38.8	ug/Kg	97			77	124	
	o-Xylene	20	19.3	ug/Kg	97			77	123	
	Styrene	20	19.3	ug/Kg	97			76	124	
	Bromoform	20	19.1	ug/Kg	96			67	132	
	Isopropylbenzene	20	19.1	ug/Kg	96			68	134	
	1,1,2,2-Tetrachloroethane	20	18.8	ug/Kg	94			70	124	
	1,3-Dichlorobenzene	20	19.1	ug/Kg	96			77	121	
	1,4-Dichlorobenzene	20	19.2	ug/Kg	96			75	120	
	1,2-Dichlorobenzene	20	18.8	ug/Kg	94			78	121	

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VN0219WBL01

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM Case No.: Q1380

SAS No.: Q1380 SDG NO.: Q1380

Lab File ID: VN085784.D

Lab Sample ID: VN0219WBL01

Date Analyzed: 02/19/2025

Time Analyzed: 11:49

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) N

Instrument ID: MSVOA_N

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VN0219WBS01	VN0219WBS01	VN085785.D	02/19/2025
VN0219WBSD01	VN0219WBSD01	VN085786.D	02/19/2025
VPB192-HYD-20250214	Q1380-12	VN085787.D	02/19/2025
BP-VPB-192-EB-20250212	Q1380-01	VN085788.D	02/19/2025
BP-VPB-192-TB-20250210	Q1380-02	VN085789.D	02/19/2025
BP-VPB-192-GW-780-782	Q1380-09	VN085790.D	02/19/2025

COMMENTS: _____

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VY0219SBL01

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM Case No.: Q1380

SAS No.: Q1380 SDG NO.: Q1380

Lab File ID: VY021237.D

Lab Sample ID: VY0219SBL01

Date Analyzed: 02/19/2025

Time Analyzed: 15:21

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) Y

Instrument ID: MSVOA_Y

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VY0219SBS01	VY0219SBS01	VY021238.D	02/19/2025
BP-VPB-192-GW-725-727	Q1380-06	VY021241.D	02/19/2025

COMMENTS: _____

VOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

VY0226SBL01

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM Case No.: Q1380

SAS No.: Q1380 SDG NO.: Q1380

Lab File ID: VY021321.D

Lab Sample ID: VY0226SBL01

Date Analyzed: 02/26/2025

Time Analyzed: 10:50

GC Column: RXI-624 ID: 0.25 (mm)

Heated Purge: (Y/N) Y

Instrument ID: MSVOA_Y

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
VY0226SBS01	VY0226SBS01	VY021322.D	02/26/2025
BP-VPB-192-GW-725-727MS	Q1380-07MS	VY021342.D	02/26/2025
BP-VPB-192-GW-725-727MSD	Q1380-08MSD	VY021343.D	02/26/2025

COMMENTS: _____

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG NO.: Q1380
Lab File ID: VN085771.D BFB Injection Date: 02/18/2025
Instrument ID: MSVOA_N BFB Injection Time: 10:35
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	18.1
75	30.0 - 60.0% of mass 95	49.7
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	1.4 (1.8) 1
174	50.0 - 100.0% of mass 95	81.7
175	5.0 - 9.0% of mass 174	6 (7.4) 1
176	95.0 - 101.0% of mass 174	78.9 (96.5) 1
177	5.0 - 9.0% of mass 176	5.4 (6.9) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC100	VSTDIC100	VN085772.D	02/18/2025	11:09
VSTDICCC050	VSTDICCC050	VN085773.D	02/18/2025	11:32
VSTDIC010	VSTDIC010	VN085775.D	02/18/2025	12:20
VSTDIC005	VSTDIC005	VN085776.D	02/18/2025	12:43
VSTDIC001	VSTDIC001	VN085777.D	02/18/2025	13:07
VSTDIC020	VSTDIC020	VN085779.D	02/18/2025	14:18

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG NO.: Q1380
Lab File ID: VN085781.D BFB Injection Date: 02/19/2025
Instrument ID: MSVOA_N BFB Injection Time: 10:14
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N N

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	18.4
75	30.0 - 60.0% of mass 95	49.6
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.8
173	Less than 2.0% of mass 174	1.3 (1.7) 1
174	50.0 - 100.0% of mass 95	75.8
175	5.0 - 9.0% of mass 174	6 (7.9) 1
176	95.0 - 101.0% of mass 174	76 (100.3) 1
177	5.0 - 9.0% of mass 176	4.6 (6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VN085782.D	02/19/2025	10:51
VN0219WBL01	VN0219WBL01	VN085784.D	02/19/2025	11:49
VN0219WBS01	VN0219WBS01	VN085785.D	02/19/2025	12:49
VN0219WBSD01	VN0219WBSD01	VN085786.D	02/19/2025	13:22
VPB192-HYD-20250214	Q1380-12	VN085787.D	02/19/2025	13:57
BP-VPB-192-EB-20250212	Q1380-01	VN085788.D	02/19/2025	14:21
BP-VPB-192-TB-20250210	Q1380-02	VN085789.D	02/19/2025	14:45
BP-VPB-192-GW-780-782	Q1380-09	VN085790.D	02/19/2025	15:08
VSTDCCC050EC	VSTDCCC050	VN085792.D	02/19/2025	15:58

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG NO.: Q1380
Lab File ID: VY021019.D BFB Injection Date: 02/03/2025
Instrument ID: MSVOA_Y BFB Injection Time: 09:51
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	15.3
75	30.0 - 60.0% of mass 95	49
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	0.3 (0.4) 1
174	50.0 - 100.0% of mass 95	89.7
175	5.0 - 9.0% of mass 174	7 (7.8) 1
176	95.0 - 101.0% of mass 174	87 (97) 1
177	5.0 - 9.0% of mass 176	5.9 (6.8) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VY021020.D	02/03/2025	10:35
VSTDIC010	VSTDIC010	VY021021.D	02/03/2025	10:57
VSTDIC020	VSTDIC020	VY021022.D	02/03/2025	11:20
VSTDIC050	VSTDIC050	VY021023.D	02/03/2025	11:43
VSTDIC100	VSTDIC100	VY021024.D	02/03/2025	12:21
VSTDIC150	VSTDIC150	VY021025.D	02/03/2025	12:44

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG NO.: Q1380
Lab File ID: VY021235.D BFB Injection Date: 02/19/2025
Instrument ID: MSVOA_Y BFB Injection Time: 09:09
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	20.6
75	30.0 - 60.0% of mass 95	54.3
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.5
173	Less than 2.0% of mass 174	1.1 (1.3) 1
174	50.0 - 100.0% of mass 95	84.8
175	5.0 - 9.0% of mass 174	6.4 (7.6) 1
176	95.0 - 101.0% of mass 174	81.6 (96.3) 1
177	5.0 - 9.0% of mass 176	5.4 (6.6) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY021236.D	02/19/2025	14:55
VY0219SBL01	VY0219SBL01	VY021237.D	02/19/2025	15:21
VY0219SBS01	VY0219SBS01	VY021238.D	02/19/2025	15:49
BP-VPB-192-GW-725-727	Q1380-06	VY021241.D	02/19/2025	16:59
VSTDCCC050EC	VSTDCCC050	VY021245.D	02/19/2025	18:32

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG NO.: Q1380
Lab File ID: VY021308.D BFB Injection Date: 02/25/2025
Instrument ID: MSVOA_Y BFB Injection Time: 12:10
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	20.8
75	30.0 - 60.0% of mass 95	55.4
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.7
173	Less than 2.0% of mass 174	0.7 (0.8) 1
174	50.0 - 100.0% of mass 95	84.7
175	5.0 - 9.0% of mass 174	6.4 (7.5) 1
176	95.0 - 101.0% of mass 174	83.6 (98.8) 1
177	5.0 - 9.0% of mass 176	5.5 (6.5) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDIC005	VSTDIC005	VY021309.D	02/25/2025	12:40
VSTDIC010	VSTDIC010	VY021310.D	02/25/2025	13:03
VSTDIC020	VSTDIC020	VY021311.D	02/25/2025	13:26
VSTDIC050	VSTDIC050	VY021312.D	02/25/2025	14:48
VSTDIC150	VSTDIC150	VY021314.D	02/25/2025	15:57
VSTDIC100	VSTDIC100	VY021316.D	02/25/2025	16:49

VOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
BROMOFLUOROBENZENE (BFB)

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG NO.: Q1380
Lab File ID: VY021319.D BFB Injection Date: 02/26/2025
Instrument ID: MSVOA_Y BFB Injection Time: 09:12
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: Y/N Y

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
50	15.0 - 40.0% of mass 95	21.5
75	30.0 - 60.0% of mass 95	55.4
95	Base Peak, 100% relative abundance	100
96	5.0 - 9.0% of mass 95	6.6
173	Less than 2.0% of mass 174	1.2 (1.5) 1
174	50.0 - 100.0% of mass 95	78.5
175	5.0 - 9.0% of mass 174	6.5 (8.3) 1
176	95.0 - 101.0% of mass 174	75.9 (96.7) 1
177	5.0 - 9.0% of mass 176	5.1 (6.7) 2

1-Value is % mass 174

2-Value is % mass 176

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
VSTDCCC050	VSTDCCC050	VY021320.D	02/26/2025	10:11
VY0226SBL01	VY0226SBL01	VY021321.D	02/26/2025	10:50
VY0226SBS01	VY0226SBS01	VY021322.D	02/26/2025	11:22
BP-VPB-192-GW-725-727MS	Q1380-07MS	VY021342.D	02/26/2025	19:21
BP-VPB-192-GW-725-727MSD	Q1380-08MSD	VY021343.D	02/26/2025	19:43
VSTDCCC050EC	VSTDCCC050	VY021344.D	02/26/2025	20:06

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG NO.: Q1380
Lab File ID: VN085782.D Date Analyzed: 02/19/2025
Instrument ID: MSVOA_N Time Analyzed: 10:51
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	338951	8.22	542050	9.09	475442	11.86
UPPER LIMIT	677902	8.718	1084100	9.594	950884	12.359
LOWER LIMIT	169476	7.718	271025	8.594	237721	11.359
EPA SAMPLE NO.						
BP-VPB-192-EB-20250212	292071	8.22	537683	9.10	478583	11.86
BP-VPB-192-TB-20250210	264567	8.22	487878	9.10	434147	11.86
BP-VPB-192-GW-780-782	259651	8.22	470979	9.10	403744	11.87
VPB192-HYD-20250214	297765	8.22	540496	9.09	473230	11.86
VN0219WBL01	286037	8.22	514740	9.09	457806	11.86
VN0219WBS01	317605	8.22	521252	9.09	448364	11.86
VN0219WBSD01	291949	8.22	482343	9.09	419097	11.86

IS1 = Pentafluorobenzene

IS2 = 1,4-Difluorobenzene

IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG NO.: Q1380
Lab File ID: VN085782.D Date Analyzed: 02/19/2025
Instrument ID: MSVOA_N Time Analyzed: 10:51
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	234508	13.788				
UPPER LIMIT	469016	14.288				
LOWER LIMIT	117254	13.288				
EPA SAMPLE NO.						
BP-VPB-192-EB-20250212	200119	13.79				
BP-VPB-192-TB-20250210	174960	13.79				
BP-VPB-192-GW-780-782	157449	13.79				
VPB192-HYD-20250214	194311	13.79				
VN0219WBL01	185142	13.79				
VN0219WBS01	214627	13.79				
VN0219WBSD01	199880	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG NO.: Q1380
Lab File ID: VY021236.D Date Analyzed: 02/19/2025
Instrument ID: MSVOA_Y Time Analyzed: 14:55
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	200462	7.72	311138	8.63	266562	11.43
UPPER LIMIT	400924	8.219	622276	9.128	533124	11.926
LOWER LIMIT	100231	7.219	155569	8.128	133281	10.926
EPA SAMPLE NO.						
BP-VPB-192-GW-725-727	223938	7.71	424683	8.62	405408	11.42
VY0219SBL01	252953	7.72	478949	8.62	429832	11.43
VY0219SBS01	187422	7.71	302089	8.62	255752	11.42

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG NO.: Q1380
Lab File ID: VY021236.D Date Analyzed: 02/19/2025
Instrument ID: MSVOA_Y Time Analyzed: 14:55
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	126814	13.359				
UPPER LIMIT	253628	13.859				
LOWER LIMIT	63407	12.859				
EPA SAMPLE NO.						
BP-VPB-192-GW-725-727	161420	13.35				
VY0219SBL01	157077	13.35				
VY0219SBS01	127239	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG NO.: Q1380
Lab File ID: VY021320.D Date Analyzed: 02/26/2025
Instrument ID: MSVOA_Y Time Analyzed: 10:11
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS1 AREA #	RT #	IS2 AREA #	RT #	IS3 AREA #	RT #
12 HOUR STD	193888	7.71	308714	8.62	268805	11.42
UPPER LIMIT	387776	8.213	617428	9.116	537610	11.92
LOWER LIMIT	96944	7.213	154357	8.116	134403	10.92
EPA SAMPLE NO.						
BP-VPB-192-GW-725-727MS	157236	7.71	260195	8.62	210847	11.41
BP-VPB-192-GW-725-727MSD	137100	7.71	223153	8.62	181056	11.41
VY0226SBL01	251129	7.71	483695	8.62	470141	11.42
VY0226SBS01	184417	7.71	291138	8.62	254891	11.41

IS1 = Pentafluorobenzene
IS2 = 1,4-Difluorobenzene
IS3 = Chlorobenzene-d5

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: TETR06
Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG NO.: Q1380
Lab File ID: VY021320.D Date Analyzed: 02/26/2025
Instrument ID: MSVOA_Y Time Analyzed: 10:11
GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) Y

	IS4 AREA #	RT #				
12 HOUR STD	131935	13.347				
UPPER LIMIT	263870	13.847				
LOWER LIMIT	65967.5	12.847				
EPA SAMPLE NO.						
BP-VPB-192-GW-725-727MS	76053	13.35				
BP-VPB-192-GW-725-727MSD	63404 *	13.35				
VY0226SBL01	192201	13.35				
VY0226SBS01	128725	13.35				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area
AREA LOWER LIMIT = -50% of internal standard area
RT UPPER LIMIT = +0.50 minutes of internal standard RT
RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.
* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	
Client Sample ID:	VN0219WBL01		SDG No.:	Q1380
Lab Sample ID:	VN0219WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085784.D	1		02/19/25 11:49	VN021925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	0.50	U	0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	0.75	U	0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	3.80	U	1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	0.75	U	0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	0.50	U	0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	0.50	U	0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	0.75	U	0.26	0.75	1.00	ug/L
67-64-1	Acetone	3.80	U	1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	0.75	U	0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	0.50	U	0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	0.50	U	0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	0.50	U	0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	0.50	U	0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	2.50	U	1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	0.50	U	0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	0.75	U	0.25	0.75	1.00	ug/L
67-66-3	Chloroform	0.50	U	0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	0.50	U	0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	0.50	U	0.19	0.50	1.00	ug/L
71-43-2	Benzene	0.50	U	0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	0.50	U	0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	0.75	U	0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	0.50	U	0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	0.50	U	0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	2.50	U	0.75	2.50	5.00	ug/L
108-88-3	Toluene	0.50	U	0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	0.50	U	0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	0.50	U	0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	0.50	U	0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	2.50	U	1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	
Client Sample ID:	VN0219WBL01		SDG No.:	Q1380
Lab Sample ID:	VN0219WBL01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085784.D	1		02/19/25 11:49	VN021925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	0.50	U	0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	0.50	U	0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	0.50	U	0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	0.50	U	0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	1.00	U	0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	0.50	U	0.14	0.50	1.00	ug/L
100-42-5	Styrene	0.50	U	0.16	0.50	1.00	ug/L
75-25-2	Bromoform	0.50	U	0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	0.50	U	0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	0.50	U	0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	0.50	U	0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	0.50	U	0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	0.50	U	0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	56.1		81 - 118		112%	SPK: 50
1868-53-7	Dibromofluoromethane	54.6		80 - 119		109%	SPK: 50
2037-26-5	Toluene-d8	51.9		89 - 112		104%	SPK: 50
460-00-4	4-Bromofluorobenzene	48.1		85 - 114		96%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	286000	8.218				
540-36-3	1,4-Difluorobenzene	515000	9.094				
3114-55-4	Chlorobenzene-d5	458000	11.859				
3855-82-1	1,4-Dichlorobenzene-d4	185000	13.788				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	
Client Sample ID:	VY0219SBL01		SDG No.:	Q1380
Lab Sample ID:	VY0219SBL01		Matrix:	SOIL
Analytical Method:	SW8260		% Solid:	100
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY021237.D	1		02/19/25 15:21	VY021925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units(Dry Weight)
TARGETS							
74-87-3	Chloromethane	2.50	U	1.20	2.50	5.00	ug/Kg
75-01-4	Vinyl Chloride	2.50	U	0.77	2.50	5.00	ug/Kg
74-83-9	Bromomethane	4.00	U	1.00	4.00	5.00	ug/Kg
75-00-3	Chloroethane	2.50	U	1.00	2.50	5.00	ug/Kg
75-69-4	Trichlorofluoromethane	4.00	U	0.91	4.00	5.00	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	2.50	U	1.10	2.50	5.00	ug/Kg
75-35-4	1,1-Dichloroethene	2.50	U	0.78	2.50	5.00	ug/Kg
67-64-1	Acetone	20.0	U	6.20	20.0	25.0	ug/Kg
75-15-0	Carbon Disulfide	4.00	U	1.30	4.00	5.00	ug/Kg
1634-04-4	Methyl tert-butyl Ether	2.50	U	0.67	2.50	5.00	ug/Kg
75-09-2	Methylene Chloride	8.00	U	3.40	8.00	10.0	ug/Kg
156-60-5	trans-1,2-Dichloroethene	2.50	U	0.84	2.50	5.00	ug/Kg
75-34-3	1,1-Dichloroethane	2.50	U	0.63	2.50	5.00	ug/Kg
78-93-3	2-Butanone	20.0	U	5.70	20.0	25.0	ug/Kg
56-23-5	Carbon Tetrachloride	2.50	U	0.87	2.50	5.00	ug/Kg
156-59-2	cis-1,2-Dichloroethene	2.50	U	0.61	2.50	5.00	ug/Kg
67-66-3	Chloroform	4.00	U	0.67	4.00	5.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	2.50	U	0.78	2.50	5.00	ug/Kg
108-87-2	Methylcyclohexane	2.50	U	0.87	2.50	5.00	ug/Kg
71-43-2	Benzene	2.50	U	0.72	2.50	5.00	ug/Kg
107-06-2	1,2-Dichloroethane	2.50	U	0.61	2.50	5.00	ug/Kg
79-01-6	Trichloroethene	2.50	U	0.75	2.50	5.00	ug/Kg
78-87-5	1,2-Dichloropropane	2.50	U	0.66	2.50	5.00	ug/Kg
75-27-4	Bromodichloromethane	2.50	U	0.56	2.50	5.00	ug/Kg
108-10-1	4-Methyl-2-Pentanone	12.5	U	4.40	12.5	25.0	ug/Kg
108-88-3	Toluene	2.50	U	0.67	2.50	5.00	ug/Kg
10061-02-6	t-1,3-Dichloropropene	2.50	U	0.60	2.50	5.00	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	2.50	U	0.57	2.50	5.00	ug/Kg
79-00-5	1,1,2-Trichloroethane	2.50	U	0.84	2.50	5.00	ug/Kg
591-78-6	2-Hexanone	12.5	U	4.80	12.5	25.0	ug/Kg

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	
Client Sample ID:	VY0219SBL01		SDG No.:	Q1380
Lab Sample ID:	VY0219SBL01		Matrix:	SOIL
Analytical Method:	SW8260		% Solid:	100
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY021237.D	1		02/19/25 15:21	VY021925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units(Dry Weight)
124-48-1	Dibromochloromethane	2.50	U	0.65	2.50	5.00	ug/Kg
127-18-4	Tetrachloroethene	2.50	U	0.89	2.50	5.00	ug/Kg
108-90-7	Chlorobenzene	2.50	U	0.74	2.50	5.00	ug/Kg
100-41-4	Ethyl Benzene	2.50	U	0.62	2.50	5.00	ug/Kg
179601-23-1	m/p-Xylenes	5.00	U	1.40	5.00	10.0	ug/Kg
95-47-6	o-Xylene	2.50	U	0.70	2.50	5.00	ug/Kg
100-42-5	Styrene	2.50	U	0.60	2.50	5.00	ug/Kg
75-25-2	Bromoform	2.50	U	0.81	2.50	5.00	ug/Kg
98-82-8	Isopropylbenzene	2.50	U	0.67	2.50	5.00	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	2.50	U	1.10	2.50	5.00	ug/Kg
541-73-1	1,3-Dichlorobenzene	2.50	U	0.74	2.50	5.00	ug/Kg
106-46-7	1,4-Dichlorobenzene	2.50	U	0.80	2.50	5.00	ug/Kg
95-50-1	1,2-Dichlorobenzene	2.50	U	0.59	2.50	5.00	ug/Kg
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	54.2		71 - 136		108%	SPK: 50
1868-53-7	Dibromofluoromethane	50.2		78 - 119		100%	SPK: 50
2037-26-5	Toluene-d8	49.6		85 - 116		99%	SPK: 50
460-00-4	4-Bromofluorobenzene	47.8		79 - 119		96%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	253000	7.719				
540-36-3	1,4-Difluorobenzene	479000	8.622				
3114-55-4	Chlorobenzene-d5	430000	11.426				
3855-82-1	1,4-Dichlorobenzene-d4	157000	13.352				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	
Client Sample ID:	VY0226SBL01		SDG No.:	Q1380
Lab Sample ID:	VY0226SBL01		Matrix:	SOIL
Analytical Method:	SW8260		% Solid:	100
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY021321.D	1		02/26/25 10:50	VY022625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units(Dry Weight)
TARGETS							
74-87-3	Chloromethane	2.50	U	1.20	2.50	5.00	ug/Kg
75-01-4	Vinyl Chloride	2.50	U	0.77	2.50	5.00	ug/Kg
74-83-9	Bromomethane	4.00	U	1.00	4.00	5.00	ug/Kg
75-00-3	Chloroethane	2.50	U	1.00	2.50	5.00	ug/Kg
75-69-4	Trichlorofluoromethane	4.00	U	0.91	4.00	5.00	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	2.50	U	1.10	2.50	5.00	ug/Kg
75-35-4	1,1-Dichloroethene	2.50	U	0.78	2.50	5.00	ug/Kg
67-64-1	Acetone	20.0	U	6.20	20.0	25.0	ug/Kg
75-15-0	Carbon Disulfide	4.00	U	1.30	4.00	5.00	ug/Kg
1634-04-4	Methyl tert-butyl Ether	2.50	U	0.67	2.50	5.00	ug/Kg
75-09-2	Methylene Chloride	8.00	U	3.40	8.00	10.0	ug/Kg
156-60-5	trans-1,2-Dichloroethene	2.50	U	0.84	2.50	5.00	ug/Kg
75-34-3	1,1-Dichloroethane	2.50	U	0.63	2.50	5.00	ug/Kg
78-93-3	2-Butanone	20.0	U	5.70	20.0	25.0	ug/Kg
56-23-5	Carbon Tetrachloride	2.50	U	0.87	2.50	5.00	ug/Kg
156-59-2	cis-1,2-Dichloroethene	2.50	U	0.61	2.50	5.00	ug/Kg
67-66-3	Chloroform	4.00	U	0.67	4.00	5.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	2.50	U	0.78	2.50	5.00	ug/Kg
108-87-2	Methylcyclohexane	2.50	U	0.87	2.50	5.00	ug/Kg
71-43-2	Benzene	2.50	U	0.72	2.50	5.00	ug/Kg
107-06-2	1,2-Dichloroethane	2.50	U	0.61	2.50	5.00	ug/Kg
79-01-6	Trichloroethene	2.50	U	0.75	2.50	5.00	ug/Kg
78-87-5	1,2-Dichloropropane	2.50	U	0.66	2.50	5.00	ug/Kg
75-27-4	Bromodichloromethane	2.50	U	0.56	2.50	5.00	ug/Kg
108-10-1	4-Methyl-2-Pentanone	12.5	U	4.40	12.5	25.0	ug/Kg
108-88-3	Toluene	2.50	U	0.67	2.50	5.00	ug/Kg
10061-02-6	t-1,3-Dichloropropene	2.50	U	0.60	2.50	5.00	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	2.50	U	0.57	2.50	5.00	ug/Kg
79-00-5	1,1,2-Trichloroethane	2.50	U	0.84	2.50	5.00	ug/Kg
591-78-6	2-Hexanone	12.5	U	4.80	12.5	25.0	ug/Kg

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	
Client Sample ID:	VY0226SBL01		SDG No.:	Q1380
Lab Sample ID:	VY0226SBL01		Matrix:	SOIL
Analytical Method:	SW8260		% Solid:	100
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY021321.D	1		02/26/25 10:50	VY022625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units(Dry Weight)
124-48-1	Dibromochloromethane	2.50	U	0.65	2.50	5.00	ug/Kg
127-18-4	Tetrachloroethene	2.50	U	0.89	2.50	5.00	ug/Kg
108-90-7	Chlorobenzene	2.50	U	0.74	2.50	5.00	ug/Kg
100-41-4	Ethyl Benzene	2.50	U	0.62	2.50	5.00	ug/Kg
179601-23-1	m/p-Xylenes	5.00	U	1.40	5.00	10.0	ug/Kg
95-47-6	o-Xylene	2.50	U	0.70	2.50	5.00	ug/Kg
100-42-5	Styrene	2.50	U	0.60	2.50	5.00	ug/Kg
75-25-2	Bromoform	2.50	U	0.81	2.50	5.00	ug/Kg
98-82-8	Isopropylbenzene	2.50	U	0.67	2.50	5.00	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	2.50	U	1.10	2.50	5.00	ug/Kg
541-73-1	1,3-Dichlorobenzene	2.50	U	0.74	2.50	5.00	ug/Kg
106-46-7	1,4-Dichlorobenzene	2.50	U	0.80	2.50	5.00	ug/Kg
95-50-1	1,2-Dichlorobenzene	2.50	U	0.59	2.50	5.00	ug/Kg
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	53.1		71 - 136		106%	SPK: 50
1868-53-7	Dibromofluoromethane	49.5		78 - 119		99%	SPK: 50
2037-26-5	Toluene-d8	49.1		85 - 116		98%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.1		79 - 119		100%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	251000	7.713				
540-36-3	1,4-Difluorobenzene	484000	8.615				
3114-55-4	Chlorobenzene-d5	470000	11.42				
3855-82-1	1,4-Dichlorobenzene-d4	192000	13.346				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	
Client Sample ID:	VN0219WBS01		SDG No.:	Q1380
Lab Sample ID:	VN0219WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085785.D	1		02/19/25 12:49	VN021925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	17.2		0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	17.2		0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	17.2		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	17.0		0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	17.1		0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	17.7		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	17.4		0.26	0.75	1.00	ug/L
67-64-1	Acetone	93.8		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	16.8		0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	17.9		0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	16.8		0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	17.0		0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	17.7		0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	89.6		1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	17.5		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	17.6		0.25	0.75	1.00	ug/L
67-66-3	Chloroform	17.5		0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	17.5		0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	18.7		0.19	0.50	1.00	ug/L
71-43-2	Benzene	18.1		0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	17.7		0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	17.3		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	17.8		0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	17.6		0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	89.2		0.75	2.50	5.00	ug/L
108-88-3	Toluene	19.0		0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	18.6		0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	18.5		0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	17.8		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	94.1		1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	
Client Sample ID:	VN0219WBS01		SDG No.:	Q1380
Lab Sample ID:	VN0219WBS01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085785.D	1		02/19/25 12:49	VN021925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	18.2		0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	18.0		0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	17.9		0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	18.8		0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	38.9		0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	19.1		0.14	0.50	1.00	ug/L
100-42-5	Styrene	18.7		0.16	0.50	1.00	ug/L
75-25-2	Bromoform	18.3		0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	18.8		0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	16.6		0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	17.8		0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	17.5		0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	17.5		0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	52.7		81 - 118		105%	SPK: 50
1868-53-7	Dibromofluoromethane	52.5		80 - 119		105%	SPK: 50
2037-26-5	Toluene-d8	54.8		89 - 112		110%	SPK: 50
460-00-4	4-Bromofluorobenzene	56.0		85 - 114		112%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	318000	8.218				
540-36-3	1,4-Difluorobenzene	521000	9.094				
3114-55-4	Chlorobenzene-d5	448000	11.859				
3855-82-1	1,4-Dichlorobenzene-d4	215000	13.788				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	
Client Sample ID:	VY0219SBS01		SDG No.:	Q1380
Lab Sample ID:	VY0219SBS01		Matrix:	SOIL
Analytical Method:	SW8260		% Solid:	100
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY021238.D	1		02/19/25 15:49	VY021925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units(Dry Weight)
TARGETS							
74-87-3	Chloromethane	30.0		1.20	2.50	5.00	ug/Kg
75-01-4	Vinyl Chloride	29.0		0.77	2.50	5.00	ug/Kg
74-83-9	Bromomethane	28.3		1.00	4.00	5.00	ug/Kg
75-00-3	Chloroethane	29.1		1.00	2.50	5.00	ug/Kg
75-69-4	Trichlorofluoromethane	24.4		0.91	4.00	5.00	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	23.3		1.10	2.50	5.00	ug/Kg
75-35-4	1,1-Dichloroethene	23.7		0.78	2.50	5.00	ug/Kg
67-64-1	Acetone	130		6.20	20.0	25.0	ug/Kg
75-15-0	Carbon Disulfide	24.6		1.30	4.00	5.00	ug/Kg
1634-04-4	Methyl tert-butyl Ether	23.4		0.67	2.50	5.00	ug/Kg
75-09-2	Methylene Chloride	24.3		3.40	8.00	10.0	ug/Kg
156-60-5	trans-1,2-Dichloroethene	23.4		0.84	2.50	5.00	ug/Kg
75-34-3	1,1-Dichloroethane	23.4		0.63	2.50	5.00	ug/Kg
78-93-3	2-Butanone	120		5.70	20.0	25.0	ug/Kg
56-23-5	Carbon Tetrachloride	21.4		0.87	2.50	5.00	ug/Kg
156-59-2	cis-1,2-Dichloroethene	23.0		0.61	2.50	5.00	ug/Kg
67-66-3	Chloroform	23.3		0.67	4.00	5.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	22.7		0.78	2.50	5.00	ug/Kg
108-87-2	Methylcyclohexane	22.7		0.87	2.50	5.00	ug/Kg
71-43-2	Benzene	22.6		0.72	2.50	5.00	ug/Kg
107-06-2	1,2-Dichloroethane	22.7		0.61	2.50	5.00	ug/Kg
79-01-6	Trichloroethene	22.0		0.75	2.50	5.00	ug/Kg
78-87-5	1,2-Dichloropropane	22.6		0.66	2.50	5.00	ug/Kg
75-27-4	Bromodichloromethane	22.1		0.56	2.50	5.00	ug/Kg
108-10-1	4-Methyl-2-Pentanone	110		4.40	12.5	25.0	ug/Kg
108-88-3	Toluene	22.6		0.67	2.50	5.00	ug/Kg
10061-02-6	t-1,3-Dichloropropene	21.9		0.60	2.50	5.00	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	22.2		0.57	2.50	5.00	ug/Kg
79-00-5	1,1,2-Trichloroethane	21.7		0.84	2.50	5.00	ug/Kg
591-78-6	2-Hexanone	110		4.80	12.5	25.0	ug/Kg

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	
Client Sample ID:	VY0219SBS01		SDG No.:	Q1380
Lab Sample ID:	VY0219SBS01		Matrix:	SOIL
Analytical Method:	SW8260		% Solid:	100
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY021238.D	1		02/19/25 15:49	VY021925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units(Dry Weight)
124-48-1	Dibromochloromethane	21.4		0.65	2.50	5.00	ug/Kg
127-18-4	Tetrachloroethene	21.8		0.89	2.50	5.00	ug/Kg
108-90-7	Chlorobenzene	21.9		0.74	2.50	5.00	ug/Kg
100-41-4	Ethyl Benzene	21.9		0.62	2.50	5.00	ug/Kg
179601-23-1	m/p-Xylenes	45.1		1.40	5.00	10.0	ug/Kg
95-47-6	o-Xylene	22.4		0.70	2.50	5.00	ug/Kg
100-42-5	Styrene	22.2		0.60	2.50	5.00	ug/Kg
75-25-2	Bromoform	21.4		0.81	2.50	5.00	ug/Kg
98-82-8	Isopropylbenzene	21.3		0.67	2.50	5.00	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	20.7		1.10	2.50	5.00	ug/Kg
541-73-1	1,3-Dichlorobenzene	21.1		0.74	2.50	5.00	ug/Kg
106-46-7	1,4-Dichlorobenzene	21.1		0.80	2.50	5.00	ug/Kg
95-50-1	1,2-Dichlorobenzene	20.8		0.59	2.50	5.00	ug/Kg
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	50.3		71 - 136		101%	SPK: 50
1868-53-7	Dibromofluoromethane	48.6		78 - 119		97%	SPK: 50
2037-26-5	Toluene-d8	47.3		85 - 116		95%	SPK: 50
460-00-4	4-Bromofluorobenzene	50.7		79 - 119		101%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	187000	7.713				
540-36-3	1,4-Difluorobenzene	302000	8.622				
3114-55-4	Chlorobenzene-d5	256000	11.42				
3855-82-1	1,4-Dichlorobenzene-d4	127000	13.352				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	
Client Sample ID:	VY0226SBS01		SDG No.:	Q1380
Lab Sample ID:	VY0226SBS01		Matrix:	SOIL
Analytical Method:	SW8260		% Solid:	100
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY021322.D	1		02/26/25 11:22	VY022625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units(Dry Weight)
TARGETS							
74-87-3	Chloromethane	18.9		1.20	2.50	5.00	ug/Kg
75-01-4	Vinyl Chloride	19.3		0.77	2.50	5.00	ug/Kg
74-83-9	Bromomethane	19.2		1.00	4.00	5.00	ug/Kg
75-00-3	Chloroethane	19.0		1.00	2.50	5.00	ug/Kg
75-69-4	Trichlorofluoromethane	19.0		0.91	4.00	5.00	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	19.8		1.10	2.50	5.00	ug/Kg
75-35-4	1,1-Dichloroethene	19.0		0.78	2.50	5.00	ug/Kg
67-64-1	Acetone	94.9		6.20	20.0	25.0	ug/Kg
75-15-0	Carbon Disulfide	18.6		1.30	4.00	5.00	ug/Kg
1634-04-4	Methyl tert-butyl Ether	19.1		0.67	2.50	5.00	ug/Kg
75-09-2	Methylene Chloride	18.5		3.40	8.00	10.0	ug/Kg
156-60-5	trans-1,2-Dichloroethene	19.0		0.84	2.50	5.00	ug/Kg
75-34-3	1,1-Dichloroethane	19.1		0.63	2.50	5.00	ug/Kg
78-93-3	2-Butanone	92.9		5.70	20.0	25.0	ug/Kg
56-23-5	Carbon Tetrachloride	19.0		0.87	2.50	5.00	ug/Kg
156-59-2	cis-1,2-Dichloroethene	19.2		0.61	2.50	5.00	ug/Kg
67-66-3	Chloroform	19.3		0.67	4.00	5.00	ug/Kg
71-55-6	1,1,1-Trichloroethane	19.2		0.78	2.50	5.00	ug/Kg
108-87-2	Methylcyclohexane	18.9		0.87	2.50	5.00	ug/Kg
71-43-2	Benzene	19.3		0.72	2.50	5.00	ug/Kg
107-06-2	1,2-Dichloroethane	19.2		0.61	2.50	5.00	ug/Kg
79-01-6	Trichloroethene	19.2		0.75	2.50	5.00	ug/Kg
78-87-5	1,2-Dichloropropane	19.1		0.66	2.50	5.00	ug/Kg
75-27-4	Bromodichloromethane	19.1		0.56	2.50	5.00	ug/Kg
108-10-1	4-Methyl-2-Pentanone	93.3		4.40	12.5	25.0	ug/Kg
108-88-3	Toluene	19.3		0.67	2.50	5.00	ug/Kg
10061-02-6	t-1,3-Dichloropropene	19.0		0.60	2.50	5.00	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	19.3		0.57	2.50	5.00	ug/Kg
79-00-5	1,1,2-Trichloroethane	19.1		0.84	2.50	5.00	ug/Kg
591-78-6	2-Hexanone	93.8		4.80	12.5	25.0	ug/Kg

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	
Client Sample ID:	VY0226SBS01		SDG No.:	Q1380
Lab Sample ID:	VY0226SBS01		Matrix:	SOIL
Analytical Method:	SW8260		% Solid:	100
Sample Wt/Vol:	5	Units: g	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY021322.D	1		02/26/25 11:22	VY022625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units(Dry Weight)
124-48-1	Dibromochloromethane	19.3		0.65	2.50	5.00	ug/Kg
127-18-4	Tetrachloroethene	19.2		0.89	2.50	5.00	ug/Kg
108-90-7	Chlorobenzene	19.0		0.74	2.50	5.00	ug/Kg
100-41-4	Ethyl Benzene	19.0		0.62	2.50	5.00	ug/Kg
179601-23-1	m/p-Xylenes	38.8		1.40	5.00	10.0	ug/Kg
95-47-6	o-Xylene	19.3		0.70	2.50	5.00	ug/Kg
100-42-5	Styrene	19.3		0.60	2.50	5.00	ug/Kg
75-25-2	Bromoform	19.1		0.81	2.50	5.00	ug/Kg
98-82-8	Isopropylbenzene	19.1		0.67	2.50	5.00	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	18.8		1.10	2.50	5.00	ug/Kg
541-73-1	1,3-Dichlorobenzene	19.1		0.74	2.50	5.00	ug/Kg
106-46-7	1,4-Dichlorobenzene	19.2		0.80	2.50	5.00	ug/Kg
95-50-1	1,2-Dichlorobenzene	18.8		0.59	2.50	5.00	ug/Kg
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	46.7		71 - 136		93%	SPK: 50
1868-53-7	Dibromofluoromethane	47.4		78 - 119		95%	SPK: 50
2037-26-5	Toluene-d8	47.6		85 - 116		95%	SPK: 50
460-00-4	4-Bromofluorobenzene	47.3		79 - 119		95%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	184000	7.707				
540-36-3	1,4-Difluorobenzene	291000	8.615				
3114-55-4	Chlorobenzene-d5	255000	11.414				
3855-82-1	1,4-Dichlorobenzene-d4	129000	13.346				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	
Client Sample ID:	VN0219WBSD01		SDG No.:	Q1380
Lab Sample ID:	VN0219WBSD01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085786.D	1		02/19/25 13:22	VN021925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
74-87-3	Chloromethane	16.5		0.35	0.50	1.00	ug/L
75-01-4	Vinyl Chloride	16.8		0.34	0.75	1.00	ug/L
74-83-9	Bromomethane	17.1		1.40	3.80	5.00	ug/L
75-00-3	Chloroethane	16.2		0.56	0.75	1.00	ug/L
75-69-4	Trichlorofluoromethane	16.7		0.34	0.50	1.00	ug/L
76-13-1	1,1,2-Trichlorotrifluoroethane	17.1		0.25	0.50	1.00	ug/L
75-35-4	1,1-Dichloroethene	17.2		0.26	0.75	1.00	ug/L
67-64-1	Acetone	90.3		1.40	3.80	5.00	ug/L
75-15-0	Carbon Disulfide	15.9		0.32	0.75	1.00	ug/L
1634-04-4	Methyl tert-butyl Ether	19.2		0.16	0.50	1.00	ug/L
75-09-2	Methylene Chloride	17.2		0.32	0.50	1.00	ug/L
156-60-5	trans-1,2-Dichloroethene	16.9		0.25	0.50	1.00	ug/L
75-34-3	1,1-Dichloroethane	17.3		0.23	0.50	1.00	ug/L
78-93-3	2-Butanone	91.4		1.30	2.50	5.00	ug/L
56-23-5	Carbon Tetrachloride	17.0		0.25	0.50	1.00	ug/L
156-59-2	cis-1,2-Dichloroethene	17.6		0.25	0.75	1.00	ug/L
67-66-3	Chloroform	17.5		0.26	0.50	1.00	ug/L
71-55-6	1,1,1-Trichloroethane	17.2		0.19	0.50	1.00	ug/L
108-87-2	Methylcyclohexane	17.3		0.19	0.50	1.00	ug/L
71-43-2	Benzene	17.8		0.16	0.50	1.00	ug/L
107-06-2	1,2-Dichloroethane	18.2		0.24	0.50	1.00	ug/L
79-01-6	Trichloroethene	16.7		0.32	0.75	1.00	ug/L
78-87-5	1,2-Dichloropropane	18.1		0.19	0.50	1.00	ug/L
75-27-4	Bromodichloromethane	18.0		0.24	0.50	1.00	ug/L
108-10-1	4-Methyl-2-Pentanone	96.3		0.75	2.50	5.00	ug/L
108-88-3	Toluene	18.8		0.18	0.50	1.00	ug/L
10061-02-6	t-1,3-Dichloropropene	18.7		0.21	0.50	1.00	ug/L
10061-01-5	cis-1,3-Dichloropropene	18.8		0.18	0.50	1.00	ug/L
79-00-5	1,1,2-Trichloroethane	17.9		0.21	0.50	1.00	ug/L
591-78-6	2-Hexanone	97.6		1.10	2.50	5.00	ug/L

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	
Client Sample ID:	VN0219WBSD01		SDG No.:	Q1380
Lab Sample ID:	VN0219WBSD01		Matrix:	Water
Analytical Method:	SW8260		% Solid:	0
Sample Wt/Vol:	5	Units: mL	Final Vol:	5000 uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1
GC Column:	RXI-624	ID : 0.25	Level :	LOW
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VN085786.D	1		02/19/25 13:22	VN021925

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
124-48-1	Dibromochloromethane	18.9		0.18	0.50	1.00	ug/L
127-18-4	Tetrachloroethene	17.5		0.25	0.50	1.00	ug/L
108-90-7	Chlorobenzene	17.5		0.13	0.50	1.00	ug/L
100-41-4	Ethyl Benzene	17.9		0.16	0.50	1.00	ug/L
179601-23-1	m/p-Xylenes	37.6		0.31	1.00	2.00	ug/L
95-47-6	o-Xylene	18.5		0.14	0.50	1.00	ug/L
100-42-5	Styrene	18.1		0.16	0.50	1.00	ug/L
75-25-2	Bromoform	18.5		0.21	0.50	1.00	ug/L
98-82-8	Isopropylbenzene	18.1		0.13	0.50	1.00	ug/L
79-34-5	1,1,2,2-Tetrachloroethane	17.6		0.27	0.50	1.00	ug/L
541-73-1	1,3-Dichlorobenzene	17.4		0.24	0.50	1.00	ug/L
106-46-7	1,4-Dichlorobenzene	17.1		0.27	0.50	1.00	ug/L
95-50-1	1,2-Dichlorobenzene	17.1		0.19	0.50	1.00	ug/L
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	55.5		81 - 118		111%	SPK: 50
1868-53-7	Dibromofluoromethane	54.7		80 - 119		109%	SPK: 50
2037-26-5	Toluene-d8	55.1		89 - 112		110%	SPK: 50
460-00-4	4-Bromofluorobenzene	56.6		85 - 114		113%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	292000	8.224				
540-36-3	1,4-Difluorobenzene	482000	9.094				
3114-55-4	Chlorobenzene-d5	419000	11.859				
3855-82-1	1,4-Dichlorobenzene-d4	200000	13.788				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	02/10/25	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	02/17/25	
Client Sample ID:	BP-VPB-192-GW-725-727MS		SDG No.:	Q1380	
Lab Sample ID:	Q1380-07MS		Matrix:	SOIL	
Analytical Method:	SW8260		% Solid:	7.4	
Sample Wt/Vol:	5.07	Units: g	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY021342.D	1		02/26/25 19:21	VY022625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units(Dry Weight)
TARGETS							
74-87-3	Chloromethane	520		15.5	33.3	66.6	ug/Kg
75-01-4	Vinyl Chloride	550		10.3	33.3	66.6	ug/Kg
74-83-9	Bromomethane	550		13.7	53.3	66.6	ug/Kg
75-00-3	Chloroethane	580		13.5	33.3	66.6	ug/Kg
75-69-4	Trichlorofluoromethane	570		12.1	53.3	66.6	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	530		14.3	33.3	66.6	ug/Kg
75-35-4	1,1-Dichloroethene	560		10.4	33.3	66.6	ug/Kg
67-64-1	Acetone	2600		83.2	270	330	ug/Kg
75-15-0	Carbon Disulfide	570		17.1	53.3	66.6	ug/Kg
1634-04-4	Methyl tert-butyl Ether	570		8.90	33.3	66.6	ug/Kg
75-09-2	Methylene Chloride	590		45.4	110	130	ug/Kg
156-60-5	trans-1,2-Dichloroethene	560		11.2	33.3	66.6	ug/Kg
75-34-3	1,1-Dichloroethane	580		8.40	33.3	66.6	ug/Kg
78-93-3	2-Butanone	2500		75.7	270	330	ug/Kg
56-23-5	Carbon Tetrachloride	520		11.6	33.3	66.6	ug/Kg
156-59-2	cis-1,2-Dichloroethene	580		8.10	33.3	66.6	ug/Kg
67-66-3	Chloroform	580		8.90	53.3	66.6	ug/Kg
71-55-6	1,1,1-Trichloroethane	570		10.4	33.3	66.6	ug/Kg
108-87-2	Methylcyclohexane	310		11.6	33.3	66.6	ug/Kg
71-43-2	Benzene	540		9.60	33.3	66.6	ug/Kg
107-06-2	1,2-Dichloroethane	550		8.10	33.3	66.6	ug/Kg
79-01-6	Trichloroethene	510		10.0	33.3	66.6	ug/Kg
78-87-5	1,2-Dichloropropane	540		8.80	33.3	66.6	ug/Kg
75-27-4	Bromodichloromethane	560		7.50	33.3	66.6	ug/Kg
108-10-1	4-Methyl-2-Pentanone	2400		58.0	170	330	ug/Kg
108-88-3	Toluene	520		8.90	33.3	66.6	ug/Kg
10061-02-6	t-1,3-Dichloropropene	520		8.00	33.3	66.6	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	520		7.60	33.3	66.6	ug/Kg
79-00-5	1,1,2-Trichloroethane	540		11.2	33.3	66.6	ug/Kg
591-78-6	2-Hexanone	2400		63.8	170	330	ug/Kg

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	02/10/25	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	02/17/25	
Client Sample ID:	BP-VPB-192-GW-725-727MS		SDG No.:	Q1380	
Lab Sample ID:	Q1380-07MS		Matrix:	SOIL	
Analytical Method:	SW8260		% Solid:	7.4	
Sample Wt/Vol:	5.07	Units: g	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY021342.D	1		02/26/25 19:21	VY022625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units(Dry Weight)
124-48-1	Dibromochloromethane	540		8.70	33.3	66.6	ug/Kg
127-18-4	Tetrachloroethene	480		11.9	33.3	66.6	ug/Kg
108-90-7	Chlorobenzene	510		9.90	33.3	66.6	ug/Kg
100-41-4	Ethyl Benzene	500		8.30	33.3	66.6	ug/Kg
179601-23-1	m/p-Xylenes	960		18.0	66.6	130	ug/Kg
95-47-6	o-Xylene	500		9.30	33.3	66.6	ug/Kg
100-42-5	Styrene	340		8.00	33.3	66.6	ug/Kg
75-25-2	Bromoform	560		10.8	33.3	66.6	ug/Kg
98-82-8	Isopropylbenzene	590		8.90	33.3	66.6	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	750		14.7	33.3	66.6	ug/Kg
541-73-1	1,3-Dichlorobenzene	450		9.90	33.3	66.6	ug/Kg
106-46-7	1,4-Dichlorobenzene	460		10.7	33.3	66.6	ug/Kg
95-50-1	1,2-Dichlorobenzene	480		7.90	33.3	66.6	ug/Kg
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	40.9		71 - 136		82%	SPK: 50
1868-53-7	Dibromofluoromethane	39.9		78 - 119		80%	SPK: 50
2037-26-5	Toluene-d8	36.4	*	85 - 116		73%	SPK: 50
460-00-4	4-Bromofluorobenzene	31.4	*	79 - 119		63%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	157000	7.707				
540-36-3	1,4-Difluorobenzene	260000	8.615				
3114-55-4	Chlorobenzene-d5	211000	11.414				
3855-82-1	1,4-Dichlorobenzene-d4	76100	13.346				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	02/10/25	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	02/17/25	
Client Sample ID:	BP-VPB-192-GW-725-727MSD		SDG No.:	Q1380	
Lab Sample ID:	Q1380-08MSD		Matrix:	SOIL	
Analytical Method:	SW8260		% Solid:	7.4	
Sample Wt/Vol:	5.02	Units: g	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY021343.D	1		02/26/25 19:43	VY022625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units(Dry Weight)
TARGETS							
74-87-3	Chloromethane	750		15.6	33.6	67.3	ug/Kg
75-01-4	Vinyl Chloride	800		10.4	33.6	67.3	ug/Kg
74-83-9	Bromomethane	770		13.9	53.8	67.3	ug/Kg
75-00-3	Chloroethane	790		13.6	33.6	67.3	ug/Kg
75-69-4	Trichlorofluoromethane	810		12.2	53.8	67.3	ug/Kg
76-13-1	1,1,2-Trichlorotrifluoroethane	770		14.4	33.6	67.3	ug/Kg
75-35-4	1,1-Dichloroethene	800		10.5	33.6	67.3	ug/Kg
67-64-1	Acetone	3900		84.0	270	340	ug/Kg
75-15-0	Carbon Disulfide	790		17.2	53.8	67.3	ug/Kg
1634-04-4	Methyl tert-butyl Ether	850		9.00	33.6	67.3	ug/Kg
75-09-2	Methylene Chloride	820		45.9	110	130	ug/Kg
156-60-5	trans-1,2-Dichloroethene	790		11.3	33.6	67.3	ug/Kg
75-34-3	1,1-Dichloroethane	800		8.50	33.6	67.3	ug/Kg
78-93-3	2-Butanone	3900		76.5	270	340	ug/Kg
56-23-5	Carbon Tetrachloride	740		11.7	33.6	67.3	ug/Kg
156-59-2	cis-1,2-Dichloroethene	810		8.20	33.6	67.3	ug/Kg
67-66-3	Chloroform	810		9.00	53.8	67.3	ug/Kg
71-55-6	1,1,1-Trichloroethane	800		10.5	33.6	67.3	ug/Kg
108-87-2	Methylcyclohexane	480		11.7	33.6	67.3	ug/Kg
71-43-2	Benzene	770		9.70	33.6	67.3	ug/Kg
107-06-2	1,2-Dichloroethane	810		8.20	33.6	67.3	ug/Kg
79-01-6	Trichloroethene	730		10.1	33.6	67.3	ug/Kg
78-87-5	1,2-Dichloropropane	770		8.90	33.6	67.3	ug/Kg
75-27-4	Bromodichloromethane	790		7.50	33.6	67.3	ug/Kg
108-10-1	4-Methyl-2-Pentanone	3900		58.5	170	340	ug/Kg
108-88-3	Toluene	740		9.00	33.6	67.3	ug/Kg
10061-02-6	t-1,3-Dichloropropene	750		8.10	33.6	67.3	ug/Kg
10061-01-5	cis-1,3-Dichloropropene	740		7.70	33.6	67.3	ug/Kg
79-00-5	1,1,2-Trichloroethane	810		11.3	33.6	67.3	ug/Kg
591-78-6	2-Hexanone	3900		64.5	170	340	ug/Kg

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	02/10/25	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	02/17/25	
Client Sample ID:	BP-VPB-192-GW-725-727MSD		SDG No.:	Q1380	
Lab Sample ID:	Q1380-08MSD		Matrix:	SOIL	
Analytical Method:	SW8260		% Solid:	7.4	
Sample Wt/Vol:	5.02	Units: g	Final Vol:	5000	uL
Soil Aliquot Vol:		uL	Test:	VOCMS Group1	
GC Column:	RXI-624	ID : 0.25	Level :	LOW	
Prep Method :					

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
VY021343.D	1		02/26/25 19:43	VY022625

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units(Dry Weight)
124-48-1	Dibromochloromethane	790		8.70	33.6	67.3	ug/Kg
127-18-4	Tetrachloroethene	680		12.0	33.6	67.3	ug/Kg
108-90-7	Chlorobenzene	720		10.0	33.6	67.3	ug/Kg
100-41-4	Ethyl Benzene	710		8.30	33.6	67.3	ug/Kg
179601-23-1	m/p-Xylenes	1400		18.2	67.3	130	ug/Kg
95-47-6	o-Xylene	710		9.40	33.6	67.3	ug/Kg
100-42-5	Styrene	460		8.10	33.6	67.3	ug/Kg
75-25-2	Bromoform	820		10.9	33.6	67.3	ug/Kg
98-82-8	Isopropylbenzene	860		9.00	33.6	67.3	ug/Kg
79-34-5	1,1,2,2-Tetrachloroethane	1100		14.8	33.6	67.3	ug/Kg
541-73-1	1,3-Dichlorobenzene	660		10.0	33.6	67.3	ug/Kg
106-46-7	1,4-Dichlorobenzene	690		10.8	33.6	67.3	ug/Kg
95-50-1	1,2-Dichlorobenzene	720		7.90	33.6	67.3	ug/Kg
SURROGATES							
17060-07-0	1,2-Dichloroethane-d4	56.3		71 - 136		113%	SPK: 50
1868-53-7	Dibromofluoromethane	55.0		78 - 119		110%	SPK: 50
2037-26-5	Toluene-d8	50.1		85 - 116		100%	SPK: 50
460-00-4	4-Bromofluorobenzene	42.4		79 - 119		85%	SPK: 50
INTERNAL STANDARDS							
363-72-4	Pentafluorobenzene	137000	7.713				
540-36-3	1,4-Difluorobenzene	223000	8.616				
3114-55-4	Chlorobenzene-d5	181000	11.414				
3855-82-1	1,4-Dichlorobenzene-d4	63400	13.347				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG No.: Q1380

Instrument ID: MSVOA_N Calibration Date(s): 02/18/2025 02/18/2025

Heated Purge: (Y/N) N Calibration Time(s): 11:09 14:18

GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:	RRF100 = VN085772.D			RRF050 = VN085773.D		RRF010 = VN085775.D		
	RRF005 = VN085776.D			RRF001 = VN085777.D		RRF020 = VN085779.D		
COMPOUND	RRF100	RRF050	RRF010	RRF005	RRF001	RRF020	RRF	% RSD
Chloromethane	0.647	0.622	0.742	0.662	0.817	0.604	0.682	11.9
Vinyl Chloride	0.698	0.678	0.796	0.729	0.761	0.674	0.723	6.7
Bromomethane	0.434	0.424	0.543	0.496		0.413	0.462	12.1
Chloroethane	0.447	0.417	0.516	0.470	0.584	0.438	0.479	12.9
Trichlorofluoromethane	1.010	0.959	1.134	1.065	1.103	1.006	1.046	6.3
1,1,2-Trichlorotrifluoroethane	0.602	0.564	0.674	0.635	0.554	0.599	0.605	7.4
1,1-Dichloroethene	0.555	0.527	0.591	0.571	0.514	0.531	0.548	5.3
Acetone	0.195	0.184	0.217	0.208	0.227	0.198	0.205	7.7
Carbon Disulfide	1.565	1.473	1.804	1.604	1.848	1.455	1.625	10.2
Methyl tert-butyl Ether	1.747	1.658	1.714	1.560	1.438	1.659	1.629	7
Methylene Chloride	0.628	0.593	0.710	0.666	0.736	0.625	0.660	8.3
trans-1,2-Dichloroethene	0.585	0.549	0.621	0.599	0.623	0.566	0.590	5.1
1,1-Dichloroethane	1.097	1.035	1.216	1.125	1.073	1.089	1.106	5.6
2-Butanone	0.308	0.294	0.331	0.299	0.279	0.299	0.302	5.6
Carbon Tetrachloride	0.573	0.528	0.583	0.545	0.533	0.538	0.550	4.1
cis-1,2-Dichloroethene	0.705	0.661	0.740	0.685	0.646	0.671	0.685	4.9
Chloroform	1.120	1.070	1.269	1.164	1.171	1.141	1.156	5.7
1,1,1-Trichloroethane	1.012	0.967	1.089	1.058	1.030	1.012	1.028	4.1
Methylcyclohexane	0.560	0.492	0.446	0.405	0.373	0.453	0.455	14.5
Benzene	1.581	1.441	1.568	1.421	1.420	1.474	1.484	4.9
1,2-Dichloroethane	0.499	0.456	0.527	0.464	0.472	0.483	0.483	5.4
Trichloroethene	0.370	0.339	0.384	0.350	0.395	0.348	0.364	6.1
1,2-Dichloropropane	0.377	0.347	0.385	0.344	0.339	0.357	0.358	5.2
Bromodichloromethane	0.569	0.518	0.571	0.533	0.512	0.541	0.541	4.6
4-Methyl-2-Pentanone	0.414	0.379	0.392	0.342	0.290	0.388	0.367	12.1
Toluene	0.993	0.902	0.924	0.820	0.689	0.891	0.870	12
t-1,3-Dichloropropene	0.576	0.521	0.518	0.471	0.408	0.505	0.500	11.3
cis-1,3-Dichloropropene	0.626	0.566	0.582	0.513	0.442	0.559	0.548	11.6
1,1,2-Trichloroethane	0.361	0.331	0.370	0.346	0.324	0.344	0.346	5
2-Hexanone	0.296	0.273	0.266	0.227	0.184	0.270	0.253	16

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG No.: Q1380

Instrument ID: MSVOA_N Calibration Date(s): 02/18/2025 02/18/2025

Heated Purge: (Y/N) N Calibration Time(s): 11:09 14:18

GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF100 = VN085772.D		RRF050 = VN085773.D		RRF010 = VN085775.D		
		RRF005 = VN085776.D		RRF001 = VN085777.D		RRF020 = VN085779.D		
COMPOUND	RRF100	RRF050	RRF010	RRF005	RRF001	RRF020	RRF	% RSD
Dibromochloromethane	0.440	0.402	0.431	0.384	0.355	0.405	0.403	7.7
Tetrachloroethene	0.384	0.361	0.420	0.391	0.393	0.375	0.387	5.2
Chlorobenzene	1.183	1.110	1.212	1.139	1.109	1.119	1.145	3.8
Ethyl Benzene	2.105	1.904	1.830	1.665	1.424	1.838	1.794	12.8
m/p-Xylenes	0.809	0.741	0.725	0.610	0.481	0.722	0.682	17.2
o-Xylene	0.771	0.704	0.666	0.584	0.502	0.659	0.648	14.5
Styrene	1.327	1.203	1.060	0.885	0.744	1.133	1.059	20.2
Bromoform	0.327	0.303	0.322	0.301	0.274	0.309	0.306	6.2
Isopropylbenzene	3.851	3.596	3.623	3.067	2.933	3.484	3.425	10.3
1,1,2,2-Tetrachloroethane	1.073	1.053	1.304	1.218	1.271	1.154	1.179	8.8
1,3-Dichlorobenzene	1.772	1.635	1.876	1.697	1.748	1.675	1.734	4.9
1,4-Dichlorobenzene	1.750	1.647	1.884	1.777	1.921	1.684	1.777	6.1
1,2-Dichlorobenzene	1.660	1.574	1.768	1.642	1.744	1.617	1.667	4.5
1,2-Dichloroethane-d4	0.698	0.613	0.588	0.767		0.575	0.648	12.6
Dibromofluoromethane	0.375	0.317	0.294	0.357		0.293	0.327	11.5
Toluene-d8	1.444	1.199	1.017	1.212		1.080	1.190	13.8
4-Bromofluorobenzene	0.503	0.419	0.323	0.370		0.346	0.392	18.2

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>CHEMTECH</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1380</u>
Instrument ID:	<u>MSVOA_Y</u>	SAS No.:	<u>Q1380</u>
Heated Purge: (Y/N)	<u>Y</u>	SDG No.:	<u>Q1380</u>
GC Column:	<u>RXI-624</u>	ID:	<u>0.25</u> (mm)
		Calibration Date(s):	<u>02/03/2025</u>
		Calibration Time(s):	<u>10:35</u>

LAB FILE ID:	RRF005 = VY021020.D	RRF010 = VY021021.D	RRF020 = VY021022.D	RRF050 = VY021023.D	RRF100 = VY021024.D	RRF150 = VY021025.D		
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
Chloromethane	0.524	0.445	0.408	0.421	0.404	0.437	0.440	10.1
Vinyl Chloride	0.535	0.441	0.404	0.424	0.408	0.443	0.443	10.9
Bromomethane	0.356	0.275	0.249	0.258	0.246	0.257	0.274	15.2
Chloroethane	0.316	0.270	0.240	0.255	0.239	0.261	0.263	10.7
Trichlorofluoromethane	0.924	0.814	0.755	0.789	0.743	0.813	0.806	8
1,1,2-Trichlorotrifluoroethane	0.587	0.507	0.472	0.494	0.466	0.510	0.506	8.6
1,1-Dichloroethene	0.529	0.471	0.441	0.469	0.452	0.498	0.477	6.8
Acetone	0.087	0.083	0.078	0.077	0.070	0.083	0.080	7.5
Carbon Disulfide	1.673	1.456	1.389	1.475	1.421	1.556	1.495	7
Methyl tert-butyl Ether	1.271	1.185	1.168	1.216	1.124	1.265	1.205	4.8
Methylene Chloride	0.570	0.506	0.473	0.485	0.452	0.494	0.497	8.2
trans-1,2-Dichloroethene	0.578	0.518	0.502	0.522	0.490	0.538	0.525	5.8
1,1-Dichloroethane	1.104	0.948	0.923	0.947	0.911	0.989	0.970	7.3
2-Butanone	0.134	0.133	0.126	0.135	0.121	0.139	0.131	5.2
Carbon Tetrachloride	0.604	0.551	0.522	0.547	0.522	0.581	0.555	5.9
cis-1,2-Dichloroethene	0.662	0.584	0.577	0.599	0.575	0.626	0.604	5.7
Chloroform	1.107	1.016	0.947	0.982	0.921	1.006	0.997	6.5
1,1,1-Trichloroethane	1.067	0.901	0.867	0.912	0.863	0.949	0.927	8.2
Methylcyclohexane	0.609	0.550	0.536	0.588	0.570	0.629	0.581	6.1
Benzene	1.512	1.382	1.327	1.378	1.323	1.436	1.393	5.1
1,2-Dichloroethane	0.435	0.390	0.367	0.379	0.352	0.395	0.386	7.4
Trichloroethene	0.401	0.355	0.341	0.349	0.338	0.375	0.360	6.7
1,2-Dichloropropane	0.357	0.329	0.319	0.327	0.311	0.341	0.331	5
Bromodichloromethane	0.528	0.500	0.469	0.482	0.459	0.506	0.491	5.2
4-Methyl-2-Pentanone	0.202	0.211	0.202	0.216	0.194	0.224	0.208	5.2
Toluene	0.913	0.856	0.840	0.876	0.844	0.918	0.875	3.9
t-1,3-Dichloropropene	0.441	0.426	0.419	0.449	0.427	0.483	0.441	5.3
cis-1,3-Dichloropropene	0.528	0.516	0.502	0.525	0.501	0.560	0.522	4.2
1,1,2-Trichloroethane	0.261	0.235	0.231	0.235	0.217	0.242	0.237	6.1
2-Hexanone	0.124	0.131	0.128	0.144	0.129	0.149	0.134	7.4

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: <u>CHEMTECH</u>	Contract: <u>TETR06</u>
Lab Code: <u>CHEM</u> Case No.: <u>Q1380</u>	SAS No.: <u>Q1380</u> SDG No.: <u>Q1380</u>
Instrument ID: <u>MSVOA_Y</u>	Calibration Date(s): <u>02/03/2025</u> <u>02/03/2025</u>
Heated Purge: (Y/N) <u>Y</u>	Calibration Time(s): <u>10:35</u> <u>12:44</u>
GC Column: <u>RXI-624</u> ID: <u>0.25</u> (mm)	

LAB FILE ID:		RRF005 = VY021020.D		RRF010 = VY021021.D		RRF020 = VY021022.D		
		RRF050 = VY021023.D		RRF100 = VY021024.D		RRF150 = VY021025.D		
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF100	RRF150	RRF	% RSD
Dibromochloromethane	0.359	0.330	0.324	0.329	0.310	0.346	0.333	5.2
Tetrachloroethene	0.409	0.367	0.352	0.361	0.351	0.387	0.371	6.2
Chlorobenzene	1.237	1.115	1.054	1.080	1.048	1.146	1.113	6.4
Ethyl Benzene	2.101	1.899	1.865	1.943	1.914	2.093	1.969	5.2
m/p-Xylenes	0.772	0.723	0.709	0.733	0.711	0.773	0.737	3.9
o-Xylene	0.723	0.658	0.660	0.681	0.670	0.729	0.687	4.6
Styrene	1.154	1.118	1.105	1.156	1.128	1.208	1.145	3.2
Bromoform	0.231	0.222	0.217	0.216	0.205	0.229	0.220	4.4
Isopropylbenzene	4.122	3.695	3.673	3.809	3.871	4.263	3.906	6.1
1,1,2,2-Tetrachloroethane	0.707	0.646	0.635	0.632	0.597	0.680	0.650	6
1,3-Dichlorobenzene	1.937	1.686	1.660	1.683	1.652	1.824	1.740	6.6
1,4-Dichlorobenzene	1.906	1.702	1.632	1.665	1.599	1.778	1.714	6.6
1,2-Dichlorobenzene	1.668	1.524	1.440	1.460	1.405	1.567	1.510	6.4
1,2-Dichloroethane-d4	0.566	0.557	0.478	0.503	0.454	0.516	0.512	8.5
Dibromofluoromethane	0.351	0.326	0.301	0.313	0.297	0.333	0.320	6.4
Toluene-d8	1.295	1.237	1.139	1.202	1.158	1.288	1.220	5.3
4-Bromofluorobenzene	0.417	0.408	0.371	0.403	0.375	0.418	0.399	5.2

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name:	<u>CHEMTECH</u>	Contract:	<u>TETR06</u>
Lab Code:	<u>CHEM</u>	Case No.:	<u>Q1380</u>
Instrument ID:	<u>MSVOA_Y</u>	SAS No.:	<u>Q1380</u>
Heated Purge: (Y/N)	<u>Y</u>	SDG No.:	<u>Q1380</u>
GC Column:	<u>RXI-624</u>	Calibration Date(s):	<u>02/25/2025</u>
ID:	<u>0.25 (mm)</u>	Calibration Time(s):	<u>12:40</u>
			<u>16:49</u>

LAB FILE ID:	RRF005 = VY021309.D	RRF010 = VY021310.D	RRF020 = VY021311.D	RRF050 = VY021312.D	RRF150 = VY021314.D	RRF100 = VY021316.D		
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF150	RRF100	RRF	% RSD
Chloromethane	0.709	0.628	0.566	0.602	0.597	0.570	0.612	8.6
Vinyl Chloride	0.648	0.591	0.554	0.627	0.606	0.580	0.601	5.6
Bromomethane	0.505	0.414	0.361	0.387	0.371	0.359	0.399	13.9
Chloroethane	0.414	0.365	0.347	0.390	0.365	0.355	0.373	6.7
Trichlorofluoromethane	0.976	0.915	0.838	0.936	0.909	0.870	0.908	5.3
1,1,2-Trichlorotrifluoroethane	0.546	0.529	0.491	0.537	0.541	0.508	0.525	4.1
1,1-Dichloroethene	0.545	0.489	0.462	0.515	0.518	0.490	0.503	5.7
Acetone	0.122	0.102	0.094	0.128	0.104	0.103	0.109	11.9
Carbon Disulfide	1.800	1.652	1.540	1.734	1.678	1.620	1.671	5.4
Methyl tert-butyl Ether	1.312	1.268	1.267	1.410	1.410	1.307	1.329	4.9
Methylene Chloride	0.733	0.598	0.539	0.585	0.547	0.529	0.588	12.8
trans-1,2-Dichloroethene	0.589	0.536	0.507	0.573	0.570	0.547	0.554	5.4
1,1-Dichloroethane	1.101	1.032	0.961	1.083	1.058	1.007	1.040	4.9
2-Butanone	0.162	0.155	0.156	0.183	0.171	0.159	0.164	6.5
Carbon Tetrachloride	0.567	0.562	0.509	0.581	0.567	0.535	0.554	4.8
cis-1,2-Dichloroethene	0.649	0.610	0.588	0.665	0.656	0.626	0.632	4.7
Chloroform	1.128	1.046	0.977	1.095	1.062	1.017	1.054	5.1
1,1,1-Trichloroethane	1.020	0.956	0.889	0.986	0.979	0.927	0.960	4.8
Methylcyclohexane	0.597	0.613	0.584	0.662	0.663	0.618	0.623	5.3
Benzene	1.515	1.445	1.359	1.533	1.472	1.409	1.456	4.5
1,2-Dichloroethane	0.422	0.417	0.393	0.434	0.419	0.394	0.413	3.9
Trichloroethene	0.375	0.360	0.341	0.382	0.375	0.354	0.364	4.3
1,2-Dichloropropane	0.360	0.350	0.329	0.371	0.357	0.336	0.350	4.5
Bromodichloromethane	0.519	0.503	0.477	0.541	0.528	0.495	0.511	4.5
4-Methyl-2-Pentanone	0.220	0.227	0.239	0.267	0.264	0.238	0.243	7.9
Toluene	0.887	0.888	0.852	0.965	0.936	0.897	0.904	4.4
t-1,3-Dichloropropene	0.438	0.436	0.434	0.515	0.509	0.474	0.468	8
cis-1,3-Dichloropropene	0.541	0.526	0.528	0.591	0.581	0.546	0.552	5
1,1,2-Trichloroethane	0.250	0.251	0.238	0.267	0.254	0.237	0.249	4.4
2-Hexanone	0.133	0.146	0.155	0.183	0.178	0.162	0.159	11.9

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG No.: Q1380

Instrument ID: MSVOA_Y Calibration Date(s): 02/25/2025 02/25/2025

Heated Purge: (Y/N) Y Calibration Time(s): 12:40 16:49

GC Column: RXI-624 ID: 0.25 (mm)

LAB FILE ID:		RRF005 = VY021309.D		RRF010 = VY021310.D		RRF020 = VY021311.D		
		RRF050 = VY021312.D		RRF150 = VY021314.D		RRF100 = VY021316.D		
COMPOUND	RRF005	RRF010	RRF020	RRF050	RRF150	RRF100	RRF	% RSD
Dibromochloromethane	0.340	0.325	0.322	0.368	0.352	0.332	0.340	5.2
Tetrachloroethene	0.405	0.380	0.351	0.389	0.386	0.355	0.378	5.5
Chlorobenzene	1.171	1.111	1.039	1.176	1.159	1.078	1.122	5
Ethyl Benzene	1.938	1.941	1.826	2.130	2.115	1.979	1.988	5.8
m/p-Xylenes	0.735	0.738	0.696	0.802	0.783	0.733	0.748	5.1
o-Xylene	0.689	0.675	0.648	0.758	0.739	0.694	0.701	5.9
Styrene	1.081	1.128	1.091	1.277	1.257	1.159	1.166	7.2
Bromoform	0.221	0.221	0.211	0.246	0.236	0.214	0.225	5.9
Isopropylbenzene	3.744	3.605	3.437	3.976	4.009	3.783	3.759	5.8
1,1,2,2-Tetrachloroethane	0.700	0.649	0.621	0.697	0.689	0.624	0.663	5.5
1,3-Dichlorobenzene	1.812	1.729	1.598	1.792	1.777	1.668	1.730	4.8
1,4-Dichlorobenzene	1.825	1.710	1.593	1.751	1.744	1.629	1.709	5
1,2-Dichlorobenzene	1.599	1.499	1.408	1.575	1.564	1.446	1.515	5.1
1,2-Dichloroethane-d4	0.681	0.510	0.548	0.551	0.566	0.547	0.567	10.3
Dibromofluoromethane	0.370	0.299	0.317	0.325	0.334	0.322	0.328	7.2
Toluene-d8	1.420	1.158	1.230	1.251	1.290	1.253	1.267	6.8
4-Bromofluorobenzene	0.490	0.387	0.404	0.423	0.435	0.420	0.426	8.2

* Compounds with required minimum RRF and maximum %RSD values.
All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG No.: Q1380

Instrument ID: MSVOA_N Calibration Date/Time: 02/19/2025 10:51

Lab File ID: VN085782.D Init. Calib. Date(s): 02/18/2025 02/18/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:09 14:18

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.682	0.646	0.1	-5.28	20
Vinyl Chloride	0.723	0.697		-3.6	20
Bromomethane	0.462	0.446		-3.46	20
Chloroethane	0.479	0.437		-8.77	20
Trichlorofluoromethane	1.046	1.027		-1.82	20
1,1,2-Trichlorotrifluoroethane	0.605	0.611		0.99	20
1,1-Dichloroethene	0.548	0.559		2.01	20
Acetone	0.205	0.226		10.24	20
Carbon Disulfide	1.625	1.544		-4.99	20
Methyl tert-butyl Ether	1.629	1.743		7	20
Methylene Chloride	0.660	0.628		-4.85	20
trans-1,2-Dichloroethene	0.590	0.594		0.68	20
1,1-Dichloroethane	1.106	1.094	0.1	-1.09	20
2-Butanone	0.302	0.308		1.99	20
Carbon Tetrachloride	0.550	0.568		3.27	20
cis-1,2-Dichloroethene	0.685	0.703		2.63	20
Chloroform	1.156	1.135		-1.82	20
1,1,1-Trichloroethane	1.028	1.025		-0.29	20
Methylcyclohexane	0.455	0.542		19.12	20
Benzene	1.484	1.556		4.85	20
1,2-Dichloroethane	0.483	0.496		2.69	20
Trichloroethene	0.364	0.377		3.57	20
1,2-Dichloropropane	0.358	0.379		5.87	20
Bromodichloromethane	0.541	0.561		3.7	20
4-Methyl-2-Pentanone	0.367	0.383		4.36	20
Toluene	0.870	0.975		12.07	20
t-1,3-Dichloropropene	0.500	0.556		11.2	20
cis-1,3-Dichloropropene	0.548	0.612		11.68	20
1,1,2-Trichloroethane	0.346	0.356		2.89	20
2-Hexanone	0.253	0.281		11.07	20
Dibromochloromethane	0.403	0.435		7.94	20
Tetrachloroethene	0.387	0.415		7.24	20
Chlorobenzene	1.145	1.191	0.3	4.02	20
Ethyl Benzene	1.794	2.061		14.88	20
m/p-Xylenes	0.682	0.811		18.92	20
o-Xylene	0.648	0.758		16.98	20
Styrene	1.059	1.287		21.53	20
Bromoform	0.306	0.323	0.1	5.56	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG No.: Q1380

Instrument ID: MSVOA_N Calibration Date/Time: 02/19/2025 10:51

Lab File ID: VN085782.D Init. Calib. Date(s): 02/18/2025 02/18/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:09 14:18

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	3.425	3.921		14.48	20
1,1,2,2-Tetrachloroethane	1.179	1.058	0.3	-10.26	20
1,3-Dichlorobenzene	1.734	1.791		3.29	20
1,4-Dichlorobenzene	1.777	1.777		0	20
1,2-Dichlorobenzene	1.667	1.704		2.22	20
1,2-Dichloroethane-d4	0.648	0.646		-0.31	20
Dibromofluoromethane	0.327	0.346		5.81	20
Toluene-d8	1.190	1.302		9.41	20
4-Bromofluorobenzene	0.392	0.446		13.78	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG No.: Q1380

Instrument ID: MSVOA_N Calibration Date/Time: 02/19/2025 15:58

Lab File ID: VN085792.D Init. Calib. Date(s): 02/18/2025 02/18/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:09 14:18

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.682	0.587	0.1	-13.93	50
Vinyl Chloride	0.723	0.619		-14.39	50
Bromomethane	0.462	0.383		-17.1	50
Chloroethane	0.479	0.387		-19.21	50
Trichlorofluoromethane	1.046	0.878		-16.06	50
1,1,2-Trichlorotrifluoroethane	0.605	0.528		-12.73	50
1,1-Dichloroethene	0.548	0.485		-11.5	50
Acetone	0.205	0.209		1.95	50
Carbon Disulfide	1.625	1.344		-17.29	50
Methyl tert-butyl Ether	1.629	1.563		-4.05	50
Methylene Chloride	0.660	0.553		-16.21	50
trans-1,2-Dichloroethene	0.590	0.517		-12.37	50
1,1-Dichloroethane	1.106	0.983	0.1	-11.12	50
2-Butanone	0.302	0.305		0.99	50
Carbon Tetrachloride	0.550	0.492		-10.55	50
cis-1,2-Dichloroethene	0.685	0.618		-9.78	50
Chloroform	1.156	1.018		-11.94	50
1,1,1-Trichloroethane	1.028	0.904		-12.06	50
Methylcyclohexane	0.455	0.466		2.42	50
Benzene	1.484	1.377		-7.21	50
1,2-Dichloroethane	0.483	0.446		-7.66	50
Trichloroethene	0.364	0.323		-11.26	50
1,2-Dichloropropane	0.358	0.334		-6.7	50
Bromodichloromethane	0.541	0.500		-7.58	50
4-Methyl-2-Pentanone	0.367	0.380		3.54	50
Toluene	0.870	0.868		-0.23	50
t-1,3-Dichloropropene	0.500	0.505		1	50
cis-1,3-Dichloropropene	0.548	0.542		-1.1	50
1,1,2-Trichloroethane	0.346	0.323		-6.65	50
2-Hexanone	0.253	0.280		10.67	50
Dibromochloromethane	0.403	0.389		-3.47	50
Tetrachloroethene	0.387	0.337		-12.92	50
Chlorobenzene	1.145	1.033	0.3	-9.78	50
Ethyl Benzene	1.794	1.793		-0.06	50
m/p-Xylenes	0.682	0.704		3.23	50
o-Xylene	0.648	0.665		2.62	50
Styrene	1.059	1.140		7.65	50
Bromoform	0.306	0.290	0.1	-5.23	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG No.: Q1380

Instrument ID: MSVOA_N Calibration Date/Time: 02/19/2025 15:58

Lab File ID: VN085792.D Init. Calib. Date(s): 02/18/2025 02/18/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:09 14:18

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	3.425	3.305		-3.5	50
1,1,2,2-Tetrachloroethane	1.179	0.989	0.3	-16.11	50
1,3-Dichlorobenzene	1.734	1.516		-12.57	50
1,4-Dichlorobenzene	1.777	1.504		-15.36	50
1,2-Dichlorobenzene	1.667	1.463		-12.24	50
1,2-Dichloroethane-d4	0.648	0.661		2.01	50
Dibromofluoromethane	0.327	0.345		5.51	50
Toluene-d8	1.190	1.326		11.43	50
4-Bromofluorobenzene	0.392	0.465		18.62	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG No.: Q1380

Instrument ID: MSVOA_Y Calibration Date/Time: 02/19/2025 14:55

Lab File ID: VY021236.D Init. Calib. Date(s): 02/03/2025 02/03/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 10:35 12:44

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.440	0.431	0.1	-2.05	20
Vinyl Chloride	0.443	0.466		5.19	20
Bromomethane	0.274	0.268		-2.19	20
Chloroethane	0.263	0.286		8.74	20
Trichlorofluoromethane	0.806	0.806		0	20
1,1,2-Trichlorotrifluoroethane	0.506	0.489		-3.36	20
1,1-Dichloroethene	0.477	0.455		-4.61	20
Acetone	0.080	0.083		3.75	20
Carbon Disulfide	1.495	1.273		-14.85	20
Methyl tert-butyl Ether	1.205	1.247		3.48	20
Methylene Chloride	0.497	0.494		-0.6	20
trans-1,2-Dichloroethene	0.525	0.503		-4.19	20
1,1-Dichloroethane	0.970	0.977	0.1	0.72	20
2-Butanone	0.131	0.135		3.05	20
Carbon Tetrachloride	0.555	0.538		-3.06	20
cis-1,2-Dichloroethene	0.604	0.603		-0.17	20
Chloroform	0.997	1.001		0.4	20
1,1,1-Trichloroethane	0.927	0.926		-0.11	20
Methylcyclohexane	0.581	0.580		-0.17	20
Benzene	1.393	1.364		-2.08	20
1,2-Dichloroethane	0.386	0.384		-0.52	20
Trichloroethene	0.360	0.351		-2.5	20
1,2-Dichloropropane	0.331	0.339		2.42	20
Bromodichloromethane	0.491	0.497		1.22	20
4-Methyl-2-Pentanone	0.208	0.218		4.81	20
Toluene	0.875	0.869		-0.69	20
t-1,3-Dichloropropene	0.441	0.461		4.53	20
cis-1,3-Dichloropropene	0.522	0.538		3.07	20
1,1,2-Trichloroethane	0.237	0.237		0	20
2-Hexanone	0.134	0.144		7.46	20
Dibromochloromethane	0.333	0.339		1.8	20
Tetrachloroethene	0.371	0.356		-4.04	20
Chlorobenzene	1.113	1.109	0.3	-0.36	20
Ethyl Benzene	1.969	1.992		1.17	20
m/p-Xylenes	0.737	0.734		-0.41	20
o-Xylene	0.687	0.708		3.06	20
Styrene	1.145	1.182		3.23	20
Bromoform	0.220	0.219	0.1	-0.46	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG No.: Q1380

Instrument ID: MSVOA_Y Calibration Date/Time: 02/19/2025 14:55

Lab File ID: VY021236.D Init. Calib. Date(s): 02/03/2025 02/03/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 10:35 12:44

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	3.906	4.141		6.02	20
1,1,2,2-Tetrachloroethane	0.650	0.643	0.3	-1.08	20
1,3-Dichlorobenzene	1.740	1.736		-0.23	20
1,4-Dichlorobenzene	1.714	1.688		-1.52	20
1,2-Dichlorobenzene	1.510	1.502		-0.53	20
1,2-Dichloroethane-d4	0.512	0.470		-8.2	20
Dibromofluoromethane	0.320	0.299		-6.56	20
Toluene-d8	1.220	1.094		-10.33	20
4-Bromofluorobenzene	0.399	0.376		-5.76	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG No.: Q1380

Instrument ID: MSVOA_Y Calibration Date/Time: 02/19/2025 18:32

Lab File ID: VY021245.D Init. Calib. Date(s): 02/03/2025 02/03/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 10:35 12:44

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.440	0.469	0.1	6.59	50
Vinyl Chloride	0.443	0.496		11.96	50
Bromomethane	0.274	0.311		13.5	50
Chloroethane	0.263	0.322		22.43	50
Trichlorofluoromethane	0.806	0.844		4.72	50
1,1,2-Trichlorotrifluoroethane	0.506	0.518		2.37	50
1,1-Dichloroethene	0.477	0.484		1.47	50
Acetone	0.080	0.095		18.75	50
Carbon Disulfide	1.495	1.357		-9.23	50
Methyl tert-butyl Ether	1.205	1.416		17.51	50
Methylene Chloride	0.497	0.568		14.29	50
trans-1,2-Dichloroethene	0.525	0.540		2.86	50
1,1-Dichloroethane	0.970	1.082	0.1	11.55	50
2-Butanone	0.131	0.161		22.9	50
Carbon Tetrachloride	0.555	0.556		0.18	50
cis-1,2-Dichloroethene	0.604	0.666		10.27	50
Chloroform	0.997	1.114		11.73	50
1,1,1-Trichloroethane	0.927	0.983		6.04	50
Methylcyclohexane	0.581	0.577		-0.69	50
Benzene	1.393	1.472		5.67	50
1,2-Dichloroethane	0.386	0.422		9.33	50
Trichloroethene	0.360	0.362		0.56	50
1,2-Dichloropropane	0.331	0.367		10.88	50
Bromodichloromethane	0.491	0.541		10.18	50
4-Methyl-2-Pentanone	0.208	0.254		22.11	50
Toluene	0.875	0.924		5.6	50
t-1,3-Dichloropropene	0.441	0.497		12.7	50
cis-1,3-Dichloropropene	0.522	0.577		10.54	50
1,1,2-Trichloroethane	0.237	0.265		11.81	50
2-Hexanone	0.134	0.169		26.12	50
Dibromochloromethane	0.333	0.364		9.31	50
Tetrachloroethene	0.371	0.361		-2.69	50
Chlorobenzene	1.113	1.160	0.3	4.22	50
Ethyl Benzene	1.969	2.074		5.33	50
m/p-Xylenes	0.737	0.768		4.21	50
o-Xylene	0.687	0.734		6.84	50
Styrene	1.145	1.245		8.73	50
Bromoform	0.220	0.242	0.1	10	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG No.: Q1380

Instrument ID: MSVOA_Y Calibration Date/Time: 02/19/2025 18:32

Lab File ID: VY021245.D Init. Calib. Date(s): 02/03/2025 02/03/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 10:35 12:44

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	3.906	3.853		-1.36	50
1,1,2,2-Tetrachloroethane	0.650	0.684	0.3	5.23	50
1,3-Dichlorobenzene	1.740	1.731		-0.52	50
1,4-Dichlorobenzene	1.714	1.677		-2.16	50
1,2-Dichlorobenzene	1.510	1.520		0.66	50
1,2-Dichloroethane-d4	0.512	0.533		4.1	50
Dibromofluoromethane	0.320	0.323		0.94	50
Toluene-d8	1.220	1.195		-2.05	50
4-Bromofluorobenzene	0.399	0.423		6.01	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG No.: Q1380

Instrument ID: MSVOA_Y Calibration Date/Time: 02/26/2025 10:11

Lab File ID: VY021320.D Init. Calib. Date(s): 02/25/2025 02/25/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 12:40 16:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.612	0.581	0.1	-5.07	20
Vinyl Chloride	0.601	0.586		-2.5	20
Bromomethane	0.399	0.370		-7.27	20
Chloroethane	0.373	0.363		-2.68	20
Trichlorofluoromethane	0.908	0.866		-4.63	20
1,1,2-Trichlorotrifluoroethane	0.525	0.496		-5.52	20
1,1-Dichloroethene	0.503	0.487		-3.18	20
Acetone	0.109	0.113		3.67	20
Carbon Disulfide	1.671	1.628		-2.57	20
Methyl tert-butyl Ether	1.329	1.308		-1.58	20
Methylene Chloride	0.588	0.533		-9.35	20
trans-1,2-Dichloroethene	0.554	0.540		-2.53	20
1,1-Dichloroethane	1.040	1.021	0.1	-1.83	20
2-Butanone	0.164	0.162		-1.22	20
Carbon Tetrachloride	0.554	0.536		-3.25	20
cis-1,2-Dichloroethene	0.632	0.633		0.16	20
Chloroform	1.054	1.038		-1.52	20
1,1,1-Trichloroethane	0.960	0.936		-2.5	20
Methylcyclohexane	0.623	0.578		-7.22	20
Benzene	1.456	1.415		-2.82	20
1,2-Dichloroethane	0.413	0.396		-4.12	20
Trichloroethene	0.364	0.355		-2.47	20
1,2-Dichloropropane	0.350	0.341		-2.57	20
Bromodichloromethane	0.511	0.494		-3.33	20
4-Methyl-2-Pentanone	0.243	0.234		-3.7	20
Toluene	0.904	0.892		-1.33	20
t-1,3-Dichloropropene	0.468	0.468		0	20
cis-1,3-Dichloropropene	0.552	0.543		-1.63	20
1,1,2-Trichloroethane	0.249	0.236		-5.22	20
2-Hexanone	0.159	0.159		0	20
Dibromochloromethane	0.340	0.332		-2.35	20
Tetrachloroethene	0.378	0.359		-5.03	20
Chlorobenzene	1.122	1.091	0.3	-2.76	20
Ethyl Benzene	1.988	1.972		-0.81	20
m/p-Xylenes	0.748	0.735		-1.74	20
o-Xylene	0.701	0.696		-0.71	20
Styrene	1.166	1.176		0.86	20
Bromoform	0.225	0.217	0.1	-3.56	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG No.: Q1380

Instrument ID: MSVOA_Y Calibration Date/Time: 02/26/2025 10:11

Lab File ID: VY021320.D Init. Calib. Date(s): 02/25/2025 02/25/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 12:40 16:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	3.759	3.742		-0.45	20
1,1,2,2-Tetrachloroethane	0.663	0.639	0.3	-3.62	20
1,3-Dichlorobenzene	1.730	1.676		-3.12	20
1,4-Dichlorobenzene	1.709	1.633		-4.45	20
1,2-Dichlorobenzene	1.515	1.469		-3.04	20
1,2-Dichloroethane-d4	0.567	0.544		-4.06	20
Dibromofluoromethane	0.328	0.320		-2.44	20
Toluene-d8	1.267	1.257		-0.79	20
4-Bromofluorobenzene	0.426	0.420		-1.41	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG No.: Q1380

Instrument ID: MSVOA_Y Calibration Date/Time: 02/26/2025 20:06

Lab File ID: VY021344.D Init. Calib. Date(s): 02/25/2025 02/25/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 12:40 16:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.612	0.566	0.1	-7.52	50
Vinyl Chloride	0.601	0.577		-3.99	50
Bromomethane	0.399	0.382		-4.26	50
Chloroethane	0.373	0.374		0.27	50
Trichlorofluoromethane	0.908	0.871		-4.07	50
1,1,2-Trichlorotrifluoroethane	0.525	0.494		-5.91	50
1,1-Dichloroethene	0.503	0.489		-2.78	50
Acetone	0.109	0.098		-10.09	50
Carbon Disulfide	1.671	1.626		-2.69	50
Methyl tert-butyl Ether	1.329	1.439		8.28	50
Methylene Chloride	0.588	0.567		-3.57	50
trans-1,2-Dichloroethene	0.554	0.566		2.17	50
1,1-Dichloroethane	1.040	1.051	0.1	1.06	50
2-Butanone	0.164	0.166		1.22	50
Carbon Tetrachloride	0.554	0.520		-6.14	50
cis-1,2-Dichloroethene	0.632	0.663		4.91	50
Chloroform	1.054	1.085		2.94	50
1,1,1-Trichloroethane	0.960	0.958		-0.21	50
Methylcyclohexane	0.623	0.557		-10.59	50
Benzene	1.456	1.421		-2.4	50
1,2-Dichloroethane	0.413	0.419		1.45	50
Trichloroethene	0.364	0.352		-3.3	50
1,2-Dichloropropane	0.350	0.341		-2.57	50
Bromodichloromethane	0.511	0.511		0	50
4-Methyl-2-Pentanone	0.243	0.248		2.06	50
Toluene	0.904	0.897		-0.77	50
t-1,3-Dichloropropene	0.468	0.465		-0.64	50
cis-1,3-Dichloropropene	0.552	0.534		-3.26	50
1,1,2-Trichloroethane	0.249	0.254		2.01	50
2-Hexanone	0.159	0.166		4.4	50
Dibromochloromethane	0.340	0.345		1.47	50
Tetrachloroethene	0.378	0.354		-6.35	50
Chlorobenzene	1.122	1.085	0.3	-3.3	50
Ethyl Benzene	1.988	1.913		-3.77	50
m/p-Xylenes	0.748	0.722		-3.48	50
o-Xylene	0.701	0.694		-1	50
Styrene	1.166	1.170		0.26	50
Bromoform	0.225	0.225	0.1	0	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG No.: Q1380

Instrument ID: MSVOA_Y Calibration Date/Time: 02/26/2025 20:06

Lab File ID: VY021344.D Init. Calib. Date(s): 02/25/2025 02/25/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 12:40 16:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	3.759	3.702		-1.52	50
1,1,2,2-Tetrachloroethane	0.663	0.666	0.3	0.45	50
1,3-Dichlorobenzene	1.730	1.669		-3.53	50
1,4-Dichlorobenzene	1.709	1.629		-4.68	50
1,2-Dichlorobenzene	1.515	1.475		-2.64	50
1,2-Dichloroethane-d4	0.567	0.529		-6.7	50
Dibromofluoromethane	0.328	0.298		-9.15	50
Toluene-d8	1.267	1.131		-10.73	50
4-Bromofluorobenzene	0.426	0.383		-10.09	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

LAB CHRONICLE

OrderID:	Q1380	OrderDate:	2/17/2025 3:59:00 PM
Client:	Tetra Tech NUS, Inc.	Project:	NWIRP Bethpage CTO WE13 - VPB-192
Contact:	Ernie Wu	Location:	#12 COC Ref. #2 Soil, VOA Ref. #3 Water

LabID	ClientID	Matrix	Test	Method	Sample Date	Prep Date	Anal Date	Received
Q1380-01	BP-VPB-192-EB-2025 0212	Water			02/12/25			02/17/25
			SVOC-SIMGroup1	8270-Modified		02/18/25	02/20/25	
Q1380-09	BP-VPB-192-GW-780- 782	Water			02/13/25			02/17/25
			SVOC-SIMGroup1	8270-Modified		02/18/25	02/20/25	
Q1380-12	VPB192-HYD-202502 14	Water			02/14/25			02/17/25
			SVOC-SIMGroup1	8270-Modified		02/18/25	02/20/25	



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Hit Summary Sheet
SW-846

SDG No.: Q1380
Client: Tetra Tech NUS, Inc.

Sample ID	Client ID	Matrix	Parameter	Concentration	C	MDL	LOD	RDL	Units
Client ID :				0.000					
Total Svoc :					0.00				
Total Concentration:					0.00				



SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	02/12/25
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005	Date Received:	02/17/25
Client Sample ID:	BP-VPB-192-EB-20250212	SDG No.:	Q1380
Lab Sample ID:	Q1380-01	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	890 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036475.D	1	02/18/25 11:40	02/20/25 11:24	PB166758

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.22	U	0.080	0.22	0.22	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.31		30 - 150		76%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.43		30 - 150		108%	SPK: 0.4
367-12-4	2-Fluorophenol	0.14		20 - 110		34%	SPK: 0.4
13127-88-3	Phenol-d6	0.078		10 - 160		20%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.30		55 - 111		76%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.38		53 - 106		96%	SPK: 0.4
118-79-6	2,4,6-Tribromophenol	0.35		46 - 122		87%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.58	*	58 - 132		146%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1490		7.746			
1146-65-2	Naphthalene-d8	3320		10.541			
15067-26-2	Acenaphthene-d10	2120		14.388			
1517-22-2	Phenanthrene-d10	4760		17.136			
1719-03-5	Chrysene-d12	4680		21.322			
1520-96-3	Perylene-d12	4460		23.592			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	02/13/25
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005	Date Received:	02/17/25
Client Sample ID:	BP-VPB-192-GW-780-782	SDG No.:	Q1380
Lab Sample ID:	Q1380-09	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	380 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036476.D	1	02/18/25 11:40	02/20/25 12:00	PB166758

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.53	UM	0.18	0.53	0.53	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.35		30 - 150		88%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.41		30 - 150		103%	SPK: 0.4
367-12-4	2-Fluorophenol	0.25		20 - 110		62%	SPK: 0.4
13127-88-3	Phenol-d6	0.18		10 - 160		44%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.36		55 - 111		91%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.47	*	53 - 106		118%	SPK: 0.4
118-79-6	2,4,6-Tribromophenol	0.36		46 - 122		90%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.50		58 - 132		126%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1680		7.746			
1146-65-2	Naphthalene-d8	3840		10.541			
15067-26-2	Acenaphthene-d10	2410		14.388			
1517-22-2	Phenanthrene-d10	5510		17.124			
1719-03-5	Chrysene-d12	5350		21.322			
1520-96-3	Perylene-d12	4860		23.59			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	02/14/25
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005	Date Received:	02/17/25
Client Sample ID:	VPB192-HYD-20250214	SDG No.:	Q1380
Lab Sample ID:	Q1380-12	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	940 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036479.D	1	02/18/25 11:40	02/20/25 13:48	PB166758

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.21	U	0.070	0.21	0.21	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.023	*	30 - 150		6%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.34		30 - 150		86%	SPK: 0.4
367-12-4	2-Fluorophenol	0.062	*	20 - 110		16%	SPK: 0.4
13127-88-3	Phenol-d6	0.0030	*	10 - 160		1%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.31		55 - 111		79%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.39		53 - 106		97%	SPK: 0.4
118-79-6	2,4,6-Tribromophenol	0.30		46 - 122		76%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.45		58 - 132		112%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1730		7.746			
1146-65-2	Naphthalene-d8	4020		10.541			
15067-26-2	Acenaphthene-d10	2590		14.388			
1517-22-2	Phenanthrene-d10	5760		17.136			
1719-03-5	Chrysene-d12	5470		21.322			
1520-96-3	Perylene-d12	4500		23.587			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



QC SUMMARY

Surrogate Summary

SW-846

SDG No.: Q1380

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
PB166758BL	PB166758BL	2-Methylnaphthalene-d10	0.4	0.35	87		30	150
		Fluoranthene-d10	0.4	0.42	104		30	150
		2-Fluorophenol	0.4	0.40	100		20	110
		Phenol-d6	0.4	0.35	86		10	160
		Nitrobenzene-d5	0.4	0.37	93		55	111
		2-Fluorobiphenyl	0.4	0.35	88		53	106
		2,4,6-Tribromophenol	0.4	0.19	48		46	122
		Terphenyl-d14	0.4	0.42	104		58	132
PB166758BS	PB166758BS	2-Methylnaphthalene-d10	0.4	0.52	129		30	150
		Fluoranthene-d10	0.4	0.41	101		30	150
		2-Fluorophenol	0.4	0.41	101		20	110
		Phenol-d6	0.4	0.39	98		10	160
		Nitrobenzene-d5	0.4	0.39	97		55	111
		2-Fluorobiphenyl	0.4	0.47	118	*	53	106
		2,4,6-Tribromophenol	0.4	0.37	93		46	122
		Terphenyl-d14	0.4	0.43	108		58	132
Q1380-01	BP-VPB-192-EB-20250212	2-Methylnaphthalene-d10	0.4	0.31	76		30	150
		Fluoranthene-d10	0.4	0.43	108		30	150
		2-Fluorophenol	0.4	0.14	34		20	110
		Phenol-d6	0.4	0.078	20		10	160
		Nitrobenzene-d5	0.4	0.30	76		55	111
		2-Fluorobiphenyl	0.4	0.38	96		53	106
		2,4,6-Tribromophenol	0.4	0.35	87		46	122
		Terphenyl-d14	0.4	0.58	146	*	58	132
Q1380-09	BP-VPB-192-GW-780-782	2-Methylnaphthalene-d10	0.4	0.35	88		30	150
		Fluoranthene-d10	0.4	0.41	103		30	150
		2-Fluorophenol	0.4	0.25	62		20	110
		Phenol-d6	0.4	0.18	44		10	160
		Nitrobenzene-d5	0.4	0.36	91		55	111
		2-Fluorobiphenyl	0.4	0.47	118	*	53	106
		2,4,6-Tribromophenol	0.4	0.36	90		46	122
		Terphenyl-d14	0.4	0.50	126		58	132
Q1380-10MS	BP-VPB-192-GW-780-782MS	2-Methylnaphthalene-d10	0.4	0.32	80		30	150
		Fluoranthene-d10	0.4	0.45	112		30	150
		2-Fluorophenol	0.4	0.24	60		20	110
		Phenol-d6	0.4	0.22	54		10	160
		Nitrobenzene-d5	0.4	0.32	80		55	111
		2-Fluorobiphenyl	0.4	0.44	110	*	53	106
		2,4,6-Tribromophenol	0.4	0.40	101		46	122
		Terphenyl-d14	0.4	0.54	135	*	58	132
Q1380-11MSD	BP-VPB-192-GW-780-782MSD	2-Methylnaphthalene-d10	0.4	0.32	80		30	150
		Fluoranthene-d10	0.4	0.47	116		30	150
		2-Fluorophenol	0.4	0.25	62		20	110
		Phenol-d6	0.4	0.23	57		10	160
		Nitrobenzene-d5	0.4	0.32	80		55	111
		2-Fluorobiphenyl	0.4	0.45	113	*	53	106
		2,4,6-Tribromophenol	0.4	0.40	101		46	122
		Terphenyl-d14	0.4	0.55	138	*	58	132
Q1380-12	VPB192-HYD-20250214	2-Methylnaphthalene-d10	0.4	0.023	6	*	30	150
		Fluoranthene-d10	0.4	0.34	86		30	150
		2-Fluorophenol	0.4	0.062	16	*	20	110

Surrogate Summary

SW-846

SDG No.: Q1380

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
Q1380-12	VPB192-HYD-20250214	Phenol-d6	0.4	0.0030	1	*	10	160
		Nitrobenzene-d5	0.4	0.31	79		55	111
		2-Fluorobiphenyl	0.4	0.39	97		53	106
		2,4,6-Tribromophenol	0.4	0.30	76		46	122
		Terphenyl-d14	0.4	0.45	112		58	132



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Matrix Spike/Matrix Spike Duplicate Summary

SW-846

SDG No.: Q1380

Client: Tetra Tech NUS, Inc.

Analytical Method: SW8270-Modified

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1380-10MS	Client Sample ID:	BP-VPB-192-GW-780-782MS					DataFile:	BN036477.D		
1,4-Dioxane	1.3	0	0.83	ug/L	64	*			70	130	

Matrix Spike/Matrix Spike Duplicate Summary
SW-846

SDG No.: Q1380
Client: Tetra Tech NUS, Inc.
Analytical Method: SW8270-Modified

Parameter	Spike	Sample Result	Result	Units	Rec	Rec Qual	RPD	RPD Qual	Low	Limits High	RPD
Lab Sample ID:	Q1380-11MSD	Client Sample ID:	BP-VPB-192-GW-780-782MSD					DataFile:	BN036478.D		
1,4-Dioxane	1.3	0	0.92	ug/L	71		10		70	130	20

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: Q1380

Client: Tetra Tech NUS, Inc.

Analytical Method: 8270-Modified **DataFile:** BN036480.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Low	Limits	RPD
								Qual		High	
PB166758BS	1,4-Dioxane	0.4	0.35	ug/L	88				70	130	

4B

SEMIVOLATILE METHOD BLANK SUMMARY

EPA SAMPLE NO.

PB166758BL

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM Case No.: Q1380

SAS No.: Q1380 SDG NO.: Q1380

Lab File ID: BN036474.D

Lab Sample ID: PB166758BL

Instrument ID: BNA_N

Date Extracted: 02/18/2025

Matrix: (soil/water) Water

Date Analyzed: 02/20/2025

Level: (low/med) LOW

Time Analyzed: 10:47

THIS METHOD BLANK APPLIES TO THE FOLLOWING SAMPLES, MS AND MSD:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED
PB166758BS	PB166758BS	BN036480.D	02/20/2025
BP-VPB-192-EB-20250212	Q1380-01	BN036475.D	02/20/2025
BP-VPB-192-GW-780-782	Q1380-09	BN036476.D	02/20/2025
BP-VPB-192-GW-780-782MS	Q1380-10MS	BN036477.D	02/20/2025
BP-VPB-192-GW-780-782MSD	Q1380-11MSD	BN036478.D	02/20/2025
VPB192-HYD-20250214	Q1380-12	BN036479.D	02/20/2025

COMMENTS: _____

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q1380

SDG NO.: Q1380

Lab File ID: BN036408.D

DFTPP Injection Date: 02/10/2025

Instrument ID: BNA_N

DFTPP Injection Time: 11:46

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	51.4
68	Less than 2.0% of mass 69	0.3 (0.7) 1
69	Mass 69 relative abundance	47.7
70	Less than 2.0% of mass 69	0.3 (0.6) 1
127	10.0 - 80.0% of mass 198	48.3
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	7
275	10.0 - 60.0% of mass 198	24.7
365	Greater than 1% of mass 198	3.3
441	Present, but less than mass 443	7.1
442	Greater than 50% of mass 198	52.3
443	15.0 - 24.0% of mass 442	10.5 (20.1) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDICC0.1	SSTDICC0.1	BN036409.D	02/10/2025	12:25
SSTDICC0.2	SSTDICC0.2	BN036410.D	02/10/2025	13:01
SSTDICCC0.4	SSTDICCC0.4	BN036411.D	02/10/2025	13:36
SSTDICC0.8	SSTDICC0.8	BN036412.D	02/10/2025	14:12
SSTDICC1.6	SSTDICC1.6	BN036413.D	02/10/2025	14:48
SSTDICC3.2	SSTDICC3.2	BN036414.D	02/10/2025	15:24
SSTDICC5.0	SSTDICC5.0	BN036415.D	02/10/2025	16:00

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: TETR06

Lab Code: CHEM

SAS No.: Q1380 SDG NO.: Q1380

Lab File ID: BN036472.D

DFTPP Injection Date: 02/20/2025

Instrument ID: BNA_N

DFTPP Injection Time: 09:32

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	57.4
68	Less than 2.0% of mass 69	0.8 (1.5) 1
69	Mass 69 relative abundance	52.1
70	Less than 2.0% of mass 69	0.3 (0.5) 1
127	10.0 - 80.0% of mass 198	50.3
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.9
275	10.0 - 60.0% of mass 198	24.3
365	Greater than 1% of mass 198	3.5
441	Present, but less than mass 443	8.4
442	Greater than 50% of mass 198	49.8
443	15.0 - 24.0% of mass 442	9.6 (19.3) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC0.4	SSTDCCC0.4	BN036473.D	02/20/2025	10:11
PB166758BL	PB166758BL	BN036474.D	02/20/2025	10:47
BP-VPB-192-EB-20250212	Q1380-01	BN036475.D	02/20/2025	11:24
BP-VPB-192-GW-780-782	Q1380-09	BN036476.D	02/20/2025	12:00
BP-VPB-192-GW-780-782MS	Q1380-10MS	BN036477.D	02/20/2025	12:36
BP-VPB-192-GW-780-782MSD	Q1380-11MSD	BN036478.D	02/20/2025	13:12
VPB192-HYD-20250214	Q1380-12	BN036479.D	02/20/2025	13:48
PB166758BS	PB166758BS	BN036480.D	02/20/2025	14:24
SSTDCCC0.4EC	SSTDCCC0.4	BN036482.D	02/20/2025	16:08

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG NO.: Q1380

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 02/20/2025

Lab File ID: BN036473.D Time Analyzed: 10:11

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	1766	7.746	4292	10.54	2899	14.39
UPPER LIMIT	3532	8.246	8584	11.041	5798	14.887
LOWER LIMIT	883	7.246	2146	10.041	1449.5	13.887
EPA SAMPLE NO.						
01 PB166758BL	1874	7.75	3905	10.55	2305	14.40
02 BP-VPB-192-EB-20250212	1491	7.75	3323	10.54	2119	14.39
03 PB166758BS	1827	7.75	4127	10.54	2357	14.39
04 BP-VPB-192-GW-780-782	1676	7.75	3840	10.54	2412	14.39
05 BP-VPB-192-GW-780-782MS	1709	7.75	3995	10.54	2495	14.39
06 BP-VPB-192-GW-780-782MSD	1787	7.75	4250	10.53	2636	14.39
07 VPB192-HYD-20250214	1727	7.75	4024	10.54	2593	14.39

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG NO.: Q1380

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 02/20/2025

Lab File ID: BN036473.D Time Analyzed: 10:11

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	6167	17.136	5687	21.313	5826	23.586
UPPER LIMIT	12334	17.636	11374	21.813	11652	24.086
LOWER LIMIT	3083.5	16.636	2843.5	20.813	2913	23.086
EPA SAMPLE NO.						
01 PB166758BL	4889	17.15	4485	21.33	3907	23.60
02 BP-VPB-192-EB-20250212	4759	17.14	4677	21.32	4457	23.59
03 PB166758BS	5256	17.14	4823	21.32	4362	23.59
04 BP-VPB-192-GW-780-782	5513	17.12	5354	21.32	4859	23.59
05 BP-VPB-192-GW-780-782MS	5565	17.12	5523	21.31	5300	23.59
06 BP-VPB-192-GW-780-782MSD	5810	17.12	5865	21.31	5479	23.58
07 VPB192-HYD-20250214	5756	17.14	5466	21.32	4501	23.59

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.



QC SAMPLE DATA

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	
Client Sample ID:	PB166758BL		SDG No.:	Q1380
Lab Sample ID:	PB166758BL		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036474.D	1	02/18/25 11:40	02/20/25 10:47	PB166758

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.20	U	0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.35		30 - 150		87%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.42		30 - 150		104%	SPK: 0.4
367-12-4	2-Fluorophenol	0.40		20 - 110		100%	SPK: 0.4
13127-88-3	Phenol-d6	0.35		10 - 160		86%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.37		55 - 111		93%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.35		53 - 106		88%	SPK: 0.4
118-79-6	2,4,6-Tribromophenol	0.19		46 - 122		48%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.42		58 - 132		104%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1870		7.746			
1146-65-2	Naphthalene-d8	3910		10.551			
15067-26-2	Acenaphthene-d10	2310		14.398			
1517-22-2	Phenanthrene-d10	4890		17.148			
1719-03-5	Chrysene-d12	4490		21.331			
1520-96-3	Perylene-d12	3910		23.595			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.		Date Collected:	
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005		Date Received:	
Client Sample ID:	PB166758BS		SDG No.:	Q1380
Lab Sample ID:	PB166758BS		Matrix:	Water
Analytical Method:	SW8270ESIM		% Solid:	0
Sample Wt/Vol:	1000	Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:		uL	Test:	SVOC-SIMGroup1
Extraction Type :		Decanted : N	Level :	LOW
Injection Volume :		GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :				

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036480.D	1	02/18/25 11:40	02/20/25 14:24	PB166758

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.35		0.070	0.20	0.20	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.52		30 - 150		129%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.41		30 - 150		101%	SPK: 0.4
367-12-4	2-Fluorophenol	0.41		20 - 110		101%	SPK: 0.4
13127-88-3	Phenol-d6	0.39		10 - 160		98%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.39		55 - 111		97%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.47	*	53 - 106		118%	SPK: 0.4
118-79-6	2,4,6-Tribromophenol	0.37		46 - 122		93%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.43		58 - 132		108%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1830	7.746				
1146-65-2	Naphthalene-d8	4130	10.541				
15067-26-2	Acenaphthene-d10	2360	14.388				
1517-22-2	Phenanthrene-d10	5260	17.136				
1719-03-5	Chrysene-d12	4820	21.322				
1520-96-3	Perylene-d12	4360	23.589				

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	02/13/25
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005	Date Received:	02/17/25
Client Sample ID:	BP-VPB-192-GW-780-782MS	SDG No.:	Q1380
Lab Sample ID:	Q1380-10MS	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	320 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036477.D	1	02/18/25 11:40	02/20/25 12:36	PB166758

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.83		0.21	0.63	0.63	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.32		30 - 150		80%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.45		30 - 150		112%	SPK: 0.4
367-12-4	2-Fluorophenol	0.24		20 - 110		60%	SPK: 0.4
13127-88-3	Phenol-d6	0.22		10 - 160		54%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.32		55 - 111		80%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.44	*	53 - 106		110%	SPK: 0.4
118-79-6	2,4,6-Tribromophenol	0.40		46 - 122		101%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.54	*	58 - 132		135%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1710		7.746			
1146-65-2	Naphthalene-d8	4000		10.541			
15067-26-2	Acenaphthene-d10	2500		14.387			
1517-22-2	Phenanthrene-d10	5570		17.124			
1719-03-5	Chrysene-d12	5520		21.313			
1520-96-3	Perylene-d12	5300		23.586			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products

Report of Analysis

Client:	Tetra Tech NUS, Inc.	Date Collected:	02/13/25
Project:	NWIRP Bethpage CTO WE13 - VPB-192 #112G08005	Date Received:	02/17/25
Client Sample ID:	BP-VPB-192-GW-780-782MSD	SDG No.:	Q1380
Lab Sample ID:	Q1380-11MSD	Matrix:	Water
Analytical Method:	SW8270ESIM	% Solid:	0
Sample Wt/Vol:	300 Units: mL	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOC-SIMGroup1
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :			

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BN036478.D	1	02/18/25 11:40	02/20/25 13:12	PB166758

CAS Number	Parameter	Conc.	Qualifier	MDL	LOD	LOQ / CRQL	Units
TARGETS							
123-91-1	1,4-Dioxane	0.92		0.23	0.67	0.67	ug/L
SURROGATES							
7297-45-2	2-Methylnaphthalene-d10	0.32		30 - 150		80%	SPK: 0.4
93951-69-0	Fluoranthene-d10	0.47		30 - 150		116%	SPK: 0.4
367-12-4	2-Fluorophenol	0.25		20 - 110		62%	SPK: 0.4
13127-88-3	Phenol-d6	0.23		10 - 160		57%	SPK: 0.4
4165-60-0	Nitrobenzene-d5	0.32		55 - 111		80%	SPK: 0.4
321-60-8	2-Fluorobiphenyl	0.45	*	53 - 106		113%	SPK: 0.4
118-79-6	2,4,6-Tribromophenol	0.40		46 - 122		101%	SPK: 0.4
1718-51-0	Terphenyl-d14	0.55	*	58 - 132		138%	SPK: 0.4
INTERNAL STANDARDS							
3855-82-1	1,4-Dichlorobenzene-d4	1790		7.746			
1146-65-2	Naphthalene-d8	4250		10.53			
15067-26-2	Acenaphthene-d10	2640		14.387			
1517-22-2	Phenanthrene-d10	5810		17.124			
1719-03-5	Chrysene-d12	5870		21.313			
1520-96-3	Perylene-d12	5480		23.584			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products



CALIBRATION SUMMARY

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\
 Method File : 8270-SIM-BN021025.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Tue Feb 11 01:17:14 2025
 Response Via : Initial Calibration

Calibration Files

0.1 =BN036409.D 0.2 =BN036410.D 0.4 =BN036411.D 0.8 =BN036412.D 1.6 =BN036413.D 3.2 =BN036414.D 5.0 =BN036415.D

	Compound	0.1	0.2	0.4	0.8	1.6	3.2	5.0	Avg	%RSD

1) I	1,4-Dichlorobenzen...	-----ISTD-----								
2)	1,4-Dioxane	0.555	0.437	0.433	0.414	0.411	0.433	0.381	0.438	12.66
3)	n-Nitrosodimet...	0.906	0.779	0.764	0.724	0.708	0.769	0.670	0.760	9.90
4) S	2-Fluorophenol	1.009	0.954	0.936	0.920	0.914	0.999	0.885	0.945	4.80
5) S	Phenol-d6	1.134	1.007	1.032	1.062	1.099	1.267	1.164	1.109	8.00
6)	bis(2-Chloroet...	1.382	1.070	1.086	1.129	1.120	1.225	1.107	1.160	9.48
7) I	Naphthalene-d8	-----ISTD-----								
8) S	Nitrobenzene-d5	0.500	0.363	0.365	0.370	0.367	0.417	0.381	0.395	12.70
9)	Naphthalene	1.400	1.141	1.116	1.088	1.075	1.186	1.073	1.154	10.01
10)	Hexachlorobuta...	0.319	0.293	0.283	0.272	0.264	0.282	0.253	0.281	7.67
11) SURR	2-Methylnaphth...	0.647	0.583	0.602	0.588	0.597	0.668	0.618	0.615	5.19
12)	2-Methylnaphth...	0.833	0.712	0.738	0.721	0.726	0.816	0.750	0.757	6.40
13) I	Acenaphthene-d10	-----ISTD-----								
14) S	2,4,6-Tribromo...	0.196	0.181	0.186	0.184	0.195	0.226	0.219	0.198	8.90
15) S	2-Fluorobiphenyl	1.409	1.390	1.377	1.491	1.564	1.738	1.558	1.504	8.57
16)	Acenaphthylene	1.807	1.667	1.692	1.683	1.734	1.964	1.820	1.767	5.98
17)	Acenaphthene	1.245	1.125	1.146	1.128	1.175	1.273	1.169	1.180	4.89
18)	Fluorene	1.696	1.630	1.661	1.627	1.669	1.829	1.646	1.680	4.17
19) I	Phenanthrene-d10	-----ISTD-----								
20)	4,6-Dinitro-2-...	0.071	0.067	0.069	0.074	0.084	0.107		0.078	19.60
21)	4-Bromophenyl-...	0.243	0.227	0.231	0.232	0.236	0.264	0.238	0.239	5.15
22)	Hexachlorobenzene	0.305	0.296	0.284	0.287	0.289	0.317	0.285	0.295	4.11
23)	Atrazine	0.196	0.190	0.187	0.186	0.194	0.229	0.213	0.199	8.00
24)	Pentachlorophenol	0.140	0.125	0.122	0.122	0.134	0.170	0.167	0.140	14.74
25)	Phenanthrene	1.233	1.090	1.095	1.112	1.138	1.273	1.153	1.156	6.12
26)	Anthracene	0.990	0.933	0.967	0.978	1.015	1.167	1.088	1.020	7.92
27) SURR	Fluoranthene-d10	1.109	1.043	1.063	1.059	1.098	1.258	1.156	1.112	6.70
28)	Fluoranthene	1.441	1.323	1.353	1.356	1.404	1.607	1.461	1.421	6.76
29) I	Chrysene-d12	-----ISTD-----								
30)	Pyrene	1.584	1.568	1.534	1.490	1.488	1.629	1.492	1.541	3.59
31) S	Terphenyl-d14	0.860	0.847	0.852	0.829	0.834	0.913	0.843	0.854	3.27
32)	Benzo(a)anthra...	1.257	1.276	1.293	1.255	1.300	1.471	1.362	1.316	5.86
33)	Chrysene	1.449	1.456	1.360	1.414	1.404	1.527	1.366	1.425	4.08
34)	Bis(2-ethylhex...	0.902	0.875	0.777	0.745	0.761	0.861	0.819	0.820	7.45
35) I	Perylene-d12	-----ISTD-----								

Method Path : Z:\svoasrv\HPCHEM1\BNA_N\Methods\

Method File : 8270-SIM-BN021025.M

36)	Indeno(1,2,3-c...	1.182	1.289	1.378	1.390	1.446	1.630	1.471	1.398	10.13
37)	Benzo(b)fluora...	1.174	1.220	1.260	1.290	1.333	1.529	1.416	1.317	9.24
38)	Benzo(k)fluora...	1.258	1.253	1.363	1.326	1.347	1.532	1.413	1.356	7.08
39) C	Benzo(a)pyrene	1.091	1.081	1.102	1.114	1.145	1.309	1.206	1.150	7.12
40)	Dibenzo(a,h)an...	0.906	1.021	1.075	1.087	1.154	1.304	1.176	1.103	11.40
41)	Benzo(g,h,i)pe...	1.140	1.212	1.254	1.230	1.269	1.400	1.249	1.250	6.27

(#) = Out of Range

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG No.: Q1380

Instrument ID: BNA_N Calibration Date/Time: 02/20/2025 10:11

Lab File ID: BN036473.D Init. Calib. Date(s): 02/10/2025 02/10/2025

EPA Sample No.: SSTDCCC0.4 Init. Calib. Time(s): 12:25 16:00

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.615	0.585		-4.9	20.0
Fluoranthene-d10	1.112	1.062		-4.5	20.0
2-Fluorophenol	0.945	0.879		-7.0	20.0
Phenol-d6	1.109	1.046		-5.7	20.0
Nitrobenzene-d5	0.395	0.399		1.0	20.0
2-Fluorobiphenyl	1.504	1.478		-1.7	20.0
2,4,6-Tribromophenol	0.198	0.165		-16.7	20.0
Terphenyl-d14	0.854	0.795		-6.9	20.0
1,4-Dioxane	0.438	0.446		1.8	20.0

All other compounds must meet a minimum RRF of 0.010.

7C

SEMIVOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG No.: Q1380

Instrument ID: BNA_N Calibration Date/Time: 02/20/2025 16:08

Lab File ID: BN036482.D Init. Calib. Date(s): 02/10/2025 02/10/2025

EPA Sample No.: SSTDCCC0.4EC Init. Calib. Time(s): 12:25 16:00

GC Column: ZB-GR ID: 0.25 (mm)

COMPOUND	RRF	RRF0.4	MIN RRF	%D	MAX%D
2-Methylnaphthalene-d10	0.615	0.590		-4.1	50.0
Fluoranthene-d10	1.112	1.050		-5.6	50.0
2-Fluorophenol	0.945	0.903		-4.4	50.0
Phenol-d6	1.109	1.067		-3.8	50.0
Nitrobenzene-d5	0.395	0.392		-0.8	50.0
2-Fluorobiphenyl	1.504	1.565		4.1	50.0
2,4,6-Tribromophenol	0.198	0.163		-17.7	50.0
Terphenyl-d14	0.854	0.801		-6.2	50.0
1,4-Dioxane	0.438	0.416		-5.0	50.0

All other compounds must meet a minimum RRF of 0.010.