

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN021025

Review By	Rahul	Review On	2/11/2025 4:21:36 PM
Supervise By	Jagrut	Supervise On	2/11/2025 6:22:45 PM
SubDirectory	BN021025	HP Acquire Method	BNA_N, 8270_HP Processing Method bn021025

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6738,SP6736,SP6735,SP6734,SP6733,SP6732,SP6731
CCC	SP6735
Internal Standard/PEM	SP6682,1ul/100ul sample
ICV/I.BLK	SP6684
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN036408.D	10 Feb 2025 11:46		RC/JU	Ok
2	SSTDICC0.1	SSTDICC0.1	BN036409.D	10 Feb 2025 12:25		RC/JU	Ok
3	SSTDICC0.2	SSTDICC0.2	BN036410.D	10 Feb 2025 13:01		RC/JU	Ok
4	SSTDICCC0.4	SSTDICCC0.4	BN036411.D	10 Feb 2025 13:36	The Calibration is Good For DOD	RC/JU	Ok
5	SSTDICC0.8	SSTDICC0.8	BN036412.D	10 Feb 2025 14:12		RC/JU	Ok
6	SSTDICC1.6	SSTDICC1.6	BN036413.D	10 Feb 2025 14:48		RC/JU	Ok
7	SSTDICC3.2	SSTDICC3.2	BN036414.D	10 Feb 2025 15:24		RC/JU	Ok
8	SSTDICC5.0	SSTDICC5.0	BN036415.D	10 Feb 2025 16:00	Compound #20 removed from 5.0ppm	RC/JU	Ok
9	SSTDICV0.4	ICVBN021025	BN036416.D	10 Feb 2025 16:36		RC/JU	Ok
10	PB166297BL	PB166297BL	BN036417.D	10 Feb 2025 17:48	Analyzed for contamination check	RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN022025

Review By	Rahul	Review On	2/21/2025 10:57:17 AM
Supervise By	Jagrut	Supervise On	2/21/2025 11:17:01 AM
SubDirectory	BN022025	HP Acquire Method	BNA_N, 8270_HP Processing Method bn021025

STD. NAME	STD REF.#
Tune/Reschk	SP6717
Initial Calibration Stds	SP6738,SP6736,SP6735,SP6734,SP6733,SP6732,SP6731
CCC	SP6735
Internal Standard/PEM	SP6682,1ul/100ul sample
ICV/I.BLK	SP6684
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN036472.D	20 Feb 2025 09:32		RC/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN036473.D	20 Feb 2025 10:11		RC/JU	Ok
3	PB166758BL	PB166758BL	BN036474.D	20 Feb 2025 10:47		RC/JU	Ok
4	Q1380-01	BP-VPB-192-EB-20250	BN036475.D	20 Feb 2025 11:24		RC/JU	Ok
5	Q1380-09	BP-VPB-192-GW-780-7	BN036476.D	20 Feb 2025 12:00		RC/JU	Ok
6	Q1380-10MS	BP-VPB-192-GW-780-7	BN036477.D	20 Feb 2025 12:36		RC/JU	Ok,M
7	Q1380-11MSD	BP-VPB-192-GW-780-7	BN036478.D	20 Feb 2025 13:12		RC/JU	Ok
8	Q1380-12	VPB192-HYD-2025021	BN036479.D	20 Feb 2025 13:48	Surrogate Fail	RC/JU	Ok
9	PB166758BS	PB166758BS	BN036480.D	20 Feb 2025 14:24		RC/JU	Ok,M
10	SP6741	SP6741	BN036481.D	20 Feb 2025 15:32	8270 Sim Surrogate	RC/JU	Ok
11	SSTDCCC0.4	SSTDCCC0.4EC	BN036482.D	20 Feb 2025 16:08		RC/JU	Ok,M

M : Manual Integration