

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG No.: Q1380

Instrument ID: MSVOA_N Calibration Date/Time: 02/19/2025 10:51

Lab File ID: VN085782.D Init. Calib. Date(s): 02/18/2025 02/18/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:09 14:18

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.682	0.646	0.1	-5.28	20
Vinyl Chloride	0.723	0.697		-3.6	20
Bromomethane	0.462	0.446		-3.46	20
Chloroethane	0.479	0.437		-8.77	20
Trichlorofluoromethane	1.046	1.027		-1.82	20
1,1,2-Trichlorotrifluoroethane	0.605	0.611		0.99	20
1,1-Dichloroethene	0.548	0.559		2.01	20
Acetone	0.205	0.226		10.24	20
Carbon Disulfide	1.625	1.544		-4.99	20
Methyl tert-butyl Ether	1.629	1.743		7	20
Methylene Chloride	0.660	0.628		-4.85	20
trans-1,2-Dichloroethene	0.590	0.594		0.68	20
1,1-Dichloroethane	1.106	1.094	0.1	-1.09	20
2-Butanone	0.302	0.308		1.99	20
Carbon Tetrachloride	0.550	0.568		3.27	20
cis-1,2-Dichloroethene	0.685	0.703		2.63	20
Chloroform	1.156	1.135		-1.82	20
1,1,1-Trichloroethane	1.028	1.025		-0.29	20
Methylcyclohexane	0.455	0.542		19.12	20
Benzene	1.484	1.556		4.85	20
1,2-Dichloroethane	0.483	0.496		2.69	20
Trichloroethene	0.364	0.377		3.57	20
1,2-Dichloropropane	0.358	0.379		5.87	20
Bromodichloromethane	0.541	0.561		3.7	20
4-Methyl-2-Pentanone	0.367	0.383		4.36	20
Toluene	0.870	0.975		12.07	20
t-1,3-Dichloropropene	0.500	0.556		11.2	20
cis-1,3-Dichloropropene	0.548	0.612		11.68	20
1,1,2-Trichloroethane	0.346	0.356		2.89	20
2-Hexanone	0.253	0.281		11.07	20
Dibromochloromethane	0.403	0.435		7.94	20
Tetrachloroethene	0.387	0.415		7.24	20
Chlorobenzene	1.145	1.191	0.3	4.02	20
Ethyl Benzene	1.794	2.061		14.88	20
m/p-Xylenes	0.682	0.811		18.92	20
o-Xylene	0.648	0.758		16.98	20
Styrene	1.059	1.287		21.53	20
Bromoform	0.306	0.323	0.1	5.56	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG No.: Q1380

Instrument ID: MSVOA_N Calibration Date/Time: 02/19/2025 10:51

Lab File ID: VN085782.D Init. Calib. Date(s): 02/18/2025 02/18/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:09 14:18

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	3.425	3.921		14.48	20
1,1,2,2-Tetrachloroethane	1.179	1.058	0.3	-10.26	20
1,3-Dichlorobenzene	1.734	1.791		3.29	20
1,4-Dichlorobenzene	1.777	1.777		0	20
1,2-Dichlorobenzene	1.667	1.704		2.22	20
1,2-Dichloroethane-d4	0.648	0.646		-0.31	20
Dibromofluoromethane	0.327	0.346		5.81	20
Toluene-d8	1.190	1.302		9.41	20
4-Bromofluorobenzene	0.392	0.446		13.78	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.