

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG No.: Q1380

Instrument ID: MSVOA_N Calibration Date/Time: 02/19/2025 15:58

Lab File ID: VN085792.D Init. Calib. Date(s): 02/18/2025 02/18/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:09 14:18

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.682	0.587	0.1	-13.93	50
Vinyl Chloride	0.723	0.619		-14.39	50
Bromomethane	0.462	0.383		-17.1	50
Chloroethane	0.479	0.387		-19.21	50
Trichlorofluoromethane	1.046	0.878		-16.06	50
1,1,2-Trichlorotrifluoroethane	0.605	0.528		-12.73	50
1,1-Dichloroethene	0.548	0.485		-11.5	50
Acetone	0.205	0.209		1.95	50
Carbon Disulfide	1.625	1.344		-17.29	50
Methyl tert-butyl Ether	1.629	1.563		-4.05	50
Methylene Chloride	0.660	0.553		-16.21	50
trans-1,2-Dichloroethene	0.590	0.517		-12.37	50
1,1-Dichloroethane	1.106	0.983	0.1	-11.12	50
2-Butanone	0.302	0.305		0.99	50
Carbon Tetrachloride	0.550	0.492		-10.55	50
cis-1,2-Dichloroethene	0.685	0.618		-9.78	50
Chloroform	1.156	1.018		-11.94	50
1,1,1-Trichloroethane	1.028	0.904		-12.06	50
Methylcyclohexane	0.455	0.466		2.42	50
Benzene	1.484	1.377		-7.21	50
1,2-Dichloroethane	0.483	0.446		-7.66	50
Trichloroethene	0.364	0.323		-11.26	50
1,2-Dichloropropane	0.358	0.334		-6.7	50
Bromodichloromethane	0.541	0.500		-7.58	50
4-Methyl-2-Pentanone	0.367	0.380		3.54	50
Toluene	0.870	0.868		-0.23	50
t-1,3-Dichloropropene	0.500	0.505		1	50
cis-1,3-Dichloropropene	0.548	0.542		-1.1	50
1,1,2-Trichloroethane	0.346	0.323		-6.65	50
2-Hexanone	0.253	0.280		10.67	50
Dibromochloromethane	0.403	0.389		-3.47	50
Tetrachloroethene	0.387	0.337		-12.92	50
Chlorobenzene	1.145	1.033	0.3	-9.78	50
Ethyl Benzene	1.794	1.793		-0.06	50
m/p-Xylenes	0.682	0.704		3.23	50
o-Xylene	0.648	0.665		2.62	50
Styrene	1.059	1.140		7.65	50
Bromoform	0.306	0.290	0.1	-5.23	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG No.: Q1380

Instrument ID: MSVOA_N Calibration Date/Time: 02/19/2025 15:58

Lab File ID: VN085792.D Init. Calib. Date(s): 02/18/2025 02/18/2025

Heated Purge: (Y/N) N Init. Calib. Time(s): 11:09 14:18

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	3.425	3.305		-3.5	50
1,1,2,2-Tetrachloroethane	1.179	0.989	0.3	-16.11	50
1,3-Dichlorobenzene	1.734	1.516		-12.57	50
1,4-Dichlorobenzene	1.777	1.504		-15.36	50
1,2-Dichlorobenzene	1.667	1.463		-12.24	50
1,2-Dichloroethane-d4	0.648	0.661		2.01	50
Dibromofluoromethane	0.327	0.345		5.51	50
Toluene-d8	1.190	1.326		11.43	50
4-Bromofluorobenzene	0.392	0.465		18.62	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.