

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG No.: Q1380

Instrument ID: MSVOA_Y Calibration Date/Time: 02/19/2025 18:32

Lab File ID: VY021245.D Init. Calib. Date(s): 02/03/2025 02/03/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 10:35 12:44

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.440	0.469	0.1	6.59	50
Vinyl Chloride	0.443	0.496		11.96	50
Bromomethane	0.274	0.311		13.5	50
Chloroethane	0.263	0.322		22.43	50
Trichlorofluoromethane	0.806	0.844		4.72	50
1,1,2-Trichlorotrifluoroethane	0.506	0.518		2.37	50
1,1-Dichloroethene	0.477	0.484		1.47	50
Acetone	0.080	0.095		18.75	50
Carbon Disulfide	1.495	1.357		-9.23	50
Methyl tert-butyl Ether	1.205	1.416		17.51	50
Methylene Chloride	0.497	0.568		14.29	50
trans-1,2-Dichloroethene	0.525	0.540		2.86	50
1,1-Dichloroethane	0.970	1.082	0.1	11.55	50
2-Butanone	0.131	0.161		22.9	50
Carbon Tetrachloride	0.555	0.556		0.18	50
cis-1,2-Dichloroethene	0.604	0.666		10.27	50
Chloroform	0.997	1.114		11.73	50
1,1,1-Trichloroethane	0.927	0.983		6.04	50
Methylcyclohexane	0.581	0.577		-0.69	50
Benzene	1.393	1.472		5.67	50
1,2-Dichloroethane	0.386	0.422		9.33	50
Trichloroethene	0.360	0.362		0.56	50
1,2-Dichloropropane	0.331	0.367		10.88	50
Bromodichloromethane	0.491	0.541		10.18	50
4-Methyl-2-Pentanone	0.208	0.254		22.11	50
Toluene	0.875	0.924		5.6	50
t-1,3-Dichloropropene	0.441	0.497		12.7	50
cis-1,3-Dichloropropene	0.522	0.577		10.54	50
1,1,2-Trichloroethane	0.237	0.265		11.81	50
2-Hexanone	0.134	0.169		26.12	50
Dibromochloromethane	0.333	0.364		9.31	50
Tetrachloroethene	0.371	0.361		-2.69	50
Chlorobenzene	1.113	1.160	0.3	4.22	50
Ethyl Benzene	1.969	2.074		5.33	50
m/p-Xylenes	0.737	0.768		4.21	50
o-Xylene	0.687	0.734		6.84	50
Styrene	1.145	1.245		8.73	50
Bromoform	0.220	0.242	0.1	10	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG No.: Q1380

Instrument ID: MSVOA_Y Calibration Date/Time: 02/19/2025 18:32

Lab File ID: VY021245.D Init. Calib. Date(s): 02/03/2025 02/03/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 10:35 12:44

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	3.906	3.853		-1.36	50
1,1,2,2-Tetrachloroethane	0.650	0.684	0.3	5.23	50
1,3-Dichlorobenzene	1.740	1.731		-0.52	50
1,4-Dichlorobenzene	1.714	1.677		-2.16	50
1,2-Dichlorobenzene	1.510	1.520		0.66	50
1,2-Dichloroethane-d4	0.512	0.533		4.1	50
Dibromofluoromethane	0.320	0.323		0.94	50
Toluene-d8	1.220	1.195		-2.05	50
4-Bromofluorobenzene	0.399	0.423		6.01	50

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.