

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG No.: Q1380

Instrument ID: MSVOA_Y Calibration Date/Time: 02/26/2025 10:11

Lab File ID: VY021320.D Init. Calib. Date(s): 02/25/2025 02/25/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 12:40 16:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Chloromethane	0.612	0.581	0.1	-5.07	20
Vinyl Chloride	0.601	0.586		-2.5	20
Bromomethane	0.399	0.370		-7.27	20
Chloroethane	0.373	0.363		-2.68	20
Trichlorofluoromethane	0.908	0.866		-4.63	20
1,1,2-Trichlorotrifluoroethane	0.525	0.496		-5.52	20
1,1-Dichloroethene	0.503	0.487		-3.18	20
Acetone	0.109	0.113		3.67	20
Carbon Disulfide	1.671	1.628		-2.57	20
Methyl tert-butyl Ether	1.329	1.308		-1.58	20
Methylene Chloride	0.588	0.533		-9.35	20
trans-1,2-Dichloroethene	0.554	0.540		-2.53	20
1,1-Dichloroethane	1.040	1.021	0.1	-1.83	20
2-Butanone	0.164	0.162		-1.22	20
Carbon Tetrachloride	0.554	0.536		-3.25	20
cis-1,2-Dichloroethene	0.632	0.633		0.16	20
Chloroform	1.054	1.038		-1.52	20
1,1,1-Trichloroethane	0.960	0.936		-2.5	20
Methylcyclohexane	0.623	0.578		-7.22	20
Benzene	1.456	1.415		-2.82	20
1,2-Dichloroethane	0.413	0.396		-4.12	20
Trichloroethene	0.364	0.355		-2.47	20
1,2-Dichloropropane	0.350	0.341		-2.57	20
Bromodichloromethane	0.511	0.494		-3.33	20
4-Methyl-2-Pentanone	0.243	0.234		-3.7	20
Toluene	0.904	0.892		-1.33	20
t-1,3-Dichloropropene	0.468	0.468		0	20
cis-1,3-Dichloropropene	0.552	0.543		-1.63	20
1,1,2-Trichloroethane	0.249	0.236		-5.22	20
2-Hexanone	0.159	0.159		0	20
Dibromochloromethane	0.340	0.332		-2.35	20
Tetrachloroethene	0.378	0.359		-5.03	20
Chlorobenzene	1.122	1.091	0.3	-2.76	20
Ethyl Benzene	1.988	1.972		-0.81	20
m/p-Xylenes	0.748	0.735		-1.74	20
o-Xylene	0.701	0.696		-0.71	20
Styrene	1.166	1.176		0.86	20
Bromoform	0.225	0.217	0.1	-3.56	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.

VOLATILE CONTINUING CALIBRATION CHECK

Lab Name: CHEMTECH Contract: TETR06

Lab Code: CHEM Case No.: Q1380 SAS No.: Q1380 SDG No.: Q1380

Instrument ID: MSVOA_Y Calibration Date/Time: 02/26/2025 10:11

Lab File ID: VY021320.D Init. Calib. Date(s): 02/25/2025 02/25/2025

Heated Purge: (Y/N) Y Init. Calib. Time(s): 12:40 16:49

GC Column: RXI-624 ID: 0.25 (mm)

COMPOUND	RRF	RRF050	MIN RRF	%D	MAX%D
Isopropylbenzene	3.759	3.742		-0.45	20
1,1,2,2-Tetrachloroethane	0.663	0.639	0.3	-3.62	20
1,3-Dichlorobenzene	1.730	1.676		-3.12	20
1,4-Dichlorobenzene	1.709	1.633		-4.45	20
1,2-Dichlorobenzene	1.515	1.469		-3.04	20
1,2-Dichloroethane-d4	0.567	0.544		-4.06	20
Dibromofluoromethane	0.328	0.320		-2.44	20
Toluene-d8	1.267	1.257		-0.79	20
4-Bromofluorobenzene	0.426	0.420		-1.41	20

All other compounds must meet a minimum RRF of 0.010.
RRF of 1,4-Dioxane = Value should be divide by 1000.